

Sutter Instrument
One Digital Drive
Novato, CA 94949
USA

7817

Date: 07-25-2013

Page: 1 of 1

Sold To: 567

Oregon Health & Science University
Accounts Payable
P.O. Box 572
Portland, OR 97207
USA

Shipped To: 567

Sutter Instrument
1 Digital Drive
Novato, CA 94949
USA

Purchase Order No:

Order Date:

Orig. Inv. No:

Shipped Via:

Service:

Transport Terms:

Sale Order No:

Batch No:

85390

Incoterm:

Product #	Description	Unit of Measure	Qty Returned	Unit Price	Total
FG-LB-SC	SMARTSHUTTER CONTROLLER	EA	-1	\$0.00	\$0.00
Please return instruments in original packing material. If you no longer have the original box and foam, we can ship it to you for an additional charge. Put the RMA number clearly on the outside of the box. This will help us identify your instrument or product quickly.					
U.S. Made Goods / USD					

Note:

EXCHANGE PER DAN
Contact: Allan Kachelmeier <kachelme@ohsu.edu>
Problem: The shutter isn't opening on acquisition and Metamorph verified the signal is getting through.

Subtotal \$0.00

Sales Tax \$0.00

Freight Charges \$0.00

Total \$0.00