dPatch®

Digital Patch Amplifier

ELECTROPHYSIOLOGY PATCH-CLAMP SYSTEM

WITH

SutterPatch® SOFTWARE

Operation Manual



SUTTER INSTRUMENT

One Digital Drive Novato, CA 94949 USA

Voice: +1 415-883-0128 Web: www.sutter.com Fax: +1 415-883-0572 Email: info@sutter.com The dPatch system has two international safety certifications.

1) The CE mark is for compliance to health, safety and environmental protection standards for products sold within the European Economic Area:



2) The RoHS (Restriction of Hazardous Substances) Directive 2002/95/EC restricts the use of hazardous substances for electronic equipment sold within the European Union:



Copyright © 2015 - 2021 Sutter Instrument Company. All Rights Reserved.

dPatch®, IPA®, Double IPA® and SutterPatch® are registered trademarks of Sutter Instrument Company. All other product and company names are trademarks TM or registered trademarks $^{®}$ of their respective holders.

DISCLAIMER

The dPatch system consists of one electronic amplifier with integrated digitizer, and one or two headstages. The purpose of the system is for the stimulation and measurement of cellular preparations. No other use is recommended.

This instrument is designed for use in a laboratory environment. It is not intended for, nor should it be used in human experimentation or applied to humans in any way. This is not a medical device.

Do not open or attempt to repair the instrument.

Do not allow an unauthorized and/or untrained operative to use this instrument.

Any misuse will be the sole responsibility of the user/owner, and Sutter Instrument Company assumes no implied or inferred liability for direct or consequential damages from this instrument if it is operated or used in any way other than for which it is designed.

SAFETY WARNINGS AND PRECAUTIONS

Electrical

Operate the dPatch system using 100 – 240 VAC, 50 - 60 Hz line voltage. This instrument is designed for use in a laboratory environment that has low electromagnetic noise and mechanical vibration. Surge suppression is recommended at all times.



Fuse Replacement: Replace only with the same type and rating:

Line Voltage: 100 – 240 VAC		
Fuse	Manufacturer Examples	
Rating	RoHS Compliant (Lead Free)	
T2.0A, 250V	Bussmann: GMC-2-R, S506-2A Littelfuse: 239.002.P	

Table 0-1. dPatch Fuses

Type: 5 x 20 mm glass tube, Medium Time Delay (Slow Blow), RoHS compliant.

Rating: T2.0A 250V (Time Delay, 2 Amps, 250 Volts)

Examples: Bussmann: GMC-2-R, S506-2A

Littelfuse: 239.002.P

Avoiding Electrical Shock and Fire-related Injury



Always use the grounded power cord provided to connect the Sutter system's power adapter to a grounded/earthed mains outlet. This is required to protect you from injury in the event that an electrical hazard occurs.

Do not disassemble the system. Refer servicing to qualified personnel.



To prevent fire or shock hazard, do not expose the unit to rain or moisture.

Operational

Failure to comply with any of the following precautions may damage this instrument.

- This instrument is designed for operation in a laboratory environment (Pollution Degree I) that is free from mechanical vibrations, electrical noise and transients.
- Operate this instrument only according to the instructions included in this manual.



Do not operate this instrument near flammable materials. The use of any hazardous materials with this instrument is not recommended and, if undertaken, is done so at the users' own risk.



Do not operate if there is any obvious damage to any part of the instrument.

Other

- Retain the original packaging for future transport of the instrument.
- Sutter Instrument Company reserves the right to change specifications without prior notice.
- Use of this instrument is for research purposes only.

Handling Micropipettes



Failure to comply with any of the following precautions may result in injury to the users of this instrument as well as those working in the general area near the instrument.

- The micropipettes used with this instrument are very sharp and relatively fragile. Avoid contact with micropipette tips to prevent accidentally impaling yourself.
- Always dispose of micropipettes by placing them into a well-marked spill-proof "sharps" container.

TABLE OF CONTENTS

DISCLAIMER	3
SAFETY WARNINGS AND PRECAUTIONS	3
Electrical	
Operational	
Other	4
Handling Micropipettes	5
1. INTRODUCTION	14
1.1 Overview	
1.2 Software Highlights	
1.3 Experiment Structure	
2. INSTALLATION	19
2.1 Computer Requirements	
2.2 SutterPatch System Environment	
2.3 Mounting Instructions	
2.4 Electrical Connections	
2.5 Install Hardware	
2.6 Install Software	
2.7 Test System	25
2.7.1 Install Model Cell	25
2.7.2 Startup	
2.7.3 Run a Membrane Test	30
3. HARDWARE OPERATION	32
3.1 dPatch Front Panel	
3.2 dPatch Rear Panel	
3.3 Grounding	
3.4 Headstage	
3.5 Holder	
3.5.1 Assembly	
3.5.2 Chloriding Silver Wire	
3.5.3 Holder Maintenance	
3.6 dPatch Amplifier Control Panel	
3.7 Lock-In Adjustments	
3.8 Dynamic Clamp Editor	
3.9 System Integration	
3.9.1 Using Peripheral Equipment	
3.9.2 Using Multiple Sutter Amplifiers3.9.3 Using Non-Sutter Amplifiers	
3.9.4 Using Non-Sutter Data Acquisition Systems	75 75
3.10 dPatch Maintenance	
3.10.1 Storage	
3.10.2 Inspection	
3.10.3 Cleaning	
3.10.4 Calibration	
4. SOFTWARE OPERATION	79
4.1 Acquisition	

4.1.1 Acquisition Control	79
4.1.2 Acquisition Measurements & Graphs	81
4.1.3 Acquisition: Routine Scope	
4.1.4 Camera Control	
4.1.5 Free Run	
4.1.6 Membrane Test	
4.1.7 Paradigm Editor	
Amplifier	
Each SweepRoutine	
Analysis	
Camera	
Clear Key	
Execute	
Export	
Front Window	
Hide Window	
Reset Timer	
Scope Operation	
Set Axis	
Set Checkbox	142
Set Mark	143
Set Metadata	144
Set Solution	
Set Tag	
Update Inputs	
Set Variable	
Set Write Steps	
Sound	
Start New Paradigm Data	
Update Inputs	
View Last	
Write to LogAlert	
Beep	
Comment	
Wait	
Flow Control: Break	
Flow Control: Chain	
Flow Control: For Loop	
Flow Control: Jump	
Flow Control: Label	
Condition: If	
Condition: Elself	161
Condition: Else	163
4.1.8 Routine Editor	171
Routine Editor: Acquisition & Routine Parameters	
Routine Editor: Input Channels	
Routine Editor: Output Channels & Waveform	
Routine Editor: Waveform Editor	208
Routine Editor: Real Time Measurements & Graphs	221
Routine Editor: Routine Variables	229
4.1.9 Solution Editor	230
4.1.10 Template Editor	233
2 Data Analysis	236

4.2.1 Action Potential Analysis	
4.2.2 Analysis Editor	244
4.2.3 Analysis Window	251
4.2.4 Data Browser	254
4.2.5 Data Navigator	256
4.2.6 Data Table	271
4.2.7 Edit Virtual Signals	271
4.2.8 Equation Editor	
4.2.9 Igor Analyses	
4.2.10 Layout Window	
4.2.11 Metadata Review	
4.2.12 Paradigm Review	
4.2.13 Reanalysis Measurements & Graphs	
4.2.14 Reanalysis Scope	
4.2.15 3D View Window	
4.2.16 Routine Review	
4.2.17 Routine Settings	
4.2.17 Noutine Settings	
4.2.19 Single Channel Analysis	
4.2.20 Synaptic Event Analysis	
4.3 General	
4.3.1 Command Window	
4.3.2 Dashboard Panel	
4.3.3 File Import/Export	
4.3.4 Log Window	
4.3.5 Menus	
4.3.6 Sample Files	
4.3.7 Set Preferences	
i. General	
ii. Files and Namingiii. Hardware	
iv. Control Panel	
v. Scope Window	
vi. Metadata	
vii. Graphs and Layouts	
viii. Data Exportviii.	389
ix. Factory Reset	390
4.3.8 Shortcut Editor	391
4.3.9 Startup	397
5. Programming	398
5.1 Data Format	
5.2 Data Structure	
5.3 Data Paths	
5.4 User Functions	
6. TROUBLESHOOTING	
6.1 Technical Support	
6.2 Manual	
6.3 Online Help	
6.3.1 Error Messages and Notifications	
6.4 Startup Issues	401

6.6 Analysis Issues	
APPENDIX A: LIMITED WARRANTY	412
APPENDIX B: SOFTWARE LICENSE	413
APPENDIX C: ACCESSORIES	
APPENDIX D: FUSE REPLACEMENT	
APPENDIX E. TECHNICAL SPECIFICATIONS	
APPENDIX F. SutterPatch Algorithms	_
Action Potential Threshold Algorithm	431
Auto 'Cell Compensation' AlgorithmAuto 'Electrode Compensation' Algorithm	
Auto Offset Algorithm	
LockIn Computation	
Single Channel FittingStandard Error of the Mean (SEM) Algorithm	
Synaptic Event Detection Reference	
TABLE OF FIGURES	
Figure 1-1. Data Structure - Planned Paradigms	16
Figure 1-2. Data Structure – Auto-triggered Paradigms	
Figure 2-1. Rear of dPatch Cabinet	
Figure 2-2. Front of dPatch Cabinet	
Figure 2-3. Splash Screen	
Figure 2-4. Welcome Screen.	
Figure 2-5. Emulation Modes	
Figure 2-6. Dashboard.	
Figure 2-7. Acquisition Dashboard.	
Figure 2-8. Amplifier Control Panel	
Figure 2-9. Dashboard	
Figure 2-10. Acquisition Dashboard	
Figure 3-1. Front of dPatch Cabinet	32
Figure 3-2. Rear of dPatch Cabinet	33
Figure 3-3. Amplifier Control Panels	41
Figure 3-4. Headstage Monitor	45
Figure 3-5. VC Control Panel	50
Figure 3-6. CC Control Panel	55

6.5 Acquisition Issues403

Figure 3-7. Dynamic Clamp Control Panel	59
Figure 3-8. I/O Control Panel	60
Figure 3-9. Dynamic Clamp Editor	64
Figure 3-10. Calibration	76
Figure 4-1. Acquisition Control	79
Figure 4-2. Acquisition Measurements & Graphs	81
Figure 4-3. Acquisition: Routine Scope Window	82
Figure 4-4. Signal Layout	83
Figure 4-5. Axis Magnification	85
Figure 4-6. Axis Scroll Bars	87
Figure 4-7. Center Button	88
Figure 4-8. Autoscale Axes.	88
Figure 4-9. Amplitude Meters	90
Figure 4-10. Membrane Test Settings	108
Figure 4-11. Paradigm Editor	116
Figure 4-12. Step: Amplifier	123
Figure 4-13. Step: Each Sweep	130
Figure 4-14. Step: Routine	131
Figure 4-15. Step: Analysis	132
Figure 4-16. Step: Camera	133
Figure 4-17. Step: Execute	134
Figure 4-18. Step: Export	137
Figure 4-19. Step: Front Window	138
Figure 4-20. Step: Hide Window	139
Figure 4-21. Step: Scope Operation	140
Figure 4-22. Step: Set Axis.	141
Figure 4-23. Step: Checkbox	142
Figure 4-24. Step: Set Mark	143
Figure 4-25. Step: Select Metadata Group	145
Figure 4-26. Step: Set Solution	147
Figure 4-27. Step: Set Tag	147
Figure 4-28. Step: Set Variable	148
Figure 4-29. Step: Sound	151
Figure 4-30. Step: Write to Log	153
Figure 4-31. Step: Write to Log Run-Time Window	154

Figure 4-32. Step: Alert	154
Figure 4-33. Step: Comment	155
Figure 4-34. Step: Wait	156
Figure 4-35. Step: Break	156
Figure 4-36. Step: Chain	157
Figure 4-37. Step: For Loop	157
Figure 4-38. Step: Jump	158
Figure 4-39. Step: Label	158
Figure 4-40. Step: If	159
Figure 4-41. Step: Else If	161
Figure 4-42. Checkboxes	164
Figure 4-43. Paradigm Variables	164
Figure 4-44. Routine Editor	171
Figure 4-45. New Routine Pool.	174
Figure 4-46. Waveform Preview Pane	176
Figure 4-47. Routine Settings.	178
Figure 4-48. Acquisition & Routine Parameters	179
Figure 4-49. Input Channels	184
Figure 4-50. Output Channels & Waveform	200
Figure 4-51. Output Channels & Waveform with Leak	205
Figure 4-52. Waveform Editor	208
Figure 4-53. Template Waves	213
Figure 4-54. Real Time Measurement Settings	221
Figure 4-55. Real Time Graphs	227
Figure 4-56. Routine Variables.	229
Figure 4-57. Solution Editor	230
Figure 4-58. Template Editor	233
Figure 4-59. Action Potential Analysis	238
Figure 4-60. Action Potential Measurements.	242
Figure 4-61. Analysis Editor	244
Figure 4-62. Analysis Window	251
Figure 4-63. Data Browser	254
Figure 4-64. Data Navigator	256
Figure 4-65. Data Table	271
Figure 4-66. Edit Virtual Signals	272

Figure 4-67. Equation Editor	284
Figure 4-68. Reanalysis Measurements & Graphs	305
Figure 4-69. Reanalysis Scope Window	307
Figure 4-70. Navigator Pane	308
Figure 4-71. 3D Axes Definition	319
Figure 4-72. 3D View	320
Figure 4-73. Routine Review	322
Figure 4-74. Routine Settings	324
Figure 4-75. Single-Channel Scope	336
Figure 4-76, Single Channel Analysis	338
Figure 4-77. Signal Controls	339
Figure 4-78. Current Transition Controls	340
Figure 4-79. Properties of Selected Transitions	341
Figure 4-80. Find Target Transition	342
Figure 4-81. Move to Target Transition	343
Figure 4-82. Plots and Tables Controls	344
Figure 4-83. Current Amplitude Histogram Controls	345
Figure 4-84. Plots and Tables Controls.	346
Figure 4-85. Scatter Plot Controls.	348
Figure 4-86. Table Controls	349
Figure 4-87. Synaptic Event Analysis	354
Figure 4-88. Command Window	359
Figure 4-89. Dashboard	360
Figure 4-90. Dashboard - Acquire Data	361
Figure 4-91. Preferences Settings	378
Figure 4-92. Shortcut Editor	391
Figure C-1. dPatch Expansion Panel	420
Figure C-2. Ground Point.	421
TABLE OF TABLES	
Table 1-1. Software Terminology	18
Table 3-1. Headstage Specifications	34
Table 3-2. Conductance Pool	66
Table 3-3. Main Unit Calibration	76

Table 3-4. Headstage Calibration	77
Table 4-1. Other Scope Buttons	93
Table 4-2. Routine Files and Pools	173
Table 4-3. Routine Editor Buttons	175
Table 4-4. Equation Parser	293
Table 4-5. Engineering Notation	294
Table 4-6. Reanalysis Scope Window Buttons	311
Table 4-7. Metadata Parameters	335
Table D-1. dPatch Fuses	422
Table E-1. dPatch Amplifier - Physical	425
Table E-2. dPatch Headstage - Physical	425
Table E-3. dPatch Headstage - Operational	426
Table E-4. dPatch Headstage – Open Circuit Noise	426
Table E-5. dPatch Headstage – Grounded Input Noise	427
Table E-6. dPatch Headstage - Risetime	427

1. INTRODUCTION

Welcome to the newest breakthroughs in low-noise patch-clamp technology! Our expert team has created an ultra-fast electrophysiology recording system. Its breakthrough digital design extends the limits of low-noise single-channel and whole-cell recording.

Sutter Instrument Company is a leading manufacturer of innovative precision instrumentation in the neuroscience field. We have a worldwide reputation for the highest quality and performance of pipette pullers, micromanipulators, light and wavelength switchers. We are proud to apply this same commitment to the next generation of patch-clamp instrumentation.

1.1 Overview

Advanced Design

The dPatch® Digital Patch Clamp Amplifier is a complete digital microelectrode patch-clamp system for ultra-fast high-fidelity voltage- and current-clamp recordings. Most of the electronics (amplifier and digitizer) used in stimulating, compensating and recording from cells are integrated into a single printed circuit board.

The extremely high base digitization rate (5 MHz) and high-resolution ADCs support the full implementation of novel Dynamic Clamp experiments, and can also extend data resolution up to 22-bits.

The accompanying SutterPatch® software brings the controls and displays for full-featured data acquisition, data analysis, and graphics/layout together into a single unified program, including a software control panel for direct access to all of the dPatch amplifier functions.

The SutterPatch software was developed in the powerful Igor Pro system environment. Igor Pro, by WaveMetrics, Inc., is a data collection, management and analysis platform with a rich set of built-in functions and routines for scientific programs.

From concept to production, from hardware to software, this fully integrated patch-clamp system is the new standard in ultra-fast very-low-noise cellular recording systems.

1.2 Software Highlights

- Full-featured electrophysiology package.
- Single program for data acquisition, data analysis and hardware control.
- Complex experimental automation.
- Publication-quality graphics.

Convenient: All SutterPatch software is run by a single application. No need to

launch multiple programs or to move data between programs.

Comprehensive: All data recordings, analyses, graphs, layouts, configurations and con-

trols are saved in a single experiment file. This ensures that data are

always kept together with their complete contexts.

Automation: Automate your experiment using a rich set of data acquisition, data

analysis, and amplifier controls. Create complex "Paradigms" that can

respond to changing conditions via conditional steps and loops.

1.3 Experiment Structure

Experiment:

An Experiment is the highest-level structure in the SutterPatch world. An Experiment file can encompass all SutterPatch activity for the entire day, such as instructions (Paradigms), data acquisition parameters (Routines), recorded data (Series), execution settings, history, and comments. During reanalysis, data can be included from multiple experiments. Typically, one Experiment is created for each cell or each preparation recorded from.

Paradigm:

A Paradigm is a sequence of control instructions used in an Experiment. Every Experiment contains at least one Paradigm, whether pre-planned by the user, or automatically created by the system.

A loaded Paradigm "pool" file can contain multiple Paradigms for rapid access and execution. Such "planned" Paradigms can contain simple sequences or sophisticated control structures, using a rich set of operations, such as conditional "If-then" decisions, nested loops, user-defined variables, hardware commands, and data acquisition Routines.




Figure 1-1. Data Structure - Planned Paradigms
An Experiment with two "Planned" Paradigms running Routines.

However, if a Routine is manually run, an "auto-triggered" Paradigm is created to maintain

the Experiment structure. This default Paradigm ensures that each Series is associated with a Paradigm in the context of an Experiment. If an auto-triggered Paradigm is already the active Paradigm, it is used for subsequent manually-run Routines.




Figure 1-2. Data Structure – Auto-triggered Paradigms

An Experiment with manually run Routines uses an "auto-triggered" Paradigm.

A Paradigm's "data" includes all data points, variable values, and metadata tags from the course of a Paradigm. Altogether, this allows reconstruction of the exact course of an experiment. While a Paradigm could be compared to an itinerary, the Paradigm data correspond to the route a journey actually took. If conditional control is used in a Paradigm, e.g., for the number of loop cycles or a decision in an "*If-then*" step, these actions are recorded in the Paradigm metadata.

Routine Parameters:

A Routine is the set of data acquisition and online data analysis parameters that control input and output channel timing, triggering, command waveforms, display and real-time analysis.

A loaded Routine "pool" file can contain multiple Routines for rapid access and execution.

Series (Routine Data):

Recording Routine data creates a Series composed of all sweeps of data from all input signals. Multiple runs of a Routine create multiple Series of data. All Series are automatically stored in the current Experiment file.

Channel:

A Channel corresponds to a physical output (digital-to-analog), or a physical (analog-to-digital) or virtual input of the dPatch system.

Analog input channels are used to record data, and are displayed in their own panes in the Acquisition scope window. There are two dedicated internal analog input channels ('Current' and 'Voltage') for each attached headstage. General-purpose Auxiliary analog input channels ('AuxIN') allow recording from external instruments. Virtual input channels allow further creative processing of any input channels.

Analog output channels are used to send electrical stimuli, such as command waveforms, to the preparation. There is a dedicated internal analog output channel ('StimOUT') to send Stimulus commands to each attached dPatch headstage. General-purpose Auxiliary analog output channels ('AuxOUT') can send output signals to external instruments.

Digital Output bits are also referred to as digital output channels ('DigOUT').

All 'Aux' and 'DigOUT' channels are available via the included BNC "octopus" breakout cable, or the optional Patch Panel rack-mount panel.

Signal:

Named analog input and output channels are referred to as Signals. A Signal is either the scaled representation of a physical channel, or the virtual result of a computation.

Sweep:

A Sweep is the sum of all data from all Signals, acquired from time zero, for a fixed duration. In SutterPatch software, the Sweep duration is determined by the length of the longest command waveform.

Trace:

A Trace is a Sweep restricted to a single Signal. Therefore, a Sweep can be described as the collection of Traces across all Signals.

Segment:

A Segment is a user-defined section of the command waveform. Each Segment has a waveform type, amplitude and duration.

SutterPatch Metadata:

Metadata are additional information associated with stored data. These can include such information as the preparation (cell, tissue, animal), instrumentation (hardware, software), environmental parameters (temperature, atmospheric composition), stimuli (chemical compounds, light, acoustic), and other parameters. Metadata information is associated with the running of Paradigms and Routines and their resulting data.

Metadata are dynamically recorded with a timestamp during an experiment. Information that can be determined by the system, such as the connected hardware, SutterPatch version, user Login Name, or the change of a digital output level, are automatically recorded without user intervention. In addition, the user can enter values for a large number of user-defined Metadata parameters, such as identifiers for the subject animal or cell, the animal species, age and genotype, the recording solutions, and the electrodes or stimuli applied during the experiment. SutterPatch currently keeps track of ~ 600 Metadata parameters.

Terminology Comparison:

A table of equivalent terms to other electrophysiology software packages:

SutterPatch	PATCHMASTER	pCLAMP
Experiment	Compound Data	N/A
Paradigm	Protocol	Sequencing Keys
Routine	PGF Sequence	Protocol
Series	Series	Trial
Sweep	Sweep	Sweep
Signal	Signal	Signal
Trace	Trace	Trace
Segment	Segment	Epoch

Table 1-1. Software Terminology

2. INSTALLATION

2.1 Computer Requirements

Minimum Configuration

OS (Operating System): Windows: Version 10 (64-bit versions)

Most language packs are compatible.

(listed in OS: Control Panel > System)

macOS*: Version 10.11 (El Capitan) to 10.15 (Cata-

lina)

(listed in OS: Apple > About this Mac)

Virtual machines and OS emulators, such as Parallels

and VMWare Fusion, are not supported.

CPU (Central Processing Unit): Dual-core i5

*Note: At this point, WaveMetrics does not fully support

Igor Pro on Mac computers based on the Apple Silicon M1 architecture. See https://www.wavemetrics.com/news/igor-and-apple-arm-processors for technical details. Preliminary experiments indicate that SutterPatch Software runs on these computers, both under Igor Pro 8 and 9, and with each of the Sutter Amplifier Systems connected. However, as with each new technology, we cannot fully

exclude incompatibilities.

RAM (Random Access Memory): 6 GB

Hard Disk Free Space: SSD (Solid State Drive) 500 GB

The drive should be configured as the primary system

drive.

Display Resolution: XGA (1024 x 768)

Computer Ports: (1) USB 3.0 SuperSpeed port

Computer USB 3.0 add-in card adapters are not recom-

mended, as compatibility can be problematic.

To check for SuperSpeed USB 3.0 ports, look in the Windows Control Panel / Device Manager / Universal Serial Bus controller section for "USB 3.0" host controllers

Note: As the OS does not provide any version information on the computer's USB ports, you might need to test the physical USB ports for operational performance. In mixed USB environments, USB 3.0 ports might be colored blue.

Also, sometimes BIOS settings, virus scanners and/or Windows updates can put a USB port to sleep.

External USB hubs are not supported.

Recommended Configuration (for Bandwidths > 50 kHz)

CPU: As fast and powerful as possible.

*Note: At this point, WaveMetrics does not fully support Igor Pro on Mac computers based on the Apple Silicon M1 architecture. See https://www.wavemetrics.com/news/igor-and-apple-arm-processors for technical details. Preliminary experiments indicate that SutterPatch Software runs on these computers, both under Igor Pro 8 and 9, and with each of the Sutter Amplifier Systems connected. However, as with each new technology, we cannot fully exclude incompatibilities.

RAM: 16 GB (or more for very large data sets).

Display Resolution: Full HD (1920 x 1080).

High resolution displays (>96-DPI), such as Retina, 4K,

5K, Quad-HD and Ultra-HD, are not supported.

2.2 SutterPatch System Environment

The SutterPatch software runs in the Igor Pro 64-bit system environment. Igor Pro is widely used by scientists to acquire and analyze data, and to create publication-quality presentation graphics.

Igor Pro Features

- High-speed data display.
- Large data set handling.
- · Waveform arithmetic.

- Extensive set of data analyses.
- · Image display and processing.
- · High-quality presentation graphics.
- · Graphical and command-line user interfaces.
- · Automation.
- Extensibility via C and C++ modules.
- Extensive online Help and PDF manual.

2.3 Mounting Instructions

Rack Mounting: The dPatch amplifier is ready for mounting in a standard 19" wide

equipment rack in a 2U space. A rack-mount hardware kit consisting of

hex screws, washers and cage nuts is included.

Benchtop Usage: Attach the four included stick-on feet to the bottom of the dPatch ampli-

fier

2.4 Electrical Connections

AC Power: 60 Hz

50 Hz

The dPatch amplifier runs on AC power from 100 to 250 VAC - no switches need to be set.

The AC power should be as clean as possible:

- At a minimum, a surge protector should be used to protect against high-voltage spikes. If lightning strikes are a concern, the surge protector should be rated > 1000 joules and > 40 kA.
- If you experience brownouts or voltage sags, use a switching power supply (SPS) to supply clean power to your instrument.
- To protect against power interruptions, use a universal power supply (UPS) for uninterrupted clean power.

2.5 Install Hardware



Figure 2-1. Rear of dPatch Cabinet

- 1. Plug the female end of the power cord into the dPatch rear panel power receptacle.
- 2. Plug the male end of the included power cord into a grounded electrical mains outlet.
- 3. Push the dPatch power button to OFF (unlit position).



WARNING! Hot-swapping of headstages should be avoided, or a software crash with data loss may occur. Turn off the dPatch system power before altering headstage connections.



Figure 2-2. Front of dPatch Cabinet

- 4. Plug the dPatch headstage preamps into the HEADSTAGE ports on the front of the dPatch amplifier each headstage is independently tuned, so any port can be used.
- 5. Plug the included Screw Terminal Panel or optional dPatch Expansion Panel into the DIGITAL OUTPUTS port on the back of the dPatch amplifier.
- 6. Connect the supplied USB 3.0 cable to your computer's USB 3.0 port and the dPatch amplifier's rear panel USB port.
- 7. Connect the included electrode holder to the headstage. See the Holder section for holder assembly instructions.

2.6 Install Software

A. Locate the Files

Use your web browser to locate the latest versions of the SutterPatch v2.2 installer software at: https://www.sutter.com/AMPLIFIERS/SutterPatch.html

and choose the "Download" tab.

If internet access is not available, attach the included USB flash drive to your computer USB port, and use your file browser to navigate to the flash drive installer files.

B. Choose Installer File

Windows	Mac
Full: SP + Igor 9	Full: SP + Igor 9
Full: SP + Igor 8	Full: SP + Igor 8
Updater: SP (Igor 9)	
Updater: SP (Igor 8)	

Separate Full installers (4) are provided for Windows OS and macOS for Igor Pro 9 and Igor Pro 8: both the Igor Pro and SutterPatch software are updated.

It is strongly recommended to run the **Igor Pro 9 Full installer** for optimum data processing performance.

However, if you do not want to upgrade to Igor Pro 9, the last version of Igor Pro 8 (v8.0.4) still supports SutterPatch v2.2,

Additionally, multiple versions of Igor Pro and SutterPatch can be installed on your computer, each with independent settings and parameters.

The Windows Full installers install both 64-bit and 32-bit English versions of Igor Pro. Only the 64-bit version supports SutterPatch. The 32-bit version is for any Windows 3rd party applications that require it.

The macOS Full installers install a 64-bit English version of Igor Pro. For macOS 32-bit support, run Igor Pro 7.

Separate Updater installers (2) are provided for Windows OS running SutterPatch on Igor Pro 9 or Igor Pro 8: only the existing SutterPatch software is updated.

We do not recommend using the SutterPatch Updater on an existing "standalone" installation of Igor Pro. If you already have Igor Pro (without SutterPatch) on your computer, please use the full installer (for the same Igor Pro version) to update the existing version of Igor Pro and install SutterPatch v2.2.

Igor Pro has a 30-day trial period where it is fully functional and fully supports SutterPatch. After 30-days, if the Igor Pro license has not been activated, Igor Pro runs in a demo mode with limited functionality that does not support the SutterPatch application.

Note: Japanese versions of Igor Pro do not support SutterPatch.

C. Install the Software

Use your file browser to navigate to the downloaded installer file and run it:

- 1. Install the full software for 'All Users' by double-clicking on:
 - Windows: sutterpatch_win_full
 - macOS: sutterpatch_mac_full
- 2. Follow the installer prompts:
 - We recommend replacing any prior versions of Igor Pro with the latest version of Igor Pro 9, after making a backup copy of all user files and parameter files in the program folder and its sub-folders.
- Prior versions of Igor Pro can be kept if desired, as different versions of Igor Pro can coexist on the same computer.
 - If an existing version of Igor Pro is found, the Igor Pro Preferences are overwritten.
 - If an existing version of SutterPatch is found, the SutterPatch sample files are overwritten.
- 3. Upon completion, the installer will report a successful installation. The following files and folders are installed:
 - dPatch QuickStart Guide (PDF) (Includes Igor Pro 9 Serial Number and Key.)
 - Release Notes (PDF)
 - SutterPatch manual (PDF)
 - SutterPatch2 / SutterPatch folder (Includes sample Routine Pool, Paradigm Pool, and Experiment data files.)
 - SutterPatch2 / SP_Code folder
 - SutterPatch2 / SP Drivers folder (XOP)
- 4. Launch Igor Pro by clicking on its icon:



5. Activate its license as instructed.

You will need to enter the Igor Pro 9 Serial Number and Activation Key found in your dPatch Quick Start Guide.

6. "Eject" the flash drive - wait for the "Safe to Remove Hardware" prompt, and then unplug it from the computer.

2.7 Test System

2.7.1 Install Model Cell

- 1. Attach the model cell to the Headstage 1 holder and tighten the screw collar.
- 2. Plug the supplied 1 mm grounding wire into the gold sockets on the headstage and model cell.
- 3. If the headstage is not inside a Faraday cage, completely surround the model cell/headstage assembly with alternative electromagnetic shielding (such as aluminum foil), and connect the shielding material to the headstage ground a short wire with a metal alligator clip on each end makes a convenient shield-ground connector.

2.7.2 Startup

- 1. Power-on the dPatch amplifier by pressing the silver POWER button on its front it lights up as blue. (It can take a few seconds for the USB connection to be established.)
- 2. Launch the SutterPatch application by clicking on the 'Igor Pro 9' icon:



An Igor Pro splash screen displays while opening files:




Figure 2-3. Splash Screen

Then the Welcome to SutterPatch "start" window displays:



Figure 2-4. Welcome Screen

- 3. Click on the 'Start' button in the 'Welcome to SutterPatch' window, and the application begins compiling. This process may take several seconds.
- 4. Specify the Experiment file name and storage location when prompted.

5. If the dPatch amplifier is OFF, or disconnected from the computer, this dialog displays:




Figure 2-5. Emulation Modes

- a. Reconnect it, and then click 'Retry', or
- b. Click in 'Select device to emulate' and select 'dPatch'.

Note: If amplifier hardware is not attached, and 'Prompt on startup' is disabled, the program automatically starts up in the last known emulation state.

6. Next, the Dashboard panel is displayed.




Figure 2-6. Dashboard.

7. Click on the 'Acquire Data' icon, and a second level of the Dashboard is displayed.



Figure 2-7. Acquisition Dashboard.



8. Click on the 'Control Panel' icon, and the Amplifier Control Panel is displayed.

Figure 2-8. Amplifier Control Panel

- a. If "Control DEMO" is displayed in the Amplifier Control Panel title bar, you are running in a hardware emulation mode to run the physical instrument, ensure that the amplifier is powered on and its USB 3.0 cable is connected, choose "New Experiment" from the Dashboard, and select "dPatch".
- b. For the next step, make sure that the Amplifier Control Panel is in voltage-clamp mode the "VC" button at the top of the Amplifier Control Panel should be highlighted in red.

2.7.3 Run a Membrane Test

The Membrane Test is useful for a quick check of the dPatch system functionality. It tests the three basic steps necessary for recording in a whole-cell configuration.

1. Go to the Dashboard window and click on the 'Acquire Data' icon.




Figure 2-9. Dashboard

2. Click on the 'Membrane Test' icon.



Figure 2-10. Acquisition Dashboard.

The following test values assume a 5 kHz filter.

3. Test the BATH mode:

This mode simulates placing an electrode into the bath solution and sending a voltage pulse through the solution.

- a. Set the Model Cell switch: Bath
- b. Click on the Membrane Test 'Bath' button.
- c. Verify value: Pipette Resistance: $\sim 10 \text{ M}\Omega$

4. Test the SEAL mode:

This mode simulates an electrode making contact onto a cell and forming a high-resistance gigaohm seal with the membrane.

- a. Set the Model Cell switch: Seal
- b. Click on the Membrane Test 'Seal' button.
- c. Verify value: Seal Resistance: $\sim 1~G\Omega$ to $1~T\Omega$

5. Test the CELL mode:

This mode simulates an electrode breaking into a cell and achieving a successful whole-cell patch.

- a. Set the Model Cell switch: Cell
- b. Click on the Membrane Test 'Cell' button.
- c. Verify values: Series Resistance: ${\sim}10~\text{M}\Omega$ Membrane Resistance: ${\sim}500~\text{M}\Omega$

Membrane Capacitance: ~28 pF

6. For a dual-headstage system:

- a. Move the model cell, ground wires and shielding to Headstage 2.
- b. Set the Acquisition window to 'Headstage 2'.
- c. Repeat steps 3-5.

3. HARDWARE OPERATION

3.1 dPatch Front Panel

The front panel of the dPatch system is used for the headstages, external I/O connections, and a power button.



Figure 3-1. Front of dPatch Cabinet

The front panel, from left to right:

POWER:

HEADSTAGE 1 & 2:	HDMI-style Type A	For dPatch headstages.
AUXILIARY INPUTS 1 – 4:	BNCs	Analog inputs from external sources.
AUXILIARY OUTPUTS 1 & 2:	BNCs	Analog outputs to external targets.
TRIGGER IN:	BNC	Digital input trigger.
TRIGGER OUT:	BNC	Digital trigger output pulse.
		Automatically sent at the start of continuous acquisition or each triggered sweep (including Membrane Test).

Turn power to unit On / Off. Lights up blue when 'On'.

Button

3.2 dPatch Rear Panel

The rear panel of the dPatch system is used for grounding, USB connections, and signal I/O.



Figure 3-2. Rear of dPatch Cabinet

[Unlabeled]:	Power-entry receptacle	For AC power cord.	
SIGNAL GROUND:	4 mm Banana socket	Low-voltage grounding.	
EARTH GROUND:	4 mm Banana socket	Instrument grounding.	
USB-3:	USB Type B receptacle	USB 3.0 computer communication.	
AUXILIARY I/O:	HDMI-style Type A	Expansion Panel analog signals.	
FIRMWARE UPDATES:	Chip	$\label{eq:Firmware code} Firmware\ code-keep\ chip\\ installed.$	
AUXILIARY INPUTS 5 – 8:	BNCs	Analog inputs from external sources.	
AUXILIARY OUTPUTS 3 & 4:	BNCs	Analog outputs to external targets.	
DIGITAL OUTPUTS 1 – 16:	DC-37 D-Sub connector:	Digital output channels.	
		(See Appendix E for pin definitions.)	

3.3 Grounding

Proper grounding is essential for the integrity of an electrophysiology laboratory setup. It greatly affects the "noise" within your system, and hence the quality of your data recordings. Very low noise levels are especially needed for single-channel recordings. While AC (mains) line-noise (hum) can be software-filtered out of a data signal, it is much more desirable to have a well-grounded electromagnetically clean hardware environment to start with.

For a properly grounded laboratory, an electrical connection is needed from your laboratory's electrical system to an "earth" ground. If your building's electrical grid does not provide a good earth ground, you can create your own earth ground by making use of the building's plumbing system, or by inserting a heavy metal bar deep into the earth.

The equipment in a rig should all be grounded to a single point to avoid ground loops. Installing a bus bar to the earth ground also helps to prevent ground loops. Consider standardizing your setups by using a GP-17 Ground Point on each rig.

"Signal" ground is a sensitive ground for low voltages:

• BNC shields: Hard-wired to signal ground (single-ended).

• Bath ground electrode: Connect to the headstage signal ground jack.

• Shielding (Faraday cage): Connect to the rear panel SIGNAL GROUND socket.

However, due to the complexity of grounding factors, you may need to test various strategies for the best grounding configuration for your system. For example, when multiple headstages are used, one or both headstages might need to be grounded.

A grounded power cable is provided with this instrument.

3.4 Headstage

The dPatch headstage supports both voltage- and (true) current-clamp in the same headstage. Multiple capacitor and resistor feedback elements optimize both single-channel and whole-cell recordings.

Feedback Element	Recording Amplitude Range	Analog Band- width*	RMS Noise**	Pipette Compensation Range	Series Resistance Range	Cell Capacitance Range
1 pF Capacitor	±20 nA	> 500 kHz	< 0.22 pA	20 pF	N/A	N/A
500 MΩ Resistor	±20 nA	> 250 kHz	< 0.7 pA	20 pF	100 MΩ	100 pF
50 MΩ Resistor	±200 nA	> 250 kHz	< 2.3 pA	20 pF	10 ΜΩ	1000 pF

Table 3-1. Headstage Specifications

- * The headstage analog bandwidth for new headstages now achieves a full 1 MHz. Earlier headstages would need to be retrofitted to achieve this bandwidth.
- ** The headstage RMS (Root-Mean-Square) noise is measured with an "open circuit" configuration at 10 kHz with an 8-pole Bessel filter. Measuring "open-circuit", i.e., with no attachments so the headstage input is exposed to the air, provides a fairly consistent baseline for such headstage noise measurements. Conversely, measuring noise with an electrode in the bath generates the worst noise conditions.

A 1 mm gold pin signal-ground socket is on the back of the headstage.

One or two dPatch headstages can be attached in any order to either HEADSTAGE port. The dPatch headstage cable is permanently attached to a preamplifier, whose cable is the component that is actually plugged into the amplifier's headstage port.

Preamps can be positioned standing up or laying down flat, as heat conduction is not an issue, for example when positioning on an anti-vibration air table.

The dPatch headstage cable length cannot be increased.



WARNING! Hot-swapping of headstages should be avoided – data loss can occur. Turn off the amplifier's power before handling headstages.

The dPatch Amplifier uses two modes of headstage operation:

1) "Capacitive" mode: Single-channel patch-clamp recording.

Feedback element: 1 pF Capacitor.

Advantages: Less noise (thermal, correction circuitry).

Faster response (increased bandwidth).

Better linearity.

A capacitive (integrating) headstage has the benefit of lower noise over resistive headstages. However, it also requires periodic "resets" of its feedback capacitor. As current flows into the feedback capacitor, the capacitor charges up to its limit, and then discharges (resets) to zero, before starting the next round of measurements.

During the reset period, the data is not valid, however, the reset transients can be optionally left unaltered, "blanked", or "masked", during or after data acquisition, This SutterPatch software implementation is much more flexible than hardware implementations, which can only mask the data (sample and hold).

The frequency of automatic resets is proportional to the amount of current (amplitude) in the recorded signal. As the capacitor rating is 1 pF, a current of 1 pA will cause the reset to occur after 700 ms (when the capacitor reaches 70% of its maximal charge.)

2) "Resistive" mode: Whole-cell patch-clamp recording.

Feedback element: $500 \text{ M}\Omega$ and $50 \text{ M}\Omega$ resistors.

Advantages: No capacitive "resets".

Extended input range of ±200 nA.

3.5 Holder

A "holder" attaches a microelectrode (pipette) to a headstage. It provides mechanical stability for the pipette, low-noise for the electrical circuit, and chemical inertness from its physical components. Our pipette holders accept electrode glass in the range of 1.0-1.7 mm OD (Outer Diameter) using sized-by-color silicone gaskets.

The standard pipette holders included with the dPatch amplifier are composed of ultra-low-noise quartz for very low-noise recording. A suction tube projects at a right angle from the middle of the barrel

.

While polycarbonate is a proven material for patch pipette holders, it undergoes significant thermal expansion. Uneven warming can lead to motion of the pipette tip, and is often incorrectly perceived as drift in the micromanipulator. Quartz has a significantly lower thermal expansion coefficient, and virtually eliminates thermal drift.




WARNING!

Quartz is fragile, and can crack or shatter on impact. Treat your quartz pipette holder with the same care as with an optical component.

3.5.1 Assembly

The holder is assembled from 8 parts incorporated into a main barrel:

End Cap - Gasket - Silver Wire - Barrel - Tubing - Gold Pin - Pin Cap - Lockdown Ring



Dimensions are in inches [mm].

Assembly Tips

- Silver wire should be kept straight do not bend or twist it.
- Fire-polish glass electrodes on both ends to prevent scratching the silver wire or the holder barrel.

Assembly Procedure

- 1. Cut the silver wire to size. Check for proper tubing height and wire-crimping length the depth it extends into the pipette plus half the length of the barrel avoid excess or insufficient amounts.
- 2. Chloride the silver wire. (see below)
- 3. Thread the silver wire through the barrel.
- 4. Cut a small piece of clear tubing sized to fill the tiny "end-cup" in the pin-side of the barrel the end with the narrower shaft.
- 5. Thread the small piece of tubing over the silver wire into the end-cup.
- 6. Crimp the end of the silver wire just slightly over the end of the tubing.
- 7. Slide the lockdown ring over the tubing-side of the barrel, with the ring's threads facing outwards.
- 8. Insert the gold pin into the recessed end of the pin cap push it through the pin hole until it stops.
- 9. Screw the pin cap onto the barrel so that pressure from the compressed snippet of tubing ensures good electrical contact between the silver wire and the gold pin.

For the most stable configuration, before screwing the pin cap onto the barrel, solder the crimped silver wire to the end of the gold pin. Apply only a small bead of solder in the very middle of the top of the pin to avoid any excess solder interfering with the parts properly mating, as excess solder can result in air or solution leaks.

10. Find a silicone gasket with an ID (inner diameter) just greater than your pipette OD (outer diameter):

Gasket ID	$\underline{\text{Color}}$
1.1 mm	Clear
1.2 mm	Green
1.5 mm	Orange-Red
1.75 mm	Blue

Note: The rubber gasket will wear out over time and need to be periodically replaced.

- 11. Thread the gasket onto the silver wire on the end-cap side the side with the narrower shaft.
- 12. Thread the end cap onto the silver wire and loosely tighten until it makes contact with the gasket.
- 13. Carefully thread a solution-filled micropipette onto the silver wire and into the gasket, and push it into the barrel until it reaches the back end of the bore in the middle of the barrel.
- 14. Tighten the end cap onto the barrel.
- 15. Attach the holder to a headstage with the lockdown ring.

3.5.2 Chloriding Silver Wire

The silver wire should be chlorided before first-time use, and then re-chlorinated monthly, or as needed.

Chemical Method

- 1. If needed, use a razor blade or fine sandpaper to rub off any insulation.
- 2. Optionally clean the silver (Ag) wire with ETOH (ethanol) to remove finger oils.
- 3. Immerse the silver wire in common household bleach (sodium hypochlorite) in glassware for 5-30 minutes until it turns purple-gray in color.
- 4. Remove the chlorided silver wire and rinse in distilled water.
- 5. Dry for storage.

Electrochemical Method

- 1. If needed, use a razor blade or fine sandpaper to rub off any insulation.
- 2. Optionally clean the silver (Ag) wire with ETOH (ethanol) to remove finger oils.

- 3. Connect a separate silver wire to each pole (positive and negative) of a household battery (1.5 V 9 V).
- 4. Immerse the two silver wires in a solution of KCL (3 M) in glassware for 5-10 minutes. The wires should not touch each other. Bubbling around a silver wire indicates electroplating is occurring.
 - Alternatively, use HCL (1M) with a 2 hour immersion time.
- 5. The charging polarity for the wires should be reversed a few times during the process.
- 6. A fully chlorided silver wire should be purple-gray in color. Remove the chlorided silver wires and rinse in deionized water.
- 7. Dry for storage.

Re-Chloriding Silver Wire

- 1. Pass the used silver wire through a flame the wire should become bright silver in color.
 - Alternatively, use a razor blade or fine sandpaper to scrape off any existing chloride.
- 2. Chloride the wire as described above.

3.5.3 Holder Maintenance

Holders must be properly maintained for good noise performance.

Storage:

- 1. Holders should be clean and dry.
- 2. Store in a container with desiccant.

Before 1st time use:

- 1. Disassemble the holder.
- 2. Rinse the polycarbonate parts in 70% ethanol.
- 3. Blot dry.
- 4. Store in a container with desiccant overnight.

After daily use:

- 1. Rinse holders with distilled water. For more thorough cleaning, wash with ethanol.
 - Caution! Washing with soapy water can leave a film.
 - Continual cleaning with ethanol can degrade the polycarbonate parts.
 - Do not clean with methanol or strong organic solvents such as acetone.
- 2. Blot dry.

Weekly Cleaning:

- 1. At least once per week, disassemble holder.
- 2. Clean the polycarbonate parts with 10-20 s sonication in distilled water.
- 3. Blot dry.
- 4. Store in a container with desiccant overnight.

3.6 dPatch Amplifier Control Panel

This software interface controls the dPatch amplifier settings. It replaces all physical knobs, dials and meters, such as found on manually-controlled amplifiers.

Most of these settings can also be programmatically controlled in a Paradigm.

Most editable numeric fields can also be adjusted via a graphic control panel with three slider bars (for 3 significant digits) by right-clicking on the field.




Figure 3-3. Amplifier Control Panels

The size and magnification of the Control Panel can be set by right-clicking in blank space in the Control Panel. A list displays with magnification settings from 50% to 400%.

General Controls

[Headstage #] tabs [1-2]

For multiple headstages, each headstage maintains its own settings.

Clicking on a headstage tab will open its last-used active mode (VC, CC) settings; the tab and its background are set to matching colors.

Note: If a headstage is unattached while the amplifier is in use, its channel will be at "ground".

Show the second headstage in its own control panel.

V V' meter: Displays the Voltage input channel level.

T I meter: Displays the Current input channel level.

VC VC button: Switch the dPatch amplifier from Current Clamp to

Voltage Clamp mode. The button for the active mode

is highlighted in red.

A "gentle" switch is used to protect cells against transition spikes that could degrade the electrode seal and integrity of the recording. The amount of voltage to hold the current steady is automatically injected into the cell during the transition (< 100 ms), before

stepping to the new voltage command level.

Switch the dPatch amplifier from Voltage Clamp to Current Clamp mode. The button for the active mode

is highlighted in red. If Dynamic Clamp is also active,

the button includes a "DynC" icon.

A "gentle" switch is used to protect cells against transition spikes that could degrade the electrode seal and integrity of the recording. The amount of current to hold the voltage steady is automatically injected into the cell during the transition (< 100 ms), before

If the mode is switched during acquisition, tags are

inserted into the data record with the new Control

stepping to the new current command level.

Panel settings.

Reset USB button: Click to re-establish the USB connection to the amplifier.

- All USB channels are reset.
- A green button indicates that a stable USB connection to the amplifier has been established.
- It can take several seconds for the USB connection to be re-established.
- A red button indicates that there is no USB connection to the amplifier, or that SutterPatch is running in hardware emulation Demo mode.
- If the amplifier is attached while in Demo mode, start a new Experiment to exit Demo mode and run the hardware "live".

CC

CC button:



• While in Demo mode, the Acquisition Test scope windows display a "DEMO" watermark in each input signal pane.

Feedback Mode [Range]

Two headstage-feedback modes are provided to improve recording performance under different conditions: capacitive mode for very-low-noise (single-channel) voltage-clamp recordings, and resistor mode for whole-cell voltage- and current-clamp recordings.

For resistor mode, two whole-cell feedback resistors provide both regular and extended range current-clamp current-injection amplitudes.

Capacitive Mode

± 20 nA range, for very low-noise single-channel recording.

Feedback element: 1 pF Capacitor.

The headstage feedback capacitor, while providing very low-noise circuitry, also generates periodic "reset" artifacts during recording. While capacitor resets do not affect the preparation, they do affect the recorded signal.

For single-channel recordings in voltage-clamp mode, ion-channel currents are typically so small, that capacitor resets are infrequent enough to be essentially ignored within the 20 nA range.

However, in current-clamp mode, voltage responses are often large enough that many resets can occur during recordings.

When at least one reset occurs in a sweep, a small gold box flashes in the Amplifier Control Panel to the left of the Capacitive Mode/Range field. When more than one reset occurs in a sweep, the gold box will flash twice with the sweep. The number of resets for the sweep will display at the bottom of the Amplifier Control Panel as "Cap. Discharges: #".

(See the Headstages section for more details.)

The raw data includes the actual capacitive reset transients, however the transients can also be optionally hidden to reduce their impact during acquisition or analysis. Use a Virtual Input Channel (Math Type:

Bessel Filter or Smooth) to manually blank or mask transients.

For Routines that contain inter-sweep intervals, if the headstage capacitor is reset during this interval, there is no interference with the acquired data. Such capacitor resets can also be forced by running the Paradigm step 'Amplifier / Reset Cap'.

Note: The Virtual Input Channel 'Integrator Reset' controls are only visible when the Routine data are acquired in Capacitive mode.

Range \pm 20 nA Standard resistor-feedback amplitude range for wholecell voltage-clamp inputs or current-clamp outputs.

Feedback element: $500 \text{ M}\Omega$ Resistor.

Range \pm 200 nA Extended resistor-feedback amplitude range for wholecell voltage-clamp inputs or current-clamp outputs.

Feedback element: $50 \text{ M}\Omega$ Resistor.

Reset Reset the active headstage controls to default settings.

• Safe mode: Initiate with a mouse left-click.

Disable these controls:

Offset 'Lock'

VC mode

Cell Compensation Rs Correction

CC mode

Electrode Compensation Bridge Balance Correction Dynamic Holding

Reset these controls to defaults:

VC mode

I-filter

CC mode

V-filter

• Default mode: Initiate with a mouse right-click or Shift-click.

Reset all fields to default settings, except:

Liquid junction Offset value

VC mode

V-holding value

CC mode

Subtract Pipette Offset

Show / Hide Controls The vertical 'Show / Hide' button displays/hides all controls (general and tabbed) below this button.

Show / Hide Monitor

The horizontal 'Show / Hide' button displays/hides the dPatch input monitor, which displays the real-time values for both headstages Voltage and Current input channels and all auxiliary input channels. Auxiliary channels can be configured for temperature and pressure readings in dPatch Settings / Configure AuxIN Monitor.




Figure 3-4. Headstage Monitor

dPatch Settings

[Demo mode only]

Enable Headstage 1 Configure the control panel to include the first headstage.

Enable Headstage 2 Configure the control panel to include the second headstage. Note: At least one headstage always remains enabled. Live Demo Mode Updates pseudo-input signal values for the Control Panel 'V' (voltage) and 'I' (current) monitors. Also updates the demo mode Membrane Test measurement monitors. If these updates are distracting, disable here. This feature does not consume extra system resources. [Hardware and Demo modes] Enable the dPatch Control Panel 'V' and 'I' Read Signal Inputs monitors to read input signals. Configure AuxIN Monitor Configure the monitor to display slowly changing temperature and pressure readings. (see below) _____ Show Bridge Balance as % Rs compensation Show Bridge Balance Magnitude in $M\Omega$ -----Combine RsCorrection and RsPrediction Separate RsCorrection and RsPrediction Display the 'Prediction' component separately. ----Subtract Pipette Offset in Current Clamp In CC mode, the Auto Offset calibration can take up to 10 s for settling time. Show control "Stim Both" Show Auto Bridge Balance Amplitude Set the test pulse amplitude of the auto bridge function. _____

Allow Filter Bandwidth higher than 50 kHz

When disabled, the input bandwidth is limited to 50 kHz or less. This restriction against very high throughput levels helps prevent inadvertently overloading the system resources during data acquisition and analysis.

Note: At very high bandwidths, the digital display can be misleading, as sparse outliers of data can make an unmagnified signal look extremely noisy.

Subtract Capacitor Reset Transients

In capacitive mode, when a capacitor reset occurs, the current generated by the reset transient is subtracted from the signal for a fixed duration of 1.024 ms.

This setting is automatically disabled if there is no EEPROM lookup table loaded, such as due to having an outdated version of it.

[Not supported in demo mode.]

Configure AuxIN Monitor

AuxIN# Transform [1-8]

Enable an auxiliary input channel.

Label Text label for dPatch Monitor auxiliary

channels.

Unit °C (for temperature) or mPa (for pressure).

Engineering Enable to use engineering units $(m, \mu, n, etc.)$

Disable to use exponents.

Equation Allows scaling and offsets to be applied.

[List]

- Not assigned
- Bath Temperature
- Ambient Temperature
- Atmospheric Pressure

Samples to Average [1, 2, 5, 10, 20. 50]

Check Check for errors in the active equations and

Liquid junction [±250.0 mV]

Correct for the liquid junction potential in patch mode.

When a recording electrode is placed into the bath, an ionic potential difference is formed between the two dissimilar solutions in the pipette and the bath. This liquid junction potential is counterbalanced when a system Offset is applied to reach a zero-current state.

However, when a patch is formed, a liquid junction no longer exists, and thus should no longer be counterbalanced by the system Offset. The 'Liquid junction' setting corrects for this, but you will need to calculate or experimentally determine the value to enter.

Note: By convention, the liquid junction potential polarity is defined in relation of the bath to the pipette solution.

- For whole-cell and outside-out patches, enter the 'Liquid junction' value.
- For cell-attached and inside-out patches, reverse the polarity of the 'Liquid junction' value you enter.

Auto Offset [±250.0 mV]

Click on the Auto Offset button to counteract any inherent hardware circuitry electronic offsets, as well as electrode-in-the-solution chemical offsets. Clicking this button applies a compensatory potential that automatically zeroes the current signal.

The Offset value is to the right of the Auto Offset button, and is independent of the holding potential.

- Offset values can be directly typed into the numeric field, or
- For fine adjustments, use the up and down spinners to increase or decrease the setting by 0.1 mV.

The Auto Offset value is an approximation that might need further adjustment. Adjust the voltage offset in voltage-clamp mode when an electrode is initially placed into the bath solution, so that the electrode current signal is zero.

See the Control Panel Settings to disable the offset for current-clamp mode.

(See the SutterPatch Algorithms appendix for the Auto Offset algorithm.)

Lock Once an electrode offset has been applied, use the 'Lock' checkbox to prevent accidental changes to the Offset.

The 'Lock' is automatically enabled whenever a Routine starts to acquire data.

Voltage Clamp Controls

VC The 'VC' tab displays the amplifier Voltage Clamp controls.




Figure 3-5. VC Control Panel

V-holding (mV) $[\pm 750.0]$

After achieving a seal, the holding voltage is typically set to the cell's equilibrium or "resting" membrane potential (typically -60 to -80 mV for neurons.) This control is active during acquisition - changes are applied to the next sweep.

Note: The holding level is restricted to ± 750 mV, because when used in conjunction with a hardware offset and/or liquid junction potential totaling up to ± 250 mV, the command potential can reach the circuitry's analog output limit of ± 1 V.

.

Enable Use the Enable checkbox to activate V-holding.

If V-holding is not enabled, the holding level is zero volts.

The holding level value is written at the start of Routine acquisition to the metadata as 'Command Holding'.

Whenever this button is activated or de-activated during acquisition, a 'Command Holding Value' tag is inserted into the data recording and

written to the metadata.

I-Filter Low-pass 8-pole Bessel filter bandwidth.

```
[ 100, 200, 500 Hz; 1, 2, 5, 10, 20, 50, 100, 250, 500, 1000 kHz ]
```

This low-pass filter is applied to the active headstage VC-mode input signals.

dPatch headstages have a fixed hardware sampling rate of 5 MHz. The actual system sampling rate is based on downsampling the hardware sampling rate, to the sampling value calculated by applying the Routine Editor / Input Channels / Nyquist Factor to the selected filter Bandwidth.

A Routine can be set up to record with either the I-Filter setting or the Input Channels / Filter Bandwidth setting. Filter controls are locked during acquisition, as they affect the sampling rate.

Note: a) At very high bandwidths, the digital display can be misleading - an unmagnified signal can look very noisy due to sparse outliers of data.

However, due to its ultra-high digitization rates, the dPatch amplifier provides excellent bit-resolution of digitized input signals, without the need for input gain controls.

For example, decimating from an 18-bit resolution at 1 MHz, the bit resolution increases by 0.414 bits for each 10-fold reduction of the bandwidth, resulting in a better than 22-bit resolution at 1 kHz.

b) Filtering is not applied in Demo mode. Routines record demo data with the "Nyquist Factor-adjusted" sampling rate.

Tip: For experiments where the shape of the response is of interest, an input filter rate of 10 kHz is commonly used.

However, for very long stimuli, you might want to use a lower input filter rate.

Stim Both Enable synchronous stimulation for two headstages during auto compensation and correction operations.

[only displays if selected in the dPatch Settings]

Electrode Compensation: Electrode capacitance compensation.

Magnitude: [0.00 - 20.0 pF] [shared with CC mode]

Tau: $[1.00 - 10.0 \,\mu s]$

+

Opens a 2-D slider panel for simultaneous tuning of both parameters.

both parameters

Note: When dragging with a mouse, slow down when approaching panel boundaries, else overshoot or

undershoot of the values can occur.

Auto Automatically sets approximate values.

(See the SutterPatch Algorithms appendix for the algorithm.)

After making an on-cell gigaohm seal, large microelectrode capacitance spikes are visible. To remove the transients, click the 'Auto' button. Then autoscale the signal, and if needed, either adjust the 'Mag' and 'Tau' controls individually, or use the slider panel for a combo control. For a square pulse command (such as a Membrane Test 'Seal' command), the goal is to eliminate the edge-effect spike transients.

Note: The Electrode Compensation required to cancel out all transients can vary slightly between the three Feedback Range settings: Capacitive Mode, Range +/-20 nA and Range +/-200 nA. If the Feedback Range is switched during an Experiment, it is recommended to rerun the Membrane Test and fine tune the Electrode and Cell Compensation, and if applicable, when in Voltage Clamp mode, Rs Correction, or when in Current Clamp mode, Electrode Compensation and Bridge Balance Correction.

Cell Compensation: Whole-cell capacitance compensation.

Rs: Series Resistance

 $[0.10 - 100.00 \text{ M}\Omega]$

Cm: Membrane Capacitance

[1.00 - 500.00 pF]

+

Opens a 2-D slider panel for simultaneous tuning of

both parameters.

Note: When dragging with a mouse, slow down

when approaching panel boundaries, else overshoot or undershoot of the values can

occur.

Auto Automatically sets approximate values using small

"gentle" steps to avoid hyperpolarization.

(See the SutterPatch Algorithms appendix for the

'Auto Cell Compensation' algorithm.)

After breaking into a cell, i.e., going "whole cell", additional large capacitive transients are now generated by the entire membrane of the cell.

To remove the transients, click the 'Auto' button. Then zoom in on the signal, and if needed, adjust the 'Rs' and 'Cm' controls separately or use the slider panel until the signal is adequately compensated. For a square pulse command, the goal is to eliminate the edge-effect transients with minimal distortion of the response signal.

The dPatch system is optimized for real-world measurements from real electrodes, so when used with the model cell, the compensation might need several more 'Auto' adjustments to compensate the model-cell capacitance.

Warning: Disable 'Cell Compensation' (and 'Rs Correction') when the

amplifier is in 'Capacitive' mode, or when running a Membrane Test in Cell mode, or results will not be valid.

Rs Correction Whole-cell Series Resistance compensation.

"Rs Correction" is used to correct command potential voltage drops, to minimize rise-time delays and slow decay phases in the current response, and to reduce unwanted filtering effects, caused by Series resistance.

Corr: [0-99%] Correction

[disabled in Capacitive mode when "Corr" and "Pred"

are displayed separately]

Pred [0-99%] Prediction

[displays if enabled in dPatch Settings menu]

Lag: $[1-100 \,\mu s]$ (RC filter component)

Lag = $1 / (2 * \pi * Bandwidth)$

Control the speed of the Correction while avoiding

possible oscillations.

Opens a 2-D slider panel for simultaneous tuning of

both parameters.

Warning: When dragging with a mouse, slow down

when approaching panel boundaries, else undershoot or overshoot of the values can

occur.

Auto Automatically sets an approximate value for

"Correction". "Lag" is a critical factor that should be

manually set by the user.

[disabled in Capacitive mode when "Corr" and "Pred" are displayed separately]

Rs Correction requires that the Electrode and Cell Compensations are first applied.

Set the Prediction (Pred) to "supercharge" the command potential. Small transients might be visible at the start and end of the current response.

Next, increase the Correction (Corr) current injected into the membrane to sharpen the rise time. As the Corr setting is increased, the current response transients also increase in size. Avoid overshooting, as if the correction is set too high, internal feedback can lead to oscillation of the circuit, i.e., "ringing", and loss of the patch.

Reduce oscillation of the circuit by adjusting the 'Lag' setting - larger values increase the stability of the circuit, but also increase the rise time.

Remove the Correction transients in the signal by reducing the Cell Compensation 'Rs' setting until a minimum value is found. Then adjust the Cell Compensation 'Rs' setting again until the best result is achieved (a smooth exponential rise), or try over again with lower Prediction/Correction settings.

Note: Disable 'Cell Compensation' and 'Rs Correction' when the amplifier is in 'Capacitive' mode, or when running a Membrane Test in Cell mode, or results will not be valid.

Current Clamp Controls




Figure 3-6. CC Control Panel

CC The 'CC' tab displays the amplifier Current Clamp controls:

I-holding: $[\pm 20,000 \text{ pA}, \pm 200,000 \text{ pA}]$

A holding current can be preset here.

The maximum range is set by the feedback 'Range' field below the VC and CC monitors.

Enable Use the Enable checkbox to activate I-holding.

If I-holding is not enabled, the holding level is zero amperes.

The holding level value is written at the start of Routine acquisition to the metadata as 'Command Holding Value'.

Whenever this button is activated or de-activated during acquisition, a 'Command Holding Value' tag is inserted into the data recording and written to the metadata.

V-Filter Low-pass 8-pole Bessel filter bandwidth.

[100, 200, 500 Hz; 1, 2, 5, 10, 20, 50, 100, 250, 500, 1000 kHz]

This low-pass filter is applied to the active headstage CC mode input

signals, and reduces the input sampling rate.

dPatch headstages have a fixed hardware sampling rate of 5 MHz. The actual system sampling rate is based on downsampling the hardware sampling rate, to the sampling value calculated by applying the Routine Editor / Input Channels Nyquist Factor to the selected filter Bandwidth.

When recording from a Routine, either the Control Panel V-Filter setting or the Routine Editor / Input Channels / Filter Bandwidth setting can be used. During acquisition, filter controls are locked, as they affect the sampling rate.

Tip: For experiments where the shape of the response is of interest, an input filter rate of 10 kHz is commonly used.

However, for very long stimuli, you might want to use a lower input filter rate.

Note: Filtering is not applied in Demo mode. Routines record data with the "Nyquist Factor-adjusted" sampling rate.

Electrode Compensation Electrode Capacitance Compensation

Magnitude: [0.00 - 20.00 pF] [shared with VC mode]

To remove microelectrode capacitance-charging transients and reduce their filtering effects (increased signal amplitudes and rise-times), click the 'Auto' button. Then autoscale the signal, and if needed, adjust the 'Magnitude' control.

CC mode Electrode Compensation is also known as "Capacitance Neutralization".

Tip: If you consistently lose cells when switching into CC mode, consider adjusting the CC mode electrode compensation value in Set Preferences / Hardware / Stability Control.

Bridge Balance Correction

Electrode Compensation must be enabled to access Bridge Balance Correction.

Enable Bridge Balance Correction to remove voltage-drop effects from the electrode Series resistance. This occurs when command currents flow into the preparation, and voltage readings from the cell during the current flow (injection) are corrected.

Note: If you are simply recording voltage (I=0) without any current injection, then Bridge Balance can be ignored or left disabled.

If the experiment is run purely in Current Clamp mode, leave the Bridge

Balance 'Percentage' at 100%, and work with the 'Magnitude' value.

• Percentage [0-200%]

• Magnitude: $[0.00 - 100.00 \text{ M}\Omega]$

[the field format is set in the dPatch Settings menu]

Note: For currents less than |10| nA, up to 100% of 100 M Ω resistance can be compensated without saturation.

To avoid saturation by larger currents, reduce the maximum compensation percentage (or the resistance) by an equivalent factor to the current increase over |10| nA.

To manually determine the Bridge Balance value, run the Membrane Test and zoom in on the initial rising phase. With Bridge Balance Correction disabled, there is a DC shift visible at the beginning of the rising signal. Enable Bridge Balance Correction and adjust until the DC shift disappears.

However, after adjustment, there might be a small glitch at the beginning of the rise, due to electrode and headstage capacitance which doesn't entirely go away. (Further adjustments to the Electrode Compensation may need to be made.) In some cases, it can be difficult to determine the exact Bridge Balance value, but as long as the Series resistance is significantly smaller than the cell resistance, the errors are very small.

Auto The Bridge Balance 'Auto' button can be used to approxi-

mate the correction value. For larger steps, it is recommended to run the 'Auto' function twice in a row.

Auto Bridge Ampl $[\pm 5,000.0 \text{ pA}]$

[only displays if enabled in the dPatch Settings]

Set the test pulse amplitude of the auto bridge

function.

However, as a superior method of calculation, start in VC mode and use Auto 'Cell Compensation'. When you first switch into CC mode, the VC mode 'Cell Compensation' Rs value is copied into the Bridge Balance Correction 'Percentage' field at 100% (or into the 'Magnitude' field with a default scaling factor of 1:1.)

Note: If the VC and CC values are not very similar (i.e., 'Percentage' near 100%), then something is not optimal.

While the (VC) Rs value and (CC) Bridge Balance value are the same electrical property, there might be slightly different electrical environments in voltage- vs. current-clamp modes, leading to slightly different adjustments of the Rs value in the two modes. The Bridge Balance setting for 'Percentage of Rs Compensation' allows for easy handling of this

difference. When the Bridge Balance 'Auto' button is clicked, the Magnitude is divided by the VC mode Rs value for a new recalculated scaling factor. Thereafter, switching from VC to CC mode will scale the Rs value with the updated factor for the new bridge balance value. Click the Auto button anytime to reset the scaling factor.

Dynamic Holding

Enable to maintain the membrane holding potential at a set target level, without drifting over time.

Note: Dynamic Holding automatically disables when Dynamic Clamp is active.

Target $[\pm 750 \text{ mV}]$ Target voltage.

Enter the voltage level to be maintained.

Tau [10 ms - 60 s]

This affects how fast the actual signal holding level is to be adjusted to the target holding setting. "Tau" is the time it takes to reduce the difference between the

actual voltage and the target voltage by 64%.

V-hold Set the target value from the amplifier V-holding

setting.

V-mon Set the target value to the cell's existing potential, as

seen in the headstage voltage monitor.

Enable during acquisition

Enable this checkbox to force Dynamic Holding to run

during acquisition.

By default, Dynamic Holding does not run during acquisition, to avoid interfering with amplifier settings and/or acquisition results, and then automatically reactivates once acquisition has finished.

Dynamic Clamp Controls

The dPatch system can be setup to perform full dynamic-clamp experiments. Configuration is done in the Dynamic Clamp Editor via the menu item SutterPatch / Hardware Control / dPatch Dynamic Clamp.




Figure 3-7. Dynamic Clamp Control Panel

DynC The 'DynC' tab contains the "dynamic clamp" controls.

Dynamic Clamp Active

Enable / disable execution of the Conductance loaded in the Dynamic Clamp Editor.

This control cannot be disabled during acquisition.

This feature is disabled when CC-Dynamic Holding is enabled.

Note: Dynamic clamping is not applied in demo mode.

Conductance Pool

The Dynamic Clamp Pool "Conductance" name.

Model

[#1-99] The model type.

Mode

The active acquisition mode.

- During Sweeps
- Run Continuously

Update Rate

[1, 2, 5, 10, 20, 50, 100, 200, 250, 500 kHz]

Voltage Source

- Headstage 1
- Headstage 2

Update Rate Error Conductance modeling error mode indicator. When active, re-

duce the update rate in the Dynamic Clamp Editor.

G-holding 1 $[\pm 10.00 \,\mu\text{S}]$

Conductance holding level for Stream 1.

Spinners increment in 1 nS steps.

G-holding 2 [$\pm 10.00 \,\mu\text{S}$]

Conductance holding level for Stream 2.

Spinners increment in 1 nS steps.

I/O Controls




Figure 3-8. I/O Control Panel

I/O The 'I/O' tab contains the digital output and auxiliary analog output controls.

Digital Output [1-16]

This section controls the holding bit pattern generated by the Digital Outputs of the amplifier. Sixteen TTL-compatible digital channels are displayed. Switch between digital states by clicking on a channel.

• On Red dot High (+3.3 V)

• Off Black dot Low (0 V)

All low Set all digital channels 'Off'.

All high Set all digital channels 'On'.

Auxiliary Output General purpose "auxiliary" analog output channels are available.

These raw outputs can be used as holding levels, or for other external

instrumentation.

AuxOUT1 - 4 [$\pm 10,000 \text{ mV}$]

Select an auxiliary analog output channel and directly edit its voltage level, or use the spinners to change the value in

ts voltage level, or use the spinners to change the value if

1 mV increments.

All to zero Set all auxiliary outputs to zero.

Note: If an auxiliary output (holding) level is changed during continuous

acquisition, the only system notification of this is a tag in the

metadata of the recording.

3.7 Lock-In Adjustments

The "lock-in" system is used to detect very small changes in membrane capacitance measurements, such as for single-vesicle fusion and retrieval measurements. Our system uses a digital implementation of phase-sensitive detection (PSD) to make its measurements.

1) To perform lock-in analysis, a reference waveform needs to be supplied to the headstage.

A suitable sine wave is generated via the Routine Editor / Waveform Editor / Sine / Sine Wave Cycles / For LockIn. (Also see Membrane Test.)

2) The lock-in response signal is processed via a virtual input channel. Measurements, such as membrane conductance, are reported. Calculations are made using 'conductance' instead of 'resistance'.

Setup in the Routine Editor / Input Channels / Virtual#:

Math Type LockIn

Current Channel Select a (source) input channel with a current (amperage)

signal.

Trace Kind Select the LockIn measurement to display.

The selected 'Trace Kind' is automatically set as the Virtual

Channel label.

CM Computed membrane capacitance.

GM Computed membrane conductance.

GS Computed series conductance.

DC DC component of measured signal.

Real number part of the lock-in response signal.

ImagY Imaginary number part of the lock-in response signal.

Cycles to Average [1-1000]

Cycles to Skip [1-1000]

V-reversal $[\pm 1000 \text{ mV}]$

When using a calculated stimulus trace, enter the reversal potential for the ion under study, such as for (Na+) sodium spikes or (K+) potassium tail currents.

3) The lock-in phase detection can be manually tuned via the menu item SutterPatch / Hardware Control / Lock-In Adjustments. Phase adjustments are made to optimize the head-stage signals, and should be done in voltage-clamp mode.

Enable Manual Adjustments Adjustments can be made using direct field edit-

ing, spinners, or a field right-click slider panel.

Phase Delay Adjustment $[\pm 1.00 \,\mu s]$

Apply a phase delay to the calculations.

Reset to '0.00 s'.

Attenuation Adjustment [0.001 - 9.999]

Apply a gain to the calculations.

Reset to '1.000'.

List Results Display results in the Command window.

Auto Adjust Automatically adjusts the phase delay so the lock-

in absolute capacitance is equal to the cell capacitance value in the Amplifier Control Panel. Requires the Membrane Test to be running a Lock-In

test pulse pattern.

The lock-in computation is quite stable - its calibration values do not change day-to-day. However, lock-in measurements can be affected by experimental conditions, including

the amplifier itself. In particular, the electrode compensation has a large influence on the results, and needs to be properly set - run the electrode compensation on a pulse, and then disable it, before running the LockIn adjustments.

When tuning, start with the Phase Delay Adjustment, and then follow with the Attenuation Adjustment.

The sensitivity of the SutterPatch software lock-in results is as good as for a hardware lock-in amplifier. However, for higher frequencies (5 kHz and higher), the lock-in sine frequency is increasingly sensitive to the lock-in phase adjustment. In such cases, you should tune-up the lock-in adjustments more frequently.

When making absolute capacitance measurements, you can improve the consistency of the measurements by adjusting the lock-in phase adjustment to a known reference capacitance, such as with the model cell.

The SutterPatch lock-in calculations are based on the Lindau/Neher method of time-resolved capacitance measurements in single cells. (See the Algorithms appendix for the algorithm and reference.)

Note: Demo mode is not designed to respond to lock-in phase and attenuation adjustments.

3.8 Dynamic Clamp Editor

The dPatch system can now perform dynamic-clamp experiments, an extension of the current-clamp technique, with your existing dPatch amplifier and SutterPatch software.

Access the Dynamic Clamp Editor via the menu item SutterPatch / Hardware Control / dPatch Dynamic Clamp.

This all-in-one dynamic-clamp implementation uses a voltage feedback from the cell to replicate the effects of simulated voltage-gated channels, ligand-gated channels, gap-junctions and synaptic inputs on living cells. A fast feedback loop monitors the voltage of the cell under study, and calculates the corresponding simulated conductance to apply to the cell on a point-by-point basis. The ultra-fast processing even simulates fast sodium conductances and complex Markov calculations, as real-time data is processed up to a 500 kHz filter rate. The simpler leak and linear conductance clamps are also fully supported.




Figure 3-9. Dynamic Clamp Editor

[Status Field]	Notifications on edits a	and file names are displayed here.
Files and Pools	[drop-down list]	
	Most recently used list of the last 5 Conductance Pool file names.	
	Load Conductance Pool	Load a previously saved Conductance Pool file into the Conductance Pool.
	New Default Con- ductance Pool	Create a new Conductance Pool with a default Conductance.
	New Conductance Pool	Create a new Conductance Pool, either with a default Conductance, or populated from the currently loaded Conductance Pool.
	Get Sample Conductance Pool	Load the sample dPatch Conductance Pool.

Revert to Last Saved	Undo any unsaved changes to the Conductance Pool
Save Conductance Pool	Save the Conductance Pool using its existing file name and path.
Save Conductance Pool As	Save the Conductance Pool to a new file, and switch to the new file. The default file name has an increment number appended to the original file name.
Save Conductance Pool Copy	Save the Conductance Pool to a new file, but do not switch to the new file. The default file name has 'Copy of' prepended to the original file name.
Merge Conductance Pools	Insert the Conductance Pool from a saved Conductance Pool file into the loaded Conductance Pool.
Send Last Used List to Command	Copy the pathname of the last used Conductance and paste it into the Command window history.

[Pathname of the loaded Conductance Pool file.]

Conductance Description

Conductance Pool

[A list of Conductance names in the loaded Conductance Pool.]

Multiple conductance equations can be defined per headstage. One equation can have multiple conductances; set a conductance to zero to remove it.



A checkmark displays next to the active (loaded) Conductance.

Load	Load the selected Conductance; a check mark displays next to its name.
[]	Read-only notifications display.
Save Pool	Save the Conductance Pool using its existing file name.
New	Add a default Conductance to the Conductance Pool and open it for editing. The default name is "untitled" with an increment number appended.
Duplicate	Add a copy of the selected Conductance to the Conductance Pool. The Conductance name-number is appended or incremented.
Delete	Remove the selected Conductance from the Conductance Pool.
Revert	Discard any unsaved changes to the selected Conductance.

Table 3-2. Conductance Pool

Headstage # tabs [1-2]

Apply Enable to apply the model and settings.

Copy from None, or from a list of all Conductance Pool entries, by headstage.

Model These models' equations are functions of voltage. The equations are checked upon activation in the Amplifier Control Panel 'DynC' tab.

Some examples from the sample Conductance Pool:

- Goldman-Hodgkin-Katz
- HCN
- Hodgkin-Huxley Simulates sodium conductance.
- KV
- Markov Model Computing-intensive state transitions.
- Predefined (Hodgkin-Huxley)

Used for testing.

TTX

Sodium channel model with TTX blocker.

• Variable Conductance

Active Mode

- During Sweeps
- Run Continuously

Update Rate

[1, 2, 5, 10, 20, 50, 100, 200 kHz]

[add for Markov models: 250, 500 kHz]

Control the dPatch internal update rate of the conductance modeling during acquisition. In general, the faster the rate, the more accurate the modeling. The modeling is done in an internal dPatch FPGA, before any filtering is applied to the input signals.

However, models with a large number of gates or states can overload the processing and generate errors if the update rate is too high. In this case, the Control Panel DynC tab 'Update Rate Error' displays an error condition, and the Update Rate should be reduced.

Voltage Signal Source

- Headstage 1
- Headstage 2

• AuxOUT3 [Headstage 1 tab]

• AuxOUT4 [Headstage 2 tab]

Command Signal to AuxOut1 [Headstage 1 tab] Enable extra output.

Command Signal to AuxOut2 [Headstage 1 tab] Enable extra output.

[Edit Model Parameters]

This button opens a 'Dynamic Clamp Parameters' window to configure channel gates, states and equations for the selected model.

(see below)

[not enabled for predefined models]

Table / Matrix Pane

Headstage # Equation(s) Display the Equation table for the selected Headstage #,

or

Headstage # Channel < ion > [for Markov models]

Display the Markov State Matrix for the selected Headstage #

and ion channel species.

Allow Editing Allow editing of the table/matrix.

围

Copy to Layout

Copy the table/matrix to a new Layout window, or append to an

existing Layout page.

Copy to Clipboard

Copy the table/matrix to the system clipboard.

Equation Table

Tip: A negative conductances can be used to remove unwanted conductances. However, if too large, it can "rail" and damage the cell. You are responsible for the correct usage.

Channel (1st column) The ion channel species.

Gate # (2nd column) 'Gate 1' is the first defined gate. If multiple gates

are defined, they are displayed in multiple 'Gate #'

columns.

If multiple ion species are defined, the gate is pre-

pended with a letter associated with an ion species

particle type:

m Sodium channel activating particles.

n Potassium channel activating particles.

h Sodium channel inactivating particles.

a "Alpha" voltage-dependent unidirectional forward

transition rate constant.

8 "Beta" voltage-dependent unidirectional backward

transition rate constant.

.....

τ Tau rate constant.

 ∞ Infinity equation.

State Matrix (Markov)

Entries are not allowed on the diagonal cells, which are filled with a padlock icon.

To use custom Markov model equations, enter a "Q" into the first cell [Row 0, Column 1]. This will call the external function "ReplaceQMatrix", which allows you to manually calculate the Q Matrix.

Contact Sutter Instrument technical support for additional details.

Edit Model Parameters Windows

Dynamic Clamp Hodgkin-Huxley Parameters

- ◀

Close the Dynamic Clamp Parameters Window.

Channel Settings

Channel Channel [1-8]

Active Enable copying and execution of the chan-

nel in the equation.

Label Enter a custom label for the channel, such

as the ion species (Na, K, Leak, etc.)

Copy Channel Copy channel parameters from an acti-

vated state.

 $V_{reversal}$ [±200 mV]

• G maximum $[\pm 10.00 \text{ nS}]$

Restrict to prevent cell damage.

- Use Conductance Stream 1
- Use Conductance Stream 2

Gate Equation as Alpha:Beta

Tau:Infinity

Gate Settings

Gate Gate [1-16]

Ion channel gate ID.

Active Enable execution of gate in the equation.

Label Enter a custom label for the gate.

Copy Gate Copy gate parameters from an activated

state.

Exponent [1-4]

Initial State [0, 1]

α Alpha rate constant equation.

Beta rate constant equation.

or

τ Tau equation.

 ∞ Infinity equation.

Dynamic Clamp Goldman-Hodgkin-Katz Parameters

- ◀

Close the Dynamic Clamp Parameters Window.

Channel Settings

Channel Channel [1-8]

Active Enable copying and execution of the chan-

nel in the equation.

Label Enter a custom label for the channel, such

as the ion species (Na, K, Leak, etc.)

Copy Channel Copy channel parameters from an acti-

vated state.

Pmax (cm/s) [0-100 kcm/s]

Valence $[\pm 10.00 \text{ nS}]$

IC Alpha:Beta

EC Tau:Infinity

Gate Settings

Gate Gate [1-16]

Ion channel gate ID.

Active Enable execution of gate in the equation.

Label Enter a custom label for the gate.

Copy Gate Copy gate parameters from an activated

state.

Exponent [1-4]

Initial State [0, 1]

α Alpha rate constant equation.

Beta rate constant equation.

Dynamic Clamp Markov Model Parameters



Close the Dynamic Clamp Parameters window.

Channel Settings

Channel Channel [1-8]

Note: The actual number of available

channels depends upon the number of states used and the memory size of their computed state coefficients. This limit is only checked when the Conductance is loaded. Typically, a 14×14 state model supports up to 4 channels, while a 10×10 (or less) state model supports all 8 channels.

Active Enable copying and execution of the chan-

nel in the equation.

Label Enter a custom label for the channel, such

as the ion species (K, Na, Leak, etc.)

Copy Channel	Copy channel parameters from an activated state.	
$V_{reversal}$	[±200 mV]	
Number of States	[2-14]	
State Equations		
ID	[S0-S15]	
Equation	[]	
Connections	[]	
Edit State Matrix	A modified State Matrix opens using state equation IDs for easy editing.	
•	Close connection matrix.	
[]	Equation ID read-only field.	
Cell editing actions		
Shift key	+ mouse click Set cell to currently active state equation.	
Shift key	+ arrow key Set next cell to currently active state equation.	
Option \	Option \setminus ALT key + mouse click Clear cell connection.	
Command \ Control key + mouse click Revert cell to original connection.		
Double cli	ck Edit state equation ID.	
Revert	Revert matrix to original connections.	
Conductance Equations (nS)		
ID	[G0-G3]	
Equation	[]	
Initial Probability	[]	

DPATCH - OPERATION MANUAL - REV. 2.2 (2021-10)

Dynamic Clamp Variable Conductance Parameters



Close the Dynamic Clamp Parameters Window.

Conductance Settings

V-reversal $[\pm 200 \text{ mV}]$

Enable Voltage Stream 2

V-reversal $[\pm 200 \text{ mV}]$

3.9 System Integration

The dPatch system can be integrated with other suitable laboratory equipment.

3.9.1 Using Peripheral Equipment

The dPatch system can control peripheral equipment, such as:

- Cameras
- Light sources
- Pulse generators
- Solution changers
- Wavelength switchers

Auxiliary analog output signals can be used to control other instruments within a range of ±10 V. Digital outputs use TTL-compatible voltage signals. Analog and digital holding levels are set in the Amplifier Control Panel, and provide a way to quickly and easily test the behavior and operation of peripherals, without the need to create or modify Routines.

The digital command output can be formatted as either a single "bit" or a 16-bit "word", as selected in the Routine Editor / Output Channels & Waveform section; the command output patterns are configured in the Waveform Editor.

Note: The analog and digital controls in the Control Panel provide a way to quickly and easily test the behavior and operation of peripherals, without the need to create or modify Routines.

3.9.2 Using Multiple Sutter Amplifiers

The dPatch amplifier is not supported for multiple-amplifier configurations.

3.9.3 Using Non-Sutter Amplifiers

The dPatch system can also be operated as a stand-alone data acquisition system interfacing to non-Sutter amplifiers. You continue to control the Sutter digitizer via the SutterPatch software Amplifier Control Panel.

The dPatch digitizer interfaces to external amplifiers via front panel BNC connections:

ANALOG OUTPUTS 1 & 2 These two auxiliary analog output channels can be used to send stimulus waveforms to external instruments, such as non-Sutter microelectrode amplifiers.

ANALOG INPUTS 1-4 These four auxiliary analog input channels can be used to digitize signals from external instruments, such as non-Sutter microelectrode amplifiers.

The dPatch digitizer can also interface to external instrumentation via additional rear-panel analog BNC connectors, and the Digital Outputs port optional 'dPatch Expansion Panel' BNC connectors or the included Screw Terminal Board.

ANALOG OUTPUTS 3 & 4 These two auxiliary analog output channels can be used to send stimulus waveforms to external instruments, such as non-Sutter microelectrode amplifiers.

ANALOG INPUTS 5-8 These four auxiliary analog input channels can be used to digitize signals from external instruments, such as non-Sutter microelectrode amplifiers.

DIGITAL OUTPUTS

Digital output patterns can be sent via sixteen TTLcompatible digital output channels to a variety of
peripheral equipment.

Auxiliary analog and digital holding levels are set in the Amplifier Control Panel I/O tab.

Note: Sutter amplifier output levels into Sutter systems attenuate by < 0.2%. HEKA amplifier output levels into Sutter systems attenuate by 0.5%. Axon Instruments amplifier output levels into Sutter systems attenuate by 5%.

3.9.4 Using Non-Sutter Data Acquisition Systems

Due to the ultra-high bandwidth of the system, the dPatch system does not operate as a stand-alone amplifier or interface to non-Sutter data acquisition systems.

3.10 dPatch Maintenance

This unit should require minimal maintenance when operated according to specifications.

3.10.1 Storage

Pre-May 2021 dPatch systems store firmware-related information in volatile memory, which requires an internal battery for backup power. To keep this battery sufficiently charged, "power on" the dPatch system for several hours at least every four years.

Note: Pre-May 2021 dPatch units will not power up with a fully discharged battery. In this case, request a replacement firmware board with non-volatile memory from Sutter Instrument.

Current dPatch systems use non-volatile memory, which does not use a battery.

3.10.2 Inspection

Periodically inspect all cables and connections to make sure that all cables are sound and that all connections are firm and evenly seated.



Warning! Turn off the dPatch power before plugging/unplugging headstages.

3.10.3 Cleaning

Routine cleaning of the dPatch system is required to prevent excessive dust accumulations. Wipe all exterior surfaces with a dry, soft, cotton cloth.

3.10.4 Calibration

The dPatch amplifier and headstages can be recalibrated, or reset to defaults, via the SutterPatch / Hardware Control / dPatch Maintenance menu. Calibrate with the model cell attached.

Please contact Sutter Instrument Technical Support before you make any changes here!




Figure 3-10. Calibration

Copy to Clipboard Copy the calibration window to the system clipboard.

Main Unit tab

	Gain	Offset (mV)
AuxOUT 1 – 4	0.500 - 1.500	±5000
AuxIN 1 - 8	0.500 - 1.500	±5000

Table 3-3. Main Unit Calibration

Serial No Serial Number of the dPatch system.

Part No Part Number of the dPatch system.

Note [Display of a text note.]

dPatch Firmware Version Firmware code version of the dPatch system.

SutterXOP Version External Operations code version of the dPatch system.

Set Defaults The dPatch system default firmware is available for re-

trieval.

Read from EEPROM Retrieve the system firmware settings from non-volatile

user memory (Electrically Erasable Programmable Read

Only Memory).

Write to EEPROM Store the system firmware settings in non-volatile

memory.

Tools

Calibrate Aux I/O Calibrate the Auxiliary Input/Output channels when a

dPatch Expansion Panel is connected.

Status

[] Status field of maintenance operations.

Headstage 1 & 2 tabs

Active Select a numbered Headstage tab to make it the active

headstage here, as well as in the Control Panel.

Voltage Clamp	Gain	Offset (pA)
200 nA Mode	0.500 - 1.500	±5000
20 nA Mode	0.500 - 1.500	±5000.0
Capacitive Mode	0.500 - 1.500	±500.00
Voltage Clamp	Gain	Offset (mV)
Stim DAC	0.500 - 1.500	±500.000
Current Clamp	Gain	Offset (mV)
200 nA Mode	0.500 - 1.500	±5000.000
20 nA Mode	0.500 - 1.500	±500.000

Table 3-4. Headstage Calibration

Serial No Serial Number of the dPatch headstage.

Headstage Part No Part Number of the dPatch headstage.

PreAmp Part No Part Number of the dPatch pre-amplifier.

Note	[Display of a text note.]	
Set Defaults	Set the dPatch headstage firmware to defaults.	
Read from EEPROM	Retrieve the headstage firmware settings from non-volutile memory.	
Write to EEPROM	Store the headstage firmware settings in non-volatile memory.	
<u>Tools</u>		
Calibrate Headstages		
<u>Status</u>	Status field of maintenance operations.	
[]		

4. SOFTWARE OPERATION

4.1 Acquisition

SutterPatch acquisition operations.

4.1.1 Acquisition Control

The interactive acquisition controls for Routines and Paradigms are grouped into this control panel.




Figure 4-1. Acquisition Control

Set Tag

Click the Set Tag button to create a time-stamped text comment in the Paradigm metadata at any time.

Enter the comment text into the field, or select a comment from a drop-down list of recently used entries. The drop-down list is saved with the SutterPatch preferences.

When run during acquisition, the comment tag is also written to the Routine. When the data is opened in a Reanalysis scope window, the tags are only visible in the 'Time Course' and 'Concatenated' display modes. Tags are also visible in the Data Navigator's Paradigm Review and Routine Review windows.

Note: "_none_" is a special case text entry: a tag is not generated.

Clear Menu

Erase the text comments from the drop-down list.

Cycle to Next

Cycle through the drop-down list of text comments each time the Set Tag button is clicked, starting from the displayed comment. When the last comment in the list is reached, it cycles back to the first comment in the list.

Note: In emulation mode, tag timing is not accurate, and tags are not set in the first sweep.

0

Timer A running clock displays the time in "hh:mm:ss" since the last

timer reset, or since a new experiment established a USB

connection or emulation mode.

Reset the Timer to 00:00:00.

Select Paradigm The active Paradigm is selected from the loaded 'Paradigm

Pool' list. See the Paradigm Editor to load a different

Paradigm Pool.

Start Paradigm Manually run the selected Paradigm - recording in the

Acquisition: Routine scope window is stopped, and a "planned"

(user named) Paradigm is created.

Stop Paradigm Terminate the current Paradigm.

If not followed by a planned Paradigm, starting a Routine creates a new date/time-stamped "auto-triggered" Paradigm.

Pause Pause a running Paradigm.

Resume running the active Paradigm.

Select Routine The active Routine is selected from the loaded 'Routine Pool'

list. See the Routine Editor to load a different Routine Pool.

Start Routine Manually run the selected Routine - start recording and displaying

data in the Acquisition: Routine scope window.

When you click the 'Start' button, the scope window is cleared, and data recording starts after ~300 ms. When acquisition is running, the scope window updates every 200 ms.

If the Sweep Start-to-Start time is ≥ 5 s, the "Time to next sweep: # s" is reported below the Start / Stop buttons.

If Metadata prompts are configured for Routines or Paradigms, the Confirm Metadata Settings dialog displays just before recording begins.

If measurement graphs are enabled, a docked "child" Analysis window opens and plots sweep-by-sweep measurements.

If no prior auto-triggered Paradigm is running, a new date/time-

stamped Paradigm is created.

Stop Acquisition Stop recording Routine data.

When you click the 'Stop Acquisition' button, data acquisition is

halted for the Series,

Stop at End of Sweep

When 'Stop at End of Sweep' is enabled, and you click the 'Stop Acquisition' button in the middle of a sweep, the partial sweep completes before data acquisition is halted; otherwise, when disabled, the partial sweep is discarded when acquisition stops.

4.1.2 Acquisition Measurements & Graphs

Make real-time changes to the online measurements and graphs, even during data acquisition, with this dialog. Edits instantly override the loaded Routine settings for fast responses.

To access this dialog, click on the Acquisition: Routine scope window button 'Measurements / Edit Measurements'.




Figure 4-2. Acquisition Measurements & Graphs

This dialog is the same as in the Routine Editor / Real Time Measurements & Graphs dialog, with two extra buttons:

- Update Routine Pool Save your edits to the Routine.
- Revert to Original Discard any edits.

4.1.3 Acquisition: Routine Scope

The Acquisition: Routine scope window is used for viewing and recording digitized time-series data, displayed as a smooth interpolated line.




Figure 4-3. Acquisition: Routine Scope Window

The scope window is titled with the active Paradigm sequence number + Routine name.

Note: Only one scope window can be open at a time. For example, if a scope window is open for data acquisition, then opening the Membrane Test, Free-Run, or data for reanalysis, will close the Acquisition: Routine scope window, and re-open it as the new type of scope window.

Controls


Signals

The central area of the scope window graphically displays input data in up to 16 separate signal panes. Click on a signal pane to make it "active" - the Y-ais border area displays in a lighter color, and the Y-axis controls (magnify, scroll) apply to it. Non-active panes display with a darker Y-axis border area.

If multiple signals are displayed stacked on top of each other, you can vertically resize the panes by clicking and dragging them with a resizing cursor. Position the mouse cursor over a pane separator (the horizontal area between panes) to change it to the resizing cursor (a horizontal line with a vertical double-headed arrow.)

Note: Two additional data points are appended to the sweep data to support post-sweep holding levels and segment boundary rounding issues.

Signal Layout



Signal pane arrangements are set by the 'Select signal layout' button in the lower left corner of the scope window. Select how the input signals are graphically arranged:

• Stack Vertical column of signals.

• Single Only the active signal.

• [m x n] Tiled array of signals in 'm' rows and 'n' columns.

Cursors

Measurement cursors are visible as light gray vertical bars in the signal panes. Each measurement region is bounded by a start-time cursor (the left edge) and an end-time cursor (the right edge).

To move a measurement region, click and drag it with the mouse - the region briefly turns dark when selected.

To resize a measurement region, click and drag an end-time cursor (the right edge of a region.)

Magnification and Scrolling

Signals can be magnified/unmagnified using several X- and Y-axis display controls in the scope window. Any magnification applied to the signals persists during acquisition.

Magnification Combo



Click on the "+" and "-" buttons to magnify / unmagnify by steps, or click and drag the slider to smoothly zoom / unzoom the active signal. The Y-axis magnification only controls the active pane; the X-axis magnification controls all signals.

■ Lock

Enable to retain the X-axis scaling and position for the next activated or created scope window (Acquisition, Analysis, Free Run, Membrane Test).

However, any changes to the X-axis duration (rescaling or autoscaling) or position (scrolling) disables the 'Lock' option.

■ Axis Zoom




Figure 4-5. Axis Magnification

When the mouse is moved into the X- or Y-axis area, the cursor changes to a double-headed arrow. As you click and drag the mouse cursor, a dark bar displays in the axis showing the magnification area; or, scroll the mouse wheel up/down to expand/shrink the X-axes or the active Y-axis.

■ Area Zoom:

Any region of interest in a signal pane can be graphically selected and expanded.

- 1. Move the mouse cursor into a signal pane it changes into a large "+".
 - 2. Click and drag a bounding box around the desired data. (This box is also referred to as a "marquee".)
- 3. Right-click in the marquee and select the desired action:

Expand Applies to all signals.

Horiz Expand Applies to all signals.


Vert Expand Applies to active signal.

Shrink Applies to all signals.

Horiz Shrink Applies to all signals.

Vert Shrink Applies to active signal.

Axis Scroll Bars



The X-axis scroll bar is directly underneath the X-axis, while the Y-axis scroll bar is on the far right-edge of the scope window. Click and drag the scroll bar slider buttons, or use their directional buttons to move the displayed signals in the desired direction. (The size of the X- and Y-axis scroll bar slider buttons reflects the amount of signal magnification.) The Y-axis scroll bar controls the active signal pane; the X-axis scroll bar controls all panes.

Center:





Figure 4-7. Center Button

Center the Y-range of the X-axis data in the active signal pane. The Y-axis offset is automatically adjusted, while the X-axis scaling is unchanged. To center all signals, shift-click the button.

Autoscale Axes



This autoscale setting is shared by the Acquisition, Membrane Test, and Free Run Scope windows.

Click to autoscale the Y-axes of all signals to their visible sweeps data limits, and to set the X-axis range to the maximum defined sweep duration for all signals.

To continuously autoscale the Y-axes of all visible signals during acquisition, in "Windows" right-click (or Shift-click or Ctrl-click) the button, or in "macOS" Control-click the button; this also sets the X-axis range to the maximum defined sweep duration for all signals. The Autoscale button remains depressed (grayed) in this state. However, continuous autoscaling is disabled by any changes to the scope window Y-axis scaling or offset.

To include the zero amplitude in the Y-ranges, enable "Include zero when autoscaling" in Set Preferences / Scope Window / General.



Y-Autoscale.

Click to autoscale the Y-axis of the selected signal to its visible sweeps data limits.

To autoscale the Y-axes of all signals, in "Windows" Shift-click the button, or in "macOS" Control-click the button.

To include the zero amplitude in the Y-range, enable "Include zero when autoscaling" in Set Preferences / Scope Window / General.

Tip: To invert the Y-axis of the active signal, such as for data with reversed polarity from an outside-out patch, right-click in the Y-axis of the signal and select Axis Properties / Axis Range. Either reverse the Manual Range Settings / Minimum and Maximum values, or disable the Manual Range and enable the Autoscale Settings / Reverse axis.

X-Scale



Autoscale the X-axis.

Set the X-axis range to the maximum defined sweep duration for all signals.

Amplitude Meters

Amplitude meters are displayed on the right border of signal panes (excluding virtual channels). They provide visual feedback on the integrity of your data recordings, similarly to how audiometers monitor audio signals.




Figure 4-9. Amplitude Meters

For the Triggered Sweeps acquisition mode, each displayed signal has its own Y-axis amplitude meter on the inner right-side of its pane.

For the Continuous Sweep acquisition mode, or if acquisition has not yet started, these meters are completely black.

The height of the colored meter bars represents a signal's data range vs. the full recording range of the dPatch digitizer. The color of the meter bar corresponds to the data "health":

•	Green:	Good	Signal within appropriate range.
			When the recorded data are within acceptable amplitude limits, the amplitude meter is green.
•	Yellow:	Caution	Signal approaching upper limit. (within 10% of range limit)
			If too much hardware gain is applied, and there is a danger that saturation will occur (as the data are near the upper limit), the amplitude meter is yellow, as a

warning sign to decrease your hardware gain.

• Red: Danger Signal too large.

(within 1% of, or at range limit)

When an amplitude meter is displayed in red, it indicates that the data might have gone out of range and be invalid.

If too much hardware gain is applied, the recorded signal will be in danger of saturating, i.e., your data will exceed the amplitude limit of the digitizer. If the data saturates, those data points are substituted with the maximum amplitude of their input channel.

Other Buttons

	Scope Settings		
344	Marks selectively flag sweeps for later reanalysis.		
\$4.5	Set all marks in sweeps of active series		
	Clear all marks in sweeps of active series		
	Begin with all marks set (in sweeps of active series)		
	Begin with all marks cleared (in sweeps of active series)		
Start	Start recording and displaying digitized analog data in the scope window input channels.		
	When you click the 'Start' button, the scope window is cleared, the Control Panel Offset is locked, and data recording starts after ~300 ms.		
	When acquisition is running, the scope window updates every 200 ms. If the Sweep Start-to-Start time is ≥ 5 s, the "Time to next sweep: # s" is reported below the Start / Stop buttons.		
	If Metadata prompts are configured for Routines or Paradigms, the Confirm Metadata Settings dialog displays just before recording begins.		

	If measurement graphs are enabled, a docked "child" Analysis window opens and plots sweep-by-sweep measurements.	
	If no prior paradigm is running, an "Auto-triggered Paradigm" is generated and assigned a Paradigm name with the current Date/Time.	
	Stop recording data immediately.	
Stop	If in the middle of a sweep, any partial sweep in progress is not saved with the data, except for Sweep 1.	
	If external triggering is configured, after clicking 'Stop', click the 'Do Trigger' button, and then 'Stop' again.	
	If the 'Stop at End of Sweep' checkbox is enabled, then the current sweep will complete before data acquisition is stopped, and the last recorded sweep will be a complete sweep of data.	
Stop at End of Sweep	The message 'Waiting to stop' displays below the Stop / Start buttons, until the last sweep completes and acquisition stops. If no sweep is in progress, acquisition stops at the end of the next sweep to be recorded.	
	Show Cursors:	
Measurements		
	Display measurement cursors in the scope window.	
	Hide Cursors:	
	Do not display cursors in the scope window.	
	Button displays as "Measurements(H)".	
	Lock Cursors:	
	Prevent cursors from being adjusted or moved.	
	Button displays as "Measurements(L)".	
	No Measurements or Graphs	
	Analyze with Active Measurements	
	Analyze with Original Routine Measurements	
	Analyze with Routine Last Executed Measurements	
	Analyze with Saved Default Measurements	
	Save as Default Measurements	
	Edit Measurements: Open a special Reanalysis Measurements & Graphs dialog	

	where all changes apply instantly to the measurements and the graphs, even during acquisition. These edits override the loaded routine for quick interactive control.	
Wipe	The 'Wipe Scope' button clears the scope of all sweeps prior to the active sweep, and also clears any corresponding measurements from the Analysis window.	
Mark	Enable/disable to "mark/unmark" the current (or upcoming) sweep. This is useful for quality control during slow acquisition of signals.	
	See the Data Navigator 'Available Actions' or use the Reanalysis scope window to analyze or process marked sweeps.	
Do Trigger	This green button appears when acquisition is started for a Routine configured with an external trigger, and provides a manual trigger option. The message "Waiting for trigger" also displays.	
Copy to Layout	Copy all visible scope signals and analyses into a new Layout window, or append to an existing Layout page.	
Copy to Clipboard	Copy the active signal to the system clipboard or, if the 'Shift' key is pressed, the complete scope window.	
Persistence Display	Enable: All acquired sweeps of the Series are displayed (per Marks and/or Scope Preferences).	
	Disable: Previous sweeps are cleared, and an incoming data point for the new sweep replaces the time-corresponding data point of the prior sweep.	
	Applies to the Scope window and its (right-click) Parametric Plot and Amplitude Histogram Plot graphs.	
Sweep#of#	The active sweep number vs. the total number of configured sweeps is reported. If multiple cycles are set, the active sweep cycle number is inserted between them.	

Table 4-1. Other Scope Buttons

Right-click Menus

X Axis

Autoscale All Axes Scale all signals Y-axes to their data, and set the

X-axis for all signals to the maximum defined

sweep duration.

Autoscale X Axis Set the X-axis for all signals to the maximum de-

fined sweep duration.

Set X Scale...

X-min

X-max

Axis Properties... Modify the axes style and components.

Y Axis

Autoscale All Axes Scale all signals Y-axes to their data, and set the

X-axis for all signals to the maximum defined

sweep duration.

Continuous Autoscale Y Axis

Continuously scale the signal's Y-axis to its data.

Autoscale Y Axis Scale the signal's Y-axis to its data.

Full scale Y Axis Set the signal's Y-axis to its full-scale range.

Use Last Y Scale

Set Y Scale...

Y-min

Y-max

Axis Properties... Modify the axis style and components.

Hide Signal <name> Hide the selected signal in the scope window.

Show Signal <name> Only Show the selected signal in the scope window, hide

all other signals.

Stack All Signals Display all signals in a single column.

Main Window

Limited data modification menu

Right-click the blank area in a signal pane.

If you click too close to the signal data, the full data modification menu Tip: displays instead; if this occurs, click near a horizontal or vertical edge of

the signal pane.

Autoscale All Axes Scale all signals Y-axes to their data, and set all

the X-axis for all signals to the maximum defined

sweep duration.

Add Annotation Add a floating text-box label to the signal pane.

To edit or delete an annotation, double-click on it.

Parametric Plot Display a graph of X vs. Y input signals in a sepa-

rate window.

Y-signal Select an input signal for the Y-axis.

X-signal Select an input signal for the X-axis.

Plot Update the plot using the new parameters.

Copy to Layout Copy the Parametric Plot graph into a new Layout

window, or append to an existing Layout page.



Copy to Clipboard

Copy the Parametric Plot graph to the system clip-

board.

Time Range The time range of the data to be plotted.

> **Full Trace** Use the entire trace for the time range.

> Sweep Time Set relative to the start time of a sweep

> > (time zero).

Start Time Set the starting time.

> Once the Start Time is within 2 sample points of the End Time, further Start Time increments will increase the End Time by the same

amount.

End Time Set the ending time.

Once the End Time is within 2 ms of the Start Time, the End Time

cannot be decremented.

Segment Time Set the time range as a ratio of the Seg-

ment duration.

Segment Select the Segment number.

Start Ratio [0 = beginning of Segment]

Set the starting time ratio.

End Ratio [1 = end of Segment]

Set the ending time ratio.

Amplitude Histogram Plot Open a real-time histogram plot window. The am-

plitude data is plotted in "real time" as samples are acquired and binned. The window is cleared

at the start of a new Series.

Y-signal Select the input signal to be analyzed.

Time Range The time range of the data to be plotted.

Full Trace Use the entire trace for the time range.

Sweep Time Set relative to the start time of a sweep

(time zero).

Start Time Set the starting time.

End Time Set the ending time.

Segment Time Set the time range as a ratio relative to the

Segment duration.

Segment Select the Segment number.

Start Ratio [0 = beginning of Segment]

Set the starting time ratio.

End Ratio [1 = end of Segment]

Set the ending time ratio.

Histogram Bins [50, 100, 200, 500, 1000, 2000, 4000]

Select the number of bins for the amplitude range

(X-axis). Changes instantly update the plot.

Plot Refresh the plot for any Time Range / Sweep Time

settings changes.

F

Copy to Layout Copy the Amplitude Histogram Plot graph into a

new Layout window, or append to an existing Lay-

out page.

Copy to Clipboard

Copy the Amplitude Histogram Plot graph to the

system clipboard.

[graph pane]

Export Graphics Copy the signal and open in a separate window.

Saves to Windows / Graph Macros.

Toggle Cursor Info Show/Hide the Cursor Info pane to measure X-Y

data points or set a fitting range. (See the 'Signal

data' section below.)

Select 'Toggle Cursor Info' again to hide the Cursor Info pane, and any cursor symbols in the active

pane.

Colors Adjust the colors used by the active signal pane:

graph background The background of the pane.

all axes The X- and Y-axis areas.

all grids The grid lines in the pane.

all tick labels The tick labels in the X- and Y-axis areas.

all axis labels The axis labels in the X- and Y-axis areas.

Toggle Cursor Info Show the Cursor Info pane to measure X-Y data

points or set a fitting range. (See the 'Signal data'

section below.)

Colors

Hide Signal '<name>' Hide the selected signal in the scope window.

Show Signal '<name>' Only Show the selected signal in the scope window, and

hide all other signals.

Stack All Signals Display all signals in a stacked signal layout.

Marquee

Click and drag the mouse to surround a region of interest, and right-click for a context menu:

Expand Set the signal's Y-axis range from the marquee

vertical data limits, and set all signals X-axes ranges from the marquee horizontal data limits.

Horiz Expand Set all signals X-axes ranges from the marquee

horizontal data limits.

Vert Expand Set the signal's Y-axis range from the marquee

vertical data limits.

Shrink Move the signal's Y-axis current limits to the posi-

tion of the marquee vertical data limits, and move all signals X-axes current limits to the position of

the marquee horizontal data limits.

Horiz Shrink Move all signals X-axes current limits to the posi-

tion of the marquee horizontal data limits.

Vert Shrink Move the signal's Y-axis current limits to the posi-

tion of the marquee vertical data limits.

Extract Template Copy the last sweep to the Template Editor.

Extract To Graph Display the first trace in a floating window, using

all data within the X-range.

Signal Data

Full data modification menu

Right-click on or near the data to display this context menu, which includes options to modify sweeps and data points, such as the marker symbols and lines.

Tip: To manually measure X-Y data values, or to set a fitting range, open a Cursor Info pane.

Toggle Cursor Info Select 'Toggle Cursor Info' to show/hide the Cursor

Info pane, and any cursor symbols in the active

pane.

Cursor Info pane



'Options' menu

One Mover Moves All

e Mover Moves All Draggable cursor mover tool

moves all cursors together with a single control.

• All Styles Change the cursor symbol

style.

• Show Cursor Pairs Display up to 5 sets of cur-

sor symbol pairs.

Cursor A

Cursor symbol for data point 'A'.

A: Symbol letter (beginning cursor of the pair).

R1_A_IV (Default) wave name.

Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'A'.

Y: Y-axis value of data point 'A'.

 ΔY Difference of the cursor pair Y values.

Cursor B

Cursor symbol for data point 'B'.

B: Symbol letter (ending cursor of the pair).

R1_A_IV (Default) wave name.

Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'B'.

Y: Y-axis value of data point 'B'.

ΛX Difference of the cursor pair X values.

Cursor Instructions

1. Click on symbol 'A' to enable it.

- 2. Manually drag the highlighted symbol onto a data point in a signal pane, or enter the data point number in the 'pnt' field.
- 3. Click on symbol 'B' to enable it.
- 4. Manually drag the highlighted symbol onto a data point in the signal pane, or enter the data point number in the 'pnt' field. X- and Y-measurements are displayed for the cursor pair data points.
- 5. Fitting can also be applied to the cursor pair data. Right-click on the data, and select 'Quick Fit' for a list of built-in Igor fitting functions.

The fit is displayed in the graph, and the fitting information is written to the Command window.

4.1.4 Camera Control

The Camera Control window displays still pictures or live video.

Camera Select a camera name from those attached to the computer.

Start View live video. This data is not stored.

Capture Take a single picture. If live video is running, this will take a

picture while live video continues to run. The image time-

stamp is reported in the Log window.

Last Capture A thumbnail of the last picture taken in the Experiment is

displayed.

All pictures are stored in the current Experiment. To review

pictures, go to:

Data Navigator Select a Paradigm or Routine. Any

associated images are listed in the

Preview pane. Click on an image name

to display the image.

Data Browser Navigate to the Data / Images folder.

Right-click on an image name, and select

'New Image'.

[video screen]

Tip: For dark-room experiments, the window background color can be adjusted by the operating system.

Windows: In the Windows Control Panel / Appearance / Personalization window,

scroll down and select the High Contrast Black theme, or use the Magni-

fier tool with option 'Turn on color inversion' enabled.

macOS: Press 'Control-Option-Command-8' to set the System Preferences / Ac-

cessibility / Display / Invert Display colors option, or open its menu with

'Command-Option-5'.

Full-camera drivers have been successfully tested for the following camera models:

Sentech drivers:

STC-MC33USBVGA

STC-MCS231U3V

STC-MB83USBVGA

STC-MBCM401U3V

STC-MBCM401U3V-NIR

STC-HD203DV

Photometrics PVCAM drivers:

Photometrics: Prime 95B

Prime 95B 25mm

Qimaging: Electro

4.1.5 Free Run

The Acquisition: Free Run scope window simulates a one- or two-channel oscilloscope, and is a quick method of viewing repetitive data.

This window operates similarly to the Acquisition: Routine scope window, with unsupported controls removed or disabled. However, when this scope window is initially created, the Autoscale button is set to the last used state, instead of using a Preferences setting.

Additional controls were also added:



The 'Store to disk' button is green when enabled, and red when disabled.

When enabled, Free Run records streaming data as sweeps into the current Experiment, and the scope window displays a sweep counter with the number of completed sweeps vs. the maximum number of available sweeps.

Note: The button also displays "H5" when run in an HDF5 Experiment.



Move (dock) the Free Run scope and settings windows next to the Amplifier Control Panel.



Show / Hide the Free Run settings sub-window.

Note: Copying to the clipboard can temporarily pause the Free Run display for several seconds, and then the display catches up to the actual acquisition.

[0.0 s]

The duration of the last (unstored)

Free Run Settings Window

Acquired /

4	[0.0 0]	digitized data. /
Stored	$[0.0 \mathrm{\ s}]$	The duration of the last (stored)
		digitized data.
Maximum Storage Duration /	[3.28.20 h]	The maximum amount of data storage time in one recording, using the current bandwidth and number of channels. This briefly displays when starting acquisition with data storage enabled.
Maximum Remaining	[3.28.20 h	3.28.19 h] The maximum / remaining amounts of data storage time in one recording, using the current band- width and number of channels. This displays during acquisition while storing data to disk.

Signal Parameters

First Channel Select an input channel to monitor:

Active Headstage

Headstage1 Headstage2 AuxIN1 - 8 Bandwidth

[Headstage channels]

[I-Filter, V-Filter] Set by the Amplifier Control

Panel filter.

This low-pass filter is applied to a headstage input signal sampling rate.

During acquisition, filter controls are locked, as they affect the sampling rate.

Note: Filtering is not applied in Demo mode.

Demo data is acquired with a "Nyquist Factor-adjusted" sampling rate.

Sampling Rate

[Auxiliary channels]

[1, 2, 5, 10, 20, 50, 100, 200 kHz]

Set the auxiliary input channel sampling rate.

Nyquist Factor

[Headstage channels]

To manage the quantity of data being processed, the Nyquist Factor downsamples the hardware input sampling rate in relation to the filter Bandwidth.

- 2 Set a Nyquist-limited minimum sampling rate of twice the filter Bandwidth.
- 5 Set the sampling rate to five times the filter Bandwidth.
- 10 Set the sampling rate to ten times the filter Bandwidth.

[1 MHz Bandwidth]

- 2.5 Set the sampling rate to 2.5 times the filter Bandwidth.
- 5 Use the highest hardware sampling rate of 5 MHz; no downsampling of data occurs.

Warning! Processing very large data sets can overload system resources.

Integrator Reset [Headstage: Capacitive Mode]

When using the amplifier's Capacitive mode, choose how to handle undesirable data transients generated by capacitor discharges (resets).

Ignore Capacitive mode reset transients

are displayed in the data (unless the Control Panel 'Settings' is enabled for 'Subtract capacitor reset

transients'.)

Blank The data during capacitive

transients are made invisible by replacing those data points with

NaNs (Not a Number).

Mask The data during capacitive

transients are replaced by the last data value before the capacitor discharge, simulating a sample-

and-hold operation.

Blank/Mask Duration [Headstage: Capacitive Mode]

 $[50.00 \, \mu s - 1.00 \, s]$

An "integrator reset" transient should be completely overwritten by the default duration

of 500 µs. Otherwise, adjust as needed.

Second Channel Select an additional input channel to monitor:

Active Headstage

Headstage1 Headstage2 AuxIN1 - 8

Bandwidth [Headstage channels]

 $[\hbox{ I-Filter, V-Filter }] \hspace{0.5cm} \hbox{Set by the Amplifier Control}$

Panel filter.

This low-pass filter is applied to a headstage

input signal sampling rate.

During acquisition, filter controls are locked, as they affect the sampling rate.

Note: Filtering is not applied in Demo mode.

Demo data is acquired with a "Nyquist

Factor-adjusted" sampling rate.

Sampling Rate

[Auxiliary channels]

[1, 2, 5, 10, 20, 50, 100, 200 kHz]

Set the auxiliary input channel sampling rate.

Nyquist Factor

[Headstage channels]

To manage the quantity of data being processed, the Nyquist Factor downsamples the hardware input sampling rate in relation to the filter Bandwidth.

- 2 Set a Nyquist-limited minimum sampling rate of twice the filter Bandwidth.
- 5 Set the sampling rate to five times the filter Bandwidth.
- 10 Set the sampling rate to ten times the filter Bandwidth.

[1 MHz Bandwidth]

- 2.5 Set the sampling rate to 2.5 times the filter Bandwidth.
- 5 Use the highest hardware input sampling rate of 5 MHz; no downsampling of data occurs.

Warning!

Processing very large data sets can overload system resources.

Integrator Reset

[Headstage: Capacitive Mode]

When running the amplifier in Capacitive mode, choose how to handle undesirable transients in the data generated by capacitor discharges (resets).

Ignore Capacitiv

Capacitive mode reset transients are displayed in the data (unless the Control Panel 'Settings' is enabled for 'Subtract capacitor reset

transients'.

Blank The data during capacitive

transients are made invisible by replacing those data points with

NaNs (Not a Number).

Mask

The data during capacitive transients are replaced by the last data value before the capacitor discharge, simulating a sampleand-hold operation.

Blank/Mask Duration

[Headstage: Capacitive Mode]

 $[50.00 \, \mu s - 1.00 \, s]$

An "integrator reset" transient should be completely overwritten by the default value of

500 μs. Otherwise, adjust as needed.

Scope Duration [0.2, 0.5, 1.0, 2.0, 5.0, 10.0, 20.0, 50.0, 100.0 s; Other:

100.00 ms - 100.00 s

The duration of the data sweep X-axis.

4.1.6 Membrane Test

The Membrane Test is primarily used to monitor seal formation and cell health in a voltage-clamp whole-cell patch-clamp configuration.

However, current-clamp mode operation is also supported, if 'Electrode Compensation' (capacitance neutralization) is perfectly set.

Acquisition: Membrane Test

This scope window operates similarly to the Acquisition: Routine scope window. By default, this scope window's top pane displays the Current signal from the active Sutter headstage, and the pane beneath it displays the corresponding Voltage command signal.

Differences to the scope window for the Membrane Test:

- Unsupported buttons were removed: Persistence Display, Signal Layout.
- The Sweeps Counter only displays when the Membrane Test is run from a Paradigm with numbered "Repeats" configured.
- The first time the Membrane Test is opened, it is in a "stopped" state to allow setup and configuration. Thereafter, re-opening the Membrane Test automatically uses the last acquisition status of its scope window.
- During an Experiment, when the scope window is created, the Autoscale button is set to the last used state, instead of using a Preferences setting.
- There is a limited right-click menu in the channel panes.

Additional controls were also added:



Move (dock) the Membrane Test windows next to the Amplifier Control Panel.



Show/Hide the Membrane Test Settings panel.

Note: If the Membrane Test is run with the amplifier in capacitive mode, be sure to monitor the membrane resistance. A lower resistance results in higher signal currents, which in turn generates more capacitive resets by the headstage. These resets are visible in the raw data, and as a blinking button in the amplifier control panel. If the frequency of resets is too high, you might need to switch the headstage into a resistor mode.

Membrane Test Settings

Configure all Membrane Test parameters in the Membrane Test Settings panel.




Figure 4-10. Membrane Test Settings

Show Monitor Only

Monitor + Test Pulse

All Settings

Open the Membrane Test 'Preferences' settings dialog.

Membrane Test Preferences

Second Membrane Test channel

Display the controls for a second input channel in the Signal Parameters section.

MT analysis on second channel

Display the controls for a second analysis channel in the Measure Parameters section.

Analysis results always visible

Keep the real-time analysis numeric fields visible for all selected Signal Parameters channels, even if the channel is disabled in Measure Parameters.

ZAP controls Display the Zap Parameters controls section in the Set-

tings panel.

Sound controls Display the sound controls in the Measure Parameters

section.

Line Frequency Reduction options

Display Line Frequency (LFR) signal controls in the Signal Parameters section, and set the power line frequency.

A/C Line Frequency • 50 Hz

• 60 Hz

Enable Holding with Seal mode selection

When the Membrane Test 'Seal' mode is selected, the control panel 'V-holding' or 'I-holding' level is also enabled.

Running Average Trace Set display options for the "running" average of the last

"N" traces, which updates with each sweep, and is set in

'Measure Parameters'.

Line Thickness [0.25, 0.50, 0.75, 1-10]

Color Open a color selection panel.

Average Traces Set display options for the last "N" traces used in the run-

ning average trace. These only display when the Measure

Parameters 'Display' is set to 'All'.

Line Thickness [0.25, 0.50, 0.75, 1-10]

Color Open a color selection panel.

Capture Trace Set the trace type to persist in the scope display, as se-

lected by the "push pin" snapshot button in the Signal Parameters section, and set its trace display options.

Source

• Signal The Membrane Test active data

trace.

• Average The Membrane Test running aver-

age trace.

Line Thickness [0.25, 0.50, 0.75, 1-10]

Color Open a color selection panel.

The Membrane Test settings panel contains the three basic steps to patch-clamp a cell:

1) Bath With a new pipette in the bath solution, a low-resistance square pulse is

visible. The pipette resistance should be very low if the tip is not clogged.

For whole-cell patch clamping of dissociated cells, typical pipette resistances are $1-5~M\Omega.$ For brain slice recordings, pipette resistances up

to 20 M Ω or higher are used.

2) Seal When an on-cell patch is formed between the pipette and the cell, voltage

transition spikes are visible. The seal resistance increases as the seal forms. The goal is to achieve a "gigaseal" with a resistance above 1 $G\Omega$.

This configuration is used for single-channel recordings, as well as in-

side-out and outside-out patches.

3) Cell After breaking through the cell membrane and creating a whole-cell patch, membrane resistance and capacitance measures are calculated

from the resulting capacitance spikes. These values can be periodically

checked to monitor the health of the cell.

The Membrane Test calculations are displayed in real-time numeric fields:

Bath Rpipette $(M\Omega)$ Pipette Resistance meter.

[Model cell = $\sim 10 \text{ M}\Omega$]

Seal Resistance meter. Seal Resistance meter.

[Model cell = $\sim 1 \text{ G}\Omega$ to $1 \text{ T}\Omega$]

(open circuit)

Cell Rseries (M Ω) Series Resistance meter.

[Model cell = $\sim 10 \text{ M}\Omega$]

Rmembrane ($M\Omega$) Membrane Resistance meter.

[Model cell = $\sim 500 \text{ M}\Omega$]

Cmembrane (pF) Membrane Capacitance meter.

[Model cell = \sim 28 pF]

These values should be periodically checked to monitor the health of the cell. In particular, monitoring the Series Resistance is helpful, as if it increases by more than 5%, the electrode tip might be clogged.

Warning! When in Cell mode, disable 'Cell Compensation' and 'Rs Correction' (in the Amplifier Control Panel), so that Rseries and Cmembrane values can be properly calculated using uncompensated capacitance spikes. Otherwise, these fields are reported as "NaN".

Tip: 'Series Resistance' and 'Access Resistance' are equivalent terms. It is helpful to periodically monitor the health of the cell via the Series Resistance, as if it increases by more than 5%, the electrode tip might be clogged.

Note: Demo values for Cell mode can vary from the model cell values.

Reported values are dependent upon experimental variables and settings, such as cell and pipette size, solution conductivity, test pulse duration, etc.

Write To Log Click the 'Write To Log' button to write the last acquired measurements for that mode to the Log window. Valid measurements are logged for the active headstage.

[] Field displays advice and warning messages to the user.

<u>Test Pulse Parameters</u>

Pulse Type

• Single Pulse A single monopolar square pulse.

• Double Pulse A symmetrical bipolar (biphasic) square pulse.

• Triangle A train of 5 symmetrical bipolar triangular

pulses.

• Sine A train of 5 sine wave pulses.

• RMS Noise No pulse – the holding level is output. The

RMS noise of the signal is measured.

In capacitive mode recording, $500~\mu s$ of masking is automatically applied to the capacitor reset transients in the data.

Lock-in Measurements are based on a phase analy-

sis of a software lock-in amplifier.

When running the Membrane Test to measure Series Resistance and

Membrane Capacitance, the whole-cell capacitance compensation is usually disabled. However, when using Pulse

Type 'Lock-in', frequencies up to 2 kHz, can be enabled. At higher lock-in frequencies, we recommend disabling Cell Compensation, else the lock-in phase delay is affected and requires

manual adjustment.

• Lock-in 2.0 ms (500 Hz)

• Lock-in 1.0 ms (1 kHz)

• Lock-in 0.5 ms (2 kHz)

• Lock-in 0.2.ms (5 kHz)

• Lock-in 0.1 ms (10 kHz)

• Lock-in 0.048 ms (20.8 kHz, OnCell only)

• Lock-in: OnCell mode

Amplitude

VC mode: [Other, -50, -20, -10, -5, -2, -1, 0.1, 1, 2, 5, 10, 20,

50 mV]

Other: $[\pm 750.00 \text{ mV}]$ Manually entered value.

A pulse amplitude is required. Any value less than |0.1| mV (absolute) is reset to

±0.10 mV.

CC mode: [Other, -2000, -1000, -500, -200, -100, -50, -20, -10,

0.1, 10, 20, 50, 100, 200, 500, 1000, 2000 pA]

Other: [±2000.00 pA] Manually entered value.

A pulse amplitude is required. Any value less than |0.1| pA (absolute) is reset to ± 0.10 pA.

Amplitude is relative to the 'Holding' level in the Amplifier Con-

trol Panel.

Duration [Other, 0.5, 1, 2, 5, 10, 20, 50, 100, 200, 500 ms]

Other: Manually entered value.

Repetition Interval [Shortest Possible, 0.2, 0.5, 1.0, 2.0, 5.0, 10 s]

O Oval marker color toggles at the start of each sweep, to or from

green.

Zap Parameters

(This feature is only active in voltage-clamp mode.)

After a gigaohm patch has been achieved, use Zap in the Seal mode to disrupt the cell membrane, as an alternative to suction in creating a whole-cell patch.

Amplitude Set the amplitude of the square wave zap pulse.

[±750.00 mV]

Duration [0.040 - 10000.00 ms]

Zap • 1st Channel

• 2nd Channel

Both

Do It Click the 'Do It' button to send a single square wave voltage pulse

from the dPatch headstage to the preparation.

Signal Parameters

First Channel • Active Headstage The headstage that is selected or highlighted.

Headstage1

• Headstage2

Take a snapshot of the next incoming trace and "pin" it as a static "captured" trace in the scope window background. Use the captured trace to compare signals before and after compensations are applied.

or

Click the "Cut" button to clear the captured trace from the dis-

play.

LFR Off Line Frequency Reduction Off.

Replace Signal

Display the signal with LFR On.

Add Signal

Add a channel for the LFR signal.

Display Stimulus Add a channel to the scope window for the stimulus signal.

Second Channel

No Channel

Inactive Headstage

The headstage that is not selected or high-

lighted.

Bandwidth VC mode: Set by I-Filter (x.x kHz)

CC mode:

Set by V-Filter (x.x kHz)

Headstage1 - 2

Bandwidth

VC mode:

Set by I-Filter (x.x kHz)

CC mode:

Set by V-Filter (x.x kHz)

AuxIN1 - 8

Sampling Rate

[1, 2, 5, 10, 20, 50, 100, 200 kHz]



Take a snapshot of the next incoming trace and "pin" it as a static "captured" trace in the scope window background. Use the captured trace to compare signals before and after compensations are applied.

or



Click the "Cut" button to clear the captured trace from the dis-

play.

Bandwidth

[Set by I-Filter (#kHz)]

LFR

Off Line Frequency Reduction Off.

Replace Signal Display the signal with LFR On.

Add Signal

Add a channel for the LFR signal.

Display Stimulus

Add a channel to the scope window for the stimulus signal.

Measure Parameters

Measurements From This option only displays when the Membrane Test Preferences for 'Second Membrane Test channel' / 'MT analysis on second channel' is enabled.

- First Channel
- Second Channel
- **Both Channels**

Num to Average

[1, 2, 5, 10]

Number of traces to average.

Measurements are derived from the last trace average(except for Lock-in and Noise measurements). The averaged trace displays in red. The average is reset when the main mode is changed (Bath - Seal - Cell).

Display

Scope window signal display options.

•	Signal	The active signal data trac	ce is disr	olaved.

The running average data trace is dis-Average

played.

Signal + Avg The active signal trace is displayed on top

of the running average trace.

Avg + Signal The running average trace is displayed on

top of the active signal trace.

All The active signal trace is displayed on top

of the running average trace and its com-

ponent sub-traces.

Resistance Monitor

Enable / disable the Membrane Test resistance monitor, which beeps the computer speaker with a variable pitch at the beginning of each Membrane Test sweep.

- Audio Off
- First Channel
- Second Channel

Hz/MΩ Factor

[0.01 - 10000.00]

Hertz / resistance factor to control the pitch of the sound monitor beep in relation to the membrane resistance.

Volume [1.00 – 100.00 %]

Computer speaker volume.

4.1.7 Paradigm Editor

The Paradigm Editor is an advanced feature that opens up a world of complex experimental control via Paradigms and Paradigm Pools. A rich set of operators and actions are available to control and/or automate data acquisition and analysis.

The Paradigm Editor allows you to create "Planned Paradigms", which offer almost unlimited flexibility in creating and/or automating your patch-clamp experiments, such as running Routines and directly controlling amplifier settings.




Figure 4-11. Paradigm Editor

Loaded Paradigms display on the left, while loaded Paradigm Steps display on the right. A bottom section can display interactive checkbox controls and/or variables.

Controls

Start/Stop Paradigm, Set Tag and Reset Timer controls are located in the Acquisition Control panel.

Current Paradigm: The name of the currently loaded Paradigm.

Step: The highlighted Paradigm Step.

Paradigm Status Status information about Paradigm execution.

Show/Hide Paradigm Editor Controls

The Paradigm Editor controls (and checkboxes) for the Paradigm Pool and Paradigm Steps can be displayed or

hidden.

Show/Hide User Checkboxes

Checkbox controls are displayed at the bottom of the Paradigm Editor controls, for use in conditional Paradigm step execution. This display is dependent upon Show Editor

Controls.

Show/Hide Variables: A Variables table can be displayed at the bottom of the

Paradigm Editor. These paradigm variables can be utilized in

any equation.

Variable names can be edited to any label, but they are only

informational, and are not supported in equations.

Last key: The last key (or key combination) pressed on the keyboard is

displayed here, such as used in Shortcuts or the 'If' and

'ElseIf" Paradigm steps. (See sample Paradigm 'Tuning with

Keys'.)

Note: Function and Control (Ctrl/Cmd) Shortcut key

combinations are not displayed.

Files and Pools These operations affect the entire "Paradigm Pool".

New Paradigm Pool Create a new blank Paradigm Pool and optionally copy

Paradigms into it from the existing Paradigm Pool.

The suggested name is auto-incremented from the previously

loaded Paradigm Pool name.

Load Paradigm Pool Load the Paradigms of a previously saved Paradigm Pool file

into the Paradigm Pool.

Revert to Last Saved

Undo any unsaved changes to the Paradigm Pool.

Save Paradigm Pool Save the Paradigm Pool using its existing file name and path.

Save Paradigm Pool As

Save the Paradigm Pool to a new file, and switch to the new file. The default file name is the same as the original file name.

Save Paradigm Pool Copy

Save the Paradigm Pool to a new file, but do not switch to the new file. The default file name has 'Copy of' prepended to it.

Merge Paradigm Pools

Insert the Paradigms from a previously saved Paradigm Pool file into the loaded Paradigm Pool.

The file path and file name of the loaded Paradigm Pool file are displayed.]

Paradigm description: A user description of the active Paradigm.

Paradigm Pool A column of Paradigm names from the loaded Paradigm Pool.

- Click on a Paradigm name to highlight it as the active Paradigm and display its steps.
- Double-click on a Paradigm name to start execution of the Paradigm and display its steps.
- Click-and-drag a Paradigm name to change its position in the column.
- To select multiple Paradigms, use a Shift-click mouse drag, or individually Shift-click the Paradigm names. Multiple Paradigms can thus be deleted, or saved to a new Paradigm Pool.

Save Pool Save the Paradigm Pool using its existing file name.

New Paradigm Create a new blank Paradigm in the Paradigm Pool.

Rename Rename the selected Paradigm.

- Valid characters are A-Z, a-z, 0-9, and "".
- Special characters are not allowed.
- Spaces are replaced by an underscore.
- The name cannot start with a number such entries will have an 'X' prepended to the name.

Duplicate Add a copy of the selected Paradigm to the Paradigm Pool.

The Paradigm name number is appended or incremented.

Delete Remove the selected paradigm from the Paradigm Pool.

Select a paradigm and click the 'Revert' button. All editable Revert steps are reset to their originally loaded values, as long as the

DPATCH - OPERATION MANUAL - REv. 2.2 (2021-10)

Paradigm Pool has not been saved.

Paradigm Steps

A column of instructions from the active Paradigm is displayed. These instructions are sequentially run by the Paradigm.

- Click on a Paradigm step to highlight it as the active step.
- Double-click on a Paradigm step to view or edit its settings.
- Click-and-drag a Paradigm step to change its position in the column.

Note: Step values are usually in standard units, i.e., "Volts" and "Amperes".

Step Buttons

Insert

Inserts a new command Step into the Paradigm Steps column:

Amplifier

Each Sweep

Routine

Analysis

Camera

Clear Key

Execute

Export

Front Window

Hide Window

Reset Timer

Scope Operation

Set Axis

Set Checkbox

Set Mark

Set Metadata	
Set Solution	
Set Tag	
Set Variable	
Set Write Steps	
Sound	
Start New Paradigm Data	
Update Inputs	
View Last	
Write to Log	
Alert	
Beep	
Comment	
Wait	
Pause	
Flow Control	
Break	
Chain	
For Loop	
Jump	
Label	
Condition	
If	

ElseIf

Else

(See details in Insertable Steps list below.)

Edit If a highlighted Step is configurable, clicking the Edit button

(or double-clicking the step) will open it in the Paradigm Steps

Editor for configuration.

Also, if a highlighted Step's text is partially hidden, use the

Edit button to view the entire entry.

The following buttons (Toggle Skip, Duplicate, Delete, Copy, Paste) can handle multiple steps. To select multiple steps, click each step with a Shift-click.

Toggle Skip Mark a step so it is not executed.

A semicolon is prepended to the Step number to "comment out" the instructions, and a Skip status is appended to the

Step text.

Example: A 'Beep' command in Paradigm step #2:

; 2 Beep, Skip=true

The leading semicolon ";" prevents this step from being executed by the instruction queue,

and the 'Skip' status is displayed.

Duplicate Insert a copy of the selected step after the selected step.

Multiple selected steps are inserted after the last selected

step.

Revert Select a Step to be reverted, and click the Revert button.

Editable fields are reset to their originally loaded values, as

long as another Paradigm has not been loaded.

Delete the selected step.

For multi-line steps, optionally delete the step without

deleting the contents of the step.

Copy Select a step to copy to the clipboard.

Paste Select a step and paste the copied step below it.

Multiple steps are pasted as a group.

Step Execute the selected step, then move to the next step.

Executing a single step does not terminate a running Paradigm, even if it is the last step in the Paradigm.

Note: A 'For' loop is processed as a single step.

Step to End Execute the selected step and all following steps as fast as the

system allows.

Slow Speed Execute 'Step to End' at ~1 second per step.

Write to Log

• Write None Write No Steps

• Write Main Write Main Steps

Action-oriented steps are recorded in the Paradigm metadata

(visible in the Metadata 'By Event' view):

Amplifier

Break

Camera

Chain

Execute

For Each Sweep

Reset Timer

Routine

Set Checkbox

Set Solution

Set Variable

Wait

• Write All Steps

Log the main steps and additional steps into the Paradigm

metadata (visible in the Metadata 'By Event' view.)

Insertable Steps

Amplifier

Control the dPatch amplifier hardware.




Figure 4-12. Step: Amplifier

Default Setting: Amplifier, Target=VHold, Equation=-0.080

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[drop-down list] Amplifier options.

Headstage

SelectProbe (select active headstage)

[1 - 2]

Most Paradigm Step commands apply to the "active" probe, the Sutter headstage presently controlled by the Amplifier Control Panel. Select the target headstage.

For a single headstage system, the active probe is always headstage number "1".

dPatch Settings

CCMode (amplifier current clamp)

Place the amplifier into Current-Clamp

mode.

VCMode (amplifier voltage clamp)

Place the amplifier into Voltage-Clamp

mode.

Hold (IHold in CC-mode, VHold in VC-mode)

 $[\pm 0.000,000,020 \text{ A } (\pm 20 \text{ nA}), \text{ or } \pm 1.000 \text{ V}]$

Set the active headstage holding level.

IHold (amplifer holding current, A)

 $[\pm 0.000,000,020 \text{ A } (\pm 20 \text{ nA})]$

Set and enable the active headstage holding level in Current-Clamp mode.

IHoldOn (amplifier holding current On)

[0 = Off, 1 = On]

Enable the active headstage holding level in Current-Clamp mode.

VHold (amplifier holding voltage, V)

 $[\pm 0.750 \text{ V}]$

Set and enable the active headstage holding level in Voltage-Clamp mode.

VHoldOn (amplifier holding voltage On)

[0 = Off, 1 = On]

Enable the active headstage holding level in Voltage-Clamp mode.

IGain (amplifier current gain, V/A)

Set the gain for the active 'Current' input channel using standard unit numbers (V/A) or scientific notation (1 mV/pA = "1e9"). The value is converted to a preset Gain level:

- 0.5 mV/pA
- 1 mV/pA
- 2.5 mV/pA
- 5 mV/pA
- 10 mV/pA

• 25 mV/pA

To help prevent saturation, exceeding a 90% threshold between levels promotes the equation value to the next higher gain level.

VGain (amplifier voltage gain, V/V)

Set the gain for the active 'Voltage' input channel using standard unit numbers (V) or scientific notation (1 mV = "1e3"). The value is converted to a preset Gain level:

- 10 mV/mV
- 20 mV/mV
- 50 mV/mV
- 100 mV/mV
- 200 mV/mV
- 500 mV/mV

To help prevent saturation, exceeding a 90% threshold between levels promotes the equation value to the next higher gain level.

Feedback (amplifier feedback mode 0, 1 or 2)

Set the amplifier feedback mode.

•	0	Capacitive Mode
•	1	±20 nA range
•	9	+200 nA range

Filter (amplifier input filter in VC- and CC-mode, Hz)

Set the filter level of the active input channel:

•	500	(500 Hz)
•	1000	(1 kHz)
•	2000	(2 kHz)
•	5000	(5 kHz)
•	10000	(10 kHz)

• 20000 (20 kHz)

To help prevent over-filtering, exceeding a 10% threshold promotes the equation value to the next higher filter level.

IFilter

(amplifier input filter in VC-mode, Hz)

Set the filter level of the active 'Current' input channel:

•	500	(500 Hz)
•	1000	(1 kHz)
•	2000	(2 kHz)
•	5000	(5 kHz)
•	10000	(10 kHz)
•	20000	(20 kHz)

To help prevent over-filtering, exceeding a 10% threshold between levels promotes the equation value to the next higher filter level.

VFilter

(amplifier input filter in CC-mode, Hz)

Set the filter level of the active 'Voltage' input channel:

•	500	(500 Hz)
•	1000	(1 kHz)
•	2000	(2 kHz)
•	5000	(5 kHz)
•	10000	(10 kHz)
•	20000	(20 kHz)

To help prevent over-filtering, exceeding a 10% threshold between levels promotes the equation value to the next higher filter level.

Offset

(amplifier pipette offset in VC-mode, V)

 $[\pm 0.5]$

Adjust the active output channel to remove any hardware- and/or solution-

related offsets.

 $OffsetLock \quad \ (amplifier \ pipette \ offset \ lock \ On \ in \ VC-$

mode)

$$[0 = Off, 1 = On]$$

Prevent the offset value from being accidentally changed during a recording.

SubtractPipOffset

(subtract pipette offset On in CC-mode)

LiquidJunc (liquid junction potential, V)

GentleSwitchC2V

(gentle mode switch CC- to VC- mode)

GentleSwitchV2C

(gentle mode switch VC- to CC- mode)

dPatch Compensation

ECompMag (amplifier electrode compensation mag,

F)

ECompTau (amplifier electrode compensation tau, s)

ECompOn (amplifier electrode compensation On in

CC-mode)

[0 = Off, 1 = On]

CmComp (amplifier cell compensation Cm, F)

Set a cell capacitance value and enable

cell capacitance compensation.

RsComp (amplifier cell compensation Rs, Ohm)

Set a series resistance value and enable

cell capacitance compensation.

RsCompOn (amplifier cell compensation On)

[0 = Off, 1 = On]

Bridge (amplifier bridge balance, Ohm)

BridgeOn (amplifier bridge balance On)

$$[0 = Off, 1 = On]$$

dPatch Correction

RsCorr (amplifier Rs correction, fraction)

[0.00 - 1.00] Converted to a percentile

RsLag (amplifier Rs correction lag, s)

RsCorrOn (amplifier Rs correction On)

[0 = Off, 1 = On]

dPatch Auto and Reset Functions

StimBoth (stimulate both headstages during auto-

compensation On)

AutoEComp (amplifier auto electrode compensation)

AutoCellComp

(amplifier auto cell compensation)

AutoRsCorr (amplifier auto Rs correction)

AutoBridge (amplifier auto bridge compensation)

AutoOffset (amplifier auto pipette offset)

[0 = Off, 1 = On]

AutoSet and enable the offset.

Reset (reset amplifier controls)

ResetCap (reset capacitor in capacitive feedback

mode)

Avoid capacitor reset transients in data recordings by generating a reset before each sweep; use the Paradigm 'Each Sweep' step with the Amplifier step option 'ResetCap' preceding the acquisition.

A short waiting step between the capacitor reset and the acquisition should be set, as there is a small delay between sending the reset command to the dPatch headstage and the actual

reset operation; a waiting period of 1 ms should be sufficient.

Dynamic Clamp

DynClampOn (amplifier dynamic Clamp On)

DynCGHold1 (amplifier dynamic Clamp G-holding 1,

DynCGHold2 (amplifier dynamic Clamp G-holding 2, S)

Dynamic Holding

DynHoldOn (amplifier dynamic holding On)

DynHold (amplifier dynamic potential, V)

DynHoldVHold

(amplifier dynamic holding potential set to V-hold)

DynHoldVmon

(amplifier dynamic holding potential set to V-mon)

DynHoldSpeed

(amplifier dynamic holding speed)

DynHoldAqOn

(amplifier dynamic active while acquiring)

Auxiliary Output

AuxOUT1 (Auxiliary Output-1, V)

AuxOUT2 (Auxiliary Output-2, V)

AuxOUT3 (Auxiliary Output-3, V)

AuxOUT4 (Auxiliary Output-4, V)

Digital Output

DigOUTWord (Digital Output Word)

DigOUT1 - 16 (Digital Output-1 - 16)

Lock-In

LockInAdjustOn

(set LockIn adjustments On)

LockInPhaseAdj

(set LockIn phase delay adjustment)

LockInAttenAdj

(set LockIn attenuation adjustment)

[Equation field] A free-form text field. This field is evaluated

and its value passed to the "target" function.

[Errors are reported under this field.]

Values are processed in standard units

(Amperes, Volts).

Check Equation Check the equation syntax. The equation is evalu-

ated for sweep #1, and if valid, it reports "Syntax

is ok."

Insert special identifier

Acquisition, amplifier and reference settings are available for use in equations. (See list below.)

Each Sweep

Control the Paradigm operations on a "per sweep" basis of a Routine. Commands to be executed are inserted between the "EachSweep, Target" line and the "ForEachEnd" line.




Figure 4-13. Step: Each Sweep

Default Setting: ForEachSweep

Each Sweep, Target=untitled

For Each End

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[drop-down list] Select a Routine name from the loaded Routine Pool.

Edit routine '<name>'

Open for editing in the Routine Editor.

Note: When using 'Each Sweep' to record data, the minimum sweep start-to-start time is +200 ms. For faster execution times, use the 'Routine' step.

Routine

Record data from a Routine.




Figure 4-14. Step: Routine

Default Setting: Routine, Target=untitled

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[drop-down list] Select an acquisition type, or a Routine to

record a data Series.

Membrane Test

Repeats [1-999 / inf]

Indefinitely Repeat an "infinite" number of

times.

Write to Metadata

Configuration

- No change
- Bath
- Seal
- Cell

• Free Run

Duration [100 ms - 999.9 s / inf]

Indefinitely Acquire for an "infinite"

duration.

Add Channel Clear

List of input channels.

[selected channels]

• List of Routine names from the loaded Routine Pool.

[selected Routines]

Edit routine "<name>"

Open the selected Routine for editing in the Routine Editor.

Note: The time from starting this command to recording data is +300 ms.

"Single-stepping" this command (when no Paradigm is running) will create an auto-triggered Paradigm.

Analysis

Save an analysis to the Analysis Editor, or combine it with prior analyses.




Figure 4-15. Step: Analysis

Default Setting: Analysis, Operation=Save to Editor

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[Operations]

Save to Editor Save the latest analysis.

Append to Last Append to the prior analysis.

- Average with Last Average with the prior analysis.
- Show Table Display analysis as numeric table.
- Show Graph (1 8) Display analysis as visual graph.

Camera

Take a single picture and/or run a live video preview. A Camera window is opened behind the Paradigm Editor window.




Figure 4-16. Step: Camera

Default Setting: Camera, Camera=_Camera_Name, Capture =true Close the 'Paradigm Steps Editor'. Undo Remove any unsaved edits to this step. Camera Select a camera on the computer system. Take a picture when executed. Capture Live view: Configure the state of the live view: No Change Keep last settings Stop live view Stop

Clear Key

Clear the 'Last key' field in the Paradigm Editor, which holds the last-pressed keyboard key since the start of the Paradigm.

Start live view

Default Setting: ClearKey

Start

Execute

Extend the functionality of SutterPatch by running an Igor Procommand.




Figure 4-17. Step: Execute

Default Setting: Execute, Command=Beep

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Command []

Run any Igor command accepted by the Command window, including user-created Functions.

Note: Igor syntax usually requires that open/close parentheses "()" be appended to the end of a command, however exceptions include the "beep" and "print" commands, for which no parentheses are used.

Insert Special Identifier

Special references can also be used within commands:

- p[1..16] (n'th Paradigm variable)
- s[series-count, sweep-count, trace-count, routine-name]

(trace of specified series)

Reference an input trace in an open Scope window via counts of Series #, Sweep #, Trace # (scope channel position), and the Routine name.

For the Acquisition: Routine scope window, the active trace has "count" values of zero. If a count number is non-zero, it is used as a relative offset (positive or negative) from the active trace.

Any fractions in count numbers are truncated to integers.

For the Reanalysis scope window, the first Series/sweep/trace has count values of one. All counts are positive relative to the first trace.

If the Routine name is left blank, the active Routine is used.

Ex: s[0,0,0,]

The Acquisition: Routine scope window active Series, active sweep, and active trace of the active Routine.

• t[#] (trace of current sweep of current series)

Reference the input trace in the open scope window channel position "#", for the active sweep of the active Series of the active Routine.

• eq[equation]

(result of the given equation)

Copy to Command Line Append the Command text to the Command window's command line.

Expand to Command Line Append the Command entry to the

Command window's command line after processing it to be compatible with Command window execution, i.e., any variables are replaced by their values.

Example 1: Reset the Timer.

Set the Execute 'Command' to:

Paradigm_ResetTimer()

Note the open and close parentheses at the end.

Example 2: Create an FFT graph of your data.

The Paradigm Steps:

- 1. ForEachSweep
- 2. EachSweep, Target=YourRoutineName
- 3. Execute, Command=FFT/OUT=3/DEST=Voltage1_ FFT t[2]
- 4. If, Left=sweep, Operation="=", Right=1
- 5. Execute, Command=Display Voltage1_FFT
- 6. EndIf
- 7. Execute, Command=SetAxis Bottom 0,60
- 8. ResetTimer
- 9. ForEachEnd
- In Step 2: Replace "YourRoutineName" with your own Routine name, or use the sample "IV" Routine.
- In Step 3: The Igor 'FFT' command is run, and "t[2]" retrieves the Acquisition:
 Routine scope's second input trace.

Export

Export data graphs into a new or open Layout window.



Figure 4-18. Step: Export

	•
Default Setting:	Export, Signal=Layout
•	Close the 'Paradigm Steps Editor'.
Undo	Remove any unsaved edits to this step.
Graphs per page	Set the graph layout configuration for new Layout windows:
1	Graph fills entire page.
2	Graphs stacked.
3	Graphs stacked.
2 x 2	Matrix display.
2 x 4	Matrix display.
[Add Signal list]	Select signals to be exported from a list of default names.
Clear	Clear the signal field, set it to 'off".
All	Select all entries.
All Signals	Select all input signals.
[list of input signals]	
All Analyses	Select all Analysis graphs.
[list of Analysis graphs]	
[list of selec	ted signals]

User-edited names can be directly entered into

the signal field.

Note: The sequence of signals is not used for positioning in the Layout window – signal positioning is based on their scope window sequence.

Front Window

Set the specified window as the front window.




Figure 4-19. Step: Front Window

Default Setting: Front Window, Target=Scope Window Close the 'Paradigm Steps Editor'. Undo Remove any unsaved edits to this step. Front Window **Analysis Editor** Camera Window Control Panel Dashboard Data Navigator **Equation Editor** Log Window Paradigm Editor Routine Editor Scope Window Shortcut Editor

Solution Editor

Template Editor

Hide Window

Hide the specified window.




Figure 4-20. Step: Hide Window

Default Setting: HideWindow, Target=Scope Window

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Hide Window Analysis Editor

Camera Window Control Panel Dashboard

Data Navigator Equation Editor

Log Window

Paradigm Editor Routine Editor Scope Window Shortcut Editor Solution Editor Template Editor

Reset Timer

Reset the Paradigm Editor Timer to 00:00:00.

Default Setting: ResetTimer

Scope Operation

Control which scope window signals are displayed, and how the sweep display operates.



Figure 4-21. Step: Scope Operation

Default Setting: Scope, Wipe=false

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Wipe Scope Clear the scope window of all sweeps, except the

last one.

Signal list: Enable to display a list of input signals.

Add Signal Clear Clear the signal list.

All Signals Select all signals.

[List of all possible input signals]

[List of selected input signals]

You can directly edit the list. User-defined signal labels can also be used.

Persistence:

• On

Off

Autoscale A one-time autoscale of the Y-axes of all selected

signals to their incoming data, i.e., to their visible sweeps data limits, and resets the X-axes to the

full sweep duration.

No change

Center Center the active signal so the mean of the Y-axis

data is vertically centered in the signal pane. Only the Y-axis offset is automatically adjusted, not the

scaling; the X-axis is unaffected.

Set Axis

Modify the axis scaling of selected signals in the open Scope window.




Figure 4-22. Step: Set Axis

Default Setting: Axis, Axis=Autoscale, Kind=Left,

Target=Current1

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[drop-down list] [Left]

Select the Y-axis scaling to apply.

• Autocenter Center the Y-range of the X-axis data.

• Autoscale Match the axis range to the data range.

Autoscale from Zero

Display from zero to the largest absolute

value.

• Full scale Display the full range of the axis.

• Set scale Enter custom range settings:

Min.

Max.

[drop-down list] Select the axes orientation.

• Left Set the Y-axes on the signal list.

• Bottom Set the X-axes of all signals.

[drop-down list] [Bottom]

Select the X-axis scaling to apply.

• Full scale Display the full range of the axis.

• Set scale Enter a custom range setting:

Min.

Max.

[drop-down list] [only displays for 'Left' axis]

Add the selected signals to the list.

- Clear the signal list.
- All Signals Select all signals.
- [list of all available input signals]

Select individual signals.

[a list of the selected signals]

[only displays for 'Left' axis]

Signal names can be directly edited; user-defined signal labels can be used.

Set Checkbox

Set Checkbox uses simple "on / off" toggles. Checkbox status can be read by 'If" and 'ElseIf' steps to make "yes/no" decisions and control the execution path of the Paradigm. If the equation evaluates to a non-zero value, the checkbox is enabled, i.e., "on".




Figure 4-23. Step: Checkbox

Default Setting: Checkbox, Count=1, Equation=true
 Close the 'Paradigm Steps Editor'.
 Undo Remove any unsaved edits to this step.
 Checkbox
 Checkboxes [1-3] are local: they are cleared whenever a Paradigm is started.
 Checkboxes [4-6] are global: their status

persists across all Paradigms in the Experiment.

[Equation field]

A free-form text field, evaluated to a value, and applied to the Checkbox.

[Errors are reported under this field.]

Check Equation

Check the equation syntax. The equation is evaluated for sweep #1, and if valid, it reports "Syntax is ok."

- The constant "True" evaluates to '1.000'.
- The constant "False" evaluates to '0.0000'.

Insert special identifier

Acquisition, amplifier and reference settings are available for use in equations. (See list below.)

Set Mark

The 'Set Mark' step marks (or unmarks) the active sweep for later processing by the Data Navigator.



Figure 4-24. Step: Set Mark

Default Setting: SetMark, Value=Set

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Mark

- Set
- Clear
- Toggle

"Marking unmarking" a sweep marks/unmarks that sweep in all signals in the same Series. Marked sweeps are loaded into the Data Navigator as "marked".

This can be used within a conditional paradigm step to mark or unmark a sweep based upon experimental conditions.

For example, when used within a paradigm 'If' step, if the leak current is too high, unmark the sweep, else mark the sweep. This is an easy way to process just the sweeps that have a reasonable leak current.

Marks are used by the Data Navigator 'Available actions':

Action Potential Analysis Synaptic Event Analysis Average Selected Sweeps Display Signal/Sweep Export Data

Set Metadata

Define Metadata parameter values to apply to the data during acquisition.

The 'Set Metadata Paradigm Step Value' dialog opens for configuration:




Figure 4-25. Step: Select Metadata Group

Default Setting: Metadata, Value=

Select Metadata Group

To change the metadata detail level, go to Preferences > Metadata.

•	Operator	[Full detail level]
•	Preparation – Animal	[Basic detail level]
•	Preparation-Tissue	[Basic detail level]
•	Preparation-Cell	[Basic detail level]
•	Experiment	[Basic detail level]

• Electrode [Extended detail level]

• Recording Solutions [Extended detail level]

Paradigm [Full detail level]

Cell Health / Quality Control [Full detail level]
Series (= Routine Data) [Full detail level]

• Stimulus [Basic detail level]

Select Metadata Parameter

Entries depend on the selected Group.

[Metadata Parameter Info]

• Use default:

• Use last value:

• Use a previous value

• Use new value

• Equation

[< Define Equation >]

Check the equation

syntax. The equation is evaluated for sweep #1, and if valid, it reports

"Syntax is ok."

Special identifier Special functions for use in

equations. (See list

below.)

Prompt for confirmation Display a metadata prompt before

acquisition.

Set Solution

A "solution" command is used to turn solution valves 'on' or 'off' in perfusion systems. A predefined digital pattern or analog level can be automatically output with this step. Solution settings are configured and numbered in the Solution Editor.




Figure 4-26. Step: Set Solution

Default Setting: Solution, Target=Off

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Set Solution [1-24]

Select a solution number to activate its valve. The number of available solutions depends on

the Solution Editor configuration.

Set Tag

A comment tag is automatically written to the Paradigm metadata with this step. Enter the comment into the 'Tag text' field.

When run during acquisition, the comment tag is also written to the Routine metadata, and when the data is opened in a Reanalysis scope window, a black vertical cursor displays at that time point.



Figure 4-27. Step: Set Tag

Default Setting: SetTag, Text=

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Tag text Enter the comment text.

Note: The comment text for this Paradigm step

is maintained separately from the manually triggered Acquisition Control

'Set Tag' button text.

Update Inputs

Takes a single "live" reading from the Auxiliary Inputs.

Use for slowly changing variables, such as temperature.

Set Variable

Variables allow flexible control of any operation that uses equations.




Figure 4-28. Step: Set Variable

Default Setting: SetVariable, Target=Paradigm, Count=1, Equation=p[1]

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[List of targets]

Paradigm Set the value of a Paradigm Variable_p[#].

[1-16. All Variables]

When 'All Variables' is selected, if varying values are desired, enter their values into the Equation field as a comma-separated list; simple equations (those without internal commas) can also be used in place of a value.) If there are more variables than list values, the "extra" variables are unchanged. If a list value is blank, the corresponding variable is unchanged.

Paradigm_Input

Set the value of the Paradigm Editor 'Input' control.

< Routine Names >

Select a Routine and set the value of its Variable_r[#]

[1-16, All Variables]

When 'All Variables' is selected, if varying values are desired, enter their values into the Equation field as a comma-separated list; simple equations (those without internal commas) can also be used in place of a value.) If there are more variables than list values, the "extra" variables are unchanged. If a list value is blank, the corresponding variable is unchanged.

Equation

Evaluates to a value, used to set variables (or the Paradigm Editor 'Input' control.)

You can likewise set the value of a variable by inserting special identifiers; for example, 'Input' reads the 'Input' control. (See sample Paradigm 'Tuning_with_Input'.)

Check Equation

Check the equation syntax. The equation is evaluated for sweep #1, and if valid, it reports "Syntax is ok."


Insert special identifier

Acquisition, amplifier and reference settings are available for use in equations.

(See_list_below.)

Set Write Steps

Configure the level of logging Paradigm metadata.



Default Setting: SetWriteSteps, Value=Main

Close the 'Paradigm Steps Editor'.

Undo

Remove any unsaved edits to this step.

[Drop-down list]

- Write No Steps
- Write Main Steps

Action-oriented steps are recorded in the Paradigm metadata (visible in the Metadata 'By Event' view):

Amplifier

Break

Camera

Chain

Execute

For Each Sweep

Reset Timer

Routine

Set Checkbox

Set Solution

Set Variable

Wait

• Write All Steps

Log the main steps and additional steps into the Paradigm metadata (visible in the Metadata 'By Event' view.)

Sound

Output a note from the computer speaker.

The frequency can be defined by a fixed value or an equation.




Figure 4-29. Step: Sound

rigare 1 20. Step. Sound		
Default Setting:	Sound, Equation=,	Volume=1
•	Close the 'Paradign	n Steps Editor'.
Undo	Remove any unsave	ed edits to this step.
Equation (Hz)	[250 - 8000]	
	Specify as an equat	ion or fixed value.
	The sound output h response range with	nas a linear frequency hin its limits:
	< 250 Hz	two clicks
	$250~\mathrm{Hz} - 8~\mathrm{kHz}$	frequency tone
	> 8 kHz	8 kHz tone
Check Equation	_	syntax. The equation is p #1, and if valid, it reports
Insert special identifier		
	Acquisition, amplification available for use in (See_list_below_)	ier and reference settings are equations.
Play Sound	Test the sound outp	out.
Volume	$[\ 0.1 - 1.0\]$	
	Use the spinners for	r 10% increments, or directly

edit the field.

Output is via the standard sound output that Igor uses:

• Windows: Built-in speakers, or a computer sound card

with external speakers.

Note: Lower frequency tones are attenuated in

volume on lower-quality speakers.

• macOS: Built-in speakers

This paradigm step can also be utilized as an Igor programming command. For instance, using an equation, one could listen to the membrane resistance of the cell under investigation

Example: Output a note.

Enter this equation in the Command window

command line:

SutterPatch#Paradigm_PlaySound(400, 1)

Start New Paradigm Data

Stop the current Paradigm and start a new Paradigm.

This forces a new Paradigm node to be created in the Data Navigator.

Update Inputs

Read a "live" data point from all auxiliary and headstage monitor input channels. This is useful for monitoring slowly changing parameters, such as temperature, without acquiring an entire sweep of data.

View Last

Display the data from the last recording in a Reanalysis scope window.

Write to Log

Enter text to be written to the Log window.




Figure 4-30. Step: Write to Log

Default Setting: WriteLog, Alert=true, Text=text_to_write, Equation=, DoBeep Close the 'Paradigm Steps Editor'. Undo Remove any unsaved edits to this step. Do Beep Generate a beep before writing. Show Alert Display and/or edit the Alert text, then write it to the Log window. Text to send to log Equation result to append [Multiple equations in a comma-separated list can be evaluated. **Format** Time Date 1-12 digits **Check Equation** Check the equation syntax for sweep #1. The equation is evaluated, and if valid, it reports "Syntax is ok."

Insert special identifier

Acquisition, amplifier and reference settings are

available for use in equations.

(See list below.)

Paradigm: Write to log

This run-time dialog displays when the Paradigm step is executed:




Figure 4-31. Step: Write to Log Run-Time Window

Elapsed time A time counter for the Alert.

Text Edit the text message.

Continue and Write Write to the metadata.

Continue, no Write Do not write to the metadata.

Alert

Display an "Alert" dialog box that pauses Paradigm execution until manually dismissed.



Figure 4-32. Step: Alert

Default Setting: Alert, Text=alert text, DoBeep=true

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Do Beep Sound a "beep" from the computer.

Text to show in Alert

Enter a message to the user.

<u>Beep</u>

Generate a "beep" sound from the computer speaker.

Default Setting: Beep

Comment

A text message can be displayed in a floating window.



Figure 4-33. Step: Comment

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Comment Enter the comment text.

To display multiple lines of text (up to 3), use "\r" as a line separator. Enter up to 40 characters per line, with a maximum of 100 characters per comment.

Note: Text characters are from the ANSI character set.

Show in window A 'Paradigm Comment' window is displayed

paradigm ends.

with the comment text, and closes when the

Wait

Temporarily pause execution of the Paradigm for a defined duration.




Figure 4-34. Step: Wait Default Setting: Wait, Time=1 Close the 'Paradigm Steps Editor'. Undo Remove any unsaved edits to this step. Use [Given value, Variable p[1] - p[16]] Seconds to Wait [displays for "Given value"] Click the spinners for 0.1 s increments, or type in a value. The precision of the wait time is 5

Pause

Pause execution of the Paradigm until the Resume button is manually clicked.

Flow Control: Break

Use a Break step to stop the execution of a Paradigm, or to interrupt For Loop and For Each Sweep loops.



Figure 4-35. Step: Break

Default Setting: Break, Kind=Paradigm

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Break Kind: Paradigm

ForLoop

Flow Control: Chain

Use to link step execution to another Paradigm.



Figure 4-36. Step: Chain

Default Setting: Chain, Target=undefined_Paradigm,

Return=true

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Return to calling Paradigm:

Once execution of the target Paradigm has completed, return execution to this Paradigm.

Paradigm to chain to:

Paradigm execution will shift to the selected Paradigm.

For multiple Chains (or recursive calls), you can link a maximum of eight Paradigms.

Flow Control: For Loop

Use a standard programming "For loop" to repeat a set of steps.




Figure 4-37. Step: For Loop

Default Setting: ForLoop, Max=1

ForEnd

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Max. ForLoop Count Number of loop cycles to run.

Loop Indefinitely Sets 'Max. ForLoop Count' to 'inf'.

Note: A "For loop" is processed as one step.

Flow Control: Jump

Shift the Paradigm sequence to an arbitrary step. When executed, a jump occurs to the step after the target Label.



Figure 4-38. Step: Jump

Default Setting: Jump, Target=jump_label

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Jump Target Enter the Label of the step to jump to.

Flow Control: Label

Create a target step Label for a Jump step, to change the sequence of execution.



Figure 4-39. Step: Label

Default Setting: Label, Target=jump_label

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

Label name Assign a name to the Label.

Condition: If

This step allows conditional Paradigm flow control between multiple choices.




Figure 4-40. Step: If

Default Setting: If, Left=, Operation='>', Right=

EndIf

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[drop-down list] Operation selection.

Compare 2 equations

Left Equation Evaluated to a value.

Check Equation Check the equation syntax. The

equation is evaluated for sweep #1, and if valid, it reports "Syntax

is ok."

Insert special identifier Acquisition, amplifier and

reference settings are available

for use in equations. (See_list_below.)

Operation: Comparison operators.

> Greater than

>= Greater than or equal to

= Equal to

!= Not equal to

<= Less than or equal to

< Less than

Note: Be careful when

comparing two floatingpoint numbers for equality, as minor variations in resolution can affect the outcome.

Right Equation Evaluated to a value.

Check Equation Check the equation syntax. The

equation is evaluated for sweep #1, and if valid, it reports "Syntax

is ok."

Insert special identifier Acquisition, amplifier and

reference settings are available

for use in equations. (See_list_below.)

Check for key pressed The "Last key" typed on keyboard

during the current Paradigm.

The "Last key" field is cleared at

the start of a Paradigm.

Key to check for Enter a text key, or insert a

"special" key.

Insert special key Use a "non-text" key.

• Space

• Return

• Esc

Check checkbox status Select a checkbox to monitor for

"on / off" status.

Checkboxes are displayed at the bottom of the Paradigm Editor

window.

Checkbox

[1-3] Paradigm-level "local"

checkboxes, cleared at start of

Paradigm.

[4-6] Experiment-level "global"

checkboxes, persists across

Paradigms.

• Answer of yes-no alert

Do Beep Your computer beeps once when

the alert displays.

Alert Text [] Enter your alert question text.

Run-time dialog

Elapsed Time A time counter for the alert.

Yes 'Yes' button (value = 1)

No 'No' button (value = 0)

Stop Paradigm Manually abort the Paradigm.

Condition: ElseIf

Allow conditional Paradigm flow control between multiple choices.




Figure 4-41. Step: Else If

Default Setting: ElseIf, Left=, Operation='>', Right=

Close the 'Paradigm Steps Editor'.

Undo Remove any unsaved edits to this step.

[drop-down list] Operation selection.

• Compare 2 equations

Left Equation Evaluated to a value.

Check Equation Check the equation syntax. The

equation is evaluated for sweep #1, and if valid, it reports "Syntax

is ok."

Insert special identifier Acquisition, amplifier and

reference settings are available

for use in equations. (See list below.)

Operation Comparison operators.

> Greater than

>= Greater than or equal to

= Equal to

!= Not equal to

<= Less than or equal to

< Less than

Note: Be careful when

comparing two floatingpoint numbers for equality, as minor variations in resolution can affect the outcome.

Right Equation Evaluated to a value.

Check Equation Check the equation syntax. The

equation is evaluated for sweep #1, and if valid, it reports "Syntax

is ok."

Insert special identifier Acquisition, amplifier and

reference settings are available

for use in equations. (See_list_below.)

• Check for key pressed The "Last key" typed on the

keyboard during the current

Paradigm.

The "Last key" field is cleared at

the start of a Paradigm

Key to check for

Enter a text key, or insert a

"special" key.

Insert special key

Use a "non-text" key.

- Space
- Return
- Esc

• Check checkbox status

Select a checkbox to monitor for

"on / off" status.

Checkboxes are displayed at the bottom of the Paradigm Editor

window.

Checkbox

[1-3] Paradigm-level "local"

checkboxes.

[4-6] Experiment-level "global"

checkboxes.

• Answer of yes-no alert

Do Beep Your computer beeps once when

the alert displays.

Alert Text [] Enter your alert question text.

Run-time dialog

Elapsed Time A timer of how long the Alert has

been displayed.

Yes 'Yes' button (value = 1)
No 'No' button (value = 0)


Stop Paradigm Manually abort the Paradigm.

Condition: Else

This step allows Paradigm flow control to continue to the next step if the previous condition fails.

Default Setting: Else

 steps. They are visible at the bottom of the Paradigm Editor window.



Input

Routine and Paradigm variables can be set to this value. Manually enter a value, or set via the Paradigm step 'Set Variable / Insert special identifier / Paradigm Parameters'.

[-1.00 - 1.00] This value is restricted to \pm 1.00 to enable scrolling through a defined range of values. It can be rescaled when used in an equation.

Paradigm Variables

The Paradigm Variables table displays at the bottom of the Paradigm Editor.




Figure 4-43. Paradigm Variables

These variable can be used in any equation, or in the Paradigm step Execute, and persist across experiments. The table can be directly edited during non-acquisition, or set via the Paradigm step Set Variable.

•	'Close' button	Closes the Variables table.
	Variable[1 – 16]	16 columns of Paradigm Variables.
Name:	Var_p[1 - 16]	Paradigm Variable names can be edited to any text.
		Note: These names are

for display only, and are not supported in equations.

Value: [] Numeric values can be

The following acquisition, amplifier and reference settings are

manually entered, or programmatically set via the Paradigm step 'Set Variable'.

available for use in equations:

Timing

Special identifiers

Time (present date-time, s)

Timer (timer time, s)

ParadigmTime (time at start of paradigm, s)

RoutineTime (time at start of routine, s)

Paradigm Parameters

Loop (active paradigm ForLoop count)

Sweep (active paradigm EachSweep count)

Sweep count of the active sweep in the scope

window.

LastSweep (total number of sweeps in active Routine)

During acquisition, this is set according to the

Routine parameters. Once acquisition

terminates, this is replaced by the count of the last acquired sweep, i.e., the last sweep in the

stored Series.

Processing can occur before or after the last

sweep of a Series.

Example: Use in a 'ForEachSweep' loop Routine, to

compare an 'If' step equation to the

sweep number.

ForEachSweep

EachSweep, Target=IV

If, Left=sweep, Operation='=', Right=LastSweep- 1

Alert, Text=LAST SWEEP, DoBeep=true EndIf ForEachEnd

AqStopped (last acquisition was stopped)

[0 = the last acquisition completed]

[1 = the last acquisition was stopped before

completion]

Stimulant (last applied stimulant concentration)

From the Solution Editor 'Concentration'

setting, for a solution configured as a 'Chemical

Stimulant'.

Input (Input variable on paradigm window)

Hold[1..2] (holding of n'th output channel)

Headstage holding level.

p[1..16] (n'th paradigm variable)

r[1..16] (n'th routine stimulus variable)

Analysis Results

m[1..16] (n'th analysis measurement value)

gx[1..16] (n'th analysis graph x value)

The X-value of the last data point in the latest

version of graph[#].

gy[1..16] (n'th analysis graph y value)

The Y-value of the last data point in the latest

version of graph[#].

Signal Readings

AuxIN[1..8] (auxiliary input, V)

A single-point reading, such as from a slowly

changing temperature probe.

Note: This usage does not require setting up a

Routine Input Channel.

Imon (amplifier current reading, A)

Vmon (amplifier voltage reading, V)

Mean[name or count, start, width] (mean of given input signal)

'name' = signal name

'count' = window-signal position

'start' = time of start, s (of measurement range)

'width' = duration, s (of measurement range)

Headstage

ActiveProbe (active probe)

[1-2]

The highlighted headstage number in the

Amplifier Control Panel.

For a single headstage system, the active probe

is always headstage number "1".

NumProbes (number of probes)

The number of dPatch headstages attached to

the system.

dPatch Settings

CCMode (amplifier current clamp)

[0 = Off, 1 = On]

VCMode (amplifier voltage clamp)

[0 = Off, 1 = On]

Hold (IHold in CC-mode, VHold in VC-mode)

 $[\pm 0.000,020 \text{ A } (\pm 20 \text{ nA}), \text{ or } \pm 1.000 \text{ V}]$

Headstage holding level.

IHold (amplifer holding current, A)

 $[\pm 0.000,020 \text{ A } (\pm 20 \text{ nA})]$

IHoldOn (amplifier holding current On)

[0 = Off, 1 = On]

VHold (amplifier holding voltage, V)

[±1.000 V]

VHoldOn (amplifier holding voltage On)

[0 = Off, 1 = On]

IGain (amplifier current gain, V/A)

Read the gain level of the active voltage-clamp

'Current' input channel.

VGain (amplifier voltage gain, V/V)

V/V evaluates to mV/mV.

Read the gain level of the active current-clamp

'Voltage' input channel.

Feedback (amplifer feedback mode: 0, 1, 2)

• 0 Capacitive Mode

• 1 ± 20 nA range

• 2 ±87200 nA range

Filter (amplifier input filter in VC- and CC-mode, Hz)

Read the low-pass filter of the active input

channel.

IFilter (amplifier input filter in VC-mode, Hz)

Read the low-pass filter of the 'Current' input

channels.

VFilter (amplifier input filter in CC-mode, Hz)

Read the low-pass filter of the 'Voltage' input

channels.

Offset (amplifier pipette offset, V)

OffsetLock (amplifier pipette offset lock On)

[0 = Off, 1 = On]

SubtractPipOffset (subtract pipette offset On in CC-mode)

$$[0 = Off, 1 = On]$$

StimBoth (stimulate both headstages during auto-

compensation On.)

LiquidJunc (liquid junction potential, V)

GentleSwitchC2V (gentle mode switch CC- to VC- mode)

GentleSwitchV2C (gentle mode switch VC- to CC- mode)

CapResets (number of capacitor resets since last sweep

start)

dPatch Compensation

ECompMag (amplifier electrode compensation magnitude, F)

ECompTau (amplifier electrode compensation tau, s)

ECompOn (amplifier electrode compensation On in CC-

mode)

[0 = Off, 1 = On]

CmComp (amplifier cell compensation Cm, F)

RsComp (amplifier cell compensation Rs, Ohm)

RsCompOn (amplifier cell compensation Rs On)

[0 = Off, 1 = On]

Bridge (amplifier bridge balance, Ohm)

BridgeOn (amplifier bridge balance On)

dPatch Correction

RsCorr (amplifier Rs correction, fraction)

RsLag (amplifier Rs correction lag, s)

RsCorrOn (amplifier Rs correction On)

[0 = Off, 1 = On]

[0 = Off, 1 = On]

Dynamic Clamp

DynClampOn (amplifier dynamic Clamp On)

DynCGHold1 (amplifier dynamic Clamp G-holding 1, S)

DynCGHold2 (amplifier dynamic Clamp G-holding 2, S)

Dynamic Holding

DynHoldOn (amplifier dynamic holding On)

DynHold (amplifier dynamic potential, V)

DynHoldSpeed (amplifier dynamic holding speed)

DynHoldAqOn (amplifier dynamic active while acquiring)

Membrane Test

Relectr[1..2] (electrode/seal/access resistance, Ohm)

Value from last Membrane Test.

Rmemb[1..2] (membrane resistance (cell mode), Ohm)

Value from last Membrane Test.

Cmemb[1..2] (membrane capacitance (cell mode), F)

RMSNoise[1..2] (membrane test RMS noise, A)

Value from last Membrane Test.

Lock-In

LockInPhaseAdj (Lock-In phase delay adjustment, s)

LockInAttenAdj (Lock-In attenuation adjustment)

4.1.8 Routine Editor

Routines contain the settings that are in effect during data acquisition. The Routine Editor allows you to define acquisition parameters, set input and output channels, and to create stimulus waveforms and online analyses. The Routine Editor is the central place to create and manage saved Routine Pools and data acquisition settings.





Figure 4-44. Routine Editor

The Routine Editor is structured to hold one or more Routines within its Routine Pool. The Routine Pool thus provides easy access to the set of Routines used in an experiment.

Tip: SutterPatch comes with a Sample Routine Pool that contains a collection of frequently used experimental scenarios. Rather than creating a new Routine, it might be easier to Duplicate a sample Routine and modify it until it meets your particular needs.



Status Field	Notifications on edits and Routine names are displayed here.	
Files and Pools	Files and Pools [drop-down list]	
	Most recently used list of the last 5 Routine Pool files.	
	Load Routine Pool	Load the Routine Pool of a previously saved Routine Pool file.
	New Default Routine Pool	Create a new Routine Pool with a default Routine.
	New Routine Pool	Create a new Routine Pool either with a default Routine, or populated with Routines from the currently loaded Routine Pool.
	Get Sample Routine Pool	Load the Routine Pool from the dPatch sample Routine Pool file SP_RoutinePool_dPatch.spr.
	Revert to Last Saved	Undo any unsaved changes to the Routine Pool.
	Save Routine Pool	Save the Routine Pool using its existing filename and path.
	Save Routine Pool As	Save the Routine Pool to a new filename, and switch to the new file. The default filename has an increment number appended to the original filename.
	Save Routine Pool Copy	Save the Routine Pool to a new file, but do not switch to the new file. The default filename has 'Copy of' prepended to the original filename.
	Merge Routine Pools	Insert the Routines from a previously saved Routine Pool file into the loaded Routine Pool.
	Merge PatchMaster PGF File	Insert the "Routines" (PGF Sequences) from a PatchMaster PGF file into the loaded Routine Pool. Re-assign the analog input and output channels in a 'Mappings' dialog that opens.
		[The amplifier hardware must be attached to enable this option.]

Convert Routine Pool	Convert the loaded Routines (designed for other instruments) to be compatible with the attached amplifier or emulation mode. The original file is overwritten. All conversion changes are written to the Command window.
Send Last Used List to History	Copy the pathnames for the 'Files and Pools' "last used" Routines list into the Command window history.

Table 4-2. Routine Files and Pools

New Routine Pool dialog:




Figure 4-45. New Routine Pool

Create a new Routine Pool populated with a default Routine, or select Routines from the loaded pool to populate the new pool.

[File Path field]

The file path of the current Routine Pool displays on the right of the 'Files and Pools' list. If the Routine Pool has not been previously saved, this field is blank.

Routine Description

A text comment can be edited and saved with the Routine.

Routine Pool

The Routine Pool section lists the names of all currently loaded Routines. Selecting a Routine name loads it into the Routine Settings section for editing and activation. As the Routine Pool contents are held in memory, the switching times between Routines are very fast.

Double click a Routine name to edit it.

- Allowable characters are A-Z, a-z, 0-9, and "".
- Special characters and spaces are not allowed.
- The maximum length of a Routine name is 22 characters.

To select multiple Routines, use a Shift-click mouse drag, or individually Shift-click the Routine names. Multiple Routines can thus be deleted, or saved to a new Routine Pool.

Note: When a Routine is selected, if it was created for a different system (i.e., IPA or Double IPA), and the Preferences / Hardware / Routine Conversion is set to 'Prompt user', then a Routine Conversion dialog displays.

Activate	Open or refresh the Acquisition: Routine scope window with the latest Routine settings, but do not start acquisition.
	This button is re-named to "In Progress" during a recording.
Execute	Open or refresh the Acquisition scope window and immediately start recording. The latest Routine settings are applied to the scope window.
/ Convert	This button is renamed to "Convert" if the selected routine was designed for a different amplifier type than the current Experiment uses. Routine conversion changes are written to the Command window.
Save Pool	Save the Routine Pool using its existing file name.
New Routine	Add a default Routine to the Routine Pool, and open it for editing. The default Routine name is "untitled" with an increment number appended.
Duplicate	Add a copy of the selected Routine to the Routine Pool. The Routine name number is appended or incremented.
Delete	Remove the selected Routine from the Routine Pool.
Revert	Discard any unsaved changes to the selected Routine.

Table 4-3. Routine Editor Buttons

Waveform Preview

The stimulus waveform is graphically displayed at the bottom of the Routine Editor.




Figure 4-46. Waveform Preview Pane

The waveform preview and its settings are updated live to reflect changes in the Waveform Editor and Amplifier Control Panel.

Measurement regions can be manually repositioned in the Preview pane. Click and hold a measurement region to highlight it in black, then drag it to a new position, and release. This also updates its Measurement Settings / Region Timing 'Start/End Time' settings.

To change the region's duration, click-drag the region's right-edge cursor; its Region Timing 'End Time' setting is updated, while its 'Start Time' is unchanged.

The preview for the digital output 'DigOUTWord' sets its Y-axis to 'Digital State (Word)', and displays the decimal value of the selected bits.

Note: A "Cityscape" display mode is used, i.e., plotting with straight horizontal and vertical lines connecting the preview sample points (vs. smooth interpolated transitions).

Copy to Layout	Copy the stimulus graph into a new Layout window, or append to an existing Layout page.
Copy to Clipboard	Copy the stimulus graph to the system clipboard

Show / Hide Preview Expand or collapse the Preview pane.

X- and Y-axis Control

- Hover the mouse cursor over an axis line until the cursor turns into a double-headed arrow, then scroll up or down to contract/expand the axis.
- In the preview pane, click and drag the mouse cursor to surround a region of interest with a bounding box (the "marquee"). Right-click in the box and select one of the expand/shrink options.

Some key settings and display controls are listed on the right of the Preview pane.

Units are in 's', or if < 1 s., then in 'ms'.

Sweep Interval: The interval of time between the starts of consecutive triggered

sweeps (Sweep Start-to-Start Time) in the active Routine.

When set to 'Shortest', this equals the longest Sweep Duration +

200 ms.

Sweep Duration: The amount of time in a sweep during which signal recording

occurs with the active Routine.

Stim Duration: The maximum amount of time during which output stimulation

occurs in a sweep.

Set in Output Channels & Waveform / Waveform Editor / Dura-

tion.

Cycle Duration: The amount of time for a cycle.

Set in Acquisition & Routine Parameters.

Cycle Bytes: The number of bytes of data in a cycle.

Cycle Stim Points: The number of points in which output stimulation occurs in a

cycle.

Series Bytes: The number of bytes of data in the Series.

Series Duration: The amount of time for the Series.

Preview Holding: The holding level in the Amplifier Control Panel.

Show Channel: A list of output channels to preview.

Show Trace: Select how to display autoscaled sweep traces in the preview

pane:

• Time Course Display all traces in continuous linear

time.

All Sweeps Display all traces overlaid from time zero.

• Sweep # Display a trace from a single sweep.

only for 'Show Channel: All Channels'

Show Measure Regions: A list of measurement regions to preview:

None No regions displayed.

- All regions displayed.
- m[#] Select a single region to display.

Routine Settings

The Routine Settings are split into five main sections. Click on a section header or its items to open its sub-window for editing.




Figure 4-47. Routine Settings

Routine Editor: Acquisition & Routine Parameters

Acquisition timing parameters are controlled in this section, such as sweep duration and sampling rates. The settings in this section are shared by all input and output channels.




Figure 4-48. Acquisition & Routine Parameters

Trigger Action

Control how and when recordings occur.

- Clock Triggered: Start a recording timed by the SutterPatch program. Hardware trigger inputs are ignored.
- Externally Triggered Sweep:

Use an external signal from other laboratory equipment to trigger the recording of each sweep in a Series.

Connect the trigger signal to the 'Trigger In' BNC on the front panel of the dPatch amplifier.

However, if a Routine is run within a Paradigm 'Each Sweep' step, then the hardware trigger is substituted by a software trigger generated by the Paradigm.

Once a sweep has been triggered, additional triggers are locked out, until the sweep has been completed.

• Externally Triggered Series:

Use an external signal from other laboratory equipment to trigger the start of a Series, then operate like Clock Triggered. A command waveform is only generated for the first cycle.

Connect the trigger signal to the 'Trigger In' BNC on the front panel of the dPatch amplifier.

However, if a Routine is run within a Paradigm 'Each Sweep' step, then the hardware trigger is substituted by a software trigger generated by the Paradigm.

• Event Triggered:

Use an amplitude event in an input signal to trigger data acquisition of a sweep. .

Event triggering is useful to reduce extraneous data when infrequent events occur during long recordings.

The Acquisition Mode is set to 'Continuous Sweeps', to continuously monitor the signal for an event trigger.

The refractory period, i.e., the time during which another event trigger cannot occur, is the same as the 'Sweep Start-to-Start Time'.

Event Triggered Settings C

Click on the 'Event Triggered' field to open this dialog.

Input Channel To Scan Trigger on this channel.

The unit of the input channel is used for the trigger threshold.

Pre-Trigger Duration [0-70.0 ms]

Trigger Threshold [±20.000 nA] current input

[±0.75 V] voltage input

Trigger Polarity Rising Edge

Trailing Edge

Minimum Trigger Duration [20.0 µs – 70.0 ms]

Acquisition Mode

• Triggered Sweeps: Each sweep is started by an internal software trigger

from a Routine or Paradigm, or by an external hardware

trigger.

To allow for system delays, there is a short gap (~200 ms) between sweeps, which can include a few microseconds

of jitter.

• Continuous Sweeps: Uninterrupted data without time gaps between sweeps

are recorded when the Sweep Start-to-Start Time is set

to the 'Sweep Duration'.

Data are displayed as successive sweeps, not as a

continuous "rolling" display.

This option does not support:

• Pausing of sweeps during recording.

• Paradigm step 'Each Sweep'.

Very short sweeps.

Warning! Very high data-processing throughput has the potential to

overload system resources and interfere with data processing. (See the Troubleshooting Appendix

Acquisition Q&As.)

Enable Output Waveforms

Output channel waveforms can be optionally disabled.

If outputs are disabled, sweep and segment durations for analysis measurements can be configured in the Input Channels / Edit Signal / Waveform Editor.

If disabled in Continuous Sweeps mode, holding levels can be controlled via the Amplifier Control Panel. And, while metadata settings are only written at the beginning of a Routine, tags are inserted for such additional changes. ..

Also, the amplifier VC/CC mode is set here, as the Output Channels section is unavailable when outputs are disabled.

Restrict To Ensures that the matching headstage is in the proper VC/CC mode, else the Routine cannot be activated or executed.

• VC Mode The Amplifier Control Panel matching headstage must be in VC mode to run the routine.

The default setting for new routines is 'VC Mode'. This

prevents CC mode pA (10⁻¹² A) current outputs from being

accidentally overscaled by VC mode routines using mV (10⁻³ V) voltage outputs.

• CC Mode The Amplfier Control Panel matching headstage must be in CC mode to run the routine.

Note: The amplifier can be switched into any mode (VC or CC) while a recording is in progress. However, it is your own responsibility to correctly interpret data with mixed recording modes.

Number of Sweeps [1 - 65000]

The number of sweeps to record.

- Note: a) When allocating large memory blocks, if more than 1 mega-sample of memory is allocated for the Routine, it can take several minutes to load. If a dPatch amplifier is attached, a progress bar displays; if in demo mode, a message displays "Allocating acquisition buffers, please wait...".
 - b) The largest signal size that SutterPatch can record is 2.5 Gsamples, with up to 16 signals (data waves) recorded at a time. This signal limit is independent of the OS maximum file size limit.

Sweep Cycles [1 - 65000]

The number of times to automatically repeat the entire set of sweeps recorded by a single Series.

Post Stimulation Cycles [for 'Continuous Sweeps' using Output Waveforms.]

After the specified number of Sweep Cycles is reached, continue acquiring cycles with outputs disabled.

Sweep Start-to-Start Time ['s' or 'ms']

The time from the start of recording a sweep to the start of the next sweep recording. For Externally Triggered and Event Triggered acquisition modes, this serves as a "refractory period", during which time additional triggers are not accepted.

Shortest Possible [for 'Triggered Sweeps']

Set to the longest waveform duration in the Series + overhead processing time (200 ms).

Sweep Duration [for 'Continuous Sweeps']

The sweep duration is the longest waveform duration in the Series (as configured in the Waveform Editor.)

Note: Demo mode sweep start-to-start times can vary during acquisition, especially on slower computers.

Persistence Display

For a recording, display each new sweep without erasing any previous sweeps.

- Off Set the Acquisition: Routine scope window into non-persistence display mode.
- On Set the Acquisition: Routine scope window into persistence display mode.
- Keep current setting

Do not change the scope window's prior settings.

Routine Editor: Input Channels

Configure the input channels.




Figure 4-49. Input Channels

Edit Signal

[for disabled Output Waveforms]

When Output Waveforms are disabled in the Acquisition Parameters section, segment timing can still be controlled via a modified Waveform Editor. This signal editor is a "timing only" version of the Waveform Editor, and only controls the duration of segments, to facilitate segment-based measurements.

When this button is enabled, double-clicking a channel will open the signal Waveform Editor.

Channel

Enable up to sixteen Input Channels for recording data:

Current# Analog input current channels hardwired from the dPatch

headstage.

Voltage# Analog input voltage channels hardwired from the dPatch

headstage.

AuxIN[1-8] Eight auxiliary analog input channels allow you to directly

digitize and record input signals from connected non-Suttter

external equipment.

Note: In demo mode, the AuxIN channels display a ±20 mV

sine wave.

Virtual [1-10] Ten virtual channels are available.

Virtual channel data are mathematically transformed data from another input channel, or are entirely computed from an

equation.

Label

A user-editable signal name for a channel.

These labels are used in:

- Routine Settings overview for Input and Output Channels
- 'Parent Out Chan'
- Virtual channel Math Equations and Source Channels
- Scope window signal panes
- Data Navigator Preview pane
- Metadata Input Signal Name

To rename an Input Channel, first enable it, then double-click its 'Label' field, and enter the new name. If the label is used by another channel, an underscore and increment number are appended to the new label.

When a virtual input channel is enabled, a default 'Math Type' label is automatically generated for it.

Unit

The base unit of measurement.

• 'Current' Channels Fixed at 'A' for current.

• 'Voltage' Channels Fixed at 'V' for voltage.

• 'AuxIN' Channels Default is 'V', but is editable.

• 'Virtual' Channels Same as its Source channel; is only editable for

virtual Math Type 'Equation'.

Note: The resolution of the signal units are automatically set.

Parent Out Chan

The "Parent Output Channel" shows which output channel is associated with which input channel. The ouput channel timing is also used for measurements with 'Cursors Relative to Segments'.

Auxiliary Input channels allow you to change the 'Parent Out Chan'.

However, Virtual channels cannot directly select Parent Output channels. Instead, they utilize their Source channel's Parent channels.

Scope Position

The input channel panes can be repositioned in the scope window.

Hide Channel in Scope Window

The selected input channel is hidden in the scope window.

Scaling [for AuxIN and Virtual Input channels]

Equation Enter a fixed value or equation for the input channel scaling factor of

the external signal.

Factor [read-only field]

The input channel scaling factor evaluated from the equation. Raw

values are converted to input units.

Note: The dPatch digitizer uses a high-resolution 18-bit ADC with

64-bit data, so data resolution is not an issue when scaling

input signals.

Offset (V) Apply an amplitude offset to the input signal (before any scaling).

Tip: To use 'mV' units, enter: '#m' or '#e-3'

To use 'pA' units, enter: "#p' or '#e-12'

Sampling Rate [for Auxiliary input channels]

[kHz (µs)] Rate (interval) [1, 2, 5, 10, 20, 50, 100, 200 kHz]

Reduce the Auxiliary input channel signal from an initial fixed 200 kHz sampling rate.

Downsampling is done by simple averaging, i.e., at a 1 kHz bandwidth, each data point is the mean of 200 original sample points.

Filter Bandwidth [for Headstage input channels (Current#, Voltage#), or a read-only field for Virtual input channels [

[kHz]

Set the input-channel filter bandwidth using a low-pass 8-pole Bessel filter:

Use Control Panel Set the 'Current' and 'Voltage' channel bandwidths via the Amplifier Control Panel.

Zero to [100, 200, 500 Hz; 1, 2, 5, 10, 20, 50 100, 250, 500, 1000 kHz]

This low-pass filter acts upon the selected headstage input signal input sampling rate. Current and Voltage signals from the same headstage share the same filter bandwidth.

Different headstages and virtual input channels can have different filter bandwidths applied to them.

Warning! Processing very large data sets can overload system resources.

Tip: For experiments where the shape of the response is of interest (such as action potentials, minis, etc.), an input filter rate of 10 kHz is commonly used.

However, for very long stimuli, you might want to use a lower input filter rate.

Channel Timing Delays

The dPatch amplifier uses analog output channels that pass signals through fixed DAC output stimulus filters, and uses input channels that pass signals through variable-bandwidth ADC input filters. These filters impose a time delay (phase delay) on the physical input and output channels, which is reflected in the recorded response signals.

However, when headstage stimulus signals are recorded (such as 'Voltage1' in voltage-clamp mode), their output and input channels are directly processed as digital streams. Thus, there are no analog filter delays in recorded stimulus signals.

This means that recorded response signals are time-delayed compared to recorded stimulus signals. Therefore, when analyzing data that requires precise timing between headstage stimulus and response signals, these timing delays should be taken into account.

Typical dPatch Filter Delays

<u>Input Bandwidth</u>	<u>Total Filter Delay</u>
$1000~\mathrm{kHz}$	$16.2~\mu s$
$500~\mathrm{kHz}$	$17.0~\mu s$
$250~\mathrm{kHz}$	$18.5\mu \mathrm{s}$
$10~\mathrm{kHz}$	$85.0~\mu s$

Note: In Demo mode, demo data uses the Nyquist Factor sampling rate timing instead of the Filter Bandwidth timing.

Keep Filter Setting at Series End

[for Headstage input channels]

When a filter bandwidth is selected in the Routine, it replaces the Amplifier Control Panel I-Filter or V-Filter setting while the Routine is acquiring Series data. When acquisition stops, the Control Panel filter normally returns to its original setting.

Enable this option to keep the Routine filter value as the Control Panel filter value after acquisition stops.

Nyquist Factor

[for Headstage input channels (Current#, Voltage, or a read-only field for Virtual channels from Parent Out channels]

To manage the quantity of data being processed, the hardware headstage sampling rate is automatically adjusted, downsampled from an initial fixed 5 MHz rate to a "Nyquist Factor" mulitple of the 'Filter Bandwidth'.

According to the Nyquist sampling theorem, the input sampling rate should oversample the input filter rate at a minimum of 2x, However, the Nyquist Factor is typically implemented at 5x - 10x for cellular responses, due to their complex shapes.

- 2 The sampling rate is set to twice (2x) the filter rate; the minimum Nyquist-limited sampling rate must be at least double the filter rate.
- 5 The sampling rate is set to five times (5x) the filter rate.
- 10 The sampling rate is set to ten times (10x) the filter rate.

For the 1 MHz Filter Bandwidth

- 2.5 The sampling rate is set to two-and-a-half times (2.5x) the filter rate; this downsamples the initial 5 MHz data by 50%.
- 5 The sampling rate is five times (5x) the filter rate; this uses the maximum sampling rate of 5 MHz, i.e., no downsampling occurs before filtering.

Warning! Processing very large data sets can overload system resources.

Sampling Interval

```
[read-only field]
```

The duration of a single data sample.

Integrator Reset

[for Math Types 'BesselFilter' and 'Smooth' in amplifier Capacitive mode]

- Ignore Capacitive-mode transients are displayed in the data.
- Blank The data during capacitive transients are made invisible by replacing those data points with NaNs (Not a Number).
- Mask The data during capacitive transients are replaced by the last data value before the transient discharge, simulating a sample-and-hold operation.

Blank/Mask Duration

```
[50.00 \, \mu s - 1.00 \, s]
```

The default value of $500~\mu s$ should be sufficient to encompass the reset transient duration.

Mode Switch Acquisition

only shows if enabled in the Routine Settings menu

[for Current channels in VC mode; for Voltage channels in CC mode]

Enable and/or configure fast (μs vs. ms) VC <> CC mode switching during acquisition.

- OFF Disable fast mode switching.
- ON Enable fast mode switching.
- Edit Configure fast mode switching.

Mode Switching Settings

Mode Switch Path VC >> CC

First Switch Type

• Fixed Position

Switch At $[10.0 \,\mu\text{s} - 56.0 \,\mu\text{s}]$

Relative To Threshold

Delay From Threshold [$10.0 \mu s - 56.0 \mu s$]

Threshold $[\pm 20.0 \text{ nA}]$

Threshold Polarity

- Rising Edge
- Trailing Edge

Second Switch

- OFF
- ON

(Mode Switch Path VC >> CC >> VC)

Switch At $[10.0 \,\mu\text{s} - 56.0 \,\mu\text{s}]$

Virtual Input Channels

Virtual input channels allow you to perform a variety of mathematical transformations on input signals in real time. To enable a virtual signal, highlight a signal name. When a virtual input channel is enabled, its configuration fields are ungrayed.

Math Type Apply a data transformation to a virtual input channel.

BaselineSubtract

BesselFilter

CycleAverage

Differentiate

DownSample

Equation

Integrate

Leak

LineFreq

LockIn

Smooth

Stimulus

SweepAverage

SweepSubtract

• BaselineSubtract

Subtract a fixed value from all data points in an input trace.

This is useful for adjusting an offset or resetting a baseline.

Post-analysis is limited to marked sweeps if the Reanalysis scope Measurements button / Edit Virtual Signals. Unmarked sweeps are set to NaNs.

Source Channel Select an input channel to process.

Baseline From Select how to calculate the subtraction

value.

• Value Subtract a fixed value.

Value Spinner adjusts in 1 pA or 1 mV

increments.

• Trace Subtract the average of the entire input

trace.

• Sweep Time Subtract the average of the data between

the Start Time and End Time.

Start Time Set the starting time of the data

to be averaged.

End Time Set the ending time of the data to

be averaged.

• Segment #s Subtract the average of a Segment.

Start Ratio Set the starting time of the data

to be averaged, as a ratio relative

to the starting time of the

Segment duration.

Start Time [derived value]

End Ratio Set the ending time of the data to

be averaged, as a ratio relative to the ending time of the Segment

duration.

End Time [derived value]

• BesselFilter A frequency-domain filter with excellent response characteristics for preserving the shape of a biological signal.

Source Channel Select an input channel to filter.

Filter Bandwidth Select a frequency range.

• LowPass Allow signal frequencies less than the

cutoff frequency, and block all higher frequencies, such as high-frequency

noise.

• HighPass Allow signal frequencies greater than

the cutoff frequency, and block all lower

frequencies.

Filter Order [1, 2, 4, 8]

Number of "poles" in the filter. A higher number provides a sharper (more

accurate) response, but consumes more processing time and system resources.

Cutoff Frequency [$0.01 \text{ Hz to} < \frac{1}{2} \text{ the sampling rate}$]

Restrict frequencies from this boundary

point onwards.

Apply Filter Delay Correction

Correct the signal for estimated digital

filtering delays by shifting the signal

forwards in time.

Integrator Reset For amplifier Capacitive mode.

• Ignore Capacitive-mode transients are

displayed in the data.

• Blank The data during capacitive transients

are made invisible by replacing those data points with NaNs (Not a Number).

The data during capacitive transients

are replaced by the last data value before the transient discharge, simulating a sample-and-hold operation. Blank/Mask Duration $[50.00 \, \mu s - 1.00 \, s]$ The default value of 500 µs should be sufficient to encompass the reset transient duration. CycleAverage Apply averaging across cycles for each numbered sweep. Post-analysis can be limited to marked sweeps via the Reanalysis scope Measurements button / Edit Virtual Signals. Source Channel Select an input channel to average. Differentiate Apply differentiation to an input signal. The instantaneous rate of change in the signal is displayed. Source Channel Select an input channel to differentiate. Downsample Apply downsampling to an input signal, i.e., reduce the sampling rate of the signal data. Source Channel Select an input channel to downsample. Source Sampling Interval [read-only field (μs)] Reduction Factor [2-100]Only whole numbers are used; non-whole numbers are rounded up or down. Sampling Rate [read-only field (Hz)] New sampling rate of the reduced signal. Equation Specify an equation to process an input signal. Post-analysis can be limited to marked sweeps via the Reanalysis scope Measurements button / Edit Virtual Signals. Source Channel Select an input channel to process. Equation [1 Click field to access the 'Specify math equation' editor.

Mask

Note: The full equation is always visible

as a tool tip, by hovering the mouse cursor over the 'Math

Equation' field.

Specify math equation for virtual signal

[<equation>] A free-form text field.

Errors are reported under this field.

Check Equation Check the equation syntax. The

equation is evaluated for sweep #1, and

if valid, it reports "Syntax is ok".

Insert special identifier

A limited set of identifiers are available for virtual equation traces. However, special references can also be used within commands.

- p[#] n'th paradigm variable
- s[series-count, sweep-count, trace-count, routine name]

Access an arbitrary input trace (data wave) via counts of Series #, Sweep #, Trace # (scope channel position), and the Routine name.

The "current" item is the "active" trace in the scope window, and has a count value of zero.

If a "count" number is non-zero, it is used as an offset from the current count value of zero. Any fractions in count numbers are truncated to integers.

If the routine name is left blank, the current routine name is used.

Ex: s[0,0,0,]

The current series, current sweep, current trace, of the current routine.

• t[#] n'th input trace

Access the input trace (data wave) in

scope channel position "n" for the last sweep of the current Series. This numbering can differ from the Scope Position "n", if signals are re-arranged or hidden.

Tip: You can duplicate an input signal

with this.

Undo All changes in the equation editing

session are discarded.

(See the Equation Editor for more details.)

• Integrate Display the integral of the data signal. This is equivalent to the signed area under a curve.

Source Channel Select an input channel to integrate.

• Leak Remove leakage current from the data signal. This is the small passive current when the cell is in a resting state.

This feature is only enabled when the Routine includes an output channel with P/N Leak Pulse enabled.

Source Channel Select an input channel to process.

Show Leak Displays the accumulated leak currents

after the subtracted data in a sweep.

Leak Zero Segment Identify a segment with no active

cellular response to the command signal.

When set to zero, the field is set to 'OFF'. To re-display the numeric spinners, enter a non-zero number into the field.

Note: The mean of the second half of the specified segment is used to compute an averaged leak current, which is then used to correct the P/N leak average. This option reduces the influence of a constant leak-current, which is otherwise added to the leak current of the main pulse.

Enable

• LineFreq Remove AC line frequency noise (hum) from the data signal.

Source Channel Select an input channel to process.

Line Frequency 60 Hz Canada, (Caribbean),

Central America, (Japan), Mexico, (South America), South Korea, Taiwan,

USA.

Some (regions) include both 50 Hz and 60 Hz

frequencies.

50 Hz Most of rest of world.

Alternating current (AC) power contains 50 or 60 Hz oscillations that can cause sinusoidal line-frequency noise in recorded signals. This FFT-based filter reduces such noise by > 90% over 6 harmonics. The adjusted signal is displayed in real time.

Note: When using short sweeps or slow sampling, performance might improve with a larger number of sample points, such as with an increased sweep duration or filter bandwidth.

Measure cell characteristics (such as membrane capacitance) with high signal-to-noise sensitivity, using a dual-phase software lock-in amplifier.

Note: This feature is only enabled when the Routine includes an output channel with a waveform Segment set to 'Sine / Sine Wave Cycles / For LockIn.'

Calculations are made using 'conductance' instead of 'resistance'.

Current Channel Select a (source) input channel with a "current"

signal.

Trace Kind Select the LockIn measurement to display.

The selected 'Trace Kind' is automatically set as

the Virtual Channel label.

CM Computed membrane

capacitance.

GM Computed membrane

conductance.

GS Computed series conductance.

DC DC component of measured

signal.

RealY Real number part of the lock-in

response signal.

ImagY Imaginary number part of the

lock-in response signal.

Cycles to Average [1-1000]

Cycles to Skip [1-1000]

V-reversal $[\pm 1000 \text{ mV}]$

When using a calculated stimulus trace, enter the reversal potential for the ion under study, such as for (Na+) sodium spikes or (K+)

potassium tail currents.

Note: See the SutterPatch Algorithms appendix for the reference and math used in the LockIn computation.

• Smooth the data with a "moving average" noise-reduction filter.

Source Channel Select an input channel to smooth.

Smoothing Type:

• Gausian A standard filter with excellent 10 – 90%

rise-time response.

Smooth Operations [1-32767]

of smoothing operations to perform.

• Boxcar A fast time-domain filter with excellent 0

- 100% rise-time response.

Smooth Repetitions [1-32767]

of smoothing repetitions to perform.

Boxcar Window Points [1-99]

of points in boxcar sliding window.

Note: For best performance, only odd

values are used.

Integrator Reset For amplifier Capacitive mode 'Current'

channels.

• Ignore Capacitive-mode transients are

displayed in the data.

• Blank The data during capacitive transients

are made invisible by replacing those data points with NaNs (Not a Number).

• Mask The data during capacitive transients

are replaced by the last data value before the transient discharge, simulating a

sample-and-hold operation.

Blank/Mask Duration [50.00 us - 1.00 s]

The default value of 500 µs should be sufficient to cover the reset transient

duration.

• Stimulus Replicate the command waveform.

Source Channel Select an input channel – the waveform

from its 'Parent Out Chan' is used.

• SweepAverage Average the input traces.

Post-analysis can be limited to marked sweeps, via the Reanalysis scope Measurements button / Edit Virtual Signals.

Source Channel Select an input channel to average.

Average Type Cumulative Average all processed

sweeps together.

Run Average Average the last "N"

sweeps.

Number of Sweeps [for Run Average]

Start Sweep Sweep number to start sweep averaging.

Set Sweep < Start Sweep To NAN

Sweeps prior to the Start Sweep are set by default to the initial source sweep. Enable to set these pre-sweeps to NaNs.

End Sweep number to end sweep averaging.

Set Sweep > End Sweep To NAN

Sweeps after the End Sweep are set by default to the initial source sweep. Enable to set these post-sweeps to NaNs.

• SweepSubtract Subtract a sweep from the input trace.

Post-analysis can be limited to marked sweeps via the Reanalysis scope Measurements button / Edit Virtual Signals. Unmarked sweeps are set to NaNs.

Source Channel Select an input channel to process.

Reference Sweep Select a sweep to be subtracted from all

other sweeps. If the sweep does not yet

exist, no subtraction occurs.

Routine Editor: Output Channels & Waveform

Configure the output channels and command waveform.




Figure 4-50. Output Channels & Waveform

[HW Status field] Hardware information on the selected channel is displayed.

Digital Outputs set as

The digital output channels (bits) can be set individually or as a group.

• Individual bits Each DigOUT (1-16) bit is individually set in its own Waveform Editor table.

The waveform preview uses the bit's binary word value for its Y-axis value, i.e., if bit 3 is 'HIGH', it has a "word" value of 4.

• 16-bit Word The 16-bit digital output pattern is controlled by a single

decimal number (0 - 65,535), which is also the waveform preview amplitude value.

The waveform preview uses the binary bit pattern word value for its Y-axis value, i.e., if bits '1' and '3' are 'HIGH', the word has a value of 5.

Edit Waveform

Click the Edit Waveform button to access the Waveform Editor table and create a stimulus waveform. (See the Waveform Editor section below.)

Channel

Click on the Output Channel checkboxes to enable analog and digital output channels in the Routine. Click on an enabled channel name to highlight and select it – the channel output parameters are displayed for configuration. Double-click an enabled channel name to open its stimulus waveform in the

StimOUT

Waveform Editor.

The default StimOUT channnels are hardwired to the dPatch headstages.

For StimOUT channels, the actual DAC output signal is passed through a 20 kHz low-pass filter before entering the headstage.

AuxOUT

The AuxOUT auxiliary analog output channels can be used to send stimulus waveforms to external instruments.

HS#GStream

Each headstage supports a dynamic clamp output channel.

DigOUT

The digital outputs are available as either a single 16-bit "word", or as 16 individual 1-bit channels, as set in the Acquisition & Parameters section.

Tip: If a signal is connected to the front panel 'COMMAND IN' BNC, that signal is summed with the StimOUT waveform that is sent to the headstage.

Label

A user-defined signal name for the channel.

These are used in:

- 'Copy Channel'
- Waveform Preview pane 'Show Channel'
- Metadata: Output Signal Name

To rename an Output Channel, first enable it, then double-click its Label field and enter the new name. If the same label is reused for another channel, an underscore and increment number will be appended to the new label.

Copy channel

Copies one channel's waveform to another channel, for output channels of the same type (i.e., "Stim", "Aux", or Digital). If a channel is enabled, then highlighting another or blank channel of the same type ungrays the 'Copy channel' field, and changes it from 'OFF' to 'None', with a drop-down list of available channels to copy from.

Restrict to

Ensures that the matching headstage is in the proper VC/CC mode, else the Routine cannot be activated or executed.

• VC Mode The Amplifier Control Panel matching headstage must be in VC mode to run the routine.

The default setting for new routines is 'VC Mode'. This prevents CC mode pA $(10^{-12} \, \text{A})$ current outputs from being accidentally overscaled by VC mode routines using mV $(10^{-3} \, \text{V})$ voltage outputs.

• CC Mode The Amplfier Control Panel matching headstage must be in CC mode to run the routine.

Dynamic Clamp output channels are automatically restricted to CC mode.

The dPatch amplifier can be switched into any mode (VC or CC) while a recording is in progress. In this case, tags are inserted into the data record with the new Control Panel settings. However, it is your own responsibility to correctly interpret data with mixed recording modes.

Unit

Enter the base unit of measurement. The signal unit resolution is automatically adjusted.

• StimOUT Channels

Fixed at 'V' for voltage clamp and "other" experiment types; fixed at 'A' for current clamp experiments.

• AuxOUT & HS#GStream Channels

Can be edited to any setting.

Scaling [for AuxOUT channels]

Equation (V/V)

Apply a scaling factor as a fixed value or an equation.

(See the 'Equation Editor' for more details.)

Factor Read-only field of Equation value.

Offset $[\pm 10.000 \text{ V}]$

Apply an offset to the output channel (before any scaling.)

Relative to Holding

If 'Relative to Holding' is enabled, the headstage output signal is the command waveform summed with the "holding" level in the Amplifier Control Panel. If the holding level is set to '0', this setting has no effect.

For the Auxiliary channels, the command waveform is summed with the Amplifier Control Panel I/O Auxiliary Output settings.

For digital channels, the command waveform is relative to the Amplifier Control Panel I/O Digital Output settings.

Note: When enabled, this setting is updated "live" by holding level changes in the Amplifier Control Panel.

Sweep End Output Settings

Control how the amplifier output levels (including I/O Auxiliary and Digital Output levels) are handled when the system is not acquiring data.

Inter-Sweep

This is the time between sweeps - after a sweep ends, but before the next sweep starts.

• Use Waveform Value Set the output signals to their last values in the

command waveform, at the end of a sweep.

Use to avoid generating a short (potentially disruptive) glitch in your preparation, caused by

returning to holding levels at the end of a

sweep.

• Use Holding Set the output signals to the Amplifier Control

Panel "holding" levels, at the end of a sweep.

This ensures that your cells are kept in a

resting state as much as possible, and that each output sweep starts from the same holding level.

Routine End

This is the time after the Routine ends, until the next Routine starts.

• Use Waveform Value Set the output signals to their last values in the command waveform, at the end of a Routine.

Use to avoid generating a short (potentially disruptive) glitch in your preparation, caused by returning to the holding levels at the end of a Routine.

• Return to Holding Set the output signals to the Amplifier Control Panel "holding" levels, at the end of a Routine.

This ensures that your cells are kept in a resting state as much as possible.

Note: In demo mode, after changing holding levels, if 'Relative to Holding' is enabled, then the 'Return to Holding' levels are updated immediately, otherwise they are only updated when a Routine is activated.

Displays the P/N Leak Pulses section.




Figure 4-51. Output Channels & Waveform with Leak

Endogenous leak currents can flow, even while a cell is in its resting state, from conditions such as an imperfect or leaky seal, or via existing ion channels, and affect response amplitudes. Online P/N leak subtraction automates the removal of such currents from the data.

If endogenous leak conductance is an issue with your cell type, and/or high temporal resolution is required along with a need to reduce capacitive transients (e.g., with voltage-gated sodium currents), click 'Enable P/N Leak Pulses' and configure its settings below.

A "leak pulse" is a replica of the stimulus waveform, and is used to record a fraction of the leakage current. In this technique, leak pulses are generated, and the responses are averaged, scaled, and subtracted from the main response to remove the effects of leakage.

Note: The sub-pulses are stored as part of the sweep. This ensures that if any events occur during the sub-pulses or between the sub- and main pulses and causes unexpected or hard-to-interpret effects, the full original recording condition can be examined.

Preview Leak Pulses

Display the leak subtraction pulses in the Routine Editor Waveform Preview panel. A leak subtraction pulse is a scaled copy of the main stimulus waveform.

Alternating Leak Polarity

You can reduce directional bias in the leak conductance by alternating the polarity of the leak subtraction pulses on a sweep-by-sweep basis, as long as no ion channels are activated.

Num Leak Pulses

Set the number of leak pulses used to average out noise and leak conductance. Adjust this number in accordance with the amount of noise in the signal. With the high precision of modern 16-bit digitizers, this number can sometimes be reduced to less than 4 leak sub-pulses.

Note: As each leak pulse replicates the stimulus waveform, larger numbers of leak pulses is not recommended, as this can greatly increase the total duration of a sweep during acquisition, and the noise in the sub- and main pulses can add up and actually increase.

The default setting of '4' Leak Pulses, when used with the default Leak Ratio (-0.250) operates equivalently to pCLAMP's default P/N setting (4 subsweeps for P/4).

Leak Ratio

Set the leak subtraction pulse size relative to the main waveform pulse, using a ratio between +1 and -1. The setting should be low enough that no electrically-gated ion channels are activated. For instance, a Leak Ratio setting of 0.25 will generate leak pulses at ¼ the amplitude of the main stimulus waveform, while a Leak Ratio of 0.2 will generate leak pulses at 1/5 the main pulse amplitude.

Note: The program scales the leak subtraction pulses based upon the Leak Ratio setting, not the number of Leak Pulses. This means that the Leak Ratio can be set independently from Num Leak Pulses, instead of those settings being

interdependent.

Tip: As an alternate way to avoid electrical activation of ion channels, use a negative ratio to reverse the polarity of the leak pulses relative to the main pulse.

Leak Hold

The leak pulses holding level can be set differently from the Routine main holding level, for flexibility in finding a suitable leak pulse voltage range. The scaled waveform amplitudes are measured relative to the Leak Hold level, but are subtracted relative to the dPatch holding level.

Set to a fixed value, or enter as an equation. (See the 'Equation Editor' for more details.)

Check Equations Check the equation syntax. The equation is

evaluated for sweep #1, and if valid, it reports

"Syntax is ok".

Insert special identifier Acquisition, amplifier and reference settings are

available for use in equations. (See list in Equation Editor.)

Undo All changes in the equation editing session are

discarded.

Leak Pulse parameters common to all D/A channels

Execute Leak Pulses

Leak pulses can be set to run before or after the main waveform pulse:

Before Main Pulse
 Sub-pulses are output relative to the Leak Hold

level. After the sub-pulses complete, the signal goes to the dPatch Holding level for the duration

of the Leak Delay before the main pulse.

• After Main Pulse After the main pulse completes, the signal goes to

the Leak Hold level for the duration of the Leak Delay setting, and then outputs sub-pulses

relative to the Leak Hold level.

Leak Delay [$1 \mu s - 3.6 ks$]

If a settling time is needed between the leak pulses and the main waveform pulse, Leak Delay will insert a time delay between the execution of the leak pulses and the main pulse. Provide enough time to avoid interference of the leak pulses with any active currents or inactivation of ion channels.

When leak pulses occur before the main pulse, Leak Delay uses the amplifier's Holding level; when leak pulses occur after the main pulse, Leak Delay uses the Leak Pulses 'Leak Hold' level.

Enable A/C Line Frequency adjustment

The effect of AC line-frequency noise (hum) can be automatically reduced during P/N leak subtraction recording:

• 60 Hz Enable the reduction of 60 cycle AC line noise.

• 50 Hz Enable the reduction of 50 cycle AC line noise.

This Line Frequency adjustment automatically calculates the proper interpulse interval for the P/N sub-pulses, so that they are counter-phased to the line frequency of the main output signal, which reduces hum without filtering the signal.

Routine Editor: Waveform Editor

Click the 'Edit Waveform' button to open the Waveform Editor and design a command waveform for the selected output channel.




Figure 4-52. Waveform Editor

Close button Use this button to close the Waveform Editor window.

Used segments [# out of 50]

Up to 50 contiguous segments can be configured in a waveform.

Actions

OFF Unused segments are labeled as 'OFF'.

Tip: A segment with a Duration of '0' ms is

equivalent to 'OFF'. This is a convenient way to

skip a segment instead of deleting it.

Stored Enable a segment for stimulation and recording.

Not In Leak If P/N LeakPulses are enabled, this will optionally

exclude the segment from being generated in the P/N

Leak Subtraction output wave.

This is useful for inactivation or recovery studies, when

commands do not change for long periods of time.

Insert a default Segment into the current position, and

increment the position of the following Segments, i.e..

move them to the right.

Copy To copy a segment, click the segment's Actions list and

select 'Copy'. A copy is inserted as the next segment.

To copy multiple segments, select the segments to be copied. Then, for the segment to be inserted before, click its Actions list, select 'Copy', and enter the number of times to copy the segments - the selected segments are inserted before the "Copy" segment.

Delete To remove a segment, select its 'Delete' Action.

If there is only one segment, it cannot be deleted there is always at least one segment enabled.

To remove multiple segments, select the desired segments. Then, click any segment's Actions list and select 'Delete'. All selected segments are deleted.

To select multiple segments, in Windows use Ctrl-click, or in macOS use Command \(\mathbb{H}\)-click, to highlight each segment, or use Shift-click to highlight a range of segments.

Any following segments shift their Segment #'s down by the number of deleted segments.

Waveform Select the waveform shape.

For Waveform types Sine / Chirp / Squarewave /

Template / Triangle, a 'Parameters' field displays below this field, to allow quick access to their parameters.

Step The waveform amplitude rapidly jumps from a pre-

existing level to the new level within one sample point, and stays at the new level for the duration of the segment. The resulting waveform shape looks like a

step.

The first segment consists of a Step waveform set to the holding level amplitude (Set to Hold).

Ramp A straight line connects the last point of the preceeding

segment (or holding level) to the last point of the

current segment.

Sine The waveform is a sinusoidal wave.

Sine Wave Cycles Multiple One or more cycles.

Single One cycle, where

the Cycle Duration is equal to the Segment Duration.

For LockIn For sensitive

capacitance measurements.

A corresponding virtual input 'LockIn' channel also needs to be enabled.

LockIn sine waves cannot be mixed

with non-LockIn sine waves.

Multiple sine LockIn segments in a waveform share the same settings (except duration) for each segment. Amplitude

 $[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$

Amplitude of the first peak from the sine wave baseline.

- Value
- Var_r[1] [16]

Tip: To offset a sine wave from the default baseline (0 units), set the segment Amplitude value, or enable Routine Editor / Output Channel 'Relative to Holding'.

For LockIn measurements, the larger the sine wave amplitude, the better the signal-to-noise ratio for the measurements, just be sure to avoid the activation range of voltage-gated ion channels.

Cycle Duration

[for LockIn only]

One cycle length (ms).

A preset list of cycle durations:

$2.0~\mathrm{ms}$	(500 Hz)
1.0 ms	(1 kHz)
$0.5~\mathrm{ms}$	(2 kHz)
$0.2~\mathrm{ms}$	(5 kHz)
0.1 ms	(10 kHz)
$0.048~\mathrm{ms}$	(20.83 kHz)

Ramp Increment

 $[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$

Apply the sine wave onto a ramp "baseline".

Not enabled for a Single cycle.

Segment Duration Sine wave duration (ms).

Squarewave	The waveform genera	ates a train of rectangular pulses.
E qual o man o	1110 ((0101111 8011010	ares a train or rectaing mar purses.

$$[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$$

Increment the baseline amplitude for each succesive pulse.

Step1 Amplitude [
$$\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}$$
]

Amplitude of first pulse.

- Value
- Var_r[1] [16]

- Value
- Var_r[1] [16]

Step2 Amplitude [
$$\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}$$
]

Amplitude of second pulse.

- Value
- Var_r[1] [16]

- Value
- Var_r[1] [16]

=

Segment Duration Square wave duration (ms).

Chirp This waveform generates a sinusoidal wave that changes its frequency over time.

Chirp Type Linear A linear change in

frequency.

Geometric A geometric change

in frequency.

Note: For a geometric

chirp, a minimal frequency spread is enforced: the End Frequency has to be at least double the

Start Frequency, or half or less of the Start Frequency.

Amplitude $[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$

Start Frequency [1 - 50000 Hz]

End Frequency [1-50000 Hz]

Segment Duration Chirp wave duration (ms).

Template Assign an arbitrary waveform to a segment.




Figure 4-53. Template Waves

[Status field] The Routine segment # and output sampling rate are displayed.

If the template sampling rate does not match a Routine sampling rate, the template data are interpolated to match the Routine sampling rate.

Note: Changes made in the Template Editor are only applied to Routines when the modified template in the Template Editor Pool list is copied into the Routine.

Template Editor Pool

Lists the templates loaded in the Template Editor, plus any extracted templates.

--Copy the selected template wave from the Template Editor Pool into a Routine and Segment. This button is enabled if the 'Number of Sweeps' allows more Segment templates. • Copy the selected template wave from the Template Editor Pool into the Routine. Up to 16 template waves can be loaded. 4 Copy the selected template wave from the Routine into the Template Editor Pool. Copied In Routine Lists the templates copied from the Template Editor Pool and loaded into the Routine. Each output channel can have a maximum of 16 template waves loaded in its Routine. Each template can be used in multiple segments, and each segment can use multiple templates. While the most used case will probably be a single template paired with a single segment, the possibilities are endless. Note: To avoid unnecessary increase in the size of the Routine Pools, only include templates that are actually going to be used in in a Segment. Duration (ms) The duration of the template trace. Copy the selected template in the Routine into the Segment. Remove the template from the Segment, or remove an unused Routine template. Used In Segment Lists the loaded templates that are actually used in the Routine. Each segment can use multiple templates.

> If multiple templates are copied into the Routine Segment, they will be executed in sequential or-

der, one template per sweep.

Sweeps Number of sweeps in the Routine.

Sweeps are assigned to templates in sequential order. If the number of sweeps is greater than the number of templates, the sweep number cycles back to the first template and continues incrementing the templates, etc.

Template Pool Selection

A preview of the selected template signal.

Set to Template Duration

Set the segment duration to match the template (sweep) duration.

Segment Duration The Segment Duration can be manually adjusted

here.

When typing in a value, if no unit type is entered, the unit type defaults to seconds (s). If you enter a number followed by an 'm' or 'ms', the unit type is milliseconds (ms).

Scale In Routine Template

Factor Set a scaling factor for the ampli-

tude of the template signal.

Note: When using templates in Igor Pro 8, close the Scope and Data Navigator windows before saving the Experiment, else the Ex-

periment might not re-open properly.

Triangle The waveform generates a train of triangular pulses.

Base Amplitude Increment

 $[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$

Increment the baseline amplitude

for each succesive pulse.

Peak Amplitude [$\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}$]

Amplitude of the triangle pulse.

Ramp1 Width Duration of the initial phase.

Ramp2 Width Duration of the secondary phase.

Segment Duration Duration of the triangle train.

Membrane Test

The Membrane Test runs in 'Cell' mode.

The waveform applies a predefined negative pulse step (-5 mV) with a 50% duty cycle for Routine-based measurements.

Use with the Real Time Measurements 'Analysis Functions':

- MT Series Resistance
- MT Membrane Capacitance
- MT Membrane Resistance

Warning!

For valid results, 'Cell Compensation' and Rs Correction should be disabled in the Amplifier Control Panel.

Amplitude (analog) Set the waveform amplitude for a Segment.

For the Chirp, Sine, Squarewave and Triangle waveforms, this is used as a baseline offset.

For Auxiliary output channels, when the Output Channel / Scaling Factor is not "1", i.e., when scaling is applied to the signal, then a non-editable scaled output field is also displayed below the amplitude value field.

Set to Hold

Use the Amplifier Control Panel holding level for the Segment amplitude.

For voltage-clamp experiments, records the leak current along with the actual holding voltage.

For current-clamp experiments, records the actual cell potential along with the actual holding current.

Avoid using the last Segment for this, as poststimulation data might be recorded, such as from tail currents.

Tip: To help intepret your data, record an initial baseline in Segment 1, and/or a final baseline in the last segment.

Value

 $[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$

Use a single number for the segment amplitude.

Value List

Set an arbitrary segment amplitude for each sweep.

[Sweep | Value] [# | $\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}$]

For each numbered sweep, enter

a value.

Fill Remaining List Copy the active value to all

remaining sweeps in the list.

Segment Duration Adjust the duration of the

Segment.

Number of Sweeps Adjust the number of sweeps in

the Routine.

Value+Increment Increment the segment amplitude for each sweep.

Start Value $[\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}]$

Holding

• Value

Increment Value [$\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}$]

Segment Duration Adjust the duration of the

Segment.

Number of Sweeps Adjust the number of sweeps in

the Routine.

Equation Specify the segment amplitude as an equation.

[Equation field] [$\pm 0.75 \text{ V}, \pm 20.0 \text{ nA}$]

A free-form text field.

Errors are reported under this

field.

Check Equations Check the equation syntax. The

equation is evaluated for sweep #1, and if valid, it reports "Syntax

is ok".

Insert special identifier

Acquisition, amplifier and reference settings are available

for use in equations.

(See list in Equation Editor.)

Undo All changes in the equation

editing session are discarded.

(See the Equation Editor for more details.)

Note: Computing an equation for an output wave consumes significant computing power, as every data point needs to be computed by the CPU. While a slight update delay in such operations is expected, for computers with marginal computing power, the "beach ball" icon displays while the computer is unresponsive and busy processing.

Var_r[1] – [16] Variable labels are displayed if the Routine Variables table is enabled.

Amplitude (digital) Digital settings are displayed if digital outputs are enabled.

Set a digital level for an individual bit.

• LOW = 0

HIGH = 1

Bit Word Values are the decimal number of a 16-bit word (0 - 65,535), also displayed as a bit pattern.

Duration Set the segment duration.

Bit

Set to Hold Use the Amplifier Control Panel to set the output

holding levels.

Value [0-12 ks]

Use a single number for the segment duration.

When typing in a value, if no unit type is entered, the unit type defaults to seconds (s). If you enter a number followed by an 'm' or 'ms', the unit type is

milliseconds (ms).

Value List Set an arbitrary segment duration for each sweep from

a list of numbers.

[Sweep | Value] [# | 0 - 12 ks]

For each numbered sweep, enter

a value.

(Blank lines are removed.)

Fill Remaining List Copy the active value to all

remaining sweeps in the list.

Number of Sweeps Adjust the number of sweeps in

the Routine.

Value+Increment Increment the segment duration for each sweep.

Start value [ms]
Increment value [ms]

Number of Sweeps Adjust the number of sweeps in

the Routine.

Equation

Specify segment duration as an equation.

[Equation field] A free-form text field.

[Errors are reported under this field.]

Check Equation Check the equation syntax. The

equation is evaluated for sweep #1, and

if valid, it reports "Syntax is ok".

Insert special identifier

Acquisition, amplifier and reference settings are available for use in

equations.

(See list in Equation Editor.)

Undo All changes in the equation editing

session are discarded.

(See the Equation Editor for more details.)

Var_r[1] – [16] Variable labels are displayed if the Routine Variables

table is enabled.

Segment Controls

To copy or delete a Segment, click on its background area, and its color turns gold. This enables the additional Actions items: Copy and Delete.

Standard mouse behaviour is used to select multiple Segments:

• Individual Segments: Windows: Ctrl-click

macOS: Command (光)-click

• Range of Segments Both: Shift-click

To step through the Segments, click on the left or right Segment arrow buttons.

To scroll through the Segments, use the slider at the bottom of the Waveform Editor.

Routine Editor: Real Time Measurements & Graphs

Online analyses are configured in the Real Time Measurements & Graphs dialog. Measurement regions display in the Acquisition: Routine scope window, and their associated analyses are plotted in an Analysis sub-window during acquisition.




Figure 4-54. Real Time Measurement Settings

Disable Execution of Measurements and Graphs

Block execution of all measurements and analyses with one click.

Measurement Settings

On Enable an analysis to run.

ID Measurement regions are identified with an ID number: m[#] [1-16]

Label These measurement labels display in the Waveform Preview and scope windows,

and can be used in equations. A Label name is automatically generated from the

Analysis Function; double-click to manually edit.

Copy Copy to the selected Measurement another Measurement's settings.

Clear All Measurements

All measurements and their settings are cleared.

Analysis

Be sure to set appropriate Region Timing (below) for the following analyses.

Analysis Function

Select a predefined Analysis statistic for each measurement:

Absolute area Negative area values are converted to positive and

summed with the positive area values.

Absolute peak Largest absolute value.

AP Duration Action potential duration (by percentiles).

Area Signed area - negative values negate positive values.

Decay tau Time constant of 'Decay time'.

Decay time 10 - 90% decay (fall) time of "peak to end".

Frequency Number of threshold crossings per second (Hz).

Max slope Maximum slope of simple linear regression fit.

Max value Value of largest positive sample.

Mean Arithmetic mean of the samples.

Min value Value of smallest sample.

MT Series Resistance*

Membrane test 'Rseries' value of a patch.

MT Membrane Capacitance*

Membrane test Cmembrane' value of a whole-cell patch.

MT Membrane Resistance*

Membrane test 'Rmembrane' value of a whole-cell patch.

Rise tau Time constant of 'Rise time'.

Rise time 10 - 90% rise time of "start to peak".

RMS noise Root-Mean-Square noise.

Segment amplitude Base amplitude of the selected segment.

Segment duration Duration of the selected segment.

Slope Slope of a simple linear regression fit.

Std deviation Standard deviation of the samples: $\sqrt{\text{variance}}$

Time of absolute peak

Time from sweep start to largest absolute value.

Time of max Time from sweep start to largest sample.

Time of min Time from sweep start to minimum sample.

Time to threshold Time from sweep start to first threshold crossing.

Variance of the samples.

Weighted tau Weighted time constant.

Area / Peak - y0 (based on end of measurement).

* Only available when an output channel segment uses a Membrane Test waveform. The 'Signal to Analyze' is restricted to the headstage inputs.

These analyses can be directly plotted, or used in more complex equations. (See the Equation Editor section for more details.)

Many other SutterPatch settings and readings can be plotted, without defining an Analysis measurement, through the Graph Settings axes source equations.

Note: The first sample point is used for any needed baselines.

[Status field] A short description of the selected Analysis.

Threshold This amplitude level needs to be crossed by the signal to trigger measurements for:

- AP Duration
- Rise/Decay time
- Rise/Decay Tau
- Frequency

Time to threshold

Polarity The direction of a Threshold crossing.

Positive
 Positive-direction threshold crossing.

• Negative Negative-direction threshold crossing.

• Largest change Use the polarity direction of the largest change for

Rise and Decay analyses.

AP Duration [for AP Duration only]

[20, 30, 40, 50, 60, 70, 80, 90, 100 %]

The action potential amplitude-percentile setting, to calculate the associated AP Duration width. The baseline is the first sample.

Enable Smoothing

[2-200]

Set the number of Gaussian smoothing operations per measurement.

Smoothen noisy data to reduce the effects of high-frequency noise on measurements. Apply averaging to the data sample points with an unweighted sliding average.

Note: Smoothing is not applied to the analyses 'Segment duration' and 'Segment amplitude'. These are fixed values not subject to modification.

Signal to Analyze

For each enabled Analysis measurement, select the signal to be measured from the list of Input Channels. A measurement is only made on one input channel, but it can be used in multiple graphs.

Region Timing

Cursors Relative to Set the measurement boundaries with left / right cursors.

Cursor Start times cannot be greater than their End times.

Warning!

Beware of boundary issues, where sharp transitions can be unexpectedly included or excluded in measurements. Due to the various input and output sampling rates and time durations of the actual signal, data points might not exactly match up with defined measurement regions.

You might need to adjust the measurement region to be one sample (or more) greater or less, than the target region, depending on whether you want to exclude or include the initial response. Otherwise, for example, a spike at the beginning of a Segment could skew measurement amplitudes to be larger, or a transition at the beginning of a Segment could be missed in a threshold crossing, thus lowering a Frequency count.

Sweep Time Set relative to the start time of a sweep (time zero). Increment by: [] The 'Start Time' or 'End Time' spinners increment by this amount. The listed time values depend upon the input filter bandwidth. **Reset Times** Set the cursor Start / End Time to the beginning and ending of the sweep. **Start Time** Set the left cursor start time (s). **End Time** Set the right cursor end time (s). Lock Time Step Width Fix the width of the measurement region. The measurement width is maintained at a constant value when the cursor 'Start Time' is updated. 1 The width of the cursors in seconds. The minimum width size is 2 sample points. Segment Time Set the time range as a ratio of the Segment duration. Uses the segment timing from the input signal's "Parent Output Channel". Out Channel [Output Channel list] [Segment #] Increment by [0.001, 0.002, 0.005, 0.01, 0.02, 0.05, 0.1, 0.2, 0.5]Increment the Start / End Ratios by a relative amount. Reset Ratios [0.0000/1.0000]

Reset the Start / End Ratios to span the entire segment.

Tip: If unwanted segment boundary issues occur, where measurements are affected by data in a neighboring segment, increase the Start Ratio or decrease the End Ratio until the issue is resolved.

Start Ratio [0 = beginning of Segment]

Set the left cursor as a ratio of the Segment dura-

tion.

Time [0.0000 s] Cursor start-time read-only field.

End Ratio [1 = end of Segment]

Set the right cursor as a ratio of the Segment du-

ration.

Time [0.0000 s] Cursor end-time read-only field.

If the Start / End Ratios extend past the boundary of a Segment, and the measurement is switched to a beginning or ending Segment, the Start / End

Ratios are reset to '0' and '1' respectively.

Cursor Time Width The width of the measurement region is reported.

Graph Settings




Figure 4-55. Real Time Graphs

On Enable a graph to configure its settings.

ID Graphs have a default ID (identification): g[1] - g[8]

Show Display this graph in an Analysis window during acquisition and analysis.

Note: If the Y-Axis 'Equation' field is blank, the graph will also be blank.

Label Double-click to manually enter a graph label for the Analysis sub-window.

Copy Use to transfer graph settings to a new graph. Highlight a Graph ID, then select

from the drop-down list of enabled graphs.

Y-Axis

Source Set up the source of the Y-axis numbers:

Equation Use an equation for a customized X-axis

plot. (See the Equation Editor for details.)

Many SutterPatch settings and readings can be plotted, without defining an Analysis measurement, by using Special IDs in the equation. • <m[#]: *Name*> Select a Measurement ID for the X-axis.

Label A Y-axis label is automatically generated from the Measurement

label. Directly edit to customize the Y-axis graph label.

Unit Select a standard unit from the drop-down list, or enter a custom

unit type.

Note: Standard unit resolutions, such as 'pA' or 'mV', are auto-

matically calculated and displayed in the graph.

Auto Range Restrict the graph Y-axis range.

Y-min Lower limit of the Y-axis.

Y-max Upper limit of the Y-axis.

X-Axis

Source Set up the source of the X-axis numbers:

• Equation Use an equation for a customized X-axis

plot. (See the Equation Editor for details.)

Many SutterPatch settings and readings can be plotted, without defining an Analysis measurement, by using Special IDs in

the equation.

• Time Use a standard time-base.

• <m[#]: *Name*> Select a Measurement ID for the X-axis.

Label Enter a customized name for the X-Axis.

Unit Select a standard unit from the drop-down list, or enter a custom

unit type.

Note: Standard unit resolutions, such as 'pA' or 'mV', are auto-

matically calculated and displayed in the graph.

Auto Range Restrict the graph X-axis range.

X-min Lower limit of the X-axis.

X-max Upper limit of the X-axis.

Up to 16 Routine Variables can be configured for use in Routines. These variables allow manual or automatic control of certain Routine settings.




Figure 4-56. Routine Variables

Routine Variables

In Use A checkmark means the Routine Variable is "active", i.e., set to a non-zero value, or is being used in a Routine setting or

equation field.

Label Var_r[1 - 16] Edit the default Routine Variable name

if desired.

Value Numeric values can be manually entered here, or

automatically set by the Paradigm step 'Set Variable'.

Routine Variables can be used in:

Input Channels / AuxIN / Scaling

Input Channels / Virtual Channel / Equation

Output Channels / Enable P/N Leak Pulses / Leak Hold

Output Channels / Waveform Editor / Amplitude

Output Channels / Waveform Editor / Duration

Measurements / AP Duration / Threshold

Measurements / Frequency / Threshold

Measurements / Time to Threshold / Threshold

If a Waveform Editor 'Amplitude' or 'Duration' is set to a Routine Variable, and then changed to a value, the Waveform Editor converts its 'Var_r[#]' settings to 'Value' settings, using the last enabled value.

4.1.9 Solution Editor

Create a named list of solution control "valves" to manage your commands to physical valves, solution changers or perfusions systems.




Figure 4-57. Solution Editor

Save Save changes to the current Solution Pool file. This button becomes ac-

tive once a solution is selected for editing.

Configure Open the Configure Solutions dialog to categorize solution types and con-

figure output channels.

Solution Pool Files A Solution Pool file (*.spo) can contain multiple defined solutions.

New Solution Pool Create a blank Solution Pool.

Load Solution Pool Load the Solutions of a previously saved Solution Pool

file into the Solution Pool.

Revert to Last Saved Undo any unsaved changes to the Solution Pool.

Save Solution Pool Save the Solution Pool using its existing file name and

path.

Save Solution Pool As Save the Solution Pool to a new file, and switch to the

new file.

Save Solution Pool Copy Save the Solution Pool to a new file, but do not switch to

the new file.

Note: Default file names are auto-incremented from the previously loaded Solution Pool name.

The file path and file name of the loaded Solution Pool file is displayed.

[off, 1-#] Select a "valve" radio button to open a valve. This activates its corresponding solution configuration. A radio button is available (ungrayed) when its name is changed from 'undefined'.

Only one "valve" can be active at a time. The number of radio buttons is set in the Configure Solutions sub-window.

When set to 'off', all configured solution outputs are set to a zero amplitude.

Name Double-click on a field to edit it; click-and-drag to move it up or down in

the table.

Concentration Enter a concentration value for the solution.

Tip: You can access the concentration value from the last-used 'Test

Compound' or 'Control' solution valve in any fields that accept the

Special Identifier 'Paradigm Parameters: Stimulant'.

Unit The unit type of the concentration.

Configure Solutions



Close Dialog button.

[# Solutions] Set the number of solutions to configure. When this number is changed, a new Solution Pool is created.

[4, 8, 12, 16, 20, 24]

Loading other Solution Pool files allows an unlimited number of solutions to be accessed in an Experiment.

Description A text note for the solution.

[Solution Type list] A list of predefined solution types.

- Initial Condition
- Washout
- Control

Index 1-4 Distinguish between different Control solutions.

- Test Compound
- Not a Solution

[Output Channel] Select a physical output channel and set its value.

- No Output
- AuxOUT1 4 [±10.000 V] Analog output voltage.
- DigOUT Word [0-65,535] Decimal value of a 16-bit digital word.
- DigOUT1 16 [Low / High] A single digital bit is set "high".

4.1.10 Template Editor

Templates allow any data waveform or portion of an existing data wave to be incorporated into a command waveform. The Template Editor can manage and manipulate such templates.




Figure 4-58. Template Editor

Note: Changes made in the Template Editor are only applied to Routines when the Routine Editor / Output Channels / Waveform Editor / Template wave' is used to copy the modified template in its Template Editor Pool list into the Routine.

Template Pool Files A Template Pool file (*.spt) can contain multiple defined Templates.

New Template Pool	Create a blank Template Pool.
Load Template Pool	Load the Templates of a previously saved Template Pool.
Revert to Last Saved	Undo any unsaved changes to the Template Pool.
Save Template Pool	Save the Template Pool using its existing file name and path.
Save Template Pool As	Save the Template Pool to a new file, and switch to the new file. The default file name is the same as the original

file name.

Save Template Pool Copy Save the Template Pool to a new file, but do not switch to

the new file. The default file name increments.

Merge Template Pools Insert the Template from a previously saved Template

Pool file into the loaded Template Pool.

[] The file path and file name of the loaded Template Pool file is displayed.

Import Select a template file (*.ibw).

> Alternatively, in a scope window or preview pane, click and drag the mouse to surround a region of interest with a bounding box (the "marquee"). Right-click in the box and select 'Extract Template'. A template with the signal name is added to the template list.

An extracted template is composed of a single sweep:

Acquisition scope window: Last sweep.

Reanalysis scope window:

Selected sweep.

Preview pane:

Last or selected sweep.

Analysis Editor

Selected wave.

The Y-axis values are copied to the template; the X-axis values are reset in the template to start at zero.

Note: 'Extract Template' is not implemented for the Data Navigator preview pane. Also, it is only valid with monotonically increasing or decreasing X-axes.

Export Export the selected template to a 1-D Igor wave file (*.ibw).

> To export a portion of a sweep, select the region of interest with the mouse, and use the marquee 'Extract Template' right-click command.

The new wave can now be exported.

Rename Edit the name of the selected template. Allowable characters are A-Z, a-

z, 0-9, and "_". Special characters are not allowed; spaces are replaced by

an underscore.

Ē

Duplicate Add a copy of the selected template to the list. The new template name's

number is appended or incremented.

Revert Discard any unsaved changes to the selected template.

Delete Remove the selected template from the list.



Copy to Layout Copy the selected template graph into a new

Layout window, or append to an existing Layout

page.



Copy to Clipboard Copy the selected template graph to the system

clipboard.

Save Pool Save the template pool using its existing file name.

Browse Create a template from the Experiment data in the Data Browser.

Define Template Properties

Update a data wave's X- and Y-axis parameters to be compatible with

SutterPatch templates.

Enter X-increment The data point time interval is changed, which also

adjusts the length of the trace.

Enter X-start The X-axis starting time for the data.

Enter Y-unit The Y-axis base unit (enclose between double quotes.)

Convert The data is interpolated to match the new sampling rate. While the

number of samples is updated, the length of the trace is unchanged.

Smooth Apply smoothing to the template.

Off

• Boxcar A fast time-domain filter with excellent 0 –

100% rise-time response.

• Gaussian A standard filter with excellent 10 – 90% rise-

time response.

Factor Adjust the template scaling factor.

Values are displayed with SI unit prefixes.

Offset Adjust the template offset.

Values are displayed with SI unit prefixes.

Do It Apply the adjustments to the template parameters.

Template Names A list of the loaded templates.

Click on a Template entry to make it the active one.

Double-click on a Template Name to rename it.

Click-and-drag a Template entry to reposition it in the list.

Template Parameters

Parameter settings description.

Show/Hide Preview

Display / Hide a preview pane with the selected template.

The preview pane X- and Y-axes can be controlled in two ways:

- Hover the mouse over an axis line until the cursor turns into a double-headed arrow, then scroll up or down to contract / expand the axis.
- In the preview, click and drag the mouse to surround the region of interest with a bounding box (the "marquee"). Right-click in the box and select one of the expand/shrink options.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor Info' from the right-click menu. (See 'Right-Click Menus' for scope windows.)

4.2 Data Analysis

Online data analysis is configured in the Routine Editor with measurement settings. These same settings are also available during offline data reanalysis.

General analyses are available in:

- Analysis Editor
- Equation Editor
- Igor Analyses
- Metadata Review
- Paradigm Editor
- Routine Editor
- Reanalysis scope

Specialized analyses are available via the Reanalysis scope window:

- Parametric Plot
- Amplitude Histogram Plot

or the Data Navigator:

- Action Potential Analysis
- Single-Channel Analysis
- Synaptic Analysis

For extra flexibility in controlling analyses, use a Paradigm 'Execute' step to run SutterPatch commands, Igor analyses or user-defined functions.

4.2.1 Action Potential Analysis

Action potentials (APs) are analyzed with this offline dialog. Access via the Reanalysis scope window 'Measurements' button or the Data Navigator (signal) 'Available actions' menu.




Figure 4-59. Action Potential Analysis

Sweep # The sweep number of the selected action potential event.

The Sweep # is set to '0' when the Average event is displayed in the Event pane.

Pre-select sweeps for processing by "marking" them in a scope window or the Data Navigator tree.

Event Time Time point when the potential of the selected event crosses the threshold.

[Event pane] A graph of the selected event, with the X-axis zero point reset to the

Threshold point.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor Info' from the right-click menu. Cursors are available for all AP graph

panes/windows. (see 'Right-Click Menus' for scope windows.)

Event [# of #] Event number vs. total events.

Cycle through the analyzed events in the Event pane, with the event

highlighted in red in the 'Sweep' pane.

or

Sweep [#] Cycle through the unanalyzed sweeps in the 'Sweep' pane.

[Sweep pane] Displays a sweep of data colored in blue, with the selected event high-

lighted in red.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor

Info' from the right-click menu. (See 'Right-Click Menus' for scope

windows.)

Start Click to find and analyze action potentials, and to display the Results

pane.

Average Click to display the averaged event (in Sweep # 0) in the Event pane.

The Average Event Amplitude and plot are displayed in the Save Re-

sults layout window.

Note: While the 'Sweep #' is set to '0' for averaged events, the last dis-

played Event # is unchanged.

Event Length (s) The event duration in the Event pane; the selected event is highlighted

in red in the Sweep pane.

Threshold (V) $[\pm 0.1000]$

This voltage level needs to be reached or exceeded for analysis of an

event to be triggered.

APD measure at: Set the Action Potential Duration percentile.

Measures the duration of an event at a percentile of the event's repolari-

zation amplitude.

[20/30/40/50/60/70/80/90/100%]

Analysis Start (s) [0+]

Set the start time of the Sweep data to be analyzed.

Analysis End (s) Set the end time of the Sweep data to be analyzed.

DPATCH - OPERATION MANUAL - REV. 2.2 (2021-10)

Save Results

The latest results are displayed in the 'Action Potential Analysis Results' Layout window and a Results table window. Separate 'Average AP' and 'Phase plot' hidden graphs are also created, and can be accessed via menu item Windows / Graphs.

A 'Results' table of all events automatically opens:

[] Row number, one row per event.

Sweep Number Sweep number the event is in.

Event Time (s) Time point of the event start.

Threshold (V) Amplitude of the event threshold.

Threshold Time (s) Time point of the "trigger" threshold time.

Peak (V) Amplitude of the event peak.

Peak Time (s) Time point of the event peak.

AP Duration (s) Duration of the action potential at the AP

repolarization percentile.

AHP (V) Peak amplitude of the After Hyper-Polari-

zation phase.

AHP Time (s) Time point of After Hyper-Polarization; the

event re-crosses the threshold amplitude at

this time.

Absolute Event Time The absolute time of the event from the

start of the recording.

Interevent Interval The time between adjacent events.

Max Slope (V/s) The maximum slope of the event.

An 'Action Potential Analysis Results' report automatically opens in a special Layout window:

Signal Pathname: The Igor Pro experiment pathname for the

analyzed signal.

Analysis Prefix: The prefix for the signal's analysis objects

in the Igor Pro 'Data: Analysis' folder.

Total time analyzed = [s]

Includes the Start / End times for all analyzed sweeps.

Number of events detected = Total number of events found.

Event Frequency = [Hz]

The average frequency of the found events.

All Sweeps analyzed Every sweep was analyzed.

or

Sweeps analyzed: [list of analyzed sweep #s]

[Event graph] [V vs. s]

A graph of the averaged event.

[Phase plot] [dV/dt (V/s) vs. V]

A graph of the phase plot for visual inspec-

tion of the derivatives.

Duplicate Results are copied to a new Results table and a new Layout window.

Plot The 'Plot sweeps' dialog displays to allow events to be plotted as overlap-

ping sweeps in a floating graph window.

Plot sweeps Enter a list of sweeps separated by a comma "," and/or a range of

sweeps separated by a dash "-".

Show/Hide the Sweep pane (below).

Displays the sweep data colored in blue, and the selected event high-

lighted in red.

Show/Hide the Results pane (on the right).

Displays the Action Potential Measurements results.

Results pane Measurement results are displayed for the event selected in the main

window.




Figure 4-60. Action Potential Measurements

Phase plot

A graph of the phase plot, for visual inspection of the derivatives.

[dV/dt (V/s) vs. V]

Y-axis vs. X-axis.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor Info' from the right-click menu. (See 'Right-Click Menus' for scope windows.)

Threshold potential = [V(s)]

Event starting amplitude (time from Threshold setting).

The actual start of the selected event (the biological starting amplitude), and its timepoint relative to the 'Threshold' setting timepoint.

See the Algorithms appendix for more information.

Peak = [V(s)]

The largest amplitude excursion of the event (and the timepoint of the 'Threshold' setting).

Two measures of an event's peak amplitude are reported, voltage and time. Time is reported relative to the 'Threshold' timepoint.

APD (%) = [%, s]

Action Potential Duration of the event at (n %) of amplitude repolarization.

AHP (min) = [V (s)]

The largest amplitude excursion of the "After HyperPolarization" phase of the event (time from 'Threshold' setting).

Two measures of the AHP amplitude are reported, voltage and time. AHP is when the action potential repolarization phase drops to its lowest point below the resting membrane potential, i.e., during the hyperpolarized refractory period of the cell.

Event = The analyzed event (of the total number of events) found in the

data.

Sweep = The sweep number of the analyzed event.

Event time = The sweep time of the start of the analyzed event.

Events found = The number of averaged events.

Event frequency = The average frequency of the found events.

4.2.2 Analysis Editor

View and manipulate the data in your Experiment's various analyses and graphs.




Figure 4-61. Analysis Editor

Select Choose how to view the data.

Table View a spreadsheet of the Analysis data.

Warning! Editing the table will permanently alter its data.

Tip: As there is no "Undo", before making any changes, use Operations / Duplicate to make a working copy of the data that can be deleted later.

Top Row: Row Column numbers and value of the target cell.

Source Row: Dimension labels:

source routine name and source step number.

Header Row: Analysis function measurement name.

Data Rows: Rows [0-n]

[Columns for time-based graphs]

First Column: Row numbers.

Second Column: " $X \setminus Y$ " column label.

[has blank rows]

Column 0: X-data (Time) from the first graph: start times of

the measurement sweeps.

Column 1: Y-data from the first graph: amplitudes of the

measurements.

Columns m - n: Another pair of X- and Y-data columns displays

for each additional time-based graph. However, the X-data columns are blank, as the Column $\mathbf{0}$

(Time) values apply instead.

[Columns for X-Y graphs]

First Column: Row numbers.

Second Column: "Time" column.

Time data from the first graph: start times of the

measurement sweeps.

Column 0: Y-data from the first graph: amplitudes of the Y-

axis measurements.

Column 1: X-data from the first graph: amplitudes of the X-

axis measurements.

Columns m - n: For each additional X-Y graph, a pair of Y- and X-

data columns repeat.

Note: The first data point is in row 0, so the last data point is in row

[n-1].

Row 'n' is a blank row that contains grayed-out cells. It is used to manually add extra rows of data to the table. Once a number is entered into one of these cells, the row ungrays and the next row

below is automatically grayed.

Graph-[1-8] Select an Analysis graph for the selected data

wave. The graph number refers to its original

Analysis window position.

Files Import or export an analysis graph file.

Export Table to text file The table data are written to a tab-delimited plain

text file. Any column header information is lost. To preserve such metadata, export to the binary

format.

Export Graphs as binary wave

Save the entire graph as a multi-dimensional Igor

Binary Wave file (*.ibw).

Export Graph X-column Save the X-column data, including labels, as a one-

dimensional Igor Binary Wave file (*.ibw).

Export Graph Y-column Save the Y-column data, including labels, as a one-

dimensional Igor Binary Wave file (*.ibw).

Import Table from text file Import numeric text data from comma- or tab-

delimited columns.

Import Graphs from binary wave

Open and display a saved graph.

Note: Import of one-dimensional Igor Binary

Wave files (*.ibw) is not supported.

Operations

Duplicate Insert a copy below the highlighted item.

Delete Analysis or Table

Delete the entire analysis wave.

Delete Single Graph Delete the selected graph.

Note: If an analysis cannot be deleted, it likely

exists in another Graph window or Layout

page - first close the other analysis instance via menu items Windows /

Graphs, or Windows / Layouts, or Windows

/ Layout Macros.

Delete Analysis or Table

Click to delete the selected analyses or tables. Hold

<Shift> to delete without verification.

Copy to Layout Copy the selected graph or table into a new Layout

window, or append to an existing Layout page.

0

Copy to Clipboard Copy the selected graph or table to the system clipboard.

₽

Options

Show Fits Display fit lines on the graph data.

Show Error Bars Display SEM error bars for averaged data.
Show Axes Color Display a background color for the axes.

Show Grid Display X & Y grid lines in a graph.

Show Markers Display data points with marker symbols.

Show Lines Display a line between data points.

Include Column Labels Column labels appear on the first line of an

exported table.

Math

Normalize: zero to maximum

Rescale the data, whereby the largest absolute point is set to 1.0 (or -1.0), while the zero point is maintained in, or relative to, the data range.

Normalize: minimum to maximum

Rescale the data, whereby the smallest point is set

to zero and the largest point is set to 1.0.

Invert Reverse the Y-axis sign of the data points.

Concatenate

To enable this button, select additional analysis waves (Shift-click in Windows.) The newly concatenated wave is inserted below the last selected analysis wave; if the last selected wave is also a concatenated wave, the additional data is instead concatenated with the last selected wave.. Time-course data are plotted relative to the loaded analysis wave's "time zero".

Average

Select an analysis wave (Shift-click in Windows) to be averaged with the loaded wave. A weighted average is performed, i.e., the number of data sets is accounted for when averaging in new data.

Two new entries are inserted into the wave list after the loaded wave:

- 1. The averaged wave.
- 2. The SEM (Standard Error of Means) data points wave.

If Options / Show Error Bars is enabled, the SEM data are used to display error bars in the corresponding averaged data graph.

(See Appendix F: SutterPatch Algorithms for the SEM algorithm.)

Overlay

Select multiple analysis waves (Shift-click in Windows) and plot them in the same graph. When only a single analysis wave is selected, all graphs in the Editor are overlaid.

The initial wave is plotted with standard "diamond" symbols, while the added selections are plotted with smaller "plus" symbols.

Copy Fit Waves

Copy all Quick Fits (from all open SutterPatch windows) into their own analysis waves in the Analysis Editor, so that they persist during the entire experiment, and are also saved with the experiment. Otherwise, uncopied Quick Fits disappear when another Analysis Editor analysis wave is selected, or the Cursor pane is hidden, or the window is closed.

Scale and add

Use to combine available analysis waves, with optional scaling and offset applied.

When the Factor is '1.00' and the Offset is '0.00', this operation will simply add the selected wave to the displayed wave.

Factor Set a scaling factor for a data wave that will be added to the displayed data.

Values are displayed with SI unit prefixes.

To subtract a data wave, change the Factor to a negative number.

Offset Set an offset for a data wave that will be added to the displayed data.

Values are displayed with SI unit prefixes.

Show/Hide Cursor Info

Open the Cursor Info pane to manually measure X-Y data values, or to set a fitting range. / Hide the Cursor Info pane, and any cursor symbols in the active pane.

Cursor Info pane



Options menu



One Mover Moves All

Draggable cursor mover tool – move all cursors together with a single control.

• All Styles Change the cursor symbol style.

• Show Cursor Pairs Display up to 5 sets of cursor symbol pairs.

Cursor A



Cursor symbol for data point 'A'.

A: Symbol letter (beginning cursor of the pair).

R1_A_IV (Default) wave name.

1

Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'A'.

Y: Y-axis value of data point 'A'.

 ΔY Difference of the cursor pair Y values.

Cursor B

☐ Cursor symbol for data point 'B'.

B: Symbol letter (ending cursor of the pair).

R1_A_IV (Default) wave name.

3

Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'B'.

Y: Y-axis value of data point 'B'.

 ΔX Difference of the cursor pair X values.

Cursor Instructions

1. Click on symbol 'A' to enable it.

- 2. Manually drag the highlighted symbol onto a data point in the graph, or enter the data point number in the 'pnt' field.
- 3. Click on symbol 'B' to enable it.
- 4. Manually drag the highlighted symbol onto a data point in the graph, or enter the data point number in the 'pnt' field.

X- and Y-measurements are displayed for the cursor pair data.

5. Fitting can also be applied to the cursor pair data. Right-click in the graph, and select 'Quick Fit' for a list of built-in Igor fitting functions.

The fit is displayed in the graph, and the fitting information is written to the Command window.

Analysis Wave Names

Loaded analysis waves available for manipulation.

Analysis Parameters

• Series: R# Series name of the wave.

• Graphs: # Number of graphs in the wave.

• Points: # Number of data points in the graph.

• Start: # Start time of the analysis wave,

or

• SEM Standard Error of the Mean wave.

• Average: # Number of graphs averaged or appended in the wave.

Tip: If the Analysis Parameters text is not fully visible, increase the width of the Analysis Editor window.

[Graph & Table pane]

Data point markers are plotted, or a numeric table is displayed.

X- and Y-axes can be magnified to be larger or smaller. Place the mouse cursor in the axes tick regions, then scroll the mouse wheel up or down. The axis tick region does not include the tick label (numbers) area.

The marquee tool is also supported in the Graph pane. Click and drag a bounding box around the region of interest, then right-click in it for magnification options.

To measure X-Y data points or set a fitting range, select 'ToggleCursor Info' from the right-click context menu. (See 'Show/Hide Cursor Info' above.)

4.2.3 Analysis Window

Scope measurements are plotted in an Analysis window docked on the right side of the scope window. An Analysis window can be resized or closed, but not undocked from the scope window.




Figure 4-62. Analysis Window

Online measurements are plotted during data acquisition in real-time. Online analysis measurements are configured in the Routine Editor 'Real Time Measurements & Graphs' section. A separate pane is created in the Analysis window for each enabled Measurement graph.

Data can be opened for offline review or analysis via the Dashboard 'View Last' button or the Data Navigator. When stored data are rerun for analysis, the data displays in a Reanalysis scope window, and the measurements are graphed in the accompanying Analysis window. The last measurements applied to the data are automatically used to reanalyze the data.

Save Analysis This button saves the displayed analyses with the Experiment.

Saved analyses are viewable in the Analysis Editor (or also the Data / Data Browser 'Data: Analysis' folder. Unmarked sweeps measurements are not visible in Analysis Editor tables.

Concatenate Append new measurements to the existing measurements in the graphs.

[when docked to a scope window]

Clear Erase all measurements from the graph display.

Copy to Layout Copy the analysis graphs into a new Layout window, or append to an existing Layout page.



Copy to Clipboard Copy the analysis graphs to the system clipboard.

[Graph panes]

The graph pane X- and Y-axes can be magnified to be larger or smaller. Place the mouse cursor in an axis tick region (do not include the tick labels or numbers), then scroll the mouse wheel up or down.

The marquee tool is also supported in the graph pane. Click and drag a bounding box around the region of interest, then right-click in it for magnification options.

To manually measure XY data values, or to set a fitting range, right-click in the graph to display a data modification menu, and select 'Toggle Cursor Info'. The 'Cursor Info' pane displays.

Select 'Toggle Cursor Info' again to hide the Cursor Info pane, and any cursor symbols in the active pane.

Cursor Info pane



'Options' menu



One Mover Moves All

Draggable cursor mover tool - move all cursors together with a single control.

• All Styles Change the cursor symbol style.

• Show Cursor Pairs Display up to 5 sets of cursor symbol pairs.

Cursor A



Cursor symbol for data point 'A'.

A:

Symbol letter (beginning cursor of the A/B pair).

R1_A_IV (Default) wave name.



Draggable cursor mover tool for the cursor pair.

pnt:

Data point number (starting from zero).

X:

X-axis value of data point 'A'.

Y: Y-axis value of data point 'A'.

 ΔY Difference of the cursor pair 'Y' values.

Cursor B

Cursor symbol for data point 'B'.

B: Symbol letter (ending cursor of the A/B pair).

R1 A IV (Default) wave name.



Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'B'.

Y: Y-axis value of data point 'B'.

 ΔX Difference of the cursor pair 'X' values.

Cursor Instructions

1. Click on symbol A to enable it.

- 2. Manually drag the highlighted symbol onto a data point in the graph, or enter the data point number in the 'pnt' field.
- 3. Click on symbol 'B' to enable it.
- 4. Manually drag the highlighted symbol onto a data point in the graph, or enter the data point number in the 'pnt' field.

X- and Y-measurements are displayed for the cursor pair data.

5. Fitting can also be applied to the cursor pair data. Right-click in the graph and select 'Quick Fit' for a list of built-in Igor fitting functions.

The fit is displayed in the graph, and the fitting information is written to the Command window.

Note: When applied to Analysis windows, the same Quick Fit is applied to all panes selected for fitting.

4.2.4 Data Browser

The Data Browser can be used to access and display all of the Experiment's data objects, such as recorded data waves, analysis graphs, layouts, images, metadata, Paradigms and Routines. Access it from the Data menu.




Figure 4-63. Data Browser

Objects are displayed in a tree structure, using a path "root" of 'SutterPatch'.

Warning! If this window is kept open during data acquisition, the Experiment can unexpectedly terminate!

Also, for the macOS, opening and closing windows can take a long time.

The 'Data' subfolder contains additional subfolders, followed by recorded data waves arranged per Signal.

ExperimentStructure The sequence of Paradigms and Routines.

<Names of data Series...>

Analysis This folder contains data measurements, including results from fits, and Event tables.

- Wave names that include "_M_" contain the status of sweep marks. A marked sweep has a value of '1', while an unmarked sweep has a value of '0'.
- Wave names that include "_A_" contain analysis measurements.
- Wave names that include "_R" contain capacitive mode resets.
- Wave names that include "_df_" contain the differentiated average action potential (phase plot) waveform.

Images This folder contains stored images that display in the preview pane.

Meta This folder contains a table of general system metadata parameters (unformatted).

Routines This folder contains limited information on the used Routines.

Right-click Menu

Display Display the first sweep of the data in a visual graph.

Edit Display the data in a numerical table.

SutterPatch signal data are stored in two-dimensional data waves, with one column per trace, and one row per sample point.

Warning! Editing data here permanently alters the raw data. Modify at your own risk!

Copy Full Path

Copy the object's path to the clipboard. This is in relation to an internal (hidden) Igor data folder, not the computer's file system. This path can be used by advanced Igor users in user

functions and executable commands.

4.2.5 Data Navigator

The Data Navigator window organizes and displays all levels of data for the current Experiment.




Figure 4-64. Data Navigator

Data tree hierarchy The hierarchy of sorting levels is displayed in this pane.

Build Hierarchy Re-organize the data tree using custom settings.

Select parameter group Organize by metadata parameters:

[Parameter availability depends on the selected 'Set Preferences / Metadata' detail level.]

All Categories

Frequently Used

Experiment Hierarchy

Tag

Operator

Preparation - Animal

Preparation - Tissue

Preparation - Cell

Experiment

Amplifier

Instrumentation and Software

Electrode

Recording Solutions

Paradigm

Cell Health / Quality Control

Series (= Routine Data)

Data Acquisition Settings

Imaging

Stimulus

Available parameter Click on a specific parameter from the selected group.

Click on the "copy" button to insert the selected

parameter above the highlighted level in the Hierarchy

pane.

Hierarchy The Hierarchy pane displays the new sorting hierarchy

for the current Experiment.

Remove the selected parameter from the list.

▶

The sorting levels can be re-organized by selecting an entry and clicking on the Up/Down keys to reposition

it, or using the 'Del' key to remove it.

Do It Click on the 'Do It' button to apply these changes to the

data tree.

Default Hierarchy Restore the default experimental hierarchy:

Paradigm > Routine (> Signal > Sweeps)

Data Tree Window

The current Experiment's data are arranged in the data tree in five default levels:

Experiment

Paradigm

Routine

Signal

Sweep

Selecting a node level in the data tree highlights it in blue.

Enable Marks

Allow data to be marked in the data tree or Available Actions menu for analysis processing.

Marks enabled

Node selected and marked.

All levels and nodes below it included for processing.

• Node selected and unmarked.

Only marked levels and nodes below it included for processing.

Marks disabled

Node selected.

All levels and nodes below it included for processing.

Sweeps "marked" during acquisition are loaded into the Data Navigator as "marked".

"Marking/unmarking" sweeps in a signal "marks/unmarks" those sweeps in all signals in the same Series.

Mark a range: Highlight a marked starting selection, shift-

click on the ending selection.

Clear a range: Highlight an unmarked starting selection,

shift-click on the ending selection.

Expand All nodes of the data tree are expanded down to the Signal level.

To expand a node to the next lower level, with a mouse click on a "+" node, or double-click the node name; with a keyboard, use the up/down arrow keys to select a "+" node and press the space key.

Collapse All

All nodes of the data tree are collapsed up to the Paradigm level.

To collapse the lower levels of a node, with a mouse click on a "-" node, or double-click the node name; with a keyboard, use the up/down arrow keys to select a "-" node and press the space key.

Last Series

The last Routine's first signal is highlighted in the data tree and displayed in the Preview pane.

Import

Select a previously saved Experiment, data file or image file to incorporate into the current experiment.

• Import Igor Experiment Files

The same data set (Experiment) cannot be imported twice.

When a Series name already exists in the Data Navigator, imported Series are renamed to avoid conflicts.

- Import SutterPatch HDF5 Data Files
- Import PatchMaster Data Files

[a Sutter amplifier needs to be running]

Mappings

Select the leak-subtraction mode when importing PatchMaster traces in SutterPatch.

Main Pulse Traces:

- Keep traces P/N-leak subtracted
- Make P/N-leak unsubtracted traces

- Don't subtract zero-current from traces
- Subtract zero-current from traces
- Import Image Files

Review

Paradigm: Its signals are displayed in a scope window as continuous data.

Note: This button changes to the Analyze button

DPATCH - OPERATION MANUAL - Rev. 2.2 (2021-10)

when a Routine, signal or sweep is selected.

Analyze Routine, Signal, Sweep:

> The Routine's signals are displayed in the Reanalysis scope window and Measurements analysis is run.

Note: This button changes to the Review button when

a Paradigm is selected or marked.

Simple preview / Full preview [for Paradigms and Routines]

> A "simple" preview displays a thumbnail sketch of the first signal's first Series (or sweep) of data, to the right of the preview information section.

A "full" preview displays all signals and Series in an attached pane.

Show Preview / Hide Preview

The displayed Preview information items are based upon the selected data tree level.

Note: The Data Navigator Preview panes do not support mouse operations.

Experiment

Experiment Name: The experiment file name.

HDF5 File Name: For optional HDF5 files.

Paradigms: Total number of

> Paradigms in the Experiment.

Routines Total number of Routines

in the Experiment.

Combined size of all data Total bytes in data waves:

waves in the Experiment.

Paradigm

Paradigm Description: Displays the Paradigm

description text.

Click to edit.

Metadata: Click to display the

Paradigm's metadata in a

docked sub-window.

Images: Open any saved images.

Simple preview

[Preview sub-pane] Displays a thumbnail

image of the first signal of

the first Routine.

Full preview

Paradigm Preview window Displays all signals and all

Routines in continuous mode in a docked window.

Routine

Routine Description: Displays the Routine

description text.

Click to edit.

Metadata: Click to display the

Routine's metadata in a docked sub-window.

Routine: Click to display the named

Routine's parameters.

Analysis: Click an analysis name to

open it into the Analysis

Editor.

Signals: Number of signals in the

Routine.

Sweeps: Number of sweeps in the

Routine.

Images: Open any saved images.

Simple preview

[Preview sub-pane] Displays a thumbnail

image of the first signal of

the selected Routine.

Full preview

Routine Preview window Displays all signals in the

Routine in continuous mode in a docked window.

Signal

[Preview pane] Displays a thumbnail

image of the selected

Signal.

Sweep

[Preview pane] Displays a thumbnail

image of the selected

Sweep.



A menu lists various actions for the marked and/or highlighted data levels. These actions are also available via a right-click on the selected data level.

Experiment

Copy Signal Data Paths Copy the Series internal Igor path

to the system clipboard.

(root:SutterPatch:Data:Series na

me)

Export Data (See Preferences)

Export all marked data in the

Experiment to file(s).

Uses the 'Set Preferences / Data

Export' options.

When saving files, and the new filename is the same as an existing filename in the target

folder:

- Choose a different folder, or

- First delete the older file via the OS file browser.

Analyze All Routines / Analyze All Marked Routines

Run the reanalysis on all or marked Routines in the Experiment.

[select options to be applied by the Analyze All Routines / Analyze All Marked Routines action]

Copy Analysis Results to Clipboard

Store Analysis Waves Append results to the

Analysis Editor pool.

Copy Analysis Graphs to Layout Page

[these options display when 'Enable Marks' is checked]

Mark All Paradigms All Paradigms in the Experiment

are marked.

Unmark All Paradigms All Paradigms in the Experiment

are unmarked.

Mark All Routines All Routines in the Paradigm are

marked.

Unmark All Routines All Routines in the Paradigm are

unmarked.

Mark All Signals All signals in the Paradigm are

marked.

Unmark All Signals All signals in the Paradigm are

unmarked.

Mark All Sweeps All sweeps in the Experiment are

marked.

Unmark All Sweeps All sweeps in the Experiment are

unmarked.

Paradigm

Single Review Window Re-use the same window for all

Paradigm Reviews.

Multiple Review Windows Create a new window for each

Paradigm Review.

Review Paradigm Display all Series (with all

signals and sweeps) from the highlighted Paradigm. Each signal displays in a pane in a Paradigm Review window.

Note: This action is not

supported with the HDF5 file preference "Keep only one Sweep in Memory".

Average Marked Sweeps The average sweep of all marked

sweeps is copied to the Analysis

Editor.

View Metadata Display the highlighted

Paradigm's metadata in the Metadata Review sub-window, docked to the right of the Data

Navigator window.

Edit Paradigm Description Alter the Paradigm Description

text(displayed in the Preview

window.) .

Copy Signal Data Paths Copy the Series internal Igor path

to the system clipboard.

(root:SutterPatch:Data:Series_na

me)

Export Data (See Preferences)

Export the marked Paradigm's data to file(s). Uses the 'Set Preferences / Data Export'

options.

When saving files, and the new filename is the same as an existing filename in the target

folder:

- Choose a different folder, or

First delete the older file via the OS file browser.

Discard Paradigm Remove the highlighted Paradigm

and its data from the Experiment. If marks are enabled, the selected Paradigm also needs to be

marked.

If the last Paradigm is discarded,

when acquiring another Paradigm in the same

Experiment, the new Paradigm

name will be incremented past the discarded Paradigm name.

Note: If any associated graphs

are still open when trying to discard the Paradigm, an error message will display. To fix, close any associated Graph windows found in the main menu Windows / Graphs.

Analyze All Routines / Analyze All Marked Routines

Run the reanalysis on all or marked Routines in the Paradigm.

[select options to be applied by the Analyze All Routines / Analyze All Marked Routines action]

Copy Analysis Results to Clipboard

Store Analysis Waves Append results to the

Analysis Editor pool.

Copy Analysis Graphs to Layout Page

[these options display when 'Enable Marks' is checked]

Mark All Routines All Routines in the Paradigm are

marked.

Unmark All Routines All Routines in the Paradigm are

unmarked.

Mark All Signals All signals in the Paradigm are

marked.

Unmark All Signals All signals in the Paradigm are

unmarked.

Mark All Sweeps All sweeps in the Paradigm are

marked.

Unmark All Sweeps All sweeps in the Paradigm are

unmarked.

Routine

Single Review Window Re-use the same window for all

Routine Reviews.

Multiple Review Windows Create a new window for each

Routine Review.

Analyze Routine Display the marked sweeps of all

signals in the highlighted Series in a Reanalysis scope window.

Review Routine Display all signals (with all

sweeps) from the highlighted Series. Each signal displays in a

pane in a Routine Review

window.

Note: This action is not

supported with the HDF5 file preference "Keep only one Sweep in Memory".

View Metadata Display the highlighted Series

metadata in a Metadata Review sub-window docked to the right of the Data Navigator window.

View Routine Settings Display the highlighted Series'

settings and preview in the Routine Settings window.

Edit Routine Description Alter the Routine Description text

(displayed in the Preview

window.) .

Copy Signal Data Paths Copy the Series internal Igor path

to the system clipboard.

(root:SutterPatch:Data:Series_na

me)

Show in Data Browser Open Igor's Data Browser window

to examine the highlighted Series'

data waves.

Export Data (See Preferences)

Export the marked Series to file(s). Uses the 'Set Preferences /

Data Export' options.

When saving files, and the new

filename is the same as an existing filename in the target folder:

- Choose a different folder, or
- First delete the older file via the OS file browser.

Discard Routine

Remove the highlighted Series and their data from the Experiment.

If the last Series is discarded, when acquiring another Series in the same Experiment, the new Series name will be incremented past the discarded Series name.

Note: If any associated graphs are still open when trying to discard the Routine, an error message will display. To fix, close any associated Graph windows found in the main menu Windows / Graphs.

Concatenate Sweeps

Combine all sweeps into one

sweep.

Restore concatenated Sweeps

Convert the concatenated sweep back to the original sweeps.

[these options display when 'Enable Marks' is checked]

Mark All Signals All signals in the Series are

marked.

Unmark All Signals All signals in the Series are

unmarked.

Mark All Sweeps All sweeps in the Series are

marked.

Unmark All Sweeps All sweeps in the Series are

unmarked.

Signal

Action Potential Analysis Analyze action potentials from

the marked sweeps of the

highlighted signal.

Single Channel Analysis Analyze single-channel Events in

the highlighted signal.

Synaptic Event Analysis Analyze synaptic events (EPSPs,

mEPSPs, etc.) from the marked sweeps of the highlighted signal.

Analyze Routine Display the marked sweeps of all

signals in the highlighted Series in a Reanalysis scope window.

Edit Signal Display all sweeps of the

highlighted signal as numeric columns in an editable table.

Display Signal Highlighted signal:

Display the marked sweeps in the signal in a graph window.

Marked signal:

Display all sweeps in the signal in a graph window.

Average All Sweeps Average all sweeps in the

highlighted signal and display in

the Analysis Editor.

or

Average Marked Sweeps Average the marked sweeps in

the highlighted signal and display

in the Analysis Editor.

View Metadata Display the highlighted signal's

metadata in a Metadata Review sub-window docked to the right of the Data Navigator window.

View Routine Settings Display the Series parameters in

the Routine Settings window.

Copy Signal Data Path Copy the Series internal Igor path

to the system clipboard.

(root:SutterPatch:Data:Series_na

me)

Show in Data Browser Open Igor's Data Browser window

to examine the highlighted

signal's data waves.

Export Data (See Preferences)

Highlighted signal:

Export the marked sweeps

in the signal.

Marked signal:

Export all sweeps in the

signal.

Uses the 'Set Preferences / Data

Export' options.

When saving files, and the new filename is the same as an existing filename in the target

folder:

- Choose a different folder, or

- First delete the older file via

the OS file browser.

[these options display when 'Enable Marks' is checked]

Mark All Sweeps All sweeps in the Series are

marked.

Unmark All Sweeps All sweeps in the Series are

unmarked.

Sweep

Extract Sweep Create a graph of the highlighted

sweep in the Analysis Editor.

Display Sweep Display the highlighted sweep in

a graph window.

Export Data (See Preferences)

Export the highlighted sweep.

To export a portion of a sweep, extract the data with the marquee tool, and then export from the Template Editor.

Uses the 'Set Preferences / Data Export' options.

[these options display when 'Enable Marks' is checked]

Mark All Sweeps All sweeps in the Series are

marked.

Unmark All Sweeps All sweeps in the Series are

unmarked.

4.2.6 Data Table

The Data Table provides direct access to the sample points in a data Series, using a spreadsheet-style presentation.

☐ Table0:R1_S1_IV R0 C0						
1000		3.233203 40 -12				
Row	R1_S1_IV[][0]	R1_S1_IV[][1]	R1_S1_IV[][2]	R1_S1_IV[][3]	R1_S1_IV[][4]	F
	0	1	2	3	4	
0	5.23321e-12	-5.63541e-14	1.63961e-12	3.32488e-12	8.47496e-13	^
1	-9.90098e-13	1.71653e-12	1.4612e-12	-2.72349e-12	-4.44117e-12	
2	-3.81879e-12	3.93903e-13	-2.90443e-12	4.29801e-13	6.7005e-13	
3	-7.61568e-12	-4.25361e-12	5.03512e-13	-2.13725e-12	-2.38818e-13	
4	1.77125e-12	-2.23609e-12	-9.99009e-13	5.83818e-13	-1.80796e-12	
5	-8.78851e-13	-1.72464e-12	5.22561e-12	-4.20016e-12	-2.72047e-12	
6	1.95622e-12	-4.37322e-12	-2.92145e-12	2.1944e-12	-1.13872e-13	
7	-2.17838e-12	-3.97302e-12	-1.08042e-12	-4.92174e-13	1.49426e-12	
8	3.4504e-12	-6.02635e-12	9.20459e-13	-2.37966e-12	-8.44283e-13	~
<						>

Figure 4-65. Data Table

Warning! Editing data permanently alters the raw data. Modify at your own risk!


Data Tables are accessed from the Data / Data Browser. Select a Series from the Data folder, then right-click the menu item 'Display'.

To allow adding data to the table, the last row of data in the table is followed by a final row of blank (gray) cells. Manually entering data into the final blank row causes a new last row of data to be created in the table, followed by a new final blank row.

4.2.7 Edit Virtual Signals

The Reanalysis scope window Measurements button provides access to the 'Edit Virtual Signals' dialog. Use it when applying different analysis scenarios to recorded data with "pseudo" input signals, in conjunction with the 'Reanalysis Measurements & Graphs' dialogs.

Virtual input channels allow you to perform a variety of mathematical transformations on input signals. To enable a virtual signal, highlight a signal name. When a virtual input channel is enabled, its configuration fields are ungrayed. Changes to the highlighted signal are saved when you click the 'Do It' button, and changes in unhighlighted signals are discarded.



 $Figure\ 4\text{-}66.\ Edit\ Virtual\ Signals$

Virtual Signal Parameters from

- [<Series name> (date/time stamp)]
- Overwrite with Original Routine

Source Signals

Label The input signal name.

New Label Double-click this field to edit the signal name,

then click the 'Do It' button.

Legal characters are A-Z, a-z, 0-9, and

underscore "_".

The label must start with a letter, be at least two characters long, and not be a duplicate of another label. Otherwise, the label will be automatically updated to a legal format.

To see if automatic updates will be applied to the label, after making edits, first click the 'Enter' key, before clicking the ungrayed 'Do It'

button.

Virtual Signals

Virtual signals can be added, edited, or removed from the scope window.

To enable a virtual signal, highlight it and click the 'Do It' button. A check mark is displayed in the first column, and the signal is added to the scope window.

To disable a virtual signal, highlight it and click the 'Delete' button. The check mark is removed from the first column, and the signal is removed from the scope window.

Signal The virtual signal name.

Label Double-click this field to edit the signal name,

then click the 'Do It' button.

Legal characters are A-Z, a-z, 0-9, and

underscore "_".

The label must start with a letter, be at least two characters long, and not be a duplicate of another label. Otherwise, the label will be automatically updated to a legal format.

To see if automatic updates will be applied to the label, after making edits, first click the 'Enter' key, before clicking the ungrayed 'Do It'

button.

Unit

The base unit of measurement from its Source signal. The resolution of the unit is automatically adjusted in the signal.

Parent Out Chan

The "Parent Output Channel" shows which output channel is associated with which input channel.

Scope Position

Signal Action

- None
- Show
- Hide
- Move to >> [signal name]

Scaling

Equation Use a value or an equation for the scaling ratio.

Factor Apply scaling to interpret the input signal data. Specify as a

numeric value or an equation.

Note: The dPatch system acquires data with a high-

resolution 16-bit ADC into 64-bit data words, so data resolution is not an issue when scaling input signals.

Offset Apply an amplitude offset to the input signal (after any

scaling).

For "mV" units, append with 'm' or 'e-3'.

For "pA" units, append with 'p' or 'e-12'.

Example: 5 picoamps using engineering notation: 5p

or in equivalent scientific E-notation: 5e-12

Scaled Offset Raw offset values converted to input units.

[for AuxIN channels]

Filter Bandwidth

The input-channel low-pass 8-pole Bessel filter bandwidth.

Nyquist Factor

To manage the quantity of data being processed, the hardware headstage sampling rate is downsampled from an initial fixed 5 MHz rate to a "Nyquist Factor" mulitple of the 'Filter Bandwidth':

- 2 The sampling rate is set to twice (2x) the filter rate; the minimum Nyquist-limited sampling rate must be at least double the filter rate.
- 5 The sampling rate is set to five times (5x) the filter rate.
- 10 The sampling rate is set to ten times (10x) the filter rate.

For the 1 MHz Filter Bandwidth:

- 2.5 The sampling rate is set to two-and-a-half times (2.5x) the filter rate; this downsamples the initial 5 MHz data by 50%.
- 5 The sampling rate is five times (5x) the filter rate; this uses the maximum sampling rate of 5 MHz, i.e., no downsampling occurs before filtering

Sampling Interval

The duration of a single data sample.

Math Type

Apply a data transformation to a virtual input signal:

• BaselineSubtract

Subtract a fixed value from all data points in an input trace.

This is useful for adjusting for an offset, or resetting a baseline.

Source Channel Select an input channel to process.

Baseline From Select how to calculate the subtraction

value.

Value Subtract a fixed value.

Value Spinner adjusts in 1 pA or 1 mV

increments.

• Trace Subtract the average of the entire input

trace.

• Sweep Time Subtract the average of the data between

the Start Time and End Time.

Start Time Set the starting time of the data to be

averaged.

End Time Set the ending time of the data to be

averaged.

• Segment #s Subtract the average of a Segment from

the input trace.

Start Ratio Set the starting time of the data to be

averaged, as a ratio relative to the starting time of the Segment

duration.

Start Time [derived value]

End Ratio Set the ending time of the data to be

averaged, as a ratio relative to the ending time of the Segment duration.

End Time [derived value]

Limit to Marked Sweeps

Enable to limit this analysis to marked

sweeps.

• BesselFilter A frequency-domain filter with excellent response characteristics for preserving the shape of a biological signal.

Source Channel Select an input channel to filter.

Filter Bandwidth Select a frequency range.

• LowPass Allow signal frequencies less than the

cutoff frequency, and block all higher frequencies, such as high frequency

noise.

• HighPass Allow signal frequencies greater than

the cutoff frequency, and block all lower

frequencies.

Filter Order

[1, 2, 4, 8]

Number of "poles" in the filter.

Cutoff Frequency (Hz)

[$100 \text{ to} < \frac{1}{2} \text{ the sampling rate}$]

Restrict frequencies from this boundary

point onwards.

Integrator Reset [for Capacitive Mode]

• Ignore Capacitive-mode transients are

displayed in the data.

• Blank The data during capacitive transients

are made invisible by replacing those data points with NaNs (Not a Number).

Blank Duration [$10 \mu s - 1 s$]

• Mask The data during capacitive transients

are replaced by the last data value before

the transient discharge, simulating a

sample-and-hold operation.

Mask Duration $[10 \,\mu\text{s} - 1 \,\text{s}]$

The default value of 500 μs should be

sufficient to encompass the reset

transient duration.

CycleAverage Apply averaging across cycles for each numbered sweep.

Source Channel Select an input channel to average.

Limit to Marked Sweeps

Enable to limit this analysis to marked

sweeps.

• Differentiate Apply differentiation to an input signal. The instantaneous rate of change in the signal is displayed.

Source Channel Select an input channel to differentiate

Equation Specify an equation to process an input signal.

> Source Channel Select an input channel to process.

Equation [Click field to access the 'Specify math

equation' editor.

Note: The full equation is always visible

as a tool tip, by hovering the mouse cursor over the 'Math

Equation' field.

Specify math equation for virtual signal

[<equation>] A free-form text field.

Errors are reported under this

field.

Check Equation Check the equation syntax. The

> equation is evaluated for sweep #1, and if valid, it reports "Syntax

is ok".

Insert special identifier

Special references can also be used within commands:

p[#] n'th Paradigm variable.

s[series-count, sweep-count, trace-count, routine

name

Access an arbitrary input trace (data wave) via counts of Series #, Sweep #, Trace # (Scope Position), and the Routine name.

The "current" item is the "active" trace in the scope window, and has a count value of zero.

If a "count" number is non-zero, it is used as an offset from the current count value of zero. Any fractions in count numbers are truncated to integers.

If the routine name is left blank, the current routine name is used.

Ex: s[0,0,0,]

The current series, current sweep, current trace, of the current routine.

t[#] nth input trace.

> Access the input trace (data wave) in Scope Position "n" for the last sweep of the current

Series.

This numbering can differ from the Scope Position "n" if signals are re-arranged or hidden.

Undo All changes in the equation

editing session are discarded.

(See the Equation Editor for more details.)

Limit to Marked Sweeps

Enable to limit this analysis to marked sweeps.

Integrate Display the integral of the data signal. This is equivalent to the signed area under a curve.

> Source Channel Select an input channel to integrate.

Leak Remove leakage current from the data signal. This is the small passive current when the cell is in a resting state.

> This feature is only enabled when the Routine includes an output channel with P/N Leak Pulse enabled.

Source Channel Select an input signal to process.

Show Leak Display the accumulated leak currents

after the subtracted data in a sweep. Display the average of the leak sub-

pulses.

Leak Zero Segment Identify a segment with no active

cellular response to the command signal.

When set to zero, the field is set to 'OFF'.

To re-display the numeric spinners, enter a non-zero number into the field.

The mean of the second half of the specified segment is used to compute an averaged leak current, which is then used to correct the P/N leak average. This option reduces the influence of a constant leak-current, which is otherwise included in the current of the main signal.

• LineFreq Remove AC line frequency noise (hum) from the data signal.

Source Channel Select an input channel for noise reduction.

2000

Line Frequency

60 Hz Canada, (Caribbean),

Central America, (Japan), Mexico, (South America), South Korea, Taiwan,

USA.

Some (regions) include both 50 Hz and 60 Hz

frequencies.

50 Hz Most of rest of world.

Alternating current (AC) power contains 50 or 60 Hz oscillations that can cause sinusoidal line-frequency noise in recorded signals. This FFT-based filter reduces such noise by > 90% over 6 harmonics. The adjusted signal is displayed in real time.

LockIn Measure cell characteristics (such as membrane capacitance) with high signal-to-noise sensitivity, using a dual-phase software lock-in amplifier.

Note: This feature is only enabled when the Routine includes an output channel with a waveform Segment set to 'Sine / Sine Wave Cycles / For LockIn'.

Calculations are made using 'conductance' (1 / resistance) insead of 'resistance'.

Current Channel Select a (source) input channel with a "current" signal.

Trace Kind

Select the LockIn measurement to

display.

The selected 'Trace Kind' is

automatically set as the Virtual Channel

label.

CMComputed membrane

capacitance.

GMComputed membrane

conductance.

GSComputed series conductance.

DCDC component of measured

signal.

RealY Real number part of the lock-in

response signal.

ImagY

Imaginary number part of the

lock-in response signal.

Cycles to Average [1-1000]

Cycles to Skip [1-1000]

V-reversal [±1000 mV]

> When using a calculated stimulus trace, enter the reversal potential for the ion under study, such as for (Na⁺) sodium spikes or (K⁺) potassium tail currents.

(see the SutterPatch Algorithms appendix)

Smooth Smooth the data with a "moving average" noise-reduction filter.

> Source Channel Select an input channel to smooth.

Smoothing Type

A standard filter with excellent 10 - 90%Gausian

rise-time response

Smooth Operations [1-32767]

of smoothing operations to perform.

• Boxcar A fast time-domain filter with excellent 0 – 100% rise-time response

Smooth Repetitions [1-32767]

of smoothing repetitions to perform.

Boxcar Window Points [1 – 99]

of points in boxcar sliding window.

Note: For best performance, only odd values are used.

Integrator Reset [for Capacitive Mode]

• Ignore Capacitive-mode transients are displayed in the data.

• Blank The data during capacitive transients are made invisible by replacing those data points with NaNs (Not a Number).

Blank Duration $[10 \mu s - 1 s]$

 Mask The data during capacitive transients are replaced by the last data value before the transient discharge, simulating a sample-and-hold operation.

Mask Duration [$10 \mu s - 1 s$]

The default value of $500 \mu s$ should be sufficient to encompass the reset transient duration.

• Stimulus Replicate the command waveform.

Source Channel Select an input channel – the waveform from its Parent Out Chan is used.

• SweepAverage

Average the input traces.

Source Channel Select an input channel to average.

Average Type Cumulative

Average all processed sweeps

together.

RunAverage

Average the last "N" sweeps.

Number of Sweeps [for Run Average]

Limit to Marked Sweeps

Enable to limit this analysis to marked

sweeps.

• SweepSubtract

Subtract a sweep from the input trace.

Source Channel Select an input channel to process.

Reference Sweep Select a sweep to be subtracted from all

other sweeps. If the sweep does not yet

exist, no subtraction occurs.

Limit to Marked Sweeps

Enable to limit this analysis to marked

sweeps.

4.2.8 Equation Editor

The Equation Editor manages simple or complex expressions that evaluate to a value. Such math equations can be used to create stimulus waveforms, or for data analysis.

Access the Equation Editor from the SutterPatch menu.




Figure 4-67. Equation Editor

Equation Pool Files Equations are created and saved in an Equation Pool file.

New Equation Pool	Create a blank Equation Pool file.			
Load Equation Pool	Load the Equations of a previously saved Equation Pool file into the Equation Pool.			
Revert to Last Saved	Undo any unsaved changes to the Equation Pool.			
Save Equation Pool	Save the Equation Pool using its existing file name and path.			
Save Equation Pool As	Save the Equation Pool to a new file, and switch to the new file. The default file name is the original file name.			
Save Equation Pool Copy	Save the Equation Pool to a new file, but do not switch to the new file. The default file name has 'Copy of' prepended to it.			
Merge Equation Pools	Insert the Equations from a previously saved Equation Pool file into the loaded Equation Pool.			
Note: Equation Pool files are simple text files (*.txt) that can be directly edited.				

New Create a blank Equation.

Duplicate Add a copy of the selected Equation to the Equation Pool.

Delete Remove the selected Equation from the Equation Pool.

Revert Select an Equation and click the Revert button. All editable steps are re-

set to their last saved settings.

Save Pool Save the Equation Pool using its existing file name.

Edit Make edits to the 'Equation' field.

Check Equation The equation is evaluated (for sweep #1), and if valid, it reports "Syntax

is ok".

Insert special identifier

Acquisition, amplifier and reference settings are available for use in

equations. (see list below)

Equations Table

Label A column of editable equation names, for easy usage in equations

in place of the actual equation. (see 'Syntax' below)

Equation A column of associated equations in free-form text fields, that

evaluate as math expressions.

Equations are limited to a maximum of 80 characters, including

white space.

Special Identifiers

These acquisition, amplifier and reference settings are available for use in equations:

Timing

Time (present date-time, s)

Timer (timer time, s)

ParadigmTime (time at start of paradigm, s)

RoutineTime (time at start of routine, s)

Paradigm Parameters

Loop (active paradigm ForLoop count)

Sweep (active paradigm EachSweep count)

LastSweep (active paradigm sweep count of last sweep)

Processing can occur before or after the last sweep of a

series.

Example: Compare 'sweep' numbers in a

ForEachSweep loop in a Paradigm 'If'

step:

ForEachSweep

EachSweep, Target=IV

If, Left=sweep, Operation='=', Right=LastSweep-1

Alert, Text=LastSweep, DoBeep=true

EndIf ForEachEnd

AqStopped (last acquisition was stopped)

The last Routine-Series did not complete by itself.

Stimulant (last applied stimulant concentration)

From the Solution Editor 'Concentration' setting for solutions configured as a 'Chemical Stimulant'.

Input (Input variable on paradigm window)

Hold[1..2] (holding of n'th output channel)

Headstage holding level.

p[1..16] (n'th paradigm variable)

r[1..16] (n'th routine stimulus variable)

Analysis Results

m[1..16] (n'th analysis measurement value)

gx[1..16] (n'th analysis graph x value)

gy[1..16] (n'th analysis graph y value)

Signal Readings

AuxIN[1..8] (auxiliary input, V)

A single-point voltage reading from an Auxiliary Input channel, such as from a slowly changing temperature

probe.

Note: This usage does not require setting up a

Routine Input Channel.

Imon (amplifier current reading, A)

In the Amplifier Control Panel (pA).

Vmon (amplifier voltage reading, V)

In the Amplifier Control Panel (mV).

Mean[name or count, start,width]

(mean of given input signal)

Headstage

ActiveProbe (active headstage)

[1-2]

The "active" probe number is the Sutter headstage presently controlled by the Amplifier Control Panel.

For a single headstage system, the active probe is

always headstage number "1".

NumProbes (number of headstages)

[1-2]

The number of dPatch headstages attached to the

system.

dPatch Settings

CCMode (amplifier current clamp)

VCMode (amplifier voltage clamp)

Hold (IHold in CC-mode, VHold in VC-mode)

 $[\pm 0.000,000,020 \text{ A } (\pm 20 \text{ nA}), \text{ or } \pm 1.000 \text{ V } (\pm 1000 \text{ mV})]$

Headstage holding level.

IHold (amplifer holding current, A)

 $[\pm 0.000,000,020 \text{ A } (\pm 20,000 \text{ pA})]$

IHoldOn (amplifier holding current On)

VHold (amplifier holding voltage, V)

[±1.000 V (±1000 mV)]

VHoldOn (amplifier holding voltage On)

IGain (amplifier current gain, V/A)

Read the gain of the active voltage-clamp 'Current'

input channel.

VGain (amplifier voltage gain, V/V)

Read the gain of the active current-clamp 'Voltage'

input channel.

Feedback (amplifier feedback mode: 0, 1, 2)

Filter (amplifier input filter in VC- and CC- mode, Hz)

Read the filter of the appropriate input channel in

either VC- or CC-mode.

IFilter (amplifier input filter in VC-mode, Hz)

Read the filter of the 'Current' input channels.

VFilter (amplifier input filter in CC-mode, Hz)

Read the filter of the 'Voltage' input channels.

Offset (amplifier pipette offset in VC-mode, V)

OffsetLock (amplifier pipette offset lock ON in VC-mode)

SubtractPipOffset (subtract pipette offset ON in VC-mode)

StimBoth (stimulate both headstages during auto-compensation

On)

LiquidJunc (liquid junction potential offset, V)

GentleSwitchC2V (gentle mode switch CC- to VC- mode)

GentleSwitchV2C (gentle mode switch VC- to CC- mode)

CapResets (number of capacitor resets since last sweep start)

dPatch Compensation

ECompMag (amplifier electrode compensation magnitude, F)

ECompTau (amplifier electrode compensation tau. s)

ECompOn (amplifier electrode compensation On in CC-mode)

CmComp (amplifier cell compensation Cm, F)

RsComp (amplifier cell compensation Rs, Ohm)

RsCompOn (amplifier cell compensation On)

Bridge (amplifier bridge balance, Ohm)

BridgeOn (amplifier bridge balance On)

dPatch Correction

RsCorr (amplifier Rs correction, fraction)

RsLag (amplifier Rs correction lag, s)

RsCorrOn (amplifier Rs correction On)

Dynamic Clamp

DynClampOn (amplifier dynamic Clamp On)

DynCGHold1 (amplifier dynamic Clamp G-holding 1, S)

DynCGHold2 (amplifier dynamic Clamp G-holding 2, S)

Dynamic Holding

DynHoldOn (amplifier dynamic holding On)

[0 = Off, 1 = On]

DynHold (amplifier dynamic holding potential, V)

DynHoldSpeed (amplifier dynamic holding speed)

DynHoldAqOn (amplifier dynamic active while acquiring)

Membrane Test

Relectr[1..2] (electrode/seal/access resistance, Ohm)

Value from last Membrane Test.

Rmemb[1..2] (membrane resistance (cell mode), Ohm)

Value from last Membrane Test.

Cmemb[1..2] (membrane capacitance (cell mode), F)

Value from last Membrane Test.

RMSNoise[1..2] (membrane test RMS noise, A)

Value from last Membrane Test.

Lock-In

LockInPhaseAdj (Lock-In phase delay adjustment, s)

LockInAttenAdj (Lock-In attenuation adjustment)

Other identifiers are forwarded to Igor Pro's 'Execute' command.

Equation Extras

Arguments

X

The "X" (or "x") specifier allows an argument to be passed to an equation. Insert "X" as the placeholder(s) in numeric expressions.

To call such an equation in other parts of the program, prepend a "#" to the equation label, and append the argument in parentheses.

Example 1 Pass the value "1.7" to the named equation

label "My_Equation":

 $\#My_Equation(1.7)$

Example 2 Send an AuxOUT voltage command to a

piezo-drive controller in distance units, us-

ing the sample conversion formula:

volts = $((micrometers + 0.08)/4.04)^1.3$

Instead of retyping this equation every time it is used, use an argument 'X' in the

equation:

volts = $((X + 0.08)/4.04)^1.3$

Label the equation as:

um2volt

Pass a distance of 10 micrometers to the labeled equation in a Routine (Routine Editor / Output Channels / Waveform Editor / Amplitude segment), or in a Paradigm (Paradigm Editor / Amplifier step / Auxiliary Output target) as:

#um2volt(10)

Constants

true 1 false 0

OFF

The following constants have 27-digit precision:

e 2.71... (Euler's number)

pi 3.14... (π)

0

Lists

Anywhere equations can be used, a list of comma-separated equations can also be used, to generate a sequence of values. If the sequence extends beyond the end of the list, the sequence wraps around and continues from the beginning of the list again, and so on.

Places used:

Paradigm Steps

Amplifier

Checkbox

Set Variable

Sound

Write Log

If

Else If

Routine Editor

Virtual Input Channel: Equation

Waveform Editor: Amplitude, Duration

Measurements: Time to Threshold

Graphs: X-Axis, Y-Axis

Example:

Create a sequence of increasing values with a 1/2/5 progression, such as might be used to increase a Routine's waveform amplitude or duration, on a per sweep basis:

1m*10^ceil(sweep/3),2m*10^ceil(sweep/3),5m*10^ceil(sweep/3)

This will generate a sequence of values of: 10m, 20m, 50m, 100m, 200m, 500m, 1000m...

The 'ceil' function rounds up any fraction to the next higher whole number, and "sweep" is a special identifier that reports the active sweep number. So, for the first 3 sweeps (1, 2, 3), "ceil(sweep/3)" generates a '1'. As '10' raised to '1' is '10', the initial number (1, 2, 5) is multiplied by '10', resulting in values of "10m, 20m, 50m".

For the next 3 sweeps (4, 5, 6), the sequence wraps around the list, and now "ceil(sweep/3)" generates a '2'. As '10' raised to '2' is '100', the initial number (1, 2, 5) is now multiplied by '100', resulting in values of "100m, 200m, 500m".

Parsing and Operators

Equation parsing is executed from left to right, processing the highest precedence level operators first, except for comparison and bitwise operators, which associate from right to left.

Precedence	Operation Type	Operator
8	Comment	;
7	Exponentiation, Arithmetic operations: Left Shift, Right Shift	^, <<, >>
6	Negation operations: Unary Negation, Logical Negation	- , !
5	Multiplication, Division, Remainder	*, /, %
4	Addition, Subtraction	+, -
3	Bitwise operations: And, Or, Nor, Xor	&, , nor, %^
2	Comparison operations: Greater Than, Greater Than or Equal, Less Than, Less Than or	>, >=, <, <=, ==, !=

	Equal, Equal To, Not Equal To	
1	Logical operations: And, Or, Conditional	&&, ,?:
0	All other operations	round, trunc, ceil, floor, exp, sqrt, ln, log, sin, cos, tan, asin, acos, atan, abs, rad, deq, noise, random

Table 4-4. Equation Parser

For expressions using Comparison and Logical operators, it is recommended to use parentheses to explicitly define the order of execution.

There are also some differences in operator processing between the SutterPatch equation parser and the Igor Pro command parser:

Comment:

SutterPatch All characters to the right of a semi-colon

are ignored.

Igor Pro: A semi-colon separates multiple commands

on the same command line. An Igor com-

ment uses a double slash: //

The Conditional operator "?:" is a shortcut for an if-else-endif expression. It evaluates as:

<expression> ? <True> : <False>

If the <expression> operand evaluates as non-zero, the <True> numeric operand is evaluated.

If the <expression> evaluates as zero, the <False> numeric operand is evaluated.

For complex expressions, only the real portion is evaluated.

Note: The ": " is a colon with two blank spaces around it.

An arithmetic left shift (<<) is the same as a bitwise left shift, whereby the least significant bit is padded with a zero. However, while an arithmetic right shift (>>) fills the most significant bit with its original value, thus preserving the sign, a bitwise right shift pads the most significant bit with a zero. A bitwise right shift can be constructed from existing operators.

Example: Shift # right by "n" bits

 $\#/2^n$

Syntax

All equations use the same syntax as Igor Pro, with a few additions:

- Three kinds of brackets [], {}, (), can be used equivalently to improve the clarity of nested expressions.
- Numeric values can be written in scientific E-notation using exponents:

5e-12 (5 picoamps)

or in equivalent engineering notation using unit prefixes:

5p (5 picoamps)

Prefix	Exponent	Prefix Name	Prefix	Exponent	Prefix Name
k	10 ³	Kilo	m	10 ⁻³	milli
М	10 ⁶	Mega	μ (or u)	10 ⁻⁶	micro
G	10 ⁹	Giga	n	10 ⁻⁹	nano
Т	10 ¹²	Tera	р	10 ⁻¹²	pico
Р	10 ¹⁵	Peta	f	10 ⁻¹⁵	femto
E	10 ¹⁸	Exa	а	10 ⁻¹⁸	atto
Z	10 ²¹	Zetta	z	10 ⁻²¹	zepto
Υ	10 ²⁴	Yotta	у	10 ⁻²⁴	yokto

Table 4-5. Engineering Notation

• Insert an equation from the Equation Editor Pool into an Equation field by entering "#" followed by the label of the equation, e.g., "#MyLabel". For variable inputs, "#MyLabel(5)" passes the argument "5" to the equation labeled "MyLabel" for evaluation.

Example: Using an LED light source

To stimulate in increments of light intensity, use an equation to transform light intensity values in Routine variables into actual stimulus values with amplitudes in volts.

Build an equation in the equation pool as follows:

equation = $\ln(r[1]) * 2.55 + 3$

The natural log of the Routine Variable r[1] is multiplied by '2.55' and added to '3'.

label = power_to_volts

In the Waveform Editor, set a Segment Amplitude field to 'Equation', and enter the equation as "#power_to_volts".

Areas of SutterPatch that use a simplified version of the Equation Editor and allow equations and equation Labels to be used:

Paradigm Editor

Steps: Amplifier, Checkbox, Set Variable, Sound, Write Log, If,

Else If

Routine Editor

Input Channels: Virtual Channels: Math Type: Equation

Output Channels: P/N Leak Pulses: Leak Hold; Waveform Editor: Ampli-

tude, Duration

Measurements: Time to Threshold, X-Axis, Y-Axis

Note: Computing an equation for a data wave consumes significant computing power, as every data point needs to be computed by the CPU. While a slight update delay in such operations is expected, for computers with marginal computing power, the "beach ball" icon displays while the computer is unresponsive and busy processing.

4.2.9 Igor Analyses

Numerous mathematical operations are found in the Analysis main menu, and are documented in the Igor Pro Help.

These built-in Igor fitting analyses are also accessible via the "Scope" window right-click menus:

Curve Fitting Create your own fitting equation.

• Quick Fit Use a pre-defined equation:

line poly [3-10] poly_XOffset [3-10]

gauss

Ior

Voight

 $exp_XOffset$

dblexp_XOffset

exp

dblexp

dblexp_peak

 \sin

HillEquation

Sigmoid

Use for Boltzmann function.

Power

LogNormal

poly2D [1-10]

Gauss2D

Fit Between Cursors

Weight from Error Bar Wave

Textbox Preferences

Example: Perform a fit on a section of a sweep:

- 1. Open the data into a Reanalysis scope window.
- 2. In the scope window, right-click 'Toggle Cursor Info' to display the cursor pane.
- 3. Drag cursors 'A' and 'B' from the cursor pane onto the data to set the fitting range.
- 4. Right-click on the data, and select Quick Fit and the fit of your choice.
- 5. Fitting results are written to the Command window.

Other built-in Igor analyses include:

- Transforms
- Fourier Transforms
- Periodogram
- Lomb Periodogram
- MultiTaperPSD

- Discrete Wavelet Transform
- Continuous Wavelet Transform
- Wagner Transform
- Short-Time Fourier Transform

- Convolve
- Correlate
- Differentiate
- Integrate
- Smooth
- Interpolate
- Filter
- Resample
- Sort
- Histogram
- Compose Expression
- Packages
- Average Waves
- Batch Curve Fitting
- Function Grapher
- Global Fit
- Igor Filter Design Laboratory
- Median XY Smoothing
- MultiPeak Fitting
- Percentiles and Box Plot
- Wave Arithmetic

4.2.10 Layout Window

The Layout window is used to prepare your data for publication. Scope window input signals, analysis graphs and other objects can be exported to a Layout window for graphical arrangement and editing.

A default Layout window is automatically created when SutterPatch is launched - display it with the menu command SutterPatch / Layout Page / Show Layout. Only one Layout window exists at a time. If no Layout window exists, it can be manually created via the 'Layout' button

located in various windows. The Layout window can also be created by running a Paradigm 'Export' step.

Note: Layout windows are sometimes created hidden behind other windows.

Clicking a Layout button appends its associated items into an existing Layout page (or a new Layout window.) Each signal and analysis graph is appended as an individual object.

The default configuration of "2 x 4" ('column' x 'row') objects per page can be changed in Preferences / Export_Graphics or the Paradigm 'Export' step, and is applied when a new Layout window is created:

- 1 single pane
- 2 2 stacked panes
- 3 3 stacked panes
- 2 x 2 matrix
- 2 x 3 matrix
- 2 x 4 matrix

Once a Layout window page is filled, additional objects are automatically appended into additional Layout pages.

A toolbar displays in the upper-left edge of the Layout window – the top two buttons reconfigure the toolbar buttons:

- Operate Mode Selection tools and object insertion mode.
- Draw Mode Drawing tools mode.

The main menu 'Layout' command also displays to 'Show' or 'Delete' the Layout window.

4.2.11 Metadata Review

"Metadata" parameters describe the system environment, attached Sutter instrumentation, Paradigm and Routine acquisition settings, and tag information.

Metadata parameters can be retrieved several different ways:

Data Navigator:

Open a Metadata Review docked sub-window.

When this docked window is open, it is linked to the Data Navigator window, where changing the Paradigm or Series node selection also updates the docked Metadata Review parameters.

- Select a Paradigm or Series and click on 'Metadata' in the Preview pane.
- Select a Paradigm or Series, then click the 'Available actions' button and select 'View Metadata'.
- Right-click a Paradigm or Series and select 'View Metadata'.

Reanalysis scope window:

Open a Metadata Review floating sub-window.

Open a Series into a Reanalysis scope window (from the Data Navigator 'Analyze' button or Action 'Analyze Routine'), and use the 'View Metadata' button to open a Metadata Review floating sub-window.

To simultaneously display metadata from two different Paradigms, use floating and docked windows:

Select the first Paradigm and the Action 'Review Paradigm'. Then, in the Paradigm Review window, click the 'View Metadata' button to open a Metadata Review floating sub-window. Next, select the second Paradigm, and select the Action 'View Metadata' to open a Metadata Review docked sub-window.

To simultaneously display metadata from two different Series, use floating and docked windows:

Select the first Series and the Action 'Analyze Routine' (or 'Review Routine'). Then, click the 'View Metadata' button to open a Metadata Review floating sub-window. Next, select the second Series, and the Action 'View Metadata' to open a Metadata Review docked sub-window.

Paradigm	[]	Displays the name of the Paradigm.
Routine	[]	Displays the name of the Series.
Signal	[]	Displays the name of the Signal [only if selected in the Data Navigator]



Copy to Clipboard

Copy the metadata settings to the system clipboard.

Metadata parameters from all detail levels are displayed (see Set Preferences / Metadata).

- By Event Events are grouped by [time-stamp] [Event #] [Event type]. Highlighted values are editable.
 - Absolute Time Display the event times in the computer system time.

• Relative Time Display the event times relative to the start of the Paradigm.

• By Parameter This allows you to look at a Parameter and how it changes

over time. Parameters are grouped into major categories.

(see below)

Parameters might also list "Prior" values. These include system and Paradigm parameters written before a Routine

starts.

Expand All All settings entries are displayed.

Two columns of information are presented (parameter name and value). If the first column's text does not fully display, either increase the width of the window, or adjust the indentation of the second column by dragging it when the mouse cursor turns into a double-headed arrow.

Collapse All All settings entries are hidden, collapsed up to the Event or

Parameter level.

'By Parameter' Metadata Categories / Parameters

<u>Tag</u>

Tag Number

Tag Creation Timestamp

Timer Time at Tag Creation

Tag Signals

Tag Source Event

Operator

Login Name

Experiment

Experiment Timestamp

Amplifier

Amplifier Sequence Number

Amplifier Manufacturer

Amplifier Model

Amplifier Revision

Amplifier Firmware Version

Amplifier Serial Number

Amplifier Channel

Number of Available Headstages

Headstage Sequence Number

Headstage Model

Headstage Revision

Headstage Serial Number

[for attached HW] [for attached HW]

Headstage Preamplifier Model Headstage Preamplifier Revision

[for attached HW]

<u>Instrumentation and Software</u>

Interface Sequence Number

Interface Manufacturer

Interface Model

Interface Revision

Interface Firmware Version

Interface Serial Number

Interface Input Channel (physical)

Interface Signal Type

Interface Number of Digital Outputs

Computer Name

Physical Computer Memory

Operating System Platform

Operating System

Software Environment

Software Environment Version

Software Environment Build

Software Environment Kind

Software Environment Serial Number

Data Acquisition Software

Data Acquisition Software Version

Data Acquisition Software Build

Data Acquisition XOP Version

Paradigm

Paradigm Data Sequence Number

Paradigm Data Base Name

Paradigm Name

Paradigm Description

Paradigm Data Start Timestamp

Series (= Routine Data)

Series Sequence Number

Series Base Name

Routine Name

Routine Acquisition Mode

Routine Description

Routine Data Start Timestamp

Routn. Completed / Terminated Early

Number of Input Signals

Sweep

Sweep Number

Data Acquisition Settings

Active Headstage

Recording Mode

Headstage Gain

Headstage Feedback Mode	
Filter Cutoff Frequency	
Filter Type	
Input Offset Voltage	[VC mode]
Input Offset Lock On/Off	[VC mode]
Input Liquid Junction Potential	
Subtract Pip. Offset in Current Clamp	[CC mode]
Input Signal Name	
Input Signal Unit	
Input Scaling Factor	[for AuxIN channels]
Input Scaling Offset	
Input Full-scale Minimum	
Input Full-scale Maximum	
Input Sampling Rate	
Auxiliary Input Signal Offset	[for AuxIN channels]
Virtual Signal Scaling Offset	[for Virtual Input channels]
Virtual Signal Math Type	[for Virtual Input channels]
Virtual Signal Equation	[for Virtual Input channels]
Virtual Signal Source Channel	[for Virtual Input channels]
Virtual Signal Source Signal Name	[for Virtual Input channels]
Virtual Signal Subtract Baseline Type	[for Virtual BaselineSubtract]
Virtual Signal Subtract Baseline Start	[for Virtual BaselineSubtract]
Virtual Signal Subtract Baseline End	[for Virtual BaselineSubtract]
Virtual Signal Filter Type	[for Virtual BesselFilter]
Virtual Signal Filter Order	[for Virtual BesselFilter]
Virtual Signal Filter Cutoff Frequency	[for Virtual BesselFilter]
Virtual Signal Integrator Reset Strategy	[for Virtual BesselFilter &
virtual Digital Integration Reset Strategy	Smoothing: Capacitive mode]
Virtual Signal Integrator Reset Duration	[for Virtual BesselFilter &
virtual Digital Integration Reset Duration	Smoothing: Capacitive mode]
Virtual Signal Leak Display On/Off	[for Virtual Leak]
Virtual Signal Leak Zero Segment	[for Virtual Leak]
Virtual Signal Line Frequency Base	[for Virtual LineFreq]
Virtual Signal LockIn Trace Kind	[for Virtual LockIn]
Virtual Signal LockIn Cycles to Average	[for Virtual LockIn]
Virtual Signal LockIn Cycles to Average Virtual Signal LockIn Cycles to Skip	[for Virtual LockIn]
Virtual Signal LockIn Cycles to Skip Virtual Signal LockIn Reversal Potential	[for Virtual LockIn]
_	[for Virtual Smoothing]
Virtual Signal Smoothing Algorithm Virtual Signal Smoothing Factor	
Virtual Signal Swoons Processed	[for Virtual Smoothing]
Virtual Signal Reference Sweep	[for Virtual SweepAverage]
Virtual Signal Reference Sweep	[for Virtual SweepSubtract]
Electrode Fast Magnitude	[VC mode]
Electrode Fast Time Constant Whole call Companyation On Off	[VC mode]
Whole-cell Compensation On/Off	[VC mode]
WC Comp – Series Resistance	[VC mode, if WC Comp. On]
WC Comp – Membrane Capacitance	[VC mode, if WC Comp. On]
Series Resistance Correction On/Off	[VC mode]
Series Resistance Prediction Value	[VC mode, if Rs Correction On
Series Resistance Correction Value	[VC mode, if Rs Correction On
Series Resistance Corr. Lag Time	[VC mode, if Rs Correction On
Capacitance Neutralization On/Off	[CC mode]

Capacitance Neutralization Mag.	[CC mode, if Cap Neut On]
Bridge Balance On/Off	[CC mode]
Bridge Balance Resistance	[CC mode, if Bridge Balance On]
Current Clamp Dynamic Hold On/Off	[CC mode]
Current Clamp Dyn. Hold Potential	[CC mode, if Dynamic Hold On]
- ·	-
Current Clamp Dynamic Hold Speed	[CC mode, if Dynamic Hold On]
CC Dynamic Hold On for Acquisition	[CC mode, if Dynamic Hold On]
Dynamic Clamp On/Off	[CC mode]
Dynamic Clamp Conductance	[CC mode, if DynClamp Active]
Dynamic Clamp Conductance Descr.	[CC mode, if DynClamp Active]
Dynamic Clamp Model]	[CC mode, if DynClamp Active]
Command Signal Name 1	
Command Signal Unit 1	
Command Full-scale Minimum 1	
Command Full-scale Maximum 1	
Command Sampling Rate 1	
Command Holding Enabled 1	
Command Holding Value 1	['0' if Holding disabled]
Auxiliary Output Signal Name 1	[if AuxOUT1 enabled]
Auxiliary Output Scaling Factor 1	[if AuxOUT1 enabled]
Auxiliary Output Offset 1	[if AuxOUT1 enabled]
Auxiliary Output Holding Value 1	['0' in D/IPA demo mode]
Command Signal Name 2	[if StimOUT 1 & 2 enabled]
Command Signal Unit 2	[if StimOUT 1 & 2 enabled]
Command Full-scale Minimum 2	[if StimOUT 1 & 2 enabled]
Command Full-scale Maximum 2	[if StimOUT 1 & 2 enabled]
Command Sampling Rate 2	[if StimOUT 1 & 2 enabled]
Command Holding Enabled 2	[if StimOUT 1 & 2 enabled]
Command Holding Value 2	['0' if Holding disabled]
Auxiliary Output Signal Name 2	[if AuxOUT2 enabled]
Auxiliary Output Signal Name 2 Auxiliary Output Scaling Factor 2	[if AuxOUT2 enabled]
Auxiliary Output Offset 2	[if AuxOUT2 enabled]
Auxiliary Output Onset 2 Auxiliary Output Holding Value 2	[II AuxOO 12 ellabled]
v 1	[if Ann OITT2 anablad]
Auxiliary Output Signal Name 3	[if AuxOUT3 enabled]
Auxiliary Output Scaling Factor 3	[if AuxOUT3 enabled]
Auxiliary Output Offset 3	[if AuxOUT3 enabled]
Auxiliary Output Holding Value 3	Fig. A. OTTMA
Auxiliary Output Signal Name 4	[if AuxOUT4 enabled]
Auxiliary Output Scaling Factor 4	[if AuxOUT4 enabled]
Auxiliary Output Offset 4	[if AuxOUT4 enabled]
Auxiliary Output Holding Value 4	
Digital Holding Pattern $(1 \rightarrow N)$	[1-16] bits
C4:	
Stimulus	F:00 1 /: / 13
Compound Group	[if Solution activated]
Compound Group Index	[if Solution activated; for Control
	and Test Compounds]
Compound Name	[if Solution activated; for Control
	and Test Compounds]
Compound Description	Lif Solution activated: for Control

[if Solution activated; for Control

Compound Description

or Test Compound]

Compound Concentration

Compound Concentration Unit

[if Solution activated; for Control

and Test Compounds]

[if Solution activated, for Control

and Test Compounds]

4.2.12 Paradigm Review

'Paradigm Review' displays data from all Series within the Paradigm in a modified reanalysis scope window, in a Time Course or Concatenated view. This view also displays tags that occur between Series.

Access this window from the Data Navigator 'Available actions' menu 'Review Paradigm'.

The state of the Autoscale button (one-time vs. continuous) applies to all Paradigm Review and Routine Review windows.

To display the Series_Signal_Routine name of the selected data at the bottom of the window, click on the data.

To open a Series into a Reanalysis scope window, right-click on the Series data, and select Analyze <Series Name> from the menu list.

(For more information on the window controls, see the Reanalysis scope section.)

4.2.13 Reanalysis Measurements & Graphs

The Reanalysis scope window Measurements button 'Edit Measurements' provides access to the 'Reanalysis Measurements & Graphs' dialog. Use it to apply different analysis scenarios to recorded data. Settings changes for input channel measurements and analysis graphs override the loaded Routine for quick interactive control.




Figure 4-68. Reanalysis Measurements & Graphs

This dialog operates similarly to the Routine Editor: Real Time Measurements & Graphs dialog, with the addition of optional configurations:

Measurement Parameters from

The name of the Series data, and the date/time of the last update to the Measurement parameters.

The current parameters can be optionally overwritten (updated) from these sources:

- Overwrite Measurements from Original Routine
- Overwrite Measurements with Active Values

The last run analysis values are remembered.

• Overwrite measurements from Saved Default

[Available when previously saved in the 'Measurements' button drop-down list.]

Analysis Examples

Example 1: Plot the mean of the data (using sample routine IV).

- 1. Set measurement m[5] to the 'Mean' analysis and select signal Current1.
- 2. Enable graph [g5].
- 3. From the graph's Y-Axis list, select m[5]. The Equation field displays:

m[5].

- 4. Set 'X-Axis' to 'time'.
- 5. Run the analysis.
- 6. An Analysis window displays a graph of the mean vs. time.

Example 2: Plot the difference between two measurements.

- 1. Set measurement m[5] to the 'Mean' analysis and select signal Current1.
- 2. Set measurement m[6] to the 'Mean' analysis, using the same signal.
- 3. Adjust the m[6] cursors Start/End times so they do not overlap with the m[5] cursors.
- 4. Enable graph [g6].
- 5. For the graph's Y-Axis, select 'Y-Equation' and enter the equation as:

$$m[5] - m[6]$$
.

- 6. Set the X-Axis to 'time'.
- 7. Run the analysis.
- 8. An Analysis window displays a graph of the difference vs. time.

4.2.14 Reanalysis Scope

This analysis version of the scope window is used to display and reanalyze stored data.




Figure 4-69. Reanalysis Scope Window

Both physical and virtual channels can be displayed and analyzed here.

A few of the window's controls are the same as in the 'Acquisition: <Routine>' scope window with slight changes, and additional controls are also included:

Navigation pane

The Navigation pane appears at the top of the Reanalysis scope window. It displays an overview of the active signal's full-scale data, with a gray box surrounding the magnification area.




Figure 4-70. Navigator Pane

The Navigation pane "magnification" box can be used to scroll through the active signal's data. Place the mouse cursor over the magnification box and it changes into a 'hand' icon; click and drag the magnification box to move it around the data of interest.

Buttons

	Scope Settings:		
	Show All Sweeps	All sweeps are visible.	
⇔	Show Marked Sweeps	Only marked sweeps are visible.	
		The text "Showing Marked" displays above the Scope window 'Mark' checkbox.	
	Set all marks in sweeps of	of active series	
	Clear all marks in sweep	s of active series	
Run Analysis	•	for the data series, and graph the results in this button is grayed out, edit the Measures.	
	Note: To stop a long-rung the bottom right corner of	ning analysis, click on the 'Abort' button in f the main screen.	
Measurements		nent cursors in the scope window.	
	Hide Cursors: Do not display cursors in the scope window.		
	Button displays as	s "Measurements(H)".	
	Lock Cursors:		
		om being adjusted or moved.	
	Button displays as	s "Measurements(L)".	
	No Measurements or Gra	phs	
	Analyze with Active Mea	surements	
	Analyze with Original Ro	outine Measurements	

	Analyze with Routine Last Executed Measurements	
	Analyze with Saved Default Measurements	
	Save as Default Measurements	
Edit Measurements: Open a special Reanalysis Measurements & Graphs dia where all changes apply instantly to the measurements a the graphs, even during acquisition. These edits override to loaded routine for quick interactive control.		
	Edit Virtual Signals: Open the virtual input signals panel for editing.	
	Action Potential Analysis	
	Analyze action potentials.	
	Synaptic Event Analysis	
	Analyze synaptic events (EPSPs, mEPSPs, etc.)	
	Single Channel Analysis	
	Analyze a single ion channel.	
View Meta	Display any extra information (metadata) associated with the displayed data Series, such as the operator, preparation details, solution information, etc., in a floating window.	
Data Navigator	Open a Data Navigator window with all of your Experiment data and metadata available for analysis in a tree structure.	
4 •	Show and analyze [Previous / Next] Routine.	
[] Mark	Enable/disable to "mark/unmark" the active sweep for display and/or analysis.	
	The Data Navigator 'Available Actions' can process marked sweeps as a group.	
	When the Scope Settings 'Show marked sweeps' is enabled, the text "Showing Marked" displays above the Scope window 'Mark' checkbox.	
Y-Autoscale	Click to autoscale the Y-axis of the selected signal to its visible sweeps data limits.	
‡	To autoscale the Y-axes of all visible signals, in "Windows" Shift-click the button, or in "macOS" Control-click the button.	

	To include the zero amplitude in the Y-ranges, enable "Include zero when autoscaling" in Set Preferences / Scope Window / General.		
	Tip: To invert the Y with reversed p the Y-axis of th Range. Either and Maximum	-axis of the selected signal, such as for data solarity from an outside-out patch, right-click in e signal and select Axis Properties / Axis reverse the Manual Range Settings / Minimum values, or disable the Manual Range and enale Settings / Reverse axis.	
Autoscale	Click to autoscale the Y-axes of all signals to their visible sweeps data limits, and to set the X-axis range to the maximum defined sweep duration for all signals.		
23	For persistent autoscaling, in "Windows" right-click (or Shift-click or Ctrl-click) the button, or in "macOS" Control-click the button (it turns dark gray).		
	To include the zero amplitude in the Y-ranges, enable "Include zero when autoscaling" in Set Preferences / Scope Window / General.		
Persistence Display	Display all sweeps (per Marks and/or Scope Preferences). When disabled, only the active sweep is displayed.		
	Applies to the Scope window and its (right-click) Parametric Plot and Amplitude Histogram Plot graphs.		
Sweeps Display	This button has 3 modes:		
	Sweeps	Each trace starts from time zero to the duration of the waveform.	
	Time Course	Display sweeps in time sequence on a single time axis. Portions without data are left blank (such as the time between triggered sweeps.)	
		Note: Emulation mode has a minimum 0.5 s interval between sweeps, both triggered and continuous. If the sweep duration is less than 0.5 s, the time between sweeps will be padded with "blank" time,	
	Concatenated	Display sweeps similarly to the Time Course mode, but any blank portions are replaced by a vertical line.	
	The 'Show 3D view of current signal' button brings up a separate 3D display window attached to the right of the Analysis window. The Sweep data are color-coded for amplitude, and their 3D graph can be rotated in any direction.		

Sweep #:

The 'Sweep #' display at the bottom of the scope window indicates the 'active sweep' number, the total number of sweeps in the Series, and either "all" or the total number of visible sweeps (per Preferences).

Table 4-6. Reanalysis Scope Window Buttons

Tags are only shown in the Time Course and Concatenated display modes. They display as vertical blue lines at the tag time points in the data. Their associated text boxes are positioned in the top-most signal pane:

Event <#> Tag Comment <Time stamp> Comment: <text>

Right-click Menus

X Axis

Autoscale All Axes Scale all signals Y-axes to their data, and set the

X-axis range for all signals to the maximum de-

fined sweep duration.

Autoscale X Axis Set the X-axis range for all signals to the maxi-

mum defined sweep duration.

Set X Scale...

X-min

X-max

Axis Properties... Modify the axes style and components.

Y Axis

Autoscale All Axes Scale all signals Y-axes to their data, and set the

X-axis range for all signals to the maximum de-

fined sweep duration.

Continuous Autoscale Y Axis

Continuously scale the signal's Y-axis to its data.

Autoscale Y Axis Scale the signal's Y-axis to its data.

Full scale Y Axis Set the signal's Y-axis to its full-scale range.

Use Last Y Scale

Set Y Scale...

Y-min

Y-max

Axis Properties... Modify the axis style and components.

Hide Signal <name> Hide the selected signal in the scope window.

Show Signal <name> Only Show the selected signal in the scope window, hide

all other signals.

Stack All Signals Display all signals in a single column.

Main Window

Limited data modification menu

Right-click in the blank area in a signal pane.

Tip: If you click too close to the signal data, the full data modification menu displays instead if this occurs, click near a horizontal or vertical edge of the signal pane.

This context menu is the same as in the Acquisition: Routine scope window (plus a couple additional items):

Show All Sweeps (with triggered sweeps)
 Show Marked Sweeps (with triggered sweeps)

Autoscale All Axes Scale all signals Y-axes to their data, and set the

X-axis range for all signals to the maximum de-

fined sweep duration.

Add Annotation Add a floating text-box label to the signal pane.

To edit or delete an annotation, double-click on it.

Parametric Plot Display a graph of X vs. Y input signals in a sepa-

rate window.

Y-signal Select an input signal for the Y-axis.

X-signal Select an input signal for the X-axis.

Plot Update the plot using the new parameters.

Copy to Layout Copy the Parametric Plot graph into a new Layout

window, or append to an existing Layout page.

Copy to Clipboard Copy the Parametric Plot graph to the system clip-

board, or if the 'Shift' key is pressed, the complete

scope window.

Time Range The time range of the data to be plotted.

Full Trace Use the entire trace for the time range.

Sweep Time Set relative to the start time of a sweep

(time zero).

Start Time Set the starting time.

Once the Start Time is within 2 sample points of the End Time, further Start Time increments will increase the End Time by the same

amount.

End Time Set the ending time.

Once the End Time is within 2 ms of the Start Time, the End Time

cannot be decremented.

Segment Time Set the time range as a ratio of the Seg-

ment duration.

Segment Select the Segment number.

Start Ratio [0 = beginning of Segment]

Set the starting time ratio.

End Ratio [1 = end of Segment]

Set the ending time ratio.

Amplitude Histogram Plot Open a histogram plot window. The amplitudes of

the data are binned and plotted. The window is

cleared at the start of a new Series.

Y-signal Select the input signal to be analyzed.

Time Range The time range of the data to be plotted.

Full Trace Use the entire trace for the time range.

Sweep Time Set relative to the start time of a sweep

(time zero).

Start Time Set the starting time.

End Time Set the ending time.

Segment Time Set the time range as a ratio relative to the

Segment duration.

Segment Select the Segment number.

Start Ratio [0 = beginning of Segment]

Set the starting time ratio.

End Ratio [1 = end of Segment]

Set the ending time ratio.

Histogram Bins [50, 100, 200, 500, 1000, 2000, 4000]

Select the number of bins for the amplitude range (X-axis). Changes instantly update the plot.

Plot Refresh the plot for any Time Range / Sweep Time

settings changes.

Copy to Layout Copy the Amplitude Histogram Plot graph into a

new Layout window, or append to an existing Lay-

out page.

Copy to Clipboard

Copy the Amplitude Histogram Plot graph to the

system clipboard.

[graph pane]

Export Graphics Copy the signal and open in a separate window.

Saves to Windows / Graph Macros.

Toggle Cursor Info Show the Cursor Info pane to measure X-Y data

points or set a fitting range. (See the 'Signal data'

section below.)

Select 'Toggle Cursor Info' again to hide the Cursor Info pane, and any cursor symbols in the active

pane

Colors Adjust the colors used by the active signal pane:

graph background The background of the pane.

all axes The X- and Y-axis areas.

all grids The grid lines in the pane.

all tick labels The tick labels in the X- and Y-axis areas.

all axis labels The axis labels in the X- and Y-axis areas.

Hide Signal '<name>' Hide the selected signal in the scope window.

Show Signal '<name>' Only Show the selected signal in the scope window, and

hide all other signals.

Stack All Signals Display all signals in a stacked signal layout.

Show All Sweeps (with triggered sweeps)

Show Marked Sweeps (with triggered sweeps)

Marquee

Click and drag the mouse to surround a region of interest, and right-click for a context menu:

Expand Set the signal's Y-axis range from the marquee

vertical data limits, and set all signals X-axes ranges from the marquee horizontal data limits.

Horiz Expand Set all signals X-axes ranges from the marquee

horizontal data limits.

Vert Expand Set the signal's Y-axis range from the marquee

vertical data limits.

Shrink Move the signal's Y-axis current limits to the posi-

tion of the marquee vertical data limits, and move all signals X-axes current limits to the position of

the marquee horizontal data limits.

Horiz Shrink Move all signals X-axes current limits to the posi-

tion of the marquee horizontal data limits.

Vert Shrink Move the signal's Y-axis current limits to the posi-

tion of the marquee vertical data limits.

Extract Template Copy the last sweep to the Template Editor.

Extract To Graph

Display the active trace in a floating window, using all data within the X-range.

Signal Data

Full data modification menu

Right-click on or near the data to display this context menu, which includes options to modify sweeps and data points, such as marker symbols and lines.

This menu is the same as in the Acquisition: Routine scope window (plus a couple of other items):

• Hide Sweep_# (with triggered sweeps)

• Show Sweep_# Only (with triggered sweeps)

Tip: To manually measure X-Y data values, or to set a fitting range, use the Cursor Info pane:

• Toggle Cursor Info Select 'Toggle Cursor Info' to show/hide the Cursor

Info pane, and any cursor symbols in the active

pane.

Cursor Info pane



Options menu



One Mover Moves All Draggable cursor mover tool

moves all cursors together with a single control.

• All Styles Change the cursor symbol

style.

• Show Cursor Pairs Display up to 5 sets of cur-

sor symbol pairs.

Cursor A

 \oplus

Cursor symbol for data point 'A'.

A: Symbol letter (beginning cursor of the pair).

R1_A_IV (Default) wave name.



Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'A'.

Y: Y-axis value of data point 'A'.

 ΔY Difference of the cursor pair Y values.

<u>Cursor B</u>

Cursor symbol for data point 'B'.

B: Symbol letter (ending cursor of the pair).

R1_A_IV (Default) wave name.



Draggable cursor mover tool for the cursor pair.

pnt: Data point number (starting from zero).

X: X-axis value of data point 'B'.

Y: Y-axis value of data point 'B'.

 ΔX Difference of the cursor pair X values.

Cursor Instructions

1. Click on symbol 'A' to enable it.

- 2. Manually drag the highlighted symbol onto a data point in the scope window, or enter the data point number in the 'pnt' field.
- 3. Click on symbol 'B' to enable it.
- 4. Manually drag the highlighted symbol onto a data point in the scope window, or enter the data point number in the 'pnt' field.

X- and Y-measurements are displayed for the cursor pair data points.

5. Fitting can also be applied to the cursor pair data. Right-click on the data, and select 'Quick Fit' for a list of built-in Igor fitting functions.

The fit is displayed in the graph, and the fitting information is written to the Command window.

Channel Timing Delays

The dPatch amplifier uses analog output and input channels that pass signals through two filters, fixed DAC output stimulus filters and variable-bandwidth ADC input filters. These filters impose a time delay (also called a phase delay) on the physical input and output channels, which is reflected in the recorded response signals.

However, when headstage stimulus signals are recorded (such as 'Voltage1' in voltage-clamp mode), their output and input channels are directly processed as digital streams. Thus, there are no analog filter delays in recorded stimulus signals.

This means that recorded response signals are time-delayed compared to the recorded stimulus signals. Therefore, when analyzing data that require precise timing between headstage stimulus and response signals, these timing delays should be taken into account:

Typical dPatch Filter Delays

Input Bandwidth	<u>Total Filter Delay</u>
$1000~\mathrm{kHz}$	$16.2~\mu s$
$500 \; \mathrm{kHz}$	$17~\mu s$
$250~\mathrm{kHz}$	$18.5~\mu s$
$10~\mathrm{kHz}$	$85~\mu s$

4 2 15 3D View Window

The Reanalysis Scope 3D View window creates a 3D representation of your data, color-coded to show amplitude variations.

Note: If the Igor/SutterPatch main window frame is not wider than the Reanalysis scope window plus its Analysis sub-window, clicking the 3D button will generate an error message, but the operation will still execute.

The axis definition in 3D View is based on the change of a waveform over the course of successive sweeps. In a two-dimensional display, the X-axis represents the Sweep Time, while the Amplitude is plotted on the vertical Y-axis. For consistency, the vertical axis in the SutterPatch 3D view is also defined as the Y-axis. In the default orientation of the 3D View, the Z axis, on which the Sweep Number is plotted, points backward and to the right.




Figure 4-71. 3D Axes Definition




Figure 4-72. 3D View

A "heat map" bar illustrates the color measurement units.

Magnification buttons are located in the upper right corner of the window for the selected axis:

DPATCH - OPERATION MANUAL - REv. 2.2 (2021-10)


• All 3 axes.

• X Sweep Time.

• Y Amplitude.

Z Time.

Zoom in Magnify.



Autoscale Set to the data limits.

X, Y and Z axis limits can be set in the bottom section of the 3D View window. Their delta value is preserved when using the scroll bars to update the visual graph (and the numeric axes limits.)

The 3D graph viewing angle can be changed with a set of 3D buttons:

= Default View	X, Y & Z axes display.
X = Right View	Y & Z axes display.

Y = Top View Z & X axes display.

Z = Front View X & Y axes display.

Alternatively, you can rotate the display in any direction by simply clicking and dragging the 3D graph. If you release the mouse button while dragging, the 3D display will rotate in the direction of the mouse drag.

Show Events Display tagged events in the 3D graph.

Show Data Points Display data points as surface dots in the 3D graph.

4.2.16 Routine Review

'Routine Review' displays the selected Series data in a modified reanalysis scope window defaulted to the 'Time Course' display mode.




Figure 4-73. Routine Review

Open this window from the Data Navigator by highlighting a Series, and selecting the 'Review Routine' command from a right-click menu or the 'Available actions' button.

The state of the Autoscale button (one-time vs. continuous) applies to all Routine Review and Paradigm Review windows.

For more information on the window controls, see the Reanalysis scope section.

New Controls

• Click in a signal pane or on the data, to display the "Series_Signal_Routine name" and "Sweep number" of the selection point, at the bottom of the window.

• To reopen a Series from the Routine Review into a Reanalysis scope window, right-click on the Series data, and select 'Analyze <Routine Name>' from the menu list.

4.2.17 Routine Settings

The Routine Settings window reports the same settings as would be seen in the Routine Editor / Routine Settings however its preview pane does not support interactive dragging of measurement regions.

Open this window from the Data Navigator by highlighting a Series, and selecting the 'View Routine Settings' command from a right-click menu or the 'Available actions' button, or by selecting the 'Routine' name in the Data Navigator preview pane.




Figure 4-74. Routine Settings

Routine Name

Displays the Routine name.

Opens the Acquisition: Routine scope window loaded with these settings.

Copy to Routine Pool

Adds this Routine to the loaded Routine Pool.

Routine Description

Displays the Routine description.

Routine Settings Listing of all settings from all sections of the Routine.

Acquisition & Routine Parameters

Input Channels

Output Channels & Waveform

Real Time Measurements & Graphs

<u>Preview panel</u> Display of the stimulus waveforms.

Copy to Layout Copy the visible stimulus waveforms into a new Layout

window, or append to an existing Layout page.

Copy to Clipboard Copy the visible stimulus waveforms to the system

clipboard.

Some key acquisition settings and display controls are listed on the right of the Preview

pane:

Show Channel Select the output signals to display.

Show Trace Select the output traces to display.

Show Measure Regions Select the measurement regions to display.

4.2.18 Set Metadata

A variety of optional experimental parameters (preparation, electrode, etc.) can be associated with an Experiment, Paradigm, or Routine as user-configurable "metadata".

Predefine the Metadata parameter values here.

Show Summary An overview of the user-defined metadata parameters.

Metadata Summary dialog

This list summarizes all Metadata parameters (from all Metadata detail levels) for which values have been defined. Double-click a line to show and edit the details for a parameter.

Metadata Parameter Parameter name.

Current Value Parameter value.

Increment Enabled If enabled, double-click to review details.

Prompt before Display metadata prompts before running:

ExptExperimentPdgmParadigm

Routine

Metadata Group

Availability is per the SutterPatch > Set Preferences > Metadata > 'Metadata detail level'.

Level 1: Basic

The default metadata groups.

Preparation – Animal Preparation – Tissue Preparation – Cell Experiment Stimulus

Level 2: Extended

Includes more metadata groups.

Electrode

Recording Solution

Level 3: Full

Includes all metadata groups.

Operator Paradigm Cell Health / Quality Control

Series (= Routine Data)

Metadata Parameter

Per the SutterPatch > Set Preferences > Metadata 'Metadata

detail level'. (See table below.)

[Parameter description]

Configuration choices for the selected parameter

• Do not write this parameter This parameter is not stored.

If this parameter was previously written in this Experiment, then its Previous Value is displayed.

• Use last value The parameter used in the previous acquisition is

written.

• Use a previous value Select from a drop-down list of the previous 20

metadata values used for acquisition in the

experiment.

• Use new value Enter a new value for the metadata parameter.

• Increment Numerically increment the value.

By

Experiment At the start of each Experiment.
Paradigm At the start of each Paradigm.
Routine At the start of each Routine.

Prefix Enter text to be prepended to the value.

Start value The initial value (including decimals and

negative numbers.)

Increment: Select an arithmetic operator [+, -, *, /].

[] Enter the incremental amount.

Suffix Enter text to be appended to the value.

Prompt for confirmation (before):

• Experiment At the start of an Experiment.

• Paradigm At the start of a named Paradigm (i.e., pre-

planned, not auto-triggered by a Routine.)

• Routine At the start of a Routine.

Confirm Metadata Settings for This dialog displays whenever an Experiment,

Paradigm or Routine is started with metadata

prompts enabled.

Write Enable to write the selected metadata

parameter with the Experiment,

Paradigm or Routine.

Metadata Parameter The name of the selected metadata

parameter.

Next Value The metadata value to write.

Update Enable so that edits made to 'Value' will

update the 'last value' for the next prompt. This field is automatically

disabled after each execution.

Prompt Disable to remove the metadata

parameter from those listed in the Confirm Metadata Settings dialog.

The 'Confirm Metadata Settings' dialog only displays if a metadata parameter

has been enabled for 'Prompt'.

BASIC	EXTENDED	FULL	GROUP / Parameters	Notes
		F	OPERATOR	
		F	Full Operator Name	
В	Е	F	PREPARATION - ANIMAL	
В	Е	F	Animal Identifier	
В	Е	F	Animal Species	Binomial species name
	E	F	Animal Strain	Strain, breed or variety characterizing the animal
	Е	F	Animal Genotype	S
В	Е	F	Animal Age	
В	E	F	Animal Age Unit	Ex.: h, d, m
В	E	F	Animal Sex / Gender	Ex.: 1: F, 2: M, 3: Undetermined
В	E	F	Animal Weight	
В	E	F	Animal Weight Unit	
В	E	F	Annual Circadian Time or Phase	
	E	F	Animal Preparation Date	ISO Date, Format: YYYY- MM-DD
	E	F	Animal Preparation Time	Time of Day, Format: hh:mm[:ss.000]
	E	F	Animal User Parameter 1 Name	
	E	F	Animal User Parameter 1	
	Е	F	Animal User Parameter 2 Name	
	E	F	Animal User Parameter 2	
	E	F	Animal User Parameter 3 Name	
	Е	F	Animal User Parameter 3	
	Е	F	Animal User Parameter 4 Name	
	Е	F	Animal User Parameter 4	
	Е	F	Animal User Parameter 5 Name	
	Е	F	Animal User Parameter 5	
В	E	F	PREPARATION - TISSUE	
В	Е	F	Tissue Preparation Identifier	
В	Е	F	Organ	
	Е	F	Organ Region	
	E	F	Preparation Method	
	E	F	Tissue Preparation Date	ISO Date, Format: YYY-MM-DD
	E	F	Tissue Preparation Time	Time of Day, Format: hh:mm[:ss.000]
	Е	F	Tissue Incubation Duration	
	Е	F	Tissue Incubation Duration Unit	
	Е	F	Tissue Incubation Temperature	
	Е	F	Tissue Incubation Temperature Unit	
	Е	F	Tissue Incubation Solution	
	Е	F	Tissue User Parameter 1 Name	
	Е	F	Tissue User Parameter 1	
	E	F	Tissue User Parameter 2 Name	
	Е	F	Tissue User Parameter 2	

	Е	F	Tissue User Parameter 3 Name	
	E	F	Tissue User Parameter 3	
	E	F	Tissue User Parameter 4 Name	
	E	F	Tissue User Parameter 4	
	E	F	Tissue User Parameter 5 Name	
	E	F	Tissue User Parameter 5	
В	E	F	PREPARATION - CELL	
В	E	F	Cell Preparation Identifier	
	E	F	Acutely Dissociated Cells	
	E	F	Cell Line	
	E	F	Slice Preparation	
	E	F	Whole-organ Preparation	
	E	F	In-situ Recording	
	E	F	Stem Cell Preparation	
	E	F	User-defined Preparation	
В	Е	F	Cell Type	
В	E	F	Cell Identifier	
В	Е	F	Cell Preparation Date	ISO Date, Format: YYY- MM-DD
В	Е	F	Cell Preparation Time	Time of Day, Format: hh:mm[:ss.000]
	Е	F	Cell Dissociation Solution	
	Е	F	Cell Preparation Dissociation Temperature	
	Е	F	Cell Prep. Dissociation Temperature Units	
В	Е	F	Cell Preparation Incubation Duration	
В	E	F	Cell Prep. Incubation Duration Unit	
В	Е	F	Cell Preparation Incubation Temperature	
В	E	F	Cell Prep. Incubation Temperature Unit	
В	Е	F	Cell Preparation Incubation Solution	
В	Е	F	Ion Channel	
	E	F	Cell Fluorescent Marker	
	E	F	Cell Diameter	
	E	F	Cell User Parameter 1 Name	
	E	F	Cell User Parameter 1	
	E	F	Cell User Parameter 2 Name	
	E	F	Cell User Parameter 2	
	E	F	Cell User Parameter 3 Name	
	Е	F	Cell User Parameter 3	
	E	F	Cell User Parameter 4 Name	
	Е	F	Cell User Parameter 4	
	Е	F	Cell User Parameter 5 Name	
	Е	F	Cell User Parameter 5	
В	Е	F	EXPERIMENT	
		F	Experiment Category 1 Name	

		F	Experiment Category 1
		F	Experiment Category 2 Name
		F	Experiment Category 2
		F	Experiment Category 3 Name
		F	Experiment Category 3
		F	Experiment Category 4 Name
		F	Experiment Category 4 Experiment Category 4
		F	Experiment Category 5 Name
		F	Experiment Category 5
В	Е	F	Experiment User Parameter 1 Name
В	E	F	Experiment User Parameter 1
В	E	F	Experiment User Parameter 2 Name
В	E	F	Experiment User Parameter 2
В	E	F	Experiment User Parameter 3 Name
В	E	F	Experiment User Parameter 3
В	E	F	Experiment User Parameter 4 Name
В	E	F	Experiment User Parameter 4
В	E	F	Experiment User Parameter 5 Name
В	E	F	Experiment User Parameter 5
	E	F	ELECTRODE
	E	F	Electrode Identifier
	E	F	Electrode Glass Manufacturer
	E	F	Electrode Glass Item Number
		F	Electrode Glass Lot Number
		F	Electrode Glass Material
		F	Electrode Glass Item Outer Diameter
		F	Electrode Glass Item Inner Diameter
		F	Filamented Glass
	Е	F	Electrode Glass Ramp Test Value
	Е	F	Pipette Puller Manufacturer
	Е	F	Pipette Puller Model
		F	Pipette Puller Serial Number
		F	Puller Filament Type
		F	Puller Filament Item Number
		F	Pull Program Number
		F	Pull Program Parameters
		F	Pull Program Air Mode
		F	Pull Program Air Pressure
		F	Puller Preheat Enabled
		F	Pull Heat-on Time
		F	Electrode Tip Diameter
		F	Electrode Taper Length
		F	Electrode Fire-polished

	F	Electrode Coated	
	F	Electrode Coating Material	
	F	Electrode Beveled	
	F F	Electrode Bevel Angle	
Е	F	Electrode User Parameter 1 Name	
E	F	Electrode User Parameter 1	
E	F	Electrode User Parameter 2 Name	
	F	Electrode User Parameter 2	
E	F	Electrode User Parameter 2 Electrode User Parameter 3 Name	
E E	F	Electrode User Parameter 3 Name Electrode User Parameter 3	
E	F	Electrode User Parameter 4 Name	
_	•	Electrode User Parameter 4	
E	F	Electrode User Parameter 5 Name	
E	F	Electrode User Parameter 5 Name Electrode User Parameter 5	
E	F	RECORDING SOLUTIONS	
E	F		
E	F	Solution Pair Identifier Solution Pair Name	
E	F		
E	F	Bath Solution Identifier	
E	F	Bath Solution Name	
	F	Bath Solution Batch	
	F	Bath Solution Composition	
	F	Bath Solution Preparation Date	
	F	Bath Solution Preparation Time	
	F	Bath Solution pH	
	F	Bath Solution pH Adjustment Agent	
	F	Bath Solution Osmolarity	
	F	Bath Solution Osmolarity Adj. Agent	
E	F	Pipette Solution Identifier	
E	F	Pipette Solution Name	
	F	Pipette Solution Batch	
	F	Pipette Solution Composition	
	F	Pipette Solution Preparation Date	
	F	Pipette Solution Preparation Time	
Е	F	Pipette Solution pH	
	F	Pipette Solution pH Adjustment Agent	
E	F	Pipette Solution Osmolarity	
	F	Pipette Solution Osmolarity Adj. Agent	
	F	Liquid Junction Potential, computed	
	F	Liquid Junction Potential, measured	
	F	Solution User Parameter 1 Name	
	F	Solution User Parameter 1	
	F	Solution User Parameter 2 Name	
	F	Solution User Parameter 2	

		F	Solution User Parameter 3 Name	
		F	Solution User Parameter 3	
		F	Solution User Parameter 3 Solution User Parameter 4 Name	
		-	Solution User Parameter 4 Name Solution User Parameter 4	
		F		
		F	Solution User Parameter 5 Name	
		F -	Solution User Parameter 5	
		F -	PARADIGM	
		F -	Bath Temperature	
		F -	Bath Temperature Unit	
		F	Ambient Temperature	
		F	Ambient Temperature Unit	
		F	Atmospheric Composition	
		F	Atmospheric Pressure	
		F	Atmospheric Pressure Unit	Of relative boundaries
		F	Atmospheric Humidity	% relative humidity ("-1" = uncontrolled)
		F	Paradigm User Comment	
		F	Paradigm User Parameter 1 Name	
		F	Paradigm User Parameter 1	
		F	Paradigm User Parameter 2 Name	
		F	Paradigm User Parameter 2	
		F	Paradigm User Parameter 3 Name	
		F	Paradigm User Parameter 3	
		F	Paradigm User Parameter 4 Name	
		F	Paradigm User Parameter 4	
		F	Paradigm User Parameter 5 Name	
		F	Paradigm User Parameter 5	
		F	CELL HEALTH / QUALITY CONTROL	
		F	Cell Health User Parameter 1 Name	
		F	Cell Health User Parameter 1	
		F	Cell Health User Parameter 2 Name	
		F	Cell Health User Parameter 2	
		F	Cell Health User Parameter 3 Name	
		F	Cell Health User Parameter 3	
		F	Cell Health User Parameter 4 Name	
		F	Cell Health User Parameter 4	
		F	Cell Health User Parameter 5 Name	
		F	Cell Health User Parameter 5	
		F	SERIES (= ROUTINE DATA)	
		F	Routine User Comment	
			IMAGING	
			Image Comment	
В	E	F	STIMULUS	
	Е	F	Key Stimulus	

	Е	F	Stimulus Duration
	E	F	Compound Group
	E	F	Compound Group Index
В	E	F	Compound Identifier
В	E	F	Compound Name
В	E	F	Compound Concentration
В	E	F	Compound Concentration Unit
D	E	F	Compound Batch
	E	F	Compound Lot
	E	F	Compound Salt Code
	E	F	Compound Solution
	E	F	Compound Vehicle / Solubility Enhancer
	E	F	Compound Vehicle Concentration
	E	F	Compound Vehicle Concentration Unit
	E	F	Compound Reservoir Identifier
	E	F	Application Tip Identifier
	E	F	Compound Plate Identifier
	E	F	Compound Plate Row
	E	F	Compound Plate Column
	E	F	Chem. Stimulus User Parameter 1 Name
	E	F	Chem. Stimulus User Parameter 1
	E	F	Chem. Stimulus User Parameter 2 Name
	E	F	Chem. Stimulus User Parameter 2
	E	F	Chem. Stimulus User Parameter 3 Name
	E	F	Chem. Stimulus User Parameter 3
	E	F	Chem. Stimulus User Parameter 4 Name
	E	F	Chem. Stimulus User Parameter 4
	Е	F	Chem. Stimulus User Parameter 5 Name
	E	F	Chem. Stimulus User Parameter 5
В	Е	F	Light Stimulus Wavelength
В	Е	F	Light Stimulus Intensity
В	Е	F	Light Stimulus Intensity Unit
	E	F	Light Stimulus User Parameter 1 Name
	E	F	Light Stimulus User Parameter 1
	Е	F	Light Stimulus User Parameter 2 Name
	E	F	Light Stimulus User Parameter 2
	Е	F	Light Stimulus User Parameter 3 Name
	Е	F	Light Stimulus User Parameter 3
	Е	F	Light Stimulus User Parameter 4 Name
	Е	F	Light Stimulus User Parameter 4
	Е	F	Light Stimulus User Parameter 5 Name
	E	F	Light Stimulus User Parameter 5
В	Е	F	Mechanical Stimulus Intensity

В	Е	F	Mechanical Stimulus Intensity Unit	
	_		Mechanical Stimulus User Parameter 1	
	E	F	Name	
	Е	F	Mechanical Stimulus User Parameter 1	
	Е	F	Mechanical Stimulus User Parameter 2 Name	
	Е	F	Mechanical Stimulus User Parameter 2	
			Mechanical Stimulus User Parameter 3	
	Е	F	Name	
	Е	F	Mechanical Stimulus User Parameter 3	
	E	F	Mechanical Stimulus User Parameter 4 Name	
	Е	F	Mechanical Stimulus User Parameter 4	
			Mechanical Stimulus User Parameter 5	
	E	F	Name	
_	E	F	Mechanical Stimulus User Parameter 5	
В	Е	F	Acoustic Stimulus Frequency	
В	Е	F	Acoustic Stimulus Intensity	
В	Е	F	Acoustic Stimulus Intensity Unit	
	Е	F	Acoust. Stimulus User Parameter 1 Name	
	Е	F	Acoust. Stimulus User Parameter 1	
	Е	F	Acoust. Stimulus User Parameter 2 Name	
	Е	F	Acoust. Stimulus User Parameter 2	
	Е	F	Acoust. Stimulus User Parameter 3 Name	
	E	F	Acoust. Stimulus User Parameter 3	
	Е	F	Acoust. Stimulus User Parameter 4 Name	
	Е	F	Acoust. Stimulus User Parameter 4	
	Е	F	Acoust. Stimulus User Parameter 5 Name	
	Е	F	Acoust. Stimulus User Parameter 5	
В	Е	F	Thermal Stimulus Temperature	
В	E	F	Thermal Stimulus Temperature Unit	°C, °F or K
	E	F	Thermal Stimulus User Parameter 1 Name	
	Е	F	Thermal Stimulus User Parameter 1	
	Е	F	Thermal Stimulus User Parameter 2 Name	
	E	F	Thermal Stimulus User Parameter 2	
	E	F	Thermal Stimulus User Parameter 3 Name	
	Е	F	Thermal Stimulus User Parameter 3	
	Е	F	Thermal Stimulus User Parameter 4 Name	
	Е	F	Thermal Stimulus User Parameter 4	
	Е	F	Thermal Stimulus User Parameter 5 Name	
	Е	F	Thermal Stimulus User Parameter 5	
В	Е	F	Electrical Stimulus Frequency	The frequency of an external electrical stimulus
В	E	F	Electrical Stimulus Intensity	The intensity of an external electrical stimulus

В	E	F	Electrical Stimulus Intensity Unit	The intensity unit of an external electrical stimulus
	E	F	Electrical Stimulus User Parameter 1 Name	
	E	F	Electrical Stimulus User Parameter 1	
	Е	F	Electrical Stimulus User Parameter 2 Name	
	E	F	Electrical Stimulus User Parameter 2	
	E	F	Electrical Stimulus User Parameter 3 Name	
	E	F	Electrical Stimulus User Parameter 3	
	E	F	Electrical Stimulus User Parameter 4 Name	
	E	F	Electrical Stimulus User Parameter 4	
	E	F	Electrical Stimulus User Parameter 5 Name	
	E	F	Electrical Stimulus User Parameter 5	
	Е	F	Other Stimulus User Parameter 1 Name	
	E	F	Other Stimulus User Parameter 1	
	E	F	Other Stimulus User Parameter 2 Name	
	E	F	Other Stimulus User Parameter 2	
	E	F	Other Stimulus User Parameter 3 Name	
	E	F	Other Stimulus User Parameter 3	
	E	F	Other Stimulus User Parameter 4 Name	
	Е	F	Other Stimulus User Parameter 4	
	E	F	Other Stimulus User Parameter 5 Name	
	Е	F	Other Stimulus User Parameter 5	

Table 4-7. Metadata Parameters

4.2.19 Single Channel Analysis

Perform analysis of low-noise currents from single ion channels.

Access single-channel analysis via:

- the Reanalysis scope window 'Measurements' button, or
- the Data Navigator (signal) 'Available actions' menu, or
- the main menu SutterPatch > Available Analysis Modules > Single Channel Analysis.

Single Channel Analysis uses a special scope window, where amplitude levels and transitions and are overlaid onto the raw data. When you click-and-drag in the scope window, the closest amplitude level is repositioned to the new amplitude. Because of this, to access the marquee tool in the scope window, hold down the shift key when you click-and-drag the mouse.

Marquee Right-click Menu

Special addition to the menu.

Set Time Range of Analysis

Sets the Single Channel Analysis 'Time Range' to 'Sweep Time', and the 'Start Time' and 'End Time' are set from the marquee range.




Figure 4-75. Single-Channel Scope

When single channel analysis is activated, a Single Channel Analysis control panel is opened, and the Reanalysis scope window active signal is overlaid with the transition levels of the first single-channel opening or closing, based upon the settings in the Single Channel Analysis control panel / Current Transition Controls section.

Note: Single-channel analysis only operates in the scope 'Sweeps' display mode; the Concatenated and Time Course display modes are not supported.

Scope window levels

- A dashed **green** "Before" line displays the amplitude of the previous transition/event, i.e., the level before the transition point. Manually adjust by dragging with the mouse.
- A dashed **brown** "After" line displays the amplitude of the selected transition/event, i.e., the level after the transition point. Manually adjust by dragging with the mouse.
- A solid **green** line displays the idealized trace of the found transitions/events.

At times, the dashed amplitude lines might superimpose onto the idealized trace.

• A vertical **red** line displays at the transition point between the two levels, with a red arrow indicating the direction of the transition.

If the initial levels are incorrect, a couple of basic settings need to be adjusted:

- 1. Determine the starting amplitude of the data before the first transition.
 - Zoom in on the scope data, so that the open and closed state amplitudes are well visualized. Or run the 'Plots and Tables' 'Current Amplitude Histogram' to find the amplitude peaks in the binned data.
- 2. Set the 'Current Transition Controls' estimated 'Amplitude' signed value for the first level in the data. (Use negative numbers for negative-going openings.)
- 3. Set the Start Level number for the initial data (0 = baseline state, 1 + = open states).
- 4. Click on the 'Find target transition' section 'Clear All' button.
- 5. The scope window resets the "Before" and "After" transition levels to proper values.




Figure 4-76, Single Channel Analysis

[Descriptive Information]

Single Channel Analysis Preferences

Dock to scope

Baseline Average Duration = x.x ms

Calculate the averaged baseline amplitude with the duration of the last baseline data (closed state) before the transition to an open state.

Mean Amplitude Duration = x.x ms

Calculate the mean amplitude using up to the duration of the open state data at that level, after a transition to that level.

Allow display compression

Signal Controls




Figure 4-77. Signal Controls

Control the display of single channel data in the scope window.

• Use individual sweeps Perform analysis on a single sweep at a time.

• Use complete signal For plotting - use found transitions from all

sweeps in the active signal.

[] Use marked sweeps when using complete signal

Include transitions only from the marked sweeps

in the active signal.

Show All Signals

X-Width: [Sweep / 1 s / 500, 200, 100, 50, 20, 10, 5, 2, 1 ms]

The scope window X-axis duration is reset to this value whenever a

'Find' or 'Move' transition operation is performed.

Time Range Full Trace

Sweep Time Start Time (s)

End Time (s)

Segment Time Start Ratio Relative to the start of a

segment.

End Ratio Relative to the end of a seg-

ment.

Current Transition Controls

Setup the basic level-detection parameters.




Figure 4-78. Current Transition Controls

Amplitude

Set to the expected transition size of the initial ion-channel level in the data.

For downwards-opening (negative) channels, use a negatively-signed value.

Start Level

The starting state of the initial ion-channel data:

Level 0 = Closed state

Level 1 = First open state

Level 2 = Second open state

Etc.

If there are multiple levels in the data, the program will try to automatically detect them. However, overlapping channel openings are treated as a single combined level.

Min. Duration

$$[0.0-\infty s]$$

The minimum duration for a "found" transition.

The increment/decrement spinners use a step size of 100 µs.

Note: Displayed values are rounded up or down to one decimal point for the scaled unit of display. For example, for values greater than 1.0 s, the spinners do not update the displayed value until a rounding threshold is reached for the last digit, i.e., '1.5499' converts to '1.5', while '1.5500' converts to '1.6'.

Track Limit

[0-1.0 nA]

The maximum (absolute) amount that the Baseline level can change while being automatically tracked.

Set manually, or set to 3 * RMS via the 'Measure RMS' button.

Threshold [50 - 90%]

The percentile of the Amplitude value (open state) that needs to

be reached by the raw data to "find" a transition.

Track Amplitude Store the amplitude of the measured event, instead of the theo-

retical (short) event, in the event transition table.

Track Baseline Automatically adjust the baseline amplitude (Level 0) based on

the prior data.

To use, the 'Track Limit' value must be > 0.

Measure RMS To measure the RMS (Root-Mean-Square) noise in the signal, ad-

> just the signal trace in the scope window, such that it shows a stretch of current without any channel activity (i.e., all channels

are closed), then click on the 'Measure RMS' button.

The RMS value is displayed at the bottom of the dialog, and a

value of "3 x RMS" populates the Track Limit field.

Properties of selected transition

View or alter how a transition is processed by the analysis.



Figure 4-79. Properties of Selected Transitions

Transition The count number of the selected (active) transition.

Time The start time of the selected transition.

The level number of the selected transition. Level

Note: If other unexpected levels are detected "below the base-

line", they are assigned a negative number. Negative levels might be detected from noise, biological artifacts, or an

incorrect initial 'Start Level' setting.

Status The operational status of the selected transition. • Accept Terminates the preceding event and starts a new open/close time.

The selected transition is included in the idealized trace and all Plots.

Reject Terminates the preceding event and starts a new open/close time.
 However, the selected transition is considered inappropriate for analysis, and is excluded from the idealized trace and all Plots.

Events that border a rejected transition are also excluded from histograms.

• Suppress Does not terminate the preceding event or start a new open/close time. A suppressed event is considered as "not having happened". The selected transition is excluded from the idealized trace and all Plots.

Find target transition

Find a transition based on the 'Current Transition Controls', and process the transition based on the 'Properties of selected transition'.

An "event" is a valid transition that is followed by another valid transition.



Figure 4-80. Find Target Transition

found The total number of found transitions (including rejected and

suppressed transitions).

Clear All Reanalyze the data - reset the number of found transitions to

zero, and move to the first found transition.

Right > Find, move to, and process the next transition.

Last >> Find and process all subsequent transitions, and move to the last

transition.

Note: Multiple open levels are handled in a simplistic fashion. It is assumed that there is only one channel open, and that it is open for the same state throughout the duration of the Event.

Example:

Level 1 openings: The Event duration is from the transition to the

Level 1 amplitude, to the next transition to a dif-

ferent Level amplitude.

Level 2 openings: The Event duration is from the transition to the

Level 2 amplitude, to the next transition to a dif-

ferent Level amplitude.

Etc.

Move to target transition

Among the 'found' (processed) transitions, move to an adjacent transition, or jump to the beginning or ending transition.



Figure 4-81. Move to Target Transition

<< First Move to the first found transition.</p>
< Left Move to the previous found transition.</p>
Right > Move to the next found transition.
Last >> Move to the last found transition.

Show/Hide Plots and Tables Controls

This button opens/closes the 'Tables and Plots' dialog, docked on the right of this dialog.

[] Total events in ... Status bar for the number of events in a particular Plot.

Single Channels: Plots and Tables

Note: All plots and histograms support measuring X-Y data points or setting a fitting range via the 'Toggle Cursor Info' right-click menu entry (or Ctrl-I). (See 'Right-Click Menus' for scope windows.)




Figure 4-82. Plots and Tables Controls




Figure 4-83. Current Amplitude Histogram Controls

A current amplitude histogram is often the first analysis performed on an uncharacterized channel, whereby all data points are binned by amplitude. It is used to determine:

- The quality of the recording.
- The number of levels in the open state.
- The first estimate of the open state amplitude(s).
- The first estimate of the baseline closed state.
- The frequency of openings.

[1, 2, 3]

Current Amplitude Histogram	Click to create a histogram plot of the raw data. No prior settings are needed to run this.
Copy to Layout	Copy the 'Current Amplitude Histogram' into a new Layout window, or append to an existing Layout page.
Copy to Clipboard	Copy the 'Current Amplitude Histogram' to the system clipboard. With the 'Shift' key pressed: as a graph, otherwise as a table.
Gauss Fit	Click to perform a Gaussian fit on the histogram. SutterPatch will automatically find and fit up to the three largest peaks. You can adjust the fitting ranges as needed with manual cursors.
	To fit additional smaller distributions, reposition the cur-

See the 'Table of Results' for the fitting components.

sors and click the Append button.

Select the number of peaks (levels) to fit.

When more than one peak is selected the

When more than one peak is selected, the graph reports amplitudes (in relation to the closed state peak) and P(open) and P(closed) values.

Cursors

Display fitting cursors in the plot window.

To reposition cursors, first create the histogram, and then enable 'Cursors' – this will open a Cursor Bar at the bottom of the plot window. Next, drag the cursor's symbol (labeled "A", "B", etc.) from the cursor bar to the new position (near to the X-axis) on the plot.

Append

Modify the 'Current Amplitude Histogram'.

Add

Add the current data to the existing 'Current Amplitude Histogram'.

Sub

Subtract the current data from the existing 'Current Amplitude Histogram'.

Number of Histogram Bins

[4000, 2000, 1000, 500, 200, 100, 50]

Distribution Controls

Create histogram plots of the found events.




Figure 4-84. Plots and Tables Controls

Duration Distribution Plot

Click to create a duration (dwell-time) histogram plot of the found events. The histogram bin count is reported as a 'Relative Frequency' (to 1.0) on the plot's Y-axis.

Copy to Layout

Copy the 'Duration Distribution Plot' into a new Layout window, or append to an existing Layout page.

Copy to Clipboard

Copy the 'Duration Distribution Plot' to the system clip-

board.

Exp Fit

Apply an exponential fit to the data.

log(x) Set the X-axis to a log scale.

sqrt(y) Use the square-root of the Y-axis data.

Open / Closed Select open or closed state data for the Distribution plots.

Amplitude Distribution Plot Click to create an amplitude histogram plot of the selected state's found events. The histogram bin count is reported as 'Frequency' on the plot's Y-axis.

The Amplitude Distribution Plot bins "transition deltas", which measures the *directional change* in amplitude for each transition (not the raw amplitude).

For example, an opening transition to 15 pA bins on the X-axis at 15 pA, while a following closing transition back to 0 pA bins on the X-axis at -15 pA, i.e., the delta of the transition's Before and After amplitudes.

The histogram bins plot as colored lines:

Open = red

Closed= blue

Copy to Layout Copy the 'Amplitude Distribution Plot' into a new Layout

window, or append to an existing Layout page.

Copy to Clipboard Copy the 'Amplitude Distribution Plot' to the system clip-

board.

Gauss Fit [1, 2, 3]

Cursors When an Amplitude Distribution Plot exists, you can ena-

ble draggable fitting cursors in the graph.

To position a cursor, drag its cursor symbol (labeled "A", "B", etc.) from the cursor bar onto the data at the desired

position.

Apply Limits Apply data limits to the events used in distribution plots.

Duration: Min [0.0-1.0 s]

Max. [0.0 - 1.0 s]

Amplitude: Min $\begin{bmatrix} -1.0 - 1.0 \text{ A} \end{bmatrix}$

Max. [-1.0 - 1.0 A]

Number of Histogram Bins [1000, 500, 200, 100, 50, 20]

Scatter Plot Controls




Figure 4-85. Scatter Plot Controls

The scatter plot uses "transition deltas" for event amplitudes, which plot on the Y-axis as the *directional change* in amplitude for each transition; the X-axis plots the duration of the Event.

For example, an opening transition to 15 pA will plot on the Y-axis at 15 pA, while a following closing transition back to 0 pA plots on the Y-axis at -15 pA, i.e., the delta of the transition's Before and After amplitudes.

Events are plotted as colored symbols:

Open = red

Closed= blue

Selected Event (transition) = green

[Scatter Plot] Click to create a scatter plot of the selected state's found events.

Copy to Layout Copy the 'Scatter Plot' into a new Layout window, or ap-

pend to an existing Layout page.

Copy to Clipboard Copy the 'Scatter Plot' to the system clipboard.

Open / Closed / Both Select which states are plotted.

Apply Limits Apply data limits to the events used in scatter plots.

Duration: Min [0.0-1.0 s]

Max. [0.0 - 1.0 s]

Amplitude: Min [-1.0-1.0 A]

Max. [-1.0 - 1.0 A]




Figure 4-86. Table Controls

Table of Transitions Click for a listing of all transitions.

Layout of the table

Top Row: [Cell address | Cell value]

Column Number Row: Column numbers [0, 1, 2, ...]

Row 0: Status The operational status of the selected transition.

1 = Accepted Terminates the preceding event

and starts a new open/close time. The selected transition is included in the idealized trace and all Plots.

2 = Rejected Terminates the preceding event

and starts a new open/close time. However, the selected transition is considered inappropriate for analysis, and is excluded from the ideal-

ized trace and all Plots.

Events that border a rejected transition are also excluded from histo-

grams.

3 = Suppressed Does not terminate the preceding

event or start a new open/close time. A suppressed event is considered as "not having happened". The selected transition is excluded from the idealized trace and all

Plots.

Row 1: Time Time of the start of the transition, i.e., the transition

point.

Row 2: Level The open or closed state level number.

Row 3: Amplitude_Before Amplitude of the level preceding the transition, i.e., the

level before the transition point.

Row 4: Amplitude_After Amplitude of the transition, i.e., the level after the transition point. Row 5: Duration Duration of the transition. Note: The last column of transition data is preset to a zero duration. Row 6: Amplitude Amplitude of the transition. Row 7: Amplitude_Valid Include / Exclude the transition for processing. Not a valid transition. Invalid = Valid A valid transition. Note: The very first column of transition data is always defined to be 'Invalid', and is excluded from processing. Also, the last two columns of transition data are excluded from Plots. -----Copy the 'Table of Transitions' into a new Layout window, Copy to Layout or append to an existing Layout page. Copy to Clipboard Copy the 'Table of Transitions' to the system clipboard. Clear Clear all transitions from the table, and reset to the first found transition. Save Copy to Disk Save the 'Table of Transitions' to an Igor wave (*.ibw) file. Load Table from Disk Load the 'Table of Transitions' from an Igor wave (*.ibw) file. Table of Results Click for a listing of all fitting results. Copy to Layout Copy the 'Table of Results' into a new Layout window, or append to an existing Layout page. Copy to Clipboard Copy the 'Table of Results' to the system clipboard.

found transition.

Clear all entries from the table, and reset to the first

Clear

Table Column Labels

Gaussian Fit

Header Row Row 0 Source Sweep_# Current Amplitude Histogram Fit, or Analysis Amplitude Distribution Fit Label1 Amplitude # Value1 Label2 p(closed) Value2 # Label3 p(open) Value3 # Label4 Gauss_y0_1 # Value4 $Gauss_A_1$ Label5 # Value5 Label6 Gauss_x0_1 Value6 # Label7 Gauss_width_1 # Value7 Label8 $Success_1$ # Value8 Label9 Gauss_y0_2 Value9 # Label10 Gauss_A_2 Value10 # Label11 $Gauss_x0_2$ Value11 # Label12 Gauss_width_2 # Value12 Label13 Success_2 Value13 # Label14 Gauss_y0_3

#

Value14

Label15 Gauss_A_3

Value15 #

Label16 Gauss_x0_3

Value16 #

Label17 Gauss_width_3

Value17 #

Label18 Success_3

Value18 #

Linear Exponential Fit

<u>Header Row 0</u> Row 0

Source Sweep_#

Analysis Duration Distribution Fit

Label1 Exp_y0

Value1 #

Label2 Exp_A

Value2 #

Label3 Exp_Tau

Value3 #

Label4 Exp_0

Value4 #

Label5 Success

Value5 #

Logarithmic Exponential Fit

<u>Header Row 0</u>

Source Sweep_#

Analysis Duration Distribution Fit

Label1 LogNormal_k0

Value1 #

Label2 LogNormal_k1

Value2

Label3 LogNormal_k2

Value3 #

Label4 LogNormal_k3

Value4 #

Label5 Success

Value5 #

4.2.20 Synaptic Event Analysis

Post-synaptic potentials and currents from excitatory and inhibitory events (EPSPs, EPSCs, IPSPs, IPSCs) are analyzed with this application module. Access via the Reanalysis scope window 'Measurements' button or the Data Navigator (signal) 'Available actions' menu.

Spontaneous miniature events (mEPSPs, etc.), which generate small and often overlapping Events, are detected with an innovative deconvolution algorithm. This technique finds Events with high temporal fidelity, while also improving the signal-to-noise ratio (SNR).




Figure 4-87. Synaptic Event Analysis

Sweep # The sweep number of the displayed data.

The Sweep # is set to '0' when the Average event is displayed in the Event pane (or when the template is initially created prior to analysis.)

Select arbitrary sweeps for processing by "marking" them in the Data Navigator tree, or in a scope window during acquisition or reanalysis.

Event Time The time(s) of the event threshold crossing relative to the start of the

sweep.

Amplitude The amplitude averaged around the peak by ± 1 ms.

[Event pane] A graph of the selected event overlaid by the template, with the X-axis

zero point reset to the template starting point.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor

Info' from the right-click menu. (See Right-Click Menus' for scope

windows.)

Display Template Display the ideal event's template on top of the selected event in the

graph - its Y-axis displays on the right edge of the graph.

Tip: To match the template to the data, hover the mouse cursor over

the right Y-axis, and use the mouse wheel to rescale the tem-

plate.

Remove Drift (High Pass Filter)

A 1 Hz high-pass filter is applied to the signal to remove baseline drift.

Template Menu Open the Template sub-panel to configure a template.

Create a template of a typical event as a double-exponential curve. The

data will be deconvolved to this template for further analysis.

Event Polarity [1, -1] 1 = positive

-1 = negative

Rise Time (μ s) [$10-5{,}000$] Time constant (τ) for the rising phase of

the template event.

Decay Time (us) $\begin{bmatrix} 100 - 100,000 \end{bmatrix}$ Time constant (τ) for the falling phase of

the template event.

Create Template Click to create a custom event template.

Use Average Click to use the event Average as the event template.

Detection Menu Open the Levels sub-panel to configure detection levels.

Threshold (xSD) A detection threshold represents the "Event Strength". A lower

("weaker") number finds more events, while a higher ("stronger")

number finds less events. Adjust this threshold based on empirical testing of your data.

[0.1-10] Lower# = more events (false-positives)

Higher # = less events (false negatives)

Note: The default threshold is set to 4 times the standard deviation of a Gaussian fit to an all-points histogram of the (Fourier) deconvolved data signal.

Ampl Threshold [5e-13-1e-6]

Set an amplitude threshold for the minimum size of events.

Decay tau (us) [<,>] [500-1,000,000]

Set the decay tau as "less than" or "greater than" the tau value.

Concatenate Sweeps Combine all sweeps into a single pseudo-sweep before processing.

This sometimes improve performance of the detection algorithm

signal-to-noise ratio with shorter sweeps.

Analysis Start (s) Sweep time to start looking for an event threshold.

Analysis End (s) Sweep time to stop looking for an event threshold.

Event The current event number vs. total number of events.

Remove Delete the current event from the analysis.

Sweep Show / Hide the Sweep preview pane.

A right-click menu is available in the pane, as well as the click-and-drag

marquee for magnification and extraction.

[Sweep pane] Displays a sweep of data colored in blue, with the selected event colored

in red.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor Info' from the right-click menu. (See 'Right-Click Menus' for scope

windows.)

To resize or extract data, click and drag a box around the data with the mouse marquee tool, and right-click for the marquee menu. The special menu option 'Add Mini' allows you to manually classify a raw data selection as an event during manual detection, or to include an event missed by the template detection. The new event is highlighted in red, and included in new Results tables.

Note: Manually detected events do not have an 'Event Strength' entry in the Results table, as an algorithm was not used to detect them.

Start Click to find and analyze synaptic events.

Click to display the averaged event in the Event pane. Average

The Sweep # is set to '0'.

A scatter plot of the Event Strength vs. Current is displayed in a sub-Threshold Analysis

panel.

To measure X-Y data points or set a fitting range, select 'Toggle Cursor Info' from the right-click menu. (See 'Right-Click Menus' for scope

windows.)

Save Results Results are displayed in a Layout page and a table.

Synaptic Event Analysis Results

Results are displayed in their own Layout window, accessible via

Windows / Layouts.

Signal Pathname: The path from the Igor internal root direc-

tory is displayed (see Data Browser.)

Total time analyzed = (s)

Includes the Start / End times for all

sweeps.

Number of events detected =

Total number of Events found.

Event Frequency = (Hz)

Average Event Amplitude = (pA)

±1 ms peak average.

Standard Deviation of Event Amplitude = (pA)

Graphs:

Cumulative probability vs. Amplitude.

Amplitude (Average) vs. Time. Frequency vs. Sweep Number.

Amplitude vs. Sweep Number.

The individual graphs are also ac-Note:

cessible via Windows / Graphs.

Results Table A table of columns is created.

> [blank] Row number with one row per Event.

Sweep Number The sweep number the Event is in.

Event Time (s) 'Time to event' from the start of the sweep.

Event Strength (xSD) A measure of how well the signal matches

the template. (Lower is weaker, higher is

stronger.)

Event Amplitude (A) The Event peak amplitude ± 1 ms average.

Event Integral (A*s)

10-90% Rise Time (s)

Event Decay Tau (s)

Absolute Event Time (s) A continuous time scale from the start

of acquisition, i.e., from the clicking of the 'Start' button, prior to the initial

Sweep/Series external trigger.

Inter Event Interval (s)

Duplicate Display duplicate copies of the Results table and layout.

Plot The 'Plot sweeps' dialog displays to allow events to be plotted as overlap-

ping sweeps in a floating graph window.

Enter a list of events separated by a comma "," and/or a range of Plot sweeps

events separated by a dash "-".

(See the SutterPatch Algorithms Appendix for a reference to the Event detection algorithm.)

4.3 General

SutterPatch general operations.

Note: Hidden unminimized windows can be brought into view with the menu command Windows / Control / Retrieve All Windows.

4.3.1 Command Window

This window is an Igor Pro code interpreter, and provides programmatic interaction with SutterPatch. You can manually execute Igor Pro and user-defined assignments, functions and operations in this window.




Figure 4-88. Command Window

The Command window is labeled with the current Experiment name, and is accessed from the menu Windows / Command Window.

A history of executed commands and responses displays in the upper section of the window; some warning messages also display here.

The lower section is a command buffer with a "command line", where commands to be executed are entered. Commands can be placed into the command buffer in multiple ways:

- Manually type (or copy and paste) a line of text into the command line.
- Highlight lines in the history section, and press the Enter key to copy them into the command buffer. To select the entire history, use 'CTRL-A'.
- Use the Paradigm Editor 'Execute' Step Editor buttons 'Copy to Command Line' or 'Expand to Command Line' (for vars) to transfer the step command to the command line.

Commands in the command buffer are processed when the "Enter' key is pressed.

A maximum of 400 characters can be entered into the command buffer, however they can be spread across multiple commands on multiple lines.

Note: Igor syntax usually requires that open/close parentheses "()" be appended to the

end of a command. However, exceptions include the "beep" and "print" commands, for which no parentheses are used.

The Command window has a resizing line between the upper history section and the lower command section – the mouse cursor will change to a double-headed arrow.

For more information, see Section II-2 of the Igor Pro manual.

4.3.2 Dashboard Panel

The Dashboard panel provides a convenient gateway to key areas of the SutterPatch program.






Figure 4-89. Dashboard

\Diamond	Preferences	
	Icon Size	Large Icon
		Small Icon
	Icon Orientation	Vertical
		Horizontal





View Last Data

Open the Experiment's last recorded data Series. All sweeps (marked and unmarked) are visible in the initial display.



Set Metadata Configure metadata settings and values.



New Experiment Start a new Experiment, and/or switch the amplifier model or emulation mode.



New HDF5 File Start a new HDF5 file for the Experiment.

[Replaces the 'New Experiment' button when HDF5 is enabled in SutterPatch / Set Preferences / Files and Naming.]

Clicking the Acquire Data icon opens an adjoining secondary pane:



Figure 4-90. Dashboard - Acquire Data



Control Panel

Hardware control via the Amplifier Control Panel.



Membrane Test

Monitor seal formation and cell health.



Free Run

Run an oscilloscope-style signal monitor.



Routines

Configure Routine acquisition settings.



Paradigms

Control the execution of Paradigm commands.

Note: While a scope window is open, its icon remains depressed for Membrane Test / Free Run / Routine activation.

4.3.3 File Import/Export

Experiments & Data

The "packed" Igor Pro file format is recommended for saving an Experiment for most purposes.

• Packed experiment: (*.pxp file)

A SutterPatch Experiment is saved by default as a "packed" (Igor Pro) experiment, which includes all data, analyses, graphs, routines, paradigms, etc., in one file.

A Preferences option also allows a packed Experiment to save data to an HDF5 file, which has the advantage of an unpacked Experiment (like fast saving), without the disadvantage of much larger file sizes.

Unpacked experiment: (*.uxp file, experiment folder)

A SutterPatch Experiment can also be saved as an "unpacked" (Igor Pro) Experiment, which saves all waves, procedure windows, and notebooks as individual files in an experiment or "home" Folder, along with an instruction (*.uxp) file.to recreate the Experiment.

The advantage of an unpacked Experiment is:

 Much faster saving of Experiments that include very large numbers of waves (thousands or more), as existing data waves are not resaved to the Experiment with each new recording.

The disadvantages of an unpacked Experiment are:

- Much more disk space is used, especially for Experiments that have a lot of small waves.
- The UXP format is more "fragile", as you need to keep the Experiment file and its corresponding folder together when you copy or move the Experiment.

Saved Experiments can be re-opened into the current Experiment via the Data Navigator 'Import' button.

Note: If a SutterPatch Experiment file is opened into Igor Pro without SutterPatch running, its graphs and layouts can be displayed with the menu items Windows / Graphs, or Windows / Layouts, or Windows / Layout Macros.

Axon Files

Data can be selected and exported to the Axon Instruments pCLAMP file formats via the Data Navigator 'Available Actions' button or a right-click menu. Select the ABF or ATF file format in SutterPatch / Set Preferences / Data Export.

HDF5 Files

HDF5 is a modern efficient file format for managing and saving high volumes of data.

Enable HDF5 files (*.h5) for Experiments in SutterPatch / Set Preferences / Files and Naming: Save data to separate HDF5 file.

Multiple HDF5 files can be created during an Experiment to segregate or manage data. (See the File menu for additional options.)

HEKA Files

HEKA Elektronik PatchMaster Pulse Generator Files (*.pgf) can be opened in the Routine Editor 'Pools and Files' section and their Sequences merged with the current routine pool.

Igor Files

Data can be selected and exported to the <u>Igor Pro</u> file format (*.ibw) via the Data Navigator 'Available Actions' menu. Select the Igor Binary format in SutterPatch / Set Preferences / Data Export.

Igor binary waves (*.ibw) can be loaded into the current Experiment via Data / Load Waves / Load Igor Binary. Find the files in Data / Data Browser.

Note: If data is imported from other (non-Sutter) Igor programs, adjust the scaling of the data as needed.

Template Files

Templates can be imported or exported via the Template Editor as Igor Binary Wave files (*.ibw).

Templates can also be used to export portions of data from a sweep.

Graphs and Layouts

Import or export graph files for the current experiment via the Analysis Editor / Files options. Graph files are stored as Igor Binary Wave (*.ibw) files.

Note: Graph data for each axis can also be saved as Igor Pro 6 one-dimensional wave files, however files using this older format cannot be re-imported back into SutterPatch.

Individual graphs can be saved as Graph Macros - recall them via the Windows / Graph Macros menu.

Layout windows can also be saved as Layout Macros - recall them via the Windows / Layout Macros menu.

4.3.4 Log Window

The Log window displays time-stamped commands, responses, administrative information and error messages that provide a history of the steps having a possible influence on the execution of the experiment and its data. The Log window can also serve as a user laboratory notebook for free-form entries.

At start up, the Log window displays the SutterPatch Version and Build numbers.

The following unnamed columns are used by the Log window:

Date & Time Log entries are assigned a date/time stamp.

Day name, month name, day date, year date, time:

hours:minutes:seconds, AM/PM

Event Type Log entries are assigned an Event Type.

Data Acquisition Acquisition operations, Membrane

Test measurements.

Metadata Tags.

Paradigm Paradigm operations.

Startup SutterPatch version information.

Unknown Other operations.

Event Description A text description of the log entry.

Each row with a Data Acquisition, Metadata or Paradigm entry is appended with the name of the appropriate Routine or Paradigm; if there is no value to list, just the name of the Routine or Paradigm is displayed.

4.3.5 **Menus**

The SutterPatch main menu item contains all of the SutterPatch-specific menu items. The rest of the main menu items provide the standard Igor Pro functionality. For documentation of the non-SutterPatch features, refer to the Igor Pro online help or manual.

Window/Dialog Controls

Keyboard "Return" key = 'OK / Yes' buttons

Keyboard ESC key = 'Cancel' button

File

New Experiment

Unload the current Experiment and start a new Experiment.

It is recommended that you create one Experiment per cell, to keep file sizes manageable.

Note: Whenever a new Experiment is started, you will be asked to re-save the old Experiment, even if there were no changes to it; this is required by an internal Igor cleanup process.

Open Experiment

Open a previously saved SutterPatch Experiment (*.pxp, *.uxp) file. If a SutterPatch Experiment is opened in an Igor-only session, SutterPatch is automatically loaded.

If the SutterPatch preference for HDF5 files was enabled, a "SutterPatch Question" will ask how to load the matching HDF5 file:

Load matching HDF5 File: [pathname]

 Load in modify mode, i.e., add new data, store changes in analysis files.

Open the HDF5 file in read-write mode, i.e., the original metadata and experiment structure, analysis results, images, etc. are overwritten when closing the present experiment. However, raw data are NEVER modified.

 Load in read-only mode, i.e., don't store any changes back to the file

Routine acquisition is disabled. Anything done in this session is lost when closing the Experiment.

• Cancel loading HDF5 file

Do not open the HDF5 Experiment.

A normal Igor session is launched. The SutterPatch menu is populated with blank submenus, and the command 'Reactivate SutterPatch' to re-open the HDF5 Experiment.

Save Experiment

If the current Experiment is already named, it is immediately saved. Otherwise, a 'Save experiment as' file dialog

is displayed. If Preferences are enabled for automatic file naming, a default Experiment name is provided.

Stores the Experiment data and temporary (input / output) waves, but does no other cleanup; HDF5 files are not updated to disk.

Save Experiment As

If Preferences are disabled for automatic file naming, the last used Experiment file name or the default name is displayed for renaming and saving. This then stores the Experiment data and temporary (input / output) waves, but does no other cleanup; HDF5 files are not updated to disk.

If Preferences are enabled for automatic file naming, an incrementing Experiment name is displayed for renaming and saving. However, nothing else in the Experiment gets saved.

Recent Experiments

A list of recently used Experiments.

Exit

An Experiment file 'Save' dialog is displayed before closing the program. If an Experiment is not saved, global variables and window sizes / positions are lost.

Open SutterPatch HDF5 File

[Only displays when HDF5 file saving in SutterPatch / Set Preferences / Files and Naming is enabled.]

New HDF5 File

[Only displays when HDF5 file saving in SutterPatch / Set Preferences / Files and Naming is enabled, and data has been acquired or loaded.]

Stores all existing data into the present HDF5 file, cleans-up the Experiment, and creates a new HDF5 file, so that acquisition can continue as if you had started a "New Experiment", but without starting a new SutterPatch session (or *.pxp file).

Update HDF5 File

[Only displays when HDF5 file saving in SutterPatch / Set Preferences / Files and Naming is enabled, and data has been acquired or loaded.]

Updates the active HDF5 data file without starting a new Experiment.

Compact HDF5 File

[Only displays when HDF5 file saving in SutterPatch / Set Preferences / Files and Naming is enabled, and data has been discarded.]

Removes discarded data from an HDF5 data file without resaving the entire Experiment.

Data

Load Waves / Packages / Install HDF5 Package

Load Waves / New HDF5 Browser

Data Browser Access SutterPatch objects contained in the Experiment.

Analysis

The Analysis menu provides a wide assortment of mathematical transforms.

Curve Fitting Create custom fitting equations.

Quick Fit A variety of Igor fitting equations.

Windows

The Windows menu provides access to all window controls.

Command Window A quick code interpreter to manually process SutterPatch

and Igor commands.

Control / Retrieve All Windows

Hidden unminimized windows can be brought into view

with the menu command.

Layout

The Layout menu only displays when a Layout is the active window. Use it to modify the Layout window display and objects.

SutterPatch

Dashboard Display icons for core program functions.

Acquisition Control Open a control panel with Start/Stop and other interac-

tive acquisition controls for Routines and Paradigms.

Scope Window Bring an open scope window to the front.

Hardware Control

Amplifier Control Panel Open the hardware control panel.

Reset All Amplifier Settings Return the Amplifier Control Panel to its default

settings for all headstages.

dPatch Maintenance	Calibrate the headstages and auxiliary analog channels.	
dPatch Dynamic Cla	open the Dynamic Clamp Editor to configure dynamic clamp methods and models.	
Lock-In Adjustments	Manually tune the dPatch "lock-in amplifier" system.	
Reset USB	Re-initialize USB communication with the computer. If in Demo mode, you need to start a 'New Experiment' to access 'Reset USB'.	
Membrane Test	Open and run the scope window to monitor seal formation and cell health.	
Free Run (Scope)	Open and run the scope window in oscilloscope style.	
Reset Acquisition	Stop the Paradigm and/or data acquisition and clear corrupted acquisition settings.	
Paradigm Editor	Open the dialog to load, edit and run Paradigms.	
Routine Editor	Open the dialog to load and edit Routines.	
Template Editor	Open the dialog to manage templates.	
Equation Editor	Open the dialog to load and edit Equations.	
Solution Editor	Open the dialog to control solutions.	
Camera Control	Open the window to capture images.	
Data Navigator	Open the window to organize and display the experiment Paradigm, Routine and acquisition data in a tree structure.	
Analysis Editor	Open the dialog to manage analysis graphs.	
Layout Page	Show Layout	
	Delete Layout	
Set Metadata	Open the dialog to configure user-specified experimental information.	
Set Preferences	Open the dialog to modify the default program settings.	
Log Window	Open the window to display a history of program actions.	

Shortcut Editor Open the Shortcut Editor dialog to manage keyboard shortcuts.

Action 1 [Hold+10mV | 10pA:Right]

Action 2 [Hold-10mV | 10pA:Left]

Action 3 [Hold+1mV | 1pA:Right, shift]

Action 4 [Hold-1mV | 1pA:Left, shift]

Action 5 [View last] F2

Action 6 [Stop Acquisition] F3

Action 7 [Start Routine] F4

Action 8 [Stop Routine] F5

Action 9 [Pause Paradigm] F6

Action 10 [Resume Paradigm] F7

Action 11 [Cursor Info] F10

Available Analysis Modules

Action Potential Analysis

Synaptic Event Analysis

Single Channel Analysis

Help

Igor Help Browser Igor and SutterPatch Help Topics.

About SutterPatch SutterPatch version and contact information.

Scope Right-click Menus

Different areas of the scope windows support additional functionality through "right-click" menus in Windows, or "Command-click" menus in macOS.

Scope X-Axis

(right-click the X-axis)

- Autoscale All Axes
- Autoscale X Axis
- Set X Scale...
- Axis Properties...

Scope Y-Axis

(right-click the Y-axis)

- Autoscale All Axes
- Continuous Autoscale Y Axis
- Autoscale Y Axis
- Full Scale Y Axis
- Set Y Scale...
- Axis Properties...
- Hide Signal '<signal name>'
- Show Signal '<signal name>' Only
- Stack All Signals

Acquisition Scope main window

To display a limited data modification menu, right-click the blank area in a signal.

Note: If you click too close to the data, the full data modification menu displays instead. If you are having this issue, click near a horizontal or vertical edge of the signal pane.

- Autoscale All Axes
- Add Annotation
- Export Graphics Copy the selected signal to a Graph window.
- Toggle Cursor Info
 Manually measure X-Y data values or set a fitting range.
- Colors
- Hide Signal '<signal name>'
- Show Signal '<signal name>' Only
- Stack All Signals

Reanalysis Scope main window

To display a limited data modification menu, right-click the blank area in a signal.

Note: If you click too close to the data, the full data modification menu displays instead. If you are having this issue, click near a horizontal or vertical edge of the signal pane.

- Autoscale All Axes
- Add Annotation
- Export Graphics

Copy the selected signal to a Graph window.

• Toggle Cursor Info

Manually measure X-Y data values or set a fitting range.

- Colors
- Hide Signal '<signal name>'
- Show Signal '<signal name>' Only
- Stack All Signals
- Show All Sweeps
- Show Marked Sweeps

Signal data

To display the full data modification menu, right-click on or near the data.

- Browse < signal name>
- Edit <signal name>
- Remove Sweep_#
- Hide Sweep_#
- Duplicate Sweep_#
- Replace Wave
- Copy
- Modify Sweep_#
- Customize at Point
- Mode
- Line Style
- Line Size
- Markers
- Marker Size
- Color
- Bring to Front
- Send to Back
- Forward
- Backward
- Move to Opposite Axis
- Quick Fit
- Parametric Plot

- Amplitude Histogram Plot
- Export Graphics
- Toggle Cursor Info

 Manually measure X-Y data values or set a fitting range.
- Hide Signal '<signal name>'
- Show Signal '<signal name>' Only
- Stack All Signals
- Show All Sweeps
- Show Marked Sweeps

Scope Marquee window

(click-and-drag in a signal)

- Expand
- Horiz Expand
- Vert Expand
- Shrink
- Horiz Shrink
- Vert Shrink

Extract Template

4.3.6 Sample Files

Sample settings files (subject to change) are included in the .. / Documents / SutterPatch / Parameters folder:

Dynamic Clamp Pool Sample Files

 $SP_DynamicClampPool.spd$

Dynamic Clamp models for dPatch systems.

- 1. Predefined
- 2. HodgkinHuxley
- 3. HodgkinHuxley_1
- 4. GHK
- 5. Leak
- 6. HCN1

- 7. KV4_2
- 8. KV1 4
- 9. Markov_HH
- 10. TTX_sensitive_Na
- 11. Variable_Conductance

Equation Pool Sample Files

SP_EquationPool.txt Equations for all Sutter amplifier systems.

1. X3pi 3*pi

ElapsedTime ParadigmTime - time
 Temperature aux[1]*1.23 - 273.15

Paradigm Pool Sample Files

LockIn / LockIn_dPatch.spp Paradigms for dPatch lock-in tuning.

1. LockIn_Adjust_500Hz

2. LockIn _ Adjust_1kHz

3. LockIn_Adjust_2kHz

4. LockIn_Adjust_5kHz

5. LockIn_Adjust_10kHz

6. LockIn Adjust 20kHz

7. LockIn _DoAdjust

LockIn / LockIn_IPA.spp [for D/IPA lock-in tuning]

SP_ParadigmPool_dPatch.spp Paradigms for dPatch systems.

1. Amplifier_Setup Set initial amplifier settings.

2. Start_one_Series Start acquisition of one routine.

 $3. \hspace{0.2in} \textbf{Set_amplifier_and_start_IV} \hspace{0.2in} \textbf{Set amplifier to a known state, then start a rou-} \\$

tine.

4. Interactive_acquisition_1 Run an interactive acquisition stopping at a given

analysis condition.

5. Start_two_Series Start acquisition of two subsequent routines.

6. Start_ForEachSweep Start acquisition of a routine, individually trigger-

ing each sweep.

7. Interactive_acquisition_2 Run an interactive acquisition loop that selects between 2 routines, and manually stops via a Checkbox.

8. Tuning_with_input Use the paradigm "Input" control to increment or decrement a Routine's stimulus output.

9. Toggle_Persistence Use a Checkbox to toggle scope window trace persistence while acquiring a routine.

10. Switch_Headstages Switch between multiple headstages.

11. Tuning_with_Keys

Use the keyboard to increment or decrement a Routine's stimulus output by 10 mV.

12. CellHealth_From_CC Monitor the cell's resistance and capacitance in current clamp mode.

SP_ParadigmPool_IPA.spp [for D/IPA systems]

Routine Pool Sample Files

LockIn / LockIn DIPA.spr [for Double IPA lock-in tuning]

LockIn / LockIn_dPatch.spr Routines for dPatch lock-in tuning.

1. phase delay

2. LockIn 500Hz

3. LockIn_1kHz

4. LockIn _2kHz

5. LockIn 5kHz

6. LockIn _10kHz

7. LockIn 20kHz

LockIn / LockIn_IPA.spr [for IPA lock-in tuning]

SP_RoutinePool.spr [for IPA systems]

SP_RoutinePool_Dendrite.spr [for Dendrite systems]

SP_RoutinePool_DIPA.spr [for Double IPA systems]

SP_RoutinePool_dPatch.spr Routines for two-headstage dPatch systems

1. Amplitude Equations Equation for a variety of stimulus waveforms.

2. AT_InactRec_P4 Inactivation with leak subtraction.

3. Bowtie_Test Multi-channel input with incrementing ramp

waveforms.

4. ContinuousNoOut Acquisition without any output waveform.

5. IV I-V for voltage-clamp mode.6. IV_CC I-V for current-clamp mode.

7. IV_Continuous I-V with continuous acquisition.

8. IV_P4 I-V with four leak-subtraction pulses.

9. IV_tuning I-V for sample "tuning" paradigms.

10. IV_VC_CC IV for voltage- and current-clamp modes.

11. Multi_Test Multi-channel input with an incrementing square-

step waveform.

12. Onset_SlowActivation Onset Slow activation.

13. Recovery_Inactivation Recovery from inactivation.

14. Recovery_SlowInact Recovery from slow inactivation.

15. SS_SlowInactivation Steady-state slow activation.

16. Synaptic_Stim Synaptic stimulation.

17. Synaptic_Stim30 Synaptic stimulation for 30 s.

18. Synaptic_StimPlusDig Synaptic stimulation with digital output.

19. Template_PlusVirtual Template wave and recording virtual signals.

20. Template_SpontAct Template wave from a recorded signal.

21. Template_Test Template wave for waveform output.

22. Test_Pulse Test pulse.

SP_RoutinePool_dPatch_1HS.spr Routines for one-headstage dPatch systems.

1. Amplitude Equations Equation for a variety of stimulus waveforms.

2. AT InactRec P4 Inactivation with leak subtraction.

3. Bowtie_Test Multi-channel input with incrementing ramp

waveforms.

4. ContinuousNoOut Acquisition without any output waveform.

5. IV I-V for voltage-clamp mode.6. IV CC I-V for current-clamp mode.

7. IV_Continuous I-V with continuous acquisition.

8. IV_P4 I-V with four leak-subtraction pulses.

9. IV_tuning I-V for sample "tuning" paradigms.

10. Multi_Test Multi-channel input with an incrementing square-

step waveform.

11. Onset SlowActivation Onset Slow activation.

12. Recovery_Inactivation Recovery from inactivation.

13. Recovery_SlowInact Recovery from slow inactivation. 14. SS Inactivation Steady-state inactivation. 15. SS_SlowInactivation Steady-state slow activation. 16. Synaptic_Stim Synaptic stimulation. 17. Synaptic_Stim30 Synaptic stimulation for 30 s. 18. Synaptic_StimPlusDig Synaptic stimulation with digital output. 19. Template_PlusVirtual Template wave and recording virtual signals. 20. Template_SpontAct Template wave from a recorded signal. 21. Template_Test Template wave for waveform output.

22. Test_Pulse Test pulse.

Shortcut Pool Sample Files

13. Previous Sweep

SP_Sh	ortcutPool.spo	Shortcut Actions for all Sutter amplifier systems.
1.	VHold + 10mV	Increase the Control Panel V-holding level by 10 mV.
2.	VHold-10mV	Decrease the Control Panel V-holding level by $10 $ mV.
3.	VHold + 1mV	Increase the Control Panel V-holding level by 1 mV.
4.	VHold – 1mV	Decrease the Control Panel V-holding level by 1 mV.
5.	View last	Open the last acquired Series into a Reanalysis scope window.
6.	Stop Acquisition	Stop the acquisition of a Series.
7.	Start Acquisition	Start the acquisition of a Series.
8.	Stop Paradigm	Stop the execution of a Paradigm.
9.	Pause Paradigm	Stop the execution of a Paradigm.
10.	Resume Paradigm	Resume execution of a Paradigm.
11.	Cursor Info	Toggle On/Off
12.	Next Sweep	

Solution Pool Sample Files

SP_SolutionPool.spo Solutions for all Sutter amplifier systems.

1. undefined No solutions are configured.

Template Pool Sample Files

SP_TemplatePool.spt Templates for all Sutter amplifier systems.

1. RoutinePreview

2. Single_actionPotential

3. HodgkinHuxley

4. Noise

Experiment Sample Files

Sample data (subject to change) are included in the ... / Documents / SutterPatch / Example folder:

ActionPotentials.pxp Action potential data.

Large APs.pxp Large action potential data.

MiniExample.pxp Spontaneous miniature synaptic potential data.

4.3.7 Set Preferences

Preferences settings customize the default settings for several areas of the SutterPatch program. To access, go to the SutterPatch / Set Preferences menu.




Figure 4-91. Preferences Settings

i. General

Enable drag and drop for pool editors

The Routine Editor loads a "pool" of Routines from a Routine Pool file. These Routines can be re-arranged in the Routine Pool list by clicking and dragging with the mouse.

The Paradigm Editor operates in the same manner, and also displays a list of Paradigm Steps, which can also be re-arranged by dragging-and-dropping.

Window Position

Automatic window repositioning

When SutterPatch windows or dialogs are opened or moved, when the action is done, they are automatically repositioned to be fully visible. If a "child" window is opened, the parent window is moved to the left until the child sub-window is fully visible or the parent window reaches the left edge of the main window/screen.

Restore command window position

Enable so the Command window position is remembered for the next Experiment that opens. Otherwise, a new Experiment always returns the Command window to its default size and location.

Dual-monitor option [macOS only]

One screen Prevents windows spanning across multiple

monitors.

If a "parent" window is moved to another monitor, it fully displays in the new monitor, while any child sub-window remains behind fully displayed in the original monitor.

Note: You can also bring all unminimized windows into view with the Windows > Control > Retrieve All Windows menu command.

Feedback notice display duration

[1 - 30 s]

Control how long SutterPatch messages display for

reading, before automatically closing.

Preferences Defaults Restore default Preferences settings for all preferences.

Edit Configuration Pool Create different user Preferences configurations,

selectable for loading at the start of a new experiment.

Delete Configuration Remove the named Configuration from the list.

Add New Configuration Create a new Preferences Configuration for the

existing Preferences settings.

Description Enter the name of the new Preferences

Configuration.

Update Active Configuration (#)

The "active" Configuration is updated with the

existing Preferences settings.

This dialog opens with the active Configuration name listed. The number (#) indicates its

position in the Configuration list.

Note: The 'Files and Naming' preference for HDF5 file saving cannot be disabled via running a Configuration; it must be manually disabled by the user in Set Preferences.

Also, while SutterPatch Preferences Configuration files use the *.spc file extension, this is reported by the OS as file type "PKCS #7 Certificates".

ii. Files and Naming

Data file path: Browse to select a folder.

The data folder should reside on a local disk drive. We do not recommend using a mapped network drive, as speed/throughput bottlenecks can occur, including delays when saving experiments.

Enable automatic naming for experiment

Experiment file name example: (Maximum 34 characters)

At least one of the following file name components must be enabled.

Text: [] Legal characters are A-Z, a-z, 0-9 and "_".

Include user text in the file name. At least one character is required.

Date: YYMMDD Include the date in the file name.

Time: hhmmss Include the time in the file name.

Save to separate HDF5 file Store the Experiment data waves using the HDF5 file format, a modern efficient format for managing and saving high volumes of data.

By default, the Experiment file is stored in a "packed" (*.pxp) experiment, where all experimental information is conveniently stored in one file.

However, for very long experiments, this can result in delays when saving new data, as the entire experiment is re-saved with each additional recording.

Enable the HDF5 file option to avoid such delays. A new setting becomes active after starting a New

Experiment or a new SutterPatch session.

Then, whenever a new HDF5 file is created, SutterPatch stores all existing data into the active HDF5 file, cleans up the Experiment, and creates a new HDF5 file so that HDF5 acquisition can continue as if you had started a "New Experiment", but without starting a new SutterPatch session.

The raw signal data are stored to the HDF5 disk file during acquisition after each sweep, instead of storing all data at the end of an Experiment, which can be a time-consuming experience.

Other waves from the SutterPatch Data folder (including metadata, Experiment structure, analysis results, Routines, Log, images, Dynamic Clamp conductances) are separately stored to the HDF5 file at the end of a Routine or Experiment; items outside of the SutterPatch Data folder (such as graphs and layouts) are stored to the ".pxp" Experiment file.

Note: It is strongly advised to enable the "automatic naming" option above, so that "*.h5' HDF5 files and their parent '*.pxp' Experiment file are kept "in sync".

Keep only one Sweep in Memory

[Only displays when HDF5 is enabled.]

For the leanest operation, only hold the wave of one sweep in memory, so memory buffers do not need to be re-allocated for the Experiment.

The downside is that multi-sweep data cannot be processed online, such as subtracting the data of a "reference" sweep from other sweeps.

Update HDF5 file after each routine

[Only displays when HDF5 is enabled.]

For efficient processing, while the raw signal data are written to the HDF5 file after each sweep during acquisition, the other SutterPatch Data folder information (metadata, Experiment structure, analysis results, Routines, Log, images, Dynamic Clamp conductances) is separately written to the HDF5 file.

The non-data information are automatically written to the HDF5 file at the end of an Experiment. However, if this option is enabled, the information is also automatically written to the HDF5 file at the end of each Series.

[save options]

[Only displays when HDF5 is disabled.]

Save to temp file after each routine

The raw data are saved into a temporary file after each recording. This can help to speed up file-saving time for large Experiments composed of several smaller recordings.

The temporary file starting size is based on the starting size of the Experiment. The temporary data are re-saved to the main Experiment when the Experiment is closed and/or saved.

• Save entire experiment after each routine

This default option re-saves the entire Experiment after each recording (all data and Experiment information). This is the safest method of operation for data integrity, but can produce significant post-recording file-saving delays in larger Experiments.

• Don't save to temp file after each routine

Data and information are held in memory until the Experiment is explicitly saved; there are no file-saving delays after a recording is stopped. This provides the fastest method of operation when making multiple recordings, but is also the least secure, as data loss can occur if the computer encounters problems.

Warn when file size exceeds limit

When a recording causes the Experiment to exceed the desired limit, a notification message displays after the Routine stops.

Limit

[+10.0 MB]

Enter the limit.

Note: It is advised to disable the Igor Pro 9 'Autosave' feature to prevent delays during data acquisition. See Misc / Miscellaneous Settings / Autosave.

iii. <u>Hardware</u>

• Prompt for hardware on startup (if no Sutter hardware is found).

When a new Experiment is started, if Sutter patch-clamp hardware is not connected to the computer and turned on, you are prompted to retry the USB connection or select an emulation mode.

• If no Sutter hardware is found, emulate:

• IPA Integrated Patch Amplifier system.

• Double IPA Dual-headstage IPA system.

• dPatch Digital Patch-clamp system.

• Dendrite Data Acquisition system.

When a new Experiment is started, if Sutter patch-clamp hardware is not connected to the computer and turned on, automatically start up in the selected hardware emulation mode.

Stability Control

In CC mode reduce electrode compensation by [0.0-3.0] pF

During whole-cell patching, if the Electrode Compensation control is set too high, oscillations can occur, and the patch-clamp seal can become unstable and be lost. As the Voltage Clamp mode typically operates with higher electrode compensation values than the Current Clamp mode, this preference promotes "safe" switching between the Voltage Clamp and Current Clamp modes.

If you are routinely losing cells when switching into Current Clamp mode, increase this setting from the default '0.5' to '1' or '2'.

Note: The electrode compensation reduction is done in the background, and does not affect the Control Panel current-clamp settings.

Preserve clamp state when switching modes:

• Voltage Clamp to Current Clamp

The VC actual current level is used as the CC holding level.

Current Clamp to Voltage Clamp

The CC actual voltage level is used as the VC holding level.

The current is held at 0 A for 50 ms between modes.

iv. Control Panel

Customize the active headstage tab's color in the Control Panel.

Headstage background color

Headstage 1 Color palette displays.

Headstage2 Color palette displays.

v. Scope Window

Scope setting changes are only applied when a scope window is created.

General

Time axis unit:

• Auto-set Sweep duration < 120 s, use "s"

Sweep duration >= 120 s, use "min"

Sweep duration >= 7200 s, use "h"

• SI unit Always use standard SI base units, such as "s" for

time.

Include zero when autoscaling

When using Scope window autoscaling, the Y-axes limits are rescaled with each sweep so that all data are visible. If the signal does not cross the zero line, the Y-axes ranges are extended to the zero amplitude,

Acquisition

Maximal sweeps displayed in persistence display

[2-100]

A11

Improve data readability and system performance by restricting the number of sweeps displayed, which reduces the display processing load.

Allow Display Compression Display compression is applied to the data in all "live" scope windows (Acquisition / Free Run / Membrane

Test). This reduces the display processing load, and can improve system performance when resources are low.

When there are four times as many data points as the scope width in pixels, the number of data points plotted are reduced, as the minima and maxima from two groups of up to 50 samples are displayed per screen pixel.

Reanalysis

Persistence On

Off

Keep current setting

Maximal sweeps displayed in persistence display

[2-100](30 = default value)

The last 'N' sweeps are

displayed

[All] All sweeps are displayed.

Improve data readability by restricting the number of sweeps displayed. This can also improve system performance by reducing the display processing load.

Allow Display Compression Display compression is applied to the data in the Reanalysis scope window. This reduces the display processing load, and can improve system performance when resources are low.

> When there are four times as many data points as the scope width in pixels, the number of data points plotted are reduced, as the minima and maxima from two groups of up to 50 samples are displayed per screen pixel.

Note: Display compression is not applied to Single Channel Analysis data.

Show event tags

Display tag lines in the Reanalysis scope window; also display tag text boxes in the Continuous and Concatenated display modes.

Tag Position

- Frozen
- Movable

Tag types to show

- User
- Input-triggered
- System

Tag text box

• Relative Time Time from beginning of Series.

[hours to milliseconds]

• Absolute Time Clock time.

[hours to milliseconds]

Description User Comment, from 'Set Tag'.

Tag appearance

Color by type

• Transparent

Appearance Use a color palette to choose colors.

Active signal panel color [light gray]

Inactive signal panel color [dark gray]

Active sweep color [blue]

Inactive sweep color [red]

Drop-down color palette Click on a color square to set it as the

active color.

Other Use a color dialog with more options.

Opaque < unused >

[Preview Pane] The selected signal and sweep colors are

displayed in a preview pane.

Tip: For dark-room experiments, the window background color can be adjusted by the operating system:

• Windows: In the Control Panel / Appearance / Personalization win-

dow, scroll down and select the High Contrast Black theme, or use the Windows Magnifier tool with option

'Turn on color inversion' enabled.

• macOS: Press 'Control-Option-Command-8' to set the System

Preferences / Accessibility / Display / Invert Display colors

option, or open its menu with 'Command-Option-5'.

vi. Metadata

Metadata setup detail level: Select level for metadata setup categorization complexity.

Select which metadata groups and parameters are visible for setup (in Set Metadata and Data Navigator / Build Hierarchy.)

1 Basic Show only the most essential parameters.

2 Extended Show additional detail.

3 Full Expose all available metadata parameters.

Note: This setting does not affect Metadata Review windows - all defined metadata are always displayed.

This setting does not affect data acquisition metadata prompts – all configured prompts are always executed.

vii. Graphs and Layouts

These settings apply to Layout windows and stand-alone graph windows (not graph files).

Copy "To Clipboard" graph format:

Several popular file formats are supported:

PNG Portable Network Graphics
 PDF Portable Document Format
 TIFF Tagged Image File Format

• JPEG Joint Photographic Experts Group

Note: When pasting, not all formats are supported by other programs.

Layouts

Changes are applied when a new Layout window is created.

Default graphs per new Layout page:

- 1
- 2
- 3
- 2 x 2 (Column x Row)
- 2 x 3 (Column x Row)
- 2 x 4 (Column x Row)

Graphs

left (the Y-axis)

Tick location:

- Outside
- Crossing
- Inside
- None

Labels:

- On
- Axis only Text labels only.
- Off

Grid:

- Off
- On
- Major only

bottom (the X-axis)

Tick location:

- Outside
- Crossing

- Inside
- None

Labels:

- On
- Axis only Text labels only.
- Off

Grid:

- Off
- On
- Major only

viii. <u>Data Export</u>

File Export

- Igor Binary Format Save the signal formatted as an "Igor Binary Wave" file (*.ibw).
 - Export all selected sweeps to one file per signal (one 2D wave)
 - Create individual files for each sweep (multiple 1D waves)
 - Export all selected sweeps to one concatenated file per signal
- ABF Format Save each signal of the routine formatted as an "Axon Binary File" v.1.8 (*.abf).
 - Export all selected sweeps to one file per signal
 - Ignore unselected sweeps
 - Replace unselected sweeps with NaN
 - Create individual files for each sweep
- ATF Format Save the signal formatted as an "Axon Text File" (*.atf).

Uses the table formatting preferences below.

Table Formatting: Copy to Clipboard, Text Table Export

Cell separator

- Tab
- Comma
- Semicolon

New line

- <CR> (Igor, macos)
- <CR> <LF> (Windows)

Invalid value

- Use NaN
- Use empty string
- Use zero

Digits in table entries [3-15]

ix. Factory Reset

[Factory Reset] Click this button to reset all SutterPatch preferences and

settings to their defaults. SutterPatch will need to be exited

and restarted to complete the factory reset.

Warning! When you do a factory reset, you will lose ALL your

configuration data (including Metadata, Scope, Preferences,

etc.)

4.3.8 Shortcut Editor

Keyboard control of SutterPatch is available by configuring keyboard shortcuts.




Figure 4-92. Shortcut Editor

All assigned Control (Command), Function and Shortcut Action key combinations are automatically added to the main menu SutterPatch / Shortcuts submenu. All Control and Function Key Assignments are automatically saved when the program is closed.

The Shortcuts main menu list is shared by the Control and Function keys, so each menu item can only be assigned to a single key. Prior duplicate entries are replaced by "not assigned".

Control Key Assignments / Command Key Assignments

A Control key assignment can be easily made by clicking on its drop-down list and selecting a new item from the list of SutterPatch menu items.

Use Control keys by holding down:

• Windows: Control key: Ctrl

• macOS: Command key:

and clicking its assigned number key.

Control keys and their default SutterPatch menu assignments:

Control-1 Dashboard

Control-2 Scope Window [Scope window must be open.]

Control-3 Amplifier Control Panel

Control-4 Membrane Test
Control-5 Free Run (Scope)
Control-6 Paradigm Editor
Control-7 Routine Editor

Control-8 Data Navigator

Control-9 Set Metadata

Control-0 Acquisition Control

Other SutterPatch menu items available for assignment:

Analysis Editor

Camera Module

Dynamic Clamp

Equation Editor

Lock-In Adjustments

Log Window

Reset Acquisition

Set Preferences

Shortcut Editor

Solution Editor

Template Editor

Function Key Assignments

Computer keyboards usually include a set of Function keys [0-9] for special functionality. Configure a 'Function Key' assignment by clicking on its drop-down list and selecting a new menu item or Action item.

Function keys and Control keys share the same menu items list, and each menu item can only be only be assigned to a single key. So, after any new assignment, any duplicate key is changed to "not assigned".

Function keys can also be assigned from Shortcut Actions – see below.

Note: On some keyboards, you also need to press the 'Fn' key to use Function keys.

Also, macOS reserves nearly all Function keys for itself. In order to use Function keys for a macOS application, you must first check a checkbox in the macOS Keyboard control panel. Even then, macOS will intercept some Function keys.

Function Key F1 is not available for assignment, as it is reserved for Igor's Help function.

Default Functions

F1		Igor's help function	
F2	Action 5	View last	
F3	Action 6	Stop Acquisition	
F4	Action 7	Start Routine	
F5	Action 8	Stop Routine	
F6	Action 9	Pause Paradigm	
F7	Action 10	Resume Paradigm	
F10	Action 11	Cursor Info	[Target window must be open.]

Reset Key Assignments

Reset all Control keys to their default settings. Function key "menu" items are reset to 'not assigned', while 'Action' items are unaffected.

Shortcut Actions

Up to 10 additional custom keyboard Actions can be created, managed and stored in a file.

Displays the last executed shortcut Action.

Shortcut Pool Files

New Shortcut Pool Create a default Shortcut Pool.

Load Shortcut Pool Load the Shortcuts of a previously saved Shortcut

Pool file into the Shortcut Pool.

Revert to Last Saved Undo any unsaved changes to the Shortcut Pool.

Save Shortcut Pool Save the Shortcut Pool using its existing file name

and path.

Save Shortcut Pool As... Save the Shortcut Pool to a new file, and switch to

the new file. The default file name is the same as

the original file name.

Save Shortcut Pool Copy... Save the Shortcut Pool to a new file, but do not

switch to the new file. The default file name has

'Copy of" prepended to it.

Merge Shortcut Pools Append the Shortcut Actions from a previously

saved Shortcut Pool file into the loaded Shortcut

Actions table.

[] Shortcut Pool file path.

Shortcut Actions Table

Columns

On Enable/disable the Shortcut Action.

Key The assigned keyboard key.

Modifier The keyboard "modifier key" used in a key combi-

nation - all keys are simultaneously pressed.

Windows

• Ctrl Only for use with keys (0-9).

• Alt Keys '0 – 2' reserved by Igor for File / Recent Ex-

periments.

Shift Shift key.

• Fn Function key.

• Caps Lock Ignored.

macOS

• Command Only for use with keys 0-9.

Option Option key.Shift Shift key.

• Control Keys '0 – 2' reserved by Igor for File / Recent Ex-

periments.

• Caps Lock Ignored.

Shortcut Action An Action's instructions and settings.

Click a field in the table to highlight an Action and make it the active entry. Click-and-drag a field to reposition an Action in the table.

Predefined Shortcut Actions

1	[Hold+10mV 10 pA: Right]	(keyboard arrow key)
2	[Hold-10mV 10 pA: Left]	(keyboard arrow key)
3	[Hold+1mV 1 pA: Right, shift]	(keyboard arrow key)
4	[Hold-1mV 1 pA: Left, shift]	(keyboard arrow key)
5	[View last]	F2
6	[Stop Acquisition]	F3
7	[Start Routine]	F4
8	[Stop Routine]	F5
9	[Pause Paradigm]	F6
10	[Resume Paradigm]	F7
11	[Cursor Info]	F10
12	[Next Sweep]]	In Reanalysis Scope.
13	[Previous Sweep]	In Reanalysis Scope.

The following buttons modify the Shortcut Actions:

Insert Adds an Action to the 'Shortcut Action' list and opens its Shortcut

Actions Editor for setup.

These Actions operate similarly to Paradigm steps, with an additional Label field to name the Action in the Shortcuts menu.

Amplifier Control a dPatch amplifier's settings.

Analysis Append, average, display and save anal-

yses.

Camera Take a photo or run live video.

Cursor Info Use cursors to set a fitting range for graph-

ical data.

Data Navigator Open the Experiment's data management

center.

Execute Run an Igor or SutterPatch command.

Export Send graphs to the Layout window.

Front Window Set the specified window as the front win-

dow.

Hide Window Hide the specified window.

Paradigm Load & Run, Stop, Pause or Resume a Par-

adigm.

Reset Timer Reset the Paradigm Editor Timer to zero.

Routine Record data from a Routine.

Scope Operation Control the display of the scope window

signals and sweeps.

Select Sweep Set a sweep to be the "active" sweep. Set Axis Modify the axis scaling of a signal.

Set Checkbox Set local and global checkboxes for condi-

tional processing in Paradigm 'If' steps.

Set Mark The active sweep in the scope window is

"marked" or "unmarked" for processing by

the Data Navigator.

Set Tag Write a comment tag to the Paradigm

metadata.

Set Variable Set Paradigm or Routine variables.

Start Acquisition In an open Acquisition scope window. Stop Acquisition In an open Acquisition scope window.

View Last Display the last recording in a Reanalysis

scope window.

Write Log Write a note to the Log window.

Assign Key

This button opens the Shortcut Key Input dialog (or double-click in a "Key" or "Modifier" field) to input the desired keyboard combination for a letter, number, or symbol.

Note: Available keyboard letters, numbers, and symbols can vary from computer to computer, depending on the computer OS and Igor Pro's key usage. (Reserved keys typically open another window type, or are non-responsive.)

Keyboards often have a Function (FN) button to allow

special access to the Function keys.

Although the F1 function key is reserved in Igor Pro, it

can be assigned if used with a modifier key.

If the CAPS LOCK button is on when assigning a key, the

key is case insensitive.

Delete Key Remove the Key entry for the selected Action.

Edit Open the Shortcut Editor dialog (or double-click in a "Shortcut

Action" field) to change the Action's parameters.

Execute Run the selected Action.

Delete Remove the selected Action from the 'Shortcut Action' list.

Save Saves any changes to the current Shortcut Pool file.

4.3.9 Startup

The SutterPatch application startup sequence:

1. Click on the 'Igor Pro' icon to launch SutterPatch:



- 2. Igor Pro opens with a blank 'Command' window and an Igor Pro "splash" screen displaying SutterPatch file opening information, which then close.
- 3. Next a 'Welcome to SutterPatch' screen displays with launch options:
 - Igor Only Run Igor Pro (without launching SutterPatch).
 - Open Launch SutterPatch from a saved Experiment.
 - Start Launch SutterPatch for a new Experiment.
- 4. Click 'Start', an Igor Pro 'Command' window opens, and the Welcome screen displays a progress bar while compiling the SutterPatch code, then the Welcome screen closes.
- 5. Next, if no Sutter amplifier is detected, the 'No USB Connection' pane allows you to try re-establishing the USB connections, or to select a hardware-emulation demonstration mode:
 - IPA Single Headstage.
 - DIPA Double Headstage.
 - dPatch Double Headstage.
 - Dendrite Data Acquisition System.

In demonstration (demo) mode, the Amplifier Control Panel and Acquisition scope window are labeled with "Demo", the input and output channels use simulated data, and most SutterPatch functions are available (except dPatch Calibration and Dynamic Clamp).

6. The SutterPatch files are then initialized, the Dashboard and Acquisition Control panels display, and additional SutterPatch windows display if they were open in the prior experiment.

5. PROGRAMMING

5.1 Data Format

SutterPatch data are written in a 64-bit double-precision binary floating-point format. This supports a decimal precision of 17 significant digits.

The data are stored within an Igor Experiment (*.pxp) file.

For large data sets, an optional HDF5 file format will be available for streaming data acquisition without resaving the experiment at the end of a recording.

5.2 Data Structure

SutterPatch recorded data are stored as multidimensional data waves, and are listed per signal in the Data Browser. Select a data wave in the Data Browser and right-click to 'Edit' the Signal data in a spreadsheet-style table. The two-dimensional data wave is displayed with one row per sample point and one column per trace, with the number of data table columns increasing with the number of sweeps.

Warning! The raw data can be directly edited in the Data Browser – this is not recommended, as it permanently alters the data.

Note: While SutterPatch does not read the older Igor one-dimensional wave data-format, graph data for each axis can be separately exported to it. See the Analysis Editor / Files menu.

5.3 Data Paths

The Data Browser path references an internal Igor "root" folder, and not the computer's file system. The Data Browser right-click 'Copy Full Path' command copies a Signal's data wave path to the system clipboard.

For advanced users, the object's path name can be used in user functions and executable commands. However, when referencing an active scope window, the path name to the data wave can be substituted by "t[#]", where "#" refers to the signal position number in the scope window.

5.4 User Functions

SutterPatch functionality can be extended through the use of user-defined Functions.

To create a user Function:

- 1. Open the menu for Windows / Procedure Windows / Procedure Window.
- 2. Enter your user code into the Procedure window, following its "pragma' and "include' lines.

Example:

Note: The Function name must include trailing open/close parentheses "()".

- 3. Click on the Compile button at the bottom of the window.
- 4. Enter the Function name (including parentheses) into the Command window and press 'Enter', or use it in a Paradigm 'Execute' step.

For more information on creating your own functions, see the Igor Help topics on Programming / User-Defined Functions, and Procedure Windows.

Warning! User-defined functions only exist during the Experiment. They are not stored when the Experiment is closed. If you plan to re-use them in other Experiments, save them to a separate file, such as with a word processor.

Also, while user-defined functions are stored internally by Igor, there is no visible list, so you will need to maintain such a list manually.

6. TROUBLESHOOTING

6.1 Technical Support

Technical support is provided to customers at no charge.

Support hours: 8:00 AM - 5:00 PM PST (Pacific Standard Time).

Telephone: (+1) 415.883.0128

Fax: (+1) 415.883.0572

E-mail: info@sutter.com

Address: Sutter Instrument Company

One Digital Drive Novato, CA 94949

When contacting us for technical support, please provide your SutterPatch version and "build" numbers to help us troubleshoot your situation. These numbers are found in the Start splash screen during program loading, or in the Log Window Startup events.

For issues regarding Igor Pro features (all non-SutterPatch menu items), please contact Wavemetrics, Inc. for technical support.

6.2 Manual

The dPatch manual is installed as a PDF file along with the SutterPatch software. The latest version of the manual can be downloaded from our web site: https://mww.sutter.com/AMPLIFI-ERS/SutterPatch.html.

You can navigate through the PDF document using Table of Contents links, accessed via the Bookmarks tab on the left side of the PDF screen.

6.3 Online Help

Online help is available via the main Help menu, under 'Igor Help Browser' or 'Help Topics'. The SutterPatch Help Topics / Help File names start with "SP_". The online Help includes the same information as found in the PDF manual.

Most items in SutterPatch also include a short description as a tool tip. Hover the mouse over an item to see the tool tip.

6.3.1 Error Messages and Notifications

Some SutterPatch error messages or notifications will flash to get your attention, and automatically close after several seconds, and then write to a "History" window. To review such messages, see the Command window (menu item Windows / Command Window).

6.4 Startup Issues

6.4.1 <u>Installation Fails</u>

Problem: The SutterPatch installation on Windows fails due to language pack incompati-

bilities.

Solution: Support for foreign language packs (except Japanese) has been added. If foreign

language versions still cause problems, please contact Sutter Technical Support.

6.4.2 Startup Compiler Errors

Problem: The SutterPatch loading on Windows fails due to compiler errors.

Solution: Instead of using the SutterPatch updater, run the full SutterPatch installer.

6.4.3 Application Not Loading

Problem: The SutterPatch application does not load – the startup sequence only loads

Igor Pro.

Solution: If available, execute the Igor Pro menu command 'Macros / Autocompile'.

6.4.3 Saved Experiment Not Opening

Problem: In Igor Pro 8, the Experiment does not open due to a 'Macro Recreation

Error'.

Solution: When using templates in a Routine, close the Scope and Data Navigator

windows before saving the Experiment.

402

6.4.3 Amplifier Not Powering On

Problem: The dPatch amplifier does not power on.

Solution: Pre-May 2021 dPatch systems store firmware-related information in volatile

memory, which requires an internal battery for backup power. To keep this battery sufficiently charged, "power on" the dPatch system for several hours

at least every four years.

Units will not power up with a fully discharged battery. In this case, request a replacement firmware board with non-volatile memory from Sutter Instrument

Current dPatch systems use non-volatile memory, which does not use a battery.

6.4.4 Startup EEPROM Errors

Problem: Starting up SutterPatch and simultaneously powering on the amplifier gener-

ates an EEPROM error. Attached hardware such as headstages might use in-

correct settings.

Solution: Close SutterPatch, power cycle the amplifier, and relaunch SutterPatch.

6.4.5 Startup Odd Errors

Problem: When starting up or running SutterPatch, odd program errors display.

Solution: If after a SutterPatch update, close and re-open SutterPatch.

If after an OS update, roll back the OS software update.

6.4.6 USB Communication Fails

Problem: When starting up SutterPatch in Windows, there is no USB communication

with the computer.

Solution: Power off the amplifier, then reseat the USB cable on both ends, and then power

on the amplifier.

Or, the Windows "power plan" might have disabled the USB ports.

1) Go to the Windows Start screen, and enter "edit power plan" in the Windows Search box.

2) Click on "Change advanced power settings".

3) Scroll down to "USB settings" and click on its [+] box.

- 4) Click on the "USB selective suspend setting" [+] box.
- 5) Change the "On battery" and "Plugged in" settings from 'enabled' to 'disabled', and click "OK".

Or, a Windows 10 update can sometimes disable the computer's USB ports. Either browse the web to find a Device Manager driver solution to the problem, or restore the OS to an earlier version of Windows 10, until a Windows 10 update fix is released.

6.4.7 Sample Parameter Files Not Installed

Problem: The SutterPatch installer fails to install the sample parameter files, as access is

blocked to the Program Files or Users\..\Documents folders.

Solution: Disable any virus scanners or firewalls. If that does not help, then manually

copy the sample parameter files into the Users\..\Documents\SutterPatch

\Parameters folder.

6.5 Acquisition Issues

6.5.10 Routine Loading Delays

Problem: Selecting a Routine in the Routine Editor suspends the SutterPatch program.

Solution: Hide the Routine Editor Preview pane.

If there is a very large number of sweeps to display, the Preview pane can take a long time to redraw, and the program becomes temporarily unresponsive.

6.5.1 Acquisition Does Not Start

Problem: Unable to start an acquisition, because SutterPatch thinks a prior acquisition is

still in progress.

Solution: Use the menu command SutterPatch / Reset Acquisition to clear the acquisition

status.

6.5.10 Acquisition Start Delayed

Problem: After starting acquisition, it takes a long time for the actual recording to begin.

404

Solution: If a very large command waveform has to be generated, it can take a long time

to create the output wave, which can delay the start of acquisition.

If command stimuli are not needed, disable those output channels. This does

not affect the output of holding levels.

6.5.2 Acquisition Windows Lock Up

Problem: The scope window, Routine Editor or Paradigm Editor lock up during acquisi-

tion.

Solution: Use the menu command SutterPatch / Reset Acquisition to halt acquisition.

A combination of SutterPatch-related and computer-related issues can contribute to your system's performance. For suggestions to improve it, see the Trou-

bleshooting item Sluggish Acquisition below.

6.5.3 Acquisition Terminates

Problem: During acquisition, the recording terminates unexpectedly.

Solution: Close the Analysis / Data Browser window, if it is open.

This window can consume a large amount of system resources, which can inter-

fere with data acquisition.

6.5.4 Signal Flat

Problem: A scope window input signal is completely flat during acquisition, i.e., zero am-

plitude.

Solution: The corresponding headstage might not be attached to its port. The headstage

HDMI connectors do not lock on - make sure they do not disconnect from their

port.

Power off the dPatch system and reconnect the headstage.

6.5.5 Signal Saturated

Problem: A Scope input signal is completely saturated during acquisition.

Solution: The corresponding headstage is not attached to its port. Power off the amplifier

and reconnect the headstage cable.

If the headstage is attached, the Gain setting might be set too high. Reduce the

output gain on the Amplifier Control Panel.

Note: The headstage HDMI connectors do not lock on - make sure they do not

disconnect from their port.

6.5.5 Headstage Noise

Problem: The noise levels of the instrument suddenly and erratically increase.

Solution: If the headstages are touched, the noise level will greatly increase. Make sure

you are grounded or working in a Faraday cage.

6.5.6 Paradigm Sound Reduced

Problem: The paradigm 'Sound' step volume is attenuated at lower frequencies.

Solution: Upgrade the computer speaker, such as with add-on speakers.

6.5.7 Offset Zero Delay

Problem: The Offset button in the IPA Control Panel has a short delay before it responds.

Solution: This can occur after running the Membrane Test due to internal processing.

6.5.9 Sluggish Acquisition

Problem: Data acquisition is sluggish.

Solution: The computer's available resources need to be increased to handle the system

load.

A combination of SutterPatch-related and computer-related issues can contribute to your system's performance. Here are some suggestions to improve it:

Close: Data Browser window – in Igor Pro 8, it consumes a lot of CPU

time.

Disable: Computer screen saver, and Power Save or Sleep modes.

Disable: Scope window persistence display.

Disable: Routine Editor / Input Channels / Virtual channels.

Reduce: Routine / Acquisition & Routine Parameters / Output sampling

rate.

Close: Background software.

Remove: Software for certain license protection USB keys (dongles).

Optimize: Hard disk (defragment).

Upgrade: Computer graphics card.

Increase: Computer RAM, cache size and/or CPU speed.

6.5.9 <u>Heavy Duty Acquisition (& Analysis)</u>

Problem: MHz acquisition puts the CPU performance, memory requirements and disk

speed under stress.

Acquiring one channel for 1 minute at 5 MHz will create a sweep which is 2.4 GByte large, a very large data block. Acquiring at 5 MHz processes 1000 times the amount of data compared to 5 kHz. Such data requirements are more than 1000 times the size of data requirements of normal electrophysiological experi-

ments.

Solution: Modifications can be made to maximize the throughput:

General Acquisition

Enable: Set Preferences / Files and Naming / Save to separate HDF5 file.

For HDF5 files, data is written at the end of each sweep, instead

of at the end of the Experiment (which could take hours.)

Enable: Set Preferences / Files and Naming / Keep only one Sweep in

Memory.

Reduces memory requirements and usage.

Enable: Set Preferences / Scope Window / Y axis initial settings / Use last

y-scale.

Bypasses the extra processing time of Continuous autoscale.

Enable: Set Preferences / Scope Window / Acquisition / Allow display

compression.

Reduces display memory and processing requirements and us-

age.

Don't: Resize the Scope window.

Avoid a redraw of the window, especially during critical acquisi-

tion.

Check: SSD is the primary drive for system and data storage.

Non-SSD drives do not have the required write speed.

Check: SSD drive has enough free space for the acquisition.

Acquiring one channel for 1 minute at 5 MHz creates a sweep 2.4

GByte large.

Check: CPU is a dual-core processor (i5) or higher.

This is a base requirement for SutterPatch.

Use: Igor Pro 9.

This version is much more responsive than Igor Pro 8 during ac-

quisition.

Routines

Hide: Routine Editor Preview pane

If there is a very large number of sweeps to display, the Preview pane can take a long time to re/draw, and the program becomes

temporarily unresponsive.

Reduce: Routine Editor / Input / Channels / Nyquist Factor.

Reduce the Nyquist factor allows to acquire longer durations

Enable: Routine Editor / Real Time Measurements & Graphs / Disable

Execution of Measurements and Graphs.

Free Run

Restrict: Scope Duration to 1-2 s, when storing data to disk is enabled.

Otherwise, longer disk-write times might cause screen data to

overwrite.

Reduce: Signal Parameters / First(Second) Channel / Nyquist Factor.

Reduce the Nyquist factor allows to acquire longer durations

The maximal continuous data size is set by Igor Pro's maximal wave size of 2.5 MSamples (2,500,000 samples). This corresponds to a 500 s long (8 minutes 20 s) acquisition with a bandwidth of 1 MHz, i.e., a sampling rate of 5 MHz. One can reduce the Nyquist factor for the 1 MHz acquisition from the default value of 5 to 2.5. This allows to increase the maximal continuous acquisition dura-

tion to 1000 s (16 minutes 40 s).

Data Navigator

Hide: Preview pane.

The time to scroll through data can be excessive.

Reanalysis Scope

Enable: Set Preferences / Scope Window / Reanalysis / Allow display com-

pression.

Reduces memory and processing requirements and usage.

408

6.5.8 Post-Acquisition Delay

Problem: Every time acquisition completes, there is a delay with the program operations,

as the entire *.pxp Experiment file is resaved when a recording stops.

Solution: Create new Experiments more often, so that file sizes are smaller and more

manageable.

Or, change the file saving settings in Set Preferences / Files and Naming.

6.5.11 USB Errors

Problem: A USB communications error occurs.

Solution: Right-click in the amplifier Control Panel and select Reset USB . If the USB button does not turn from red to green, then try to isolate the problem.

a) Unplug and re-plug both ends of the USB cable from the amplifier to the computer.

- b) Try another USB cable.
- c) Try another USB port
- d) Remove any USB hubs.

6.5.12 System Freezes

Problem: The system hangs up after changing the filter or VC/CC mode selection.

Solution: Reset the USB port via the amplifier Control Panel USB reset button or the

SutterPatch / Hardware Control menu.

6.5.13 Amplifier Control Panel Issues

Problem: The dPatch Amplifier Control Panel is having odd problems.

Solution: Reset the active headstage controls to default settings by right-clicking or

Shift-clicking the Control Panel 'Reset' button, ; to reset all headstages, use the menu item SutterPatch / Hardware Control./ Reset All Amplifier Settings.

6.6 Analysis Issues

6.6.1 Analysis Not Deleted

Problem: An analysis cannot be deleted in the Analysis Editor.

Solution: The analysis is still in use, i.e., displayed in another window, such as a graph

window - close the window to allow the analysis to be deleted.

6.6.2 Signal Axes Overlay

Problem: The X-axis and units are overlaid in the scope window.

Solution: There is not enough room for the X-axis and units due to the number of signals

displayed. Switch to a tiled signal layout, or reduce the number of visible sig-

nals by right-clicking a signal and selecting 'Hide Signal'.

6.6.3 Graphs & Layouts Not Visible

Problem: Cannot see SutterPatch Experiment graphs or layouts on non-SutterPatch com-

puters.

Solution: Use the Igor menu command Windows / Graph, or Windows / Layouts or Layout

Macros, to see the object. Right-click it to modify with Igor options.

6.7 General Issues

6.7.1 <u>Buttons Unresponsive</u>

Problem: When using a slower computer in emulation (demo) acquisition mode, acquisi-

tion-related actions might be difficult, such as clicking the Stop button.

Solution: You may need to click the button more than once or hold it down longer than

usual.

A combination of SutterPatch-related and computer-related issues can contribute to your system's performance. For suggestions to improve it, see the Trou-

bleshooting item Sluggish Acquisition above.

6.7.2 Slow Display of Sweeps

Problem: When displaying a large number of sweeps, the display slows down.

Solution: Disable Persistence display in the scope window, or reduce the Preference /

Scope Window / "Maximal sweeps displayed in persistence display" setting.

6.7.3 Window Maximizing

Problem: Maximizing a window only maximizes the title bar.

Solution: Certain fixed-size windows and panels will not maximize (Action Potential Anal-

ysis, Amplifier Control Panel, Dashboard, Synaptic Event Analysis, Log, Paradigm Editor, Set Metadata, Set Preferences.) This is a reported Igor issue.

Also, if the active window is maximized, creating a new window might automati-

cally "maximize" the new window.

6.7.4 Slow Window Opening/Closing

Problem: Window opening and closing is slow on the macOS.

Solution: Close the SutterPatch Data Browser. This function consumes a lot of system

resources.

6.7.5 Windows Slowly Move Up the Screen

Problem: Various SutterPatch windows slowly creep up the screen until they get to the

top frame of the SutterPatch main window in Windows 10.

Solution: Disable the Windows 10 option Start / Settings / Devices / Mouse / 'Scroll inac-

tive windows when I hover over them'.

6.7.6 Windows Slowly Move Down the Screen

Problem: Various SutterPatch windows slowly move down the screen until they get to the

bottom frame of the SutterPatch main window in Windows 10.

Solution: Reduce the Windows 10 option Start / Settings / System / 'Scale and layout' from

125% to 100%.

6.7.7 Command Window Frozen

Problem: The Command Window is blank and/or unresponsive.

Solution: Use Ctrl-J, or click on the Amplifier Control Panel, and the Command window is

redrawn as an active window. This is a reported Igor issue.

6.7.8 File Operations Crash

Problem: In Windows 10, file opening or saving crashes SutterPatch.

Solution: Remove the Dell Backup and Recovery utility v1.8, or upgrade it to a newer ver-

sion.

6.7.9 Wrong Preferences Settings

Problem: Program preferences are non-standard or corrupted.

Solution: Reset the SutterPatch preferences to their defaults via the SutterPatch / Set

Preferences / General / Preferences Defaults button.

6.7.10 Font Size Too Large

Problem: The font size is too large when using the display scaling in 'Scale and Layout'.

Solution: This can occur on high-resolution monitors running on older versions of Win-

dows 10. Upgrade to the current version of Windows 10.

6.7.10 <u>Magnification Corrupts</u> Window

Problem: After applying right-click Expansion to a window, returning to normal magnifi-

cation corrupts the window.

Solution: Disable the Set Preferences / General / 'Automatic window repositioning'.

6.7.11 Weird Behavior

Problem: There is weird or buggy behavior with the SutterPatch program.

Solution: Reset the SutterPatch program settings to their factory defaults via the

SutterPatch / Set Preferences menu command, by performing a Factory Reset.

6.7.12 Igor Pro Features

Problem: There are a large number of standard features in Igor Pro that can be used in

conjunction with the SutterPatch application.

Solution: Refer to the Igor Help browser, or to Wavemetrics, Inc., regarding issues with

Igor Pro features.

APPENDIX A: LIMITED WARRANTY

- Sutter Instrument Company, a division of Sutter Instrument Corporation, limits the warranty on this instrument to repair and replacement of defective components for two years from date of shipment, provided the instrument has been operated in accordance with the instructions outlined in this manual.
- Abuse, misuse, or unauthorized repairs will void this warranty.
- Warranty work will be performed only at the factory.
- The cost of shipment both ways is paid for by Sutter Instrument Company during the first three months this warranty is in effect, after which the cost is the responsibility of the customer.
- The limited warranty is as stated above and no implied or inferred liability for direct or consequential damages is intended.
- An extended warranty for up to three additional years can be purchased at the time of ordering, or until the original warranty expires. For pricing and other information, please contact Sutter Instrument Company.

SutterPatch Software Licensing Agreement

IMPORTANT NOTICE

PLEASE READ THIS CONTRACT CAREFULLY. BY USING ALL OR ANY PORTION OF THIS PROPRIETARY SOFTWARE YOU ACCEPT ALL THE TERMS AND CONDITIONS OF THIS AGREEMENT. YOU AGREE THAT THIS AGREEMENT IS ENFORCEABLE LIKE ANY WRITTEN NEGOTIATED AGREEMENT SIGNED BY YOU. IF YOU DO NOT AGREE, DO NOT USE THIS SOFTWARE.

1. DEFINITIONS.

- **Section 1.0.** Defined Terms. Terms defined in this Article 1 shall have the meanings given below. Defined terms may be used in the singular or plural.
- **Section 1.1.** "Agreement" means this Software License Agreement, which includes this Agreement's terms and conditions, Schedules, Exhibits, Addenda, and Amendments, if any, which are incorporated in, and form an integral part of, this Agreement.
- **Section 1.2.** "Confidential information" means any data or information, oral or written, of Sutter Instrument Corp., including without limitation, past, present, or future research, development or business activities, including any unannounced product(s) and service(s), and including any information relating to services, developments, inventions, processes, plans, financial information, customer lists, forecasts, and projections, including the terms of this Agreement.
- **Section 1.3.** "End User" means the person or entity who is a client of Licensee and is authorized by User ID and Password to use the Product under Licensee's authorization in the ordinary course of Licensee's business.
- **Section 1.4.** "Licensee" means a party or individual whose client has paid for and is using the Product or an End User, including any evaluation licensee.
- **Section 1.5.** "Licensing Fees" mean the fees paid by a Licensee for the Product.
- **Section 1.6.** "Product" means the software programs of Sutter Instrument Corp., including without limitation, SutterPatch software and corresponding documentation, printed materials and all updates or upgrades of the above that are provided to you, including without limitation, reports, graphs, test scores, interpretations and other information.
- **Section 1.7.** "Term" means the length of time the License paid to use the Product (see current pricing schedule).

Section 1.8. "Website" means the Sutter Instrument Corp. website located at https://www.sutter.com and the Wavemetrics websites located at https://www.igorpro.net.

2. GRANT OF LICENSE.

Section 2.1. Evaluation License. If the Product is used for evaluation, trial or demonstration purposes, Sutter Instrument Corp. grants such user a license solely for the purpose of evaluating, sampling, testing, or demonstrating the Product for the timeframe specified in the evaluation period. The evaluation license is subject to the following conditions: (i) Licensee hereby agrees to indemnify, defend and hold harmless Sutter Instrument Corp. and its officers, directors, employees and suppliers from and against any and all claims, damages, liabilities, costs and expenses, including without limitation, attorneys' fees and costs, arising from or in any way related to the use of the Product, and (ii) otherwise comply with the terms of this agreement.

Section 2.2. Limited License. Subject to the terms and conditions set forth in this Agreement and payment of the license fee (see current pricing schedule), Licensor grants Licensee a non-exclusive, non-transferable, non-assignable license to access and use the Product during the Term. Licensee and Licensee's permitted End Users shall be the only individuals permitted to use the Product, and Licensee is expressly prohibited from otherwise renting, leasing, loaning, selling or otherwise distributing its license rights.

Section 2.4. Description of Features and Functions. A complete description of the features and functions of the Product is available on the Website, and by accepting this Agreement, Licensee represents he/she has read and understands the nature, features, functions and limitations of the Product and agrees to the terms and conditions thereof.

3. USE OF PRODUCT.

Section 3.1. Licensee represents and warrants not to use or permit its End Users to use the Product to (i) upload, post or otherwise transmit any content that is unlawful, harmful, threatening, abusive, harassing, tortious, defamatory, vulgar, obscene, libelous, invasive of another's privacy, hateful, or racially, ethnically or otherwise objectionable; (ii) harm minors in any way; (iii) impersonate any person or entity, falsely state or otherwise misrepresent his/her affiliation with a person or entity; (iv) forge headers or otherwise manipulate identifiers in order to disguise the origin of any content transmitted through the Service; (v) upload, post or otherwise transmit any content with no right to transmit under any law or under contractual or fiduciary relationships (such as inside information, proprietary and confidential information learned or disclosed as part of employment relationships or under nondisclosure agreements); (vi) upload, post or otherwise transmit any content that infringes any patent, trademark, trade secret, copyright or other proprietary rights of any party; (vii) upload, post or otherwise transmit any material that contains software viruses or any other computer code, files or programs designed to interrupt, destroy or limit the functionality of any computer software or hardware or telecommunications equipment; (viii) violate any applicable local, state, national or international law.

Section 3.2. Licensee agrees and shall cause each of his/her End Users to agree pursuant to a written agreement not to distribute, create, copy, duplicate, re-distribute, disassemble, decompile, reverse engineer, or sublicense the Product, and Licensee shall use best efforts to ensure the same. Licensee acknowledges, understands and agrees that Licensee shall acquire no rights or

ownership interest whatsoever in the Product by its use of the Product, which ownership rights shall reside solely with Licensor or its affiliated companies.

- **Section 3.3.** Licensee acknowledges and agrees that Licensee (and Licensee's End Users) is solely and exclusively responsible for contents, data, information and communications that the Licensee and End Users upload and transmit through the Website to use the Product.
- **Section 3.4.** Licensee agrees not to access the Product by any means other than was originally provided by Licensor for use when accessing the Product.
- **Section 3.5.** As required by United States export regulations, Licensee shall not permit export of the Product, or any direct products thereof, to any country to which export is then controlled by the United States Bureau of Export Administration, unless Licensee has that agency's prior written approval. Licensee shall require its End Users to comply with the provisions of this Section.
- **Section 3.6.** If Licensee (or Licensee's End User) is an office, branch or agency of the U.S. Government, then its use, duplication and disclosure of the Product or any part thereof, shall be subject to the following additional requirements: The Product is a "commercial item," as such term is defined in 48 C.F.R. 12.101 (Oct. 1995), consisting of "commercial computer software" and "commercial computer software documentation," as such terms are used in 48 C.F. 12.212 (September 1995). Consistent with 48 C.F.R. 12.212 and 48 C.F.R. 227.7202-1 through 227.7202-4, the U.S. Government End Users shall use the Product with only those rights set forth in this section.
- **Section 3.7.** Licensor reserves the right to include in the Product one or more software devices to ensure that Licensee's use of the Product are consistent with this Agreement.
- **Section 3.8.** Licensee shall indemnify, defend and hold harmless Licensor and its officers, directors, suppliers and licensors of Licensor from any and all claims, damages, liability, costs and expenses, including without limitation, attorneys' fees and costs arising from or in any way related to Licensee's use or misuse of the Product and any other products purchased from Licensor, including without limitation, any derivative products of Licensee utilizing the Product, the hiring or firing of employees or consultants, or any business decisions made as a result of this Product.
- **Section 3.9.** Licensee may not use, copy, modify, translate, or transfer the Product software, documentation, license key, or any of the files included with the Product software except as expressly defined in this agreement. Licensee may not attempt to unlock or bypass any copy-protection, licensing or authentication algorithm utilized by the Product software. Licensee may not remove or modify any copyright notice, nor any "About" dialog or the method by which it may be invoked.

4. FEES.

- **Section 4.1.** Payment of the applicable Licensing Fee permits a Licensee to use the Product. Payment is due in full prior to or contemporaneously with the licensing or use of the Product. All Licensing Fees are payable in U.S. dollars and are non-refundable. Licensing fees may be collected as part of a bundled price included with other products from the Licensor.
- **Section 4.2.** Where required by law, Licensor will collect from Licensee, and Licensee is responsible to pay for, sales and all other taxes associated with the order, unless Licensor is provided a

valid and correct tax exemption certificate prior to order acceptance. The payment of such tax is the sole responsibility of Licensee.

5. SUPPORT.

Section 5.1. Licensor shall provide Licensee with a technical help desk facility to report problems or questions relating to the Product. The help desk is available by email at the following address:

<u>info@sutter.com</u>, or by calling the Licensor at 415-883-0128 between 8:00am and 5:00pm, Monday through Friday, Pacific Time. Licensor will usually reply within 48 hours of receipt of email.

6. INTELLECTUAL PROPERTY.

Section 6.1. Licensee recognizes and acknowledges the exclusive right of Licensor in and to all patents, trademarks, service marks, trade names, copyrights, and other intellectual property and proprietary rights in and to the Product and software, and that such patents, trademarks, service marks, trade names, copyrights, and other intellectual property and proprietary rights are the sole and exclusive property of Licensor

Section 6.2. Licensee waives its right to contest any of Licensor's trademarks, service marks, trade names, copyrights, and other intellectual property and proprietary rights in and to the Product.

Section 6.3. Licensee shall not use such trademarks, service marks, and trade names except where permitted under this Agreement without receiving Licensor's prior written approval of such use. If such approval is granted, Licensee's right to use such trademarks, service marks, and trade names shall end upon the termination of this Agreement, unless otherwise notified by Licensor.

7. CONFIDENTIALITY.

Section 7.1. Licensee shall not disclose or otherwise make available the Product, including but not limited to, the physical media, reports, and documentation to any third party other than Licensee's End Users. Unless otherwise agreed to by the parties, the scope of this Section shall be given the broadest interpretation possible and shall include, but not be limited to, the terms and conditions of this Agreement and pricing.

Section 7.2. The obligations of non-disclosure in Section 7.1 shall not apply to information that (i) is publicly available or in the public domain at the time disclosed; (ii) is, or becomes, publicly available or enters the public domain through no fault of the party receiving such information; (iii) is rightfully communicated to the recipient by persons not bound by confidentiality obligations with respect thereto; (iv) is already in the recipient's possession free of any confidentiality obligations with respect thereto at the time of disclosure; (v) is independently developed by the recipient; (vi) is approved for release or disclosure by the disclosing party without restriction, (vii) is required by a court of competent jurisdiction.

Section 7.3. Licensee acknowledges that the Product software, in source code form, remains a confidential trade secret of Licensor and therefore Licensee agrees that it shall not modify, decompile, disassemble or reverse engineer the Software or attempt to do so.

Licensee agrees to refrain from disclosing the Product software (and to take reasonable measures with its employees to ensure they do not disclose the Software) to any person, firm or entity except as expressly permitted herein.

8. TERMINATION.

Section 8.1 This Agreement shall become effective upon the use of the Product. Licensor may terminate this Agreement at any time upon the default by Licensee of any of the provisions of this Agreement with or without notice from Licensor to Licensee.

9. NO WARRANTIES.

Section 9.1. Although Licensor will make reasonable efforts to correct any material nonconformity between the Product and its then-current documentation that is brought to Licensor's attention with written, documented reports from Licensee, LICENSOR DOES NOT AND CANNOT WARRANT THE RESULTS OBTAINED BY USING THE PRODUCT. Licensor's entire liability and Licensee's exclusive remedy under this Agreement shall be limited to the lesser of (i) the cost paid by Licensee for the software portion of the product, or (ii) \$250.

Section 9.2. THE PRODUCT IS PROVIDED "AS IS." LICENSOR SPECIFICALLY DISCLAIMS ANY AND ALL WARRANTIES TO LICENSEE OF ANY KIND, EITHER EXPRESSED OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, PERFORMANCE RESULTS AND NON INFRINGEMENT. THE ENTIRE RISK AS TO THE QUALITY AND PERFORMANCE OF THE PRODUCT IS WITH LICENSEE.

Section 9.3. IN NO EVENT WILL LICENSOR BE LIABLE TO LICENSEE FOR ANY DAMAGES, INCLUDING LOST PROFITS, LOST SAVINGS, OR OTHER INCIDENTAL OR CONSEQUENTIAL DAMAGES ARISING FROM THE USE OR INABILITY TO USE THE PRODUCT EVEN IF IT HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES, OR FOR ANY CLAIM BY ANY OTHER PARTY. Some states or jurisdictions do not allow the exclusion or limitation of incidental, consequential or special damages, or the exclusion of implied warranties, or limitations on how long an implied warranty may last, so the above limitations may not apply to Licensee. To the extent permissible, any implied warranties are limited to thirty (30) days. This warranty gives Licensee specific legal rights. Licensee may have other rights which vary from state to state or jurisdiction to jurisdiction.

Section 9.4. You acknowledge, understand and agree that your use of this Product is intended to complement your decision-making process and shall not be relied upon solely and completely to make business decisions but only in conjunction with your other business practices, policies, methods, research, investigations, and due diligence. Licensor's license to use the Product is only provided to Licensee in full reliance upon Licensee's representation and warranty to comply with this provision.

10. GENERAL PROVISIONS.

Section 10.1. Licensee warrants that Licensee has the requisite authority to enter into this Agreement.

- **Section 10.2.** This Agreement constitutes the entire understanding and agreement between Licensor and Licensee regarding its subject matter, and supersedes all previous oral and written communications, agreements, memoranda, representations, or understandings between Licensor and Licensee regarding this Agreement. No other rights or licenses are granted to Licensee, except as expressly provided herein.
- **Section 10.3.** Licensor may amend the terms of this Agreement and related company documents at any time with respect to any new releases, updates or versions of the Product, which if purchased by Licensee, will subject Licensee to the terms of the then current Licensor Software Licensing Agreement.
- **Section 10.4.** This Agreement is not transferable or assignable by Licensee under any circumstances, without the prior written consent of Licensor. This agreement shall be binding upon, and is made for the benefit of, each party, its successors, and permitted assignees (if any). For the purposes of this Agreement, any change in control of Licensee shall constitute an assignment or transfer of this Agreement. As used in this section, a change in control is defined as (i) any change in ownership of more than fifty percent (50%) of the voting interest in Licensee, whether by merger, purchase, foreclosure of a security interest or other transaction, or (ii) a sale of all or substantially all of the assets of Licensee.
- **Section 10.5.** The relationship established by this Agreement between Licensee and Licensor shall be that of licensee and licensor. Nothing contained in this Agreement shall be construed as creating a relationship of agency, joint venture or partnership between Licensee and Licensor, so that neither party shall have any right whatsoever to incur any liabilities or obligations on behalf of the other party.
- **Section 10.6.** The laws of California shall govern this Agreement. Any action or proceeding brought by either party against the other arising out of, or related to, this Agreement shall be brought only in a state or federal court of competent jurisdiction located in California and the parties hereby consent to the personal jurisdiction of said courts.
- **Section 10.7.** In the event that any provision of this Agreement is found invalid or unenforceable pursuant to a judicial decree or decision, the remainder of this Agreement shall remain valid and enforceable according to its terms.
- **Section 10.8.** The headings provided in this Agreement are for convenience and reference purposes only. In the event of a conflict between the terms and conditions listed in Articles 1 through 10, and the attached Schedules, the terms and conditions shall govern.
- **Section 10.9.** A waiver of a breach, violation, or default under this Agreement shall not be a waiver of any subsequent breach, violation or default. Failure of either party to enforce compliance with any term or condition of this Agreement shall not constitute a waiver of such term or condition.
- **Section 10.10.** All notices and communications shall be in writing and shall be deemed to have been duly given when delivered or three (3) Business Days after mailing by certified mail, return receipt requested, postage prepaid, addressed to the parties at their respective addresses or at such other addresses as the parties may designate by written notice in accordance with this section.

Section 10.11. Any amendments or addenda to this Agreement, may be executed in counterparts, each of which will be considered an original, but all counterparts together will constitute one agreement. A facsimile of a signed copy of this Agreement, or an electronic or other digital signature imprinted on this Agreement, may be relied upon as an original.

APPENDIX C: ACCESSORIES

• dPatch Expansion Panel Rack-mountable analog and digital I/O BNC panel.

• Ground Point Grounding point hardware.

dPatch Expansion Panel DPATCH-PCH

This optional rack-mountable panel provides BNC ports for easy access to all dPatch digital I/O channels and rear-panel auxiliary analog channels.



Figure C-1. dPatch Expansion Panel

dPatch Expansion Panel			
Dimensions (in)	18.8 x 2 x 3.5		
Dimensions (cm)	48 x 5 x 9		
Weight (lbs)	3.5		
Weight (kg)	1.6		
Digital Output BNCs	16		
Auxiliary Analog Output BNCs	2		
Auxiliary Analog Input BNCs	4		

Table C-1. dPatch Expansion Panel

This panel attaches to the dPatch rear panel via the included 50-pin D-sub cable for digital outputs and HDMI cable for analog I/O.

Digital channels with "high" outputs are indicated by a blue LED light.

For system grounding, this optional machined brass tower provides reliable low-resistance connections for electrophysiology setups. The base plate mounts directly to air table tops (imperial and metric) with the included ¼-20 and M6 screws. The plated connectors accept up to 9 banana plugs and 8 bare wires (up to 10 gauge). A "star" ground configuration is used to avoid ground loops.




Figure C-2. Ground Point

The Ground Point 10-item kit includes cables with banana plugs and alligator clips. For very complex rigs, two sets of cables might be needed.

APPENDIX D: FUSE REPLACEMENT

In the event that the instrument fails to power up when it is switched on, the power-line fuses should be checked to determine whether they have blown. Two fuses are located in the fuse holder in the power cord module on the rear of the amplifier.

To replace a fuse:

- 1. Unplug the power cord from the power entry module, revealing the fuse holder below.
- 2. Remove the fuse holder.
- 3. If a fuse is blown, it is recommended to replace both fuses.
- 4. Insert appropriately-rated replacement fuses (see below).
- 5. Replace the fuse holder in the power entry module and reconnect the power cord.

Mains	Fuses		
Power	(5 mm x 20 mm, glass tube)		
Source	Fuse Rating	Manufacturer Examples	
100 - 240	T2.0 A,	Bussmann: GMC-2-R, S506-2A,	
VAC	250 V (Time Delay)	Littelfuse: 239.002.P	

Table D-1. dPatch Fuses

APPENDIX E. TECHNICAL SPECIFICATIONS



General Specifications

	dPatch Amplifier – Physical
Dimensions (in.) (includes handles & connectors)	19 (W) x 11 (D) x 3.5 (H)
Dimensions (cm) (includes handles & connectors)	48.2 (W) x 28 (D) x 9 (H)
Weight (lb) (with headstages)	15
Weight (kg) (with headstages)	6.8
Case	Aluminum
Computer Communications	USB 3.0 (SuperSpeed)
BNC Channels	4 Auxiliary analog outputs (current sourcing: ± 40 mA) 8 Auxiliary analog inputs (impedance: $1 \text{ M}\Omega$) 1 Digital output trigger (current sourcing: 20 mA) 1 Digital input trigger (impedance: $1 \text{ M}\Omega$)
Rack use	19" rack-mount (2U)
Benchtop use	4 Rubber feet
Signal Ground	4 mm Banana socket
Earth Ground	4 mm Banana socket
Safety Signage	CE marking (Conformité Européenne)
Digital Outputs	DC-37 female connector

		· · · · · · · · · · · · · · · · · · ·	
	PIN	DEFINITION	
	1	Digital Output 1	
	2	Digital Output 2	
	3	Digital Output 3	
	4	Digital Output 4	
	5	Digital Output 5	
	6	Digital Output 6	
	7	Digital Output 7	
	8	Digital Output 8	
	9	Digital Output 9	
	10	Digital Output 10	
	11	Digital Output 11	
	12	Digital Output 12	
	13	Digital Output 13	
	14	Digital Output 14	
	15	Digital Output 15	
	16	Digital Output 16	
	17	+5 V	
Digital Output	18	+5 V	
Pinout	19	+5 V	
	20	Ground	
	21	Ground	
	22	Ground	
	23	Ground	
	24	Ground	
	25	Ground	
	26	Ground	
	27	Ground	
	28	Ground	
	19	Ground	
	30	Ground	
	31	Ground	
	32	Ground	
	33	Ground	
	34	Ground	
	35	Ground	
	36	+5 V	

		37	+5 V	
Pin numbering	Right to left, top to bottom.			

Table E-1. dPatch Amplifier - Physical

dPatch Headsta	age - Physical
Construction	Anodized aluminum case
Dimensions (in) [L x W x H]	4.000 x 1.375 x 0.825
Dimensions (cm) [L x W x H]	10.160 x 3.493 x 2.096
Cable Length (ft) Headstage to Preamp Preamp to Port	6 6.56
Cable Length (m) Headstage to Preamp Preamp to Port	1.83 2
Weight (lbs) w/o cable w/cable	0.21 0.294
Weight (kg) w/o cable w/cable	0.095 0.133
Ground Socket (mm)	1

Table E-2. dPatch Head stage - Physical

dPatch Headstage – Operational (Measured with 8-pole Bessel filter.)

Feedback Element	Analog Bandwidth	Amplitude Range	Electrode Compen- sation Range	Series Resistance Compensation Range	Whole-cell Capacitance Compensation Range
1 pF Capacitor	> 500 kHz	± 20 nA	0.0 - 20 pF	N/A	N/A
500 MΩ Resistor	> 250 kHz	± 20 nA	0.0 - 20 pF	$0.1-100~\mathrm{M}\Omega$	1.0 – 500 pF
50 MΩ Resistor	> 250 kHz	± 200 nA	0.0 - 20 pF	$0.1-10~\mathrm{M}\Omega$	10.0 – 1000 pF

Table E-3. dPatch Headstage - Operational

dPatch Headstage Inputs- RMS Noise (VC) (Measured with 8-pole Bessel filter, voltage clamp, open circuit.)					
Feedback Element 10 kHz 5 kHz 1 kHz					
1 pF Capacitor	220 fA	95 fA	25 fA		
500 MΩ Resistor	750 fA	475 fA	200 fA		
$50~\mathrm{M}\Omega$ Resistor	2.3 pA	1.5 pA	650 fA		

Table E-4. d Patch Headstage – Open Circuit Noise

	dPatch Headstage Inputs-RMS Noise (CC)				
(Measure	d with 8-pole Bessel filt	er, current clamp, gro	ounded input.)		
Feedback 10 kHz 5 kHz 1 kHz Element					
500 MΩ Resistor	1 μV	700 nV	400 nV		

50 MΩ Resistor	1 μV	700 nV	400 nV	
-------------------	------	--------	--------	--

Table E-5. dPatch Headstage – Grounded Input Noise

dPatch Headstage Inputs - Risetime (CC) (Measured with ± 20 nA range.) Series Resistance Rs Pipette Capaci-10-90% Risetime Bandwidthtance 0Ω 10 pF 250 ns $1.4~\mathrm{MHz}$ $100~k\Omega$ 10 pF $350~\mathrm{kHz}$ $1 \mu s$ $1~\mathrm{M}\Omega$ 10 pF $> 70~\mathrm{kHz}$ $< 5 \, \mu s$ $10~\mathrm{M}\Omega$ 10 pF $> 35~\mathrm{kHz}$ $<10 \ \mu s$ $100~\mathrm{M}\Omega$ 10 pF < 15 µs >20 kHz

Table E-6. dPatch Headstage - Risetime

dPatch Data Acquisition				
Analog I/O Channel Type	Full Differential			
Analog I/O Channel Amplitude (voltage)	±10 V			
Analog I/O Channel Amplitude (current)	±20 nA, ±200 nA			
Input Channel Resolution (from headstage)	18 bits			
Input Sample Rates (from headstage)	1 kHz - 5 MHz			
Input Filter Bandwidth (from headstage)	100 Hz – 1 MHz			
Output Channel Resolution (to headstage)	16 bits			
Output Sample Rate (to headstage)	250 kHz with 100 kHz analog smoothing filter			

Auxiliary I/O Channel Resolution	16 bits
Auxiliary In Sample Rate	200 kHz
Auxiliary In Bandwidth	1 - 200 kHz
Auxiliary Out Sample Rate	$250~\mathrm{kHz}$
Auxiliary Out Current (max)	20 mA
Digital In States	0 - 0.8 V = Low 2.0 - 5.5 V = High
Digital Out States	0 - 0.4 V = Low 2.4 - 3.3 V = High
Digital Out Sample Rate	$250~\mathrm{kHz}$
Digital In Trigger Width	Edge triggered (ns)
Digital Out Trigger Width	100 μs
Digital Out Current (max)	20 mA

Table E-7. dPatch Data Acquisition




Figure E-1. Screw Terminal Board

The included Screw Terminal Board attaches to the dPatch rear panel DIGITAL OUTPUTS connector. Digital output signal wires need to be manually screwed into the appropriate pin number. An intervening 37-pin cable (not included) can also be used to extend the length of the Screw Terminal Board.

dPatch Electrical		
Power consumption	18 Watts maximum	
Mains fuse	250V 1A Slow Blow (5 mm x 20 mm) T2.0	
Cables	Shielded grounded power line cord	
Line Voltage	100 VAC – 240 VAC	

Table E-8. dPatch Electrical

dPatch System Components

Carefully remove all components from the shipping container. The following are included for a two-headstage system:

•	(1) dPatch Amplifier	DP/E-2
•	(2) dPatch Headstages	DP-HS
•	(2) Quartz Pipette Holders	EH-Q170
•	(1) Model Cell	MCELL
•	(1) Power Cord	
•	(1) USB 3.0 Cable	
•	(1) Rack Mount Kit	RACK-PK
•	(1) Screw Terminal Board	
•	(1) Quick Start Guide	(with Igor Pro Serial #)
•	(1) USB Flash Drive	SutterPatch and Igor Pro software.

Pipette Holder Parts

•	End	Cap

•	Silicone Gaskets	(O-rings, 6 ea.)
---	------------------	------------------

<u>Gasket ID</u>	$\underline{\mathrm{Color}}$
1.1 mm	Clear
1.2 mm	Green
1.5 mm	Orange-Red
1.75 mm	Blue

- Silver Wire
- Body/Barrel
- Wire Seal
- Gold Pin
- Pin Cap
- Lockdown Ring

(standard: quartz; optional: polycarbonate) (tubing)




Figure E-2. Electrode Holder

Model Cell Parts

- Model Cell
- Connector pins with crimp
- Ground wire

APPENDIX F. SUTTERPATCH ALGORITHMS

Action Potential Threshold Algorithm

[for Action Potential Analysis]

Results pane 'Threshold potential' computation:

The Event starts when the signal slope is > 1 mV/100 μs (10 V/s), or when 25% of the maximum slope is reached, whichever is smaller.

The exact 'Threshold potential' timepoint is based upon differentials using a central differences algorithm.

Auto 'Cell Compensation' Algorithm

[for the dPatch Amplifier Control Panel]

A train of 11 square pulses is applied with a cycle length being 5 times the expected tau of the transient. The routine adjusts the pulse length, if the transient would be too short or too long. Typically, the pulse amplitude is ±5 mV, i.e. 10 mV peak-to-peak. The routine reduces the pulse amplitude, if saturation would occur. See figure 3 for an overview.

The resulting step responses are averaged, skipping the first transient and averaging the central 10 square cycles. Then the negative transients response is reversed in polarity and averaged with the positive transients response, resulting in an average composed of twenty transients.

The input is acquired with a 10 kHz filter bandwidth and a 100 kHz sampling rate.

The measured transient of the whole cell capacitance starts to rise at about $60 \mu s$, measured with an apparent pipeline delay of 5 samples.

The time of half-decay is searched, starting at the first peak. The transients half-time must end in the first half of the stimulus pulse length. Otherwise, acquisition is repeated with an adjusted pulse length.

First estimates for Rs and Cm are computed:

Rs = exp(Xintercept)

Semi-logarithmic regression from the peak, over a time range of 4 * half-decay time. The mean of the last 10% is used as the baseline.

Cm = Integral / dV

Integral from latency (i.e., pipeline samples plus stimulus-filter delay) over a time range of 4 * half-decay time.

These first estimates are iteratively improved by the method as described by Sigworth et al (1995), J Neurosci. Methods, 56:195-202.

Iterations are terminated when the improvements get less than 2%, or after 10 iterations.

Notes:

- Estimating Rs by Rs = tau * dV / Qt (as used in the membrane test), underestimates Rs by about 20 %.
- The method as described by Sigworth et al. fails to get starting estimates.

Auto 'Electrode Compensation' Algorithm

[for the dPatch Amplifier Control Panel]

This algorithm adjusts for the "Cp Fast" portion of a capacitive transient.

A train of 11 square pulses is applied with a cycle length of 2 ms each. Typically, the pulse amplitude is ±5 mV, i.e., 10 mV peak-to-peak. If saturation occurs, the routine reduces the pulse amplitude.

The resulting step responses are averaged, skipping the first transient and averaging the central 10 square cycles. Then the negative transients response is reversed in polarity and averaged with the positive transients response, resulting in an average composed of twenty transients.

The input is acquired with a 50 kHz filter bandwidth and a 500 kHz sampling rate.

The measured transient of the electrode capacitance starts to rise at about 22 μ s, measured with an apparent pipeline delay of 7 samples.

The transient terminates at about 40 µs.

The transient is integrated from the 13th sample to the 21st sample, i.e., over 8 samples (sampling interval is 2 μ s per sample). The integration baseline is the mean of the samples after the transient.

The raw Integral is converted to capacitance by:

```
capacitance = (Integral * SampleInterval ) / ( PulseAmplitude * Cur-
rentGain )
```

where

"PulseAmplitude" is the test pulse amplitude (typically 5 mV),

and

"CurrentGain" is the gain of the dPatch current input, $50~M\Omega$ or $5~M\Omega$, as defined by feedback mode (or 50e6~V/A and 5e6~V/A, since an Ohm is defined as 1~V/A.)

The computed capacitance value is used as a correction value that is added to the electrode capacitance magnitude.

The measurement is repeated until the correction value gets too small (< 2% of electrode capacitance magnitude), or after 10 iterations.

The electrode capacitance tau is optimized by acquiring the same averaged transient while changing tau and measuring the RMS value, and using the tau giving the smallest RMS value. The first iteration starts at a tau value of 1 μ s, and increments by 2 μ s. Then it is refined, as the iteration cycle is repeated starting at the minimum tau minus 2 μ s and incrementing by 0.5 μ s, followed by an iteration incrementing by 0.1 μ s.

Auto Offset Algorithm

[for the Amplifier Control Panel]

- Input is acquired with a 10 kHz filter bandwidth and a 100 kHz sampling rate.
- A 40 ms interval is acquired, of which the second half, i.e., 20 ms, is averaged and used as the offset to be nulled.
- The iteration is repeated until the offset is less than full-scale/10000, i.e. 0.2 mV in VC-mode, or after 20 iterations.

LockIn Computation

[for Routine Editor Virtual Input Channels]

LockIn paper:

Lindau, M., Neher, E. Patch-clamp techniques for time-resolved capacitance measurements in single cells. *Pflugers Arch.* **411**, 137–146 (1988). https://doi.org/10.1007/BF00582306

Math used in the LockIn computation:

```
Factor = (2.0 / SinePointsPerCycle) / sine_amplitude^2
```

A = Factor * \sum (current * stim_real) \sum over one SinePointsPerCycle

B = Factor *
$$\sum$$
 (current * stim_imag)
 \sum over one SinePointsPerCycle

DC =
$$1/\text{SinePointsPerCycle} * \sum (\text{current})$$

 $\sum \text{over one SinePointsPerCycle}$

$$\frac{\text{VC-mode}}{\text{atan}(B/A)} \frac{\text{CC-mode}}{\text{atan}(B/A)}$$

$$\text{RealY} = A \qquad A / (A^2 + B^2)$$

$$\text{ImagY} = B \qquad B / (A^2 + B^2)$$

$$\text{Omega} = (2 * pi) / \text{SineCycleDuration}$$

$$\text{Gt} = \text{Idc} / (\text{Vdc} - \text{Et})$$

$$\text{Gs} = (A^2 + B^2 - A * Gt) / (A - Gt)$$

$$\text{Gm} = \text{Gt} * \text{Gs} / (\text{Gs} - \text{Gt})$$

 $Cm = (A^2 + B^2 - A * Gt)^2 / ((A - Gt)^2 + B^2) / (Omega * B)$

Single Channel Fitting

[for Single Channel Analysis]

Math used in single channel fitting:

Gaussian Fit

$$y = y0 + A * exp(-((x-x0)/width)^2)$$

 $y0 = offset$
 $A = height of curve's peak$
 $x0 = position of center of peak$
 $width = \sqrt{2} * \sigma$
 $\sigma = standard deviation of the peak$

Linear Exponential Fit

$$y = y0 + A * exp(-(x - x0) / tau$$

Logarithmic Exponential Fit

$$y = k0 + k1 * exp(-(ln(x/k2)/k3)^2)$$

Standard Error of the Mean (SEM) Algorithm

[for Analysis Editor Error Bars]

'Standard Error of the Mean' computation:

$$SEM = \sqrt{(SumSq - Mean^2 * N)/(N-1)}$$

SumSq = sum of all squared samples

Mean = sum of all samples / N

Note: The SEM algorithm is similar to the Standard Deviation " $\sqrt{ }$ (variance)", but using 'Mean' vs. 'sum of all samples'.

Synaptic Event Detection Reference

[for Synaptic Event Analysis]

Deconvolution paper:

Pernía-Andrade AJ, Goswami SP, Stickler Y, Fröbe U, Schlögl A, Jonas P. A Deconvolution-Based Method with High Sensitivity and Temporal Resolution for Detection of Spontaneous Synaptic Currents In Vitro and In Vivo. Biophys J. 2012 Oct;103(7):1429–39.